

Purchase Description
Procurement of Optical Particle Sizer / Condensation Particle Counter

The U.S. Army TMDE Activity (USATA) has a requirement to procure one (1) Optical Particle Sizer (OPS) and one (1) Condensation Particle Counter (CPC) or equivalent to reduce risk for biosensor calibration systems. The OPS and CPC to support USATA's mission as follow:

1.0 Requirements:

- Must be compatible with existing automated test equipment and existing programming.
- Have high resolution (at least sixteen channels), user-configurable size channels viewable in real time for the OPS model.
- The instrument is desired to have a 1 liter/minute inlet flow rate to match the majority of Army calibration hardware available.

2.0 Delivery Requirements

Shipping Address: U.S. Army TMDE Activity, Attn: Miles Owen, Building 5435, Redstone Arsenal, AL 35898