

TECHNICAL EXHIBIT B
Deliverables Schedule

<u>Deliverable</u>	<u>Frequency</u>	<u># of Copies</u>	<u>Medium/Format</u>	<u>Submit To</u>
On-site support from certified ISO 17025 technician [PWS para. 5.2.2] Proof of relevant training and qualifications for each service technician performing calibrations on the specified equipment types.	The contractor shall annually submit the updated documentation to the COR	The contractor will submit an electronic copy to the COR at the start of each contract year.	Documentation shall be submitted via email	The COR
On-site support from certified ISO 17025 technician [PWS para. 5.2.2] The contractor shall include in their proposal package a summary of the ISO 17025 calibration procedures intended for use on each equipment type (analytical balances, flow meters, flow controllers, digital thermometers, and analytical weight sets). This will provide the Government with visibility into the calibration process, ensuring alignment with ISO 17025 standards and traceability to appropriate standards; detailed technical specifics are not required, but a clear overview of the documented	The contractor shall annually submit the updated documentation to the COR	The contractor will submit an electronic copy to the COR at the start of each contract year.	Documentation shall be submitted via email	The COR

<u>Deliverable</u>	<u>Frequency</u>	<u># of Copies</u>	<u>Medium/Format</u>	<u>Submit To</u>
procedures is necessary.				
Maintain ISO 17025 Accreditation [PWS para. 5.2.4]	The contractor shall annually submit the updated documentation to the COR	The contractor will submit an electronic copy to the COR at the start of each contract year.	Documentation shall be submitted via email	The COR
Deliverables [PWS para. 5.2.5.4] ISO calibration report: A detailed ISO 17025 compliant report of calibration for each instrument, including all relevant data and traceability information.	The contractor shall submit all calibration reports for each equipment to the COR within five business days of completing the calibration services	The contractor shall either submit an electronic copy to the COR or provide access to the company's database to allow the COR to download the record.	Documentation shall be submitted via email	The COR
Deliverables [PWS para. 5.2.5.4] A physical stamp affixed to each calibrated instrument indicating the new calibration date and ISO 17025 accreditation	Apply a new calibration label/stamp to each equipment immediately following calibration.	One label or stamp for each equipment	Sticker	NA