

This is a SOURCES SOUGHT announcement only. It is neither a solicitation announcement nor a request for proposal or quote and does not obligate the Government to award a contract. Responses to this Sources Sought must be in writing. The purpose of this Sources Sought Announcement is for market research, to make appropriate acquisition decisions, and to gain knowledge of potential qualified businesses capable of providing the following:

Guldmann Ceiling Lift System (Brand Name Only)
*** See Statement of Requirement below for details ***

If this is within your capability, please respond with the information below:

- Company Name
- Address
- Point of Contact
- Phone, Fax, and Email
- UEI number
- GSA Contract (as applicable)
- Capability Statement that clearly and convincingly identifies your ability to fulfill this requirement.
- References, with contract numbers, demonstrating the successful completion of work like the work required in this Sources Sought, performed in the past.
- Brochures of products being offered, to include specifications.
- Authorized service provider letter from OEM
- Country of manufacture

Interested firms are reminded that in accordance with FAR 4.203, prospective contractors shall complete electronic annual representations and certifications to be considered for award. In addition, FAR 4.203-1(b) states that prospective contractors shall be registered in the System for Award Management (www.SAM.gov) database. All Service-Disabled Veteran Owned Small Businesses (SDVOSB) and Veteran Owned Small Businesses (VOSB) shall be verified in Small Business Search at <https://search.certifications.sba.gov/>.

Please include reference number **36C25526Q0391** in all correspondence regarding this Sources Sought.

Point of Contact for this Sources Sought is:

amanda.williams16@va.gov
Department of Veterans Affairs
Network Contracting Office -15
3450 S. 4th Street Trafficway
Leavenworth, KS. 66048

STATEMENT OF REQUIREMENT

Guldmann Ceiling Lift System

OBJECTIVE

The Harry S. Truman Memorial Veteran's Hospital, Women's Clinic, in Columbia, MO, requires two (2) Guldmann Ceiling Lift Systems (Brand Name Only) to provide safe and efficient patient handling, reduce staff injuries, and improve overall patient care.

TECHNICAL REQUIREMENTS

1. Guldmann Ceiling Lift System
 - a. System must include the following:
 - i. Safe Working Load (SWL): 825 lbs
 - ii. Integrated patient scale
 - iii. Hanger bar
 - iv. One (1) sling per lift
 - v. Non-recessed continuously charging track system
 1. 3x4 configuration
 2. 10x13ft or smaller
 - vi. All required hardware, track systems, and mounting components included

DELIVERABLES

CLIN 0001 – Guldmann Ceiling Lift System (Brand Name Only)

QTY: 2 EA

- In accordance with Statement of Requirement – Technical Requirements.
- Guldmann Inc. (USA) Standard Warranty Policy (including shipping, travel & labor):
 - Hoists (ceiling lifts and mobile floor lifts): Three (3) years
 - Rails: 25 years
 - Slings (reusable/washable): Three (3) years
 - Warranty commences upon sign off and acceptance

CLIN 0002 – Installation – White Glove

QTY: 2 JB

- Installation will occur in Rooms A122 and A126.
- Installation will occur in one (1) room per day, on two (2) consecutive days.
- Above ceiling work will require permits from the facility and should be supplied to the Point of Contact (POC).
- All installation hardware and tools shall be provided by the vendor.
- Installation shall include:
 - Functional and load testing of both lifts.
 - Verification of charging system and scale calibration.
 - Correction of any deficiencies prior to final acceptance.
- Contractors must check in prior to arrival to receive contractors' badges for access.
- VA Police will provide contractor access to Rooms A122 and A126.

CLIN 0003 – Training

QTY: 1 JB

- Provide on-site, in-service training to designated VA staff on safe use, weighing function, and basic maintenance.
- Provide user manuals and maintenance guides.

DELIVERY SCHEDULE

30 days after Receipt of Offer (ARO). Installation shall occur no later than 14 days after delivery.

DELIVERY LOCATION

Harry S. Truman Memorial Veteran's Hospital
800 Hospital Dr.
Columbia, MO 65201