

**Salient Characteristics
W81K00-26-Q-A123
Surgical Lighting System
Brooke Army Medical Center, JBSA Fort Sam Houston TX**

Minimum Clinical Performance Requirements:

- Dual fixture two 24-inch diameter light heads to provide center focus control, allowing adjustments in intensity and positioning directly from the sterile handle with 110 degrees of vertical travel.
- Provide deep-cavity illumination with minimal shadows. With center sterile handle, it allows for easy adjustments in intensity, focus, and positioning
- System shall provide white light and 4500K for bright, white light-tailored to different procedures.
- System shall be equipped with energy-efficient LEDs that have a lifespan of 40,000 hours or more.
- System shall include at least 110-degree vertical travel, 360-degree rotation, and drift-free stability, making it versatile for various procedures.

Source Considerations:

- Power source: 100-240VAC, 148-240VA, 50/60 Hz, 1 phase

Biomedical Tech Considerations:

- Operator Training requirements: NO
- Consumable/Spare/Repair parts inventory considerations: NO
- Special Calibration Equipment: NO
- Quality history/Supportability considerations: YES
- Ease of Maintenance/repair considerations: YES
- Regional Support and Service considerations: YES

Certifications:

- FDA certifications or 510K numbers that the equipment/system may hold: YES
- JECETC/JAC certification that the equipment/system may hold: YES
- Conformance to National/International standards such as ASTM, ISO, CSA, UL, NFPA, etc: YES
- Conformance to MIL-STDs if they are required

Special Considerations:

- Special Clinical Settings (Hospital Environment): Orthopedic Services
- Specific recommendation: Surgical lighting system shall meet a dual fixture, and white light and 4500K for bright (be able to adjust brightness as the provider needs).