



DEPARTMENT OF THE AIR FORCE
HEADQUARTERS AIR FORCE SUSTAINMENT CENTER (AFMC)
TINKER AIR FORCE BASE OKLAHOMA

12 March 2026

MEMORANDUM FOR RECORD

SUBJECT: Item Description, M26-110 Handheld XRF Analyzer

1. The 76 MXSG is requesting utilization of funds for the purchase of a Handheld XRF Analyzer.
2. The requested equipment descriptions/specifications are outlined below.
 - a. One (1) Portable XRF Spectrometer for alloy analysis, including the latest manufacturer design and safety features per the National Electrical Code (NEC).
 - (1) Salient Characteristics:
 - (a) Must analyze elements from Magnesium (Mg) 12 through Uranium (U) 92 on the Periodic table. (That is 80 different elements. #'s 12-92)
 - (b) Must provide enhanced performance for light element analysis (Mg, Al, Si, P, and S) without the use of a helium purge or vacuum.
 - (c) Analysis time should be approximately 1-2 seconds for most applications.
 - (d) The X-ray tube must be at least 50kV, 200 μ A with a rhodium (Rh) anode.
3. The following specifics pertain to the item listed in paragraph 2:
 - a. The instrument must be able to connect to a computer to download scans and results.
 - b. Computer software for managing scans and results should be included.
4. The point of contact [REDACTED]