

1. The "Technical Specifications and Salient Characteristics" file contains specific brands in certain line items. Are we allowed to propose alternative brands for the same specializations? Or only mentioned brands should be proposed? Usually there the following phrase "or similar".

Answer: Yes, alternative or similar brands and equivalent products/models are allowed and will be considered provided that the quality of the equipment is high, and such equipment fully meets required salient characteristics, functions, specialization, and regulatory requirements. The general requirement for all equipment is that the equipment be serviceable/repairable in Ukraine, and that spare parts and expendable materials be available in Ukraine.

2. Attachment 1 lists a number of medical equipment items by specific brand and model. Could you please confirm whether equivalent products will be considered, provided they fully meet the required salient characteristics and applicable regulatory requirements?

Answer: Yes, alternative or similar brands and equivalent products/models are allowed and will be considered provided that the quality of the equipment is high, and such equipment fully meets required salient characteristics, functions, specialization, and regulatory requirements.

The general requirement for all equipment is that the equipment be serviceable/repairable in Ukraine, and that spare parts and expendable materials be available in Ukraine.

Delivery location

3. The solicitation refers to delivery to Central Ukraine, while noting that exact delivery details will be provided at award. For pricing purposes, could you please confirm whether any additional delivery assumptions, security procedures, or access limitations should be considered by quoters?

Answer: There are no additional delivery assumptions to consider. All equipment must be delivered to a centralized location on the first/second floor of the hospital. There are no specific security procedures or access restrictions.

Installation and training scope

4. The solicitation states that the contractor shall provide installation and training where applicable and specifically notes training for the hematology analyzer. Could

you kindly confirm which line items require installation, assembly, commissioning, and/or user training?

Answer: Installation, commissioning, and training required for hematology analyzer only. A hospital engineer and electrician will be present on site to provide equipment grounding and safety rubber carpet.