

Sources Sought Notice

This notice is for planning purposes only and shall not be considered an invitation for bid, request for quotation, request for proposal, or as an obligation of the government to acquire any products or services. Responses to this notice will be used for informational purposes only and only for Network Contracting Office 5. No entitlement to payment of direct or indirect costs from the government will arise as a result from a response to this announcement. No contract will be awarded from this notice.

The information requested in this notice will be used to determine available sources for **Ophthalmic Ultrasound for the Washington DC VA Medical Center.**

North American Industry Classification System (NAICS) number is **339112**. The NAICS size is **1,000**.

Contractors that deem themselves capable of meeting the requirement shall provide the below information **to Ryan Singletary, Contracting Officer**, ryan.singletary@va.gov no-later-than May 18, 2026 at 9:00AM, Eastern Time.

The Department of Veterans Affairs, Veterans Health Administration, Network Contracting Office (NCO) 5 is seeking sources for the following procurement:

STATEMENT OF WORK (SOW)

- Contract Title.** Absolu Ophthalmic Ultrasound or equal
- Background.** The DC VA Medical Center is in need of Absolu Ophthalmic Ultrasound for A-scan & B-scan in the eye clinic.
- Scope of Work.** The procurement shall include all parts, materials, labor, software/licenses, resources, and training required to implement and utilize the equipment/system to their fullest capacity. All items must be covered by manufacturer's warranty and procured through a manufacturer-approved distribution channel. Distributors and resellers must be able to document ability to provide items through manufacturer-approved distribution channels upon request.
- Specifications.**

4.1 Equipment

Line Item Description	Quantity
Ellex Absolu S Ultrasound Console with Biometry, B15 MHz, B 20MHZ, 50 MZ UBM Probes	1
Ellex Biometry A- Scan Probe for Absolu	3
Ellex DICOM Software for Absolu Ultrasound	1
Clinical Applications Training	1
Height Adjustable Absolu Cart w/ Keybaord Tray	1
Wireless 9 Function Footswitch for Absolu/ Absolu S	1

Sources Sought Notice

4.2 Required Features

- Ellex Absolu S Ultrasound Console – indicated for both A-scan & B-scan;
- High-resolution imaging with:
 - 15 MHz B-scan probe (posterior segment)
 - 50 MHz UBM probe (anterior segment)
 - Biometry A-scan probe with laser aiming
- Axial resolution:
 - 35 μ m (UBM)
 - 80 μ m (20 MHz)
 - 115 μ m (15 MHz)
- Lateral resolution:
 - 60 μ m (UBM)
 - 200 μ m (20 MHz)
 - 400 μ m (15 MHz)
- Depth of field: up to 20 mm
- Frame rate: up to 16 Hz
- Cineloop: up to 400 images
- Axial length measurement for all eye types (dense cataracts, long eyes, etc.)
- Built-in IOL formulas: SRK-T, Hoffer-Q, Holladay, Haigis, Binkhorst-II, etc.
- Post-op refractive calculation with 6 keratometric correction methods
- Pattern recognition for phakic, aphakic, and pseudo-phakic eyes
- Dimensions: 445 mm (H) x 285 mm (D) x 545 mm (W) without probes
- Weight: 10.6 kg (without probes)
- Power: 80–264 VAC, 47–63 Hz, 60 VA max
- Probes equipped with a motion sensor IMUv® that provides essential real-time position of the probe on the eye and the visualization of the explored area.
- Compatible with Zeiss FORUM – Ophthalmology software
- Training & Installation Included from OEM

5. Specific Tasks.

5.1 Delivery.

- i. All equipment shall be delivered to:
Washington DC VA Medical Center
50 Irving St NW
Washington, DC 20422
ATTN: VHAWAS Equipment Committee
- ii. Contract VA Purchase Order (PO) number shall be included with delivery information (e.g., packing slip).
Note: PO number starts with station code 688 followed by a letter and 5 digits (e.g., 688A12345)
- iii. Delivery Hours: The Warehouse is open to receive deliveries Monday through Friday, 8:00 a.m. to 4:00 p.m., excluding holidays.

Sources Sought Notice

- iv. Contractor shall coordinate delivery with the on-site Point of Contact (POC) and provide shipment tracking information.
- v. Contractor shall verify delivery date and time with the on-site POC at least 3 business days prior to scheduled delivery.
- vi. Delivery is required within 90 days from award of the contract.

5.2 On Site Procedures.

- i. COVID-19 Screening: All persons entering the facility shall be subject to screening for COVID-19 per the current CDC (Center for Disease Control) guidelines.
- ii. Mask Requirement: All contractors shall wear masks while indoors.
- iii. Contract personnel are required to identify themselves as such to avoid creating an impression in the minds of members of the public that they are Government Officials.
- iv. Time spent on site shall be for a period sufficient to complete the work set forth in the statement of work.
- v. On site visits shall be scheduled in advance with the designated on-site POC.
- vi. Contractor shall check in and out with Biomedical Engineering prior to visiting the worksite for each day of work.
- vii. The VA campus is non-smoking. Contractor personnel are required to comply with this policy.
- viii. Parking: it is the responsibility of the Contractor to park in the appropriate designated parking areas.
- ix. Contractor personnel's tool bags are subject to inspection.

5.3 Installation/Implementation.

- i. Installation/Implementation will be completed by personnel certified and knowledgeable with the designated equipment/system.
- ii. Installation shall be completed within the period of performance.
- iii. Contractor shall be responsible for any Personal Protection Equipment (PPE) required when performing work on site.
- iv. Contractor shall submit all removable media to be used on a VA system to Biomedical Engineering for scanning with anti-virus software prior to use.
- v. In the case turn-in, exchange, repair, or replacement of equipment containing hard drives used by the VA, the hard drives shall be removed from the equipment and remain in possession of the VA (this includes loaned or rented equipment).
- vi. The contractor shall take all necessary precautions to prevent damage to any government property. The contractor shall report any damages immediately and shall be assessed current replacement costs for property damaged by the contractor, unless corrective action is taken. Any damaged material (i.e., trees, shrubs, lawn/turf, curbs, gutters, sidewalks, etc.) will be replaced in a

Sources Sought Notice

timely manner or corrected by the contractor with like materials, at no extra cost to the government, upon approval of the Contracting Officer.

5.4 Training.

- i. Contractor shall be responsible for providing on-site training to clinical staff/users.
- ii. Training is to be scheduled in advance with the designated on-site POC.
- iii. Education shall be provided prior to and/or at time of installation.
- iv. Education professional shall be certified to provide instruction on the designated equipment/system.
- v. Education curriculum must include: operations and set-up, user maintenance, safety, and user troubleshooting tips.

5.5 Inspection and Acceptance.

- i. Contractor shall conduct a joint inspection with the on-site POC upon completion of installation.
- ii. In the event deficiencies are identified, the contractor shall provide the date when the identified deficiencies will be addressed if not addressed on the date of installation. The Contractor shall conduct a joint inspection with the on-site POC after addressing all deficiencies. All deficiencies identified in the joint inspections shall be corrected by the Contractor prior to Government acceptance of the item. Any disputes shall be resolved by the Contracting Officer.

5.6 Deliverables.

The Contractor shall provide the below documentation for the proposed equipment to VHAWASEquipmentRequests@va.gov within 10 business days of work completion:

- i. Service report.
- ii. Warranty information.
- iii. Electronic copy of the Operators Manual.
- iv. Electronic copies of the complete technical service manuals, including troubleshooting guides, necessary diagnostic software and equipment information, schematic diagrams, and parts lists.

6. Hours of Operation.

DC VAMC's normal business hours are Monday-Friday, 8:00 am to 4:30 pm, excluding observed Federal holidays:

New Year's Day	Labor Day
Martin Luther King's Birthday	Columbus Day
President's Day	Veteran's Day
Memorial Day	Thanksgiving Day
Juneteenth	Christmas Day

Sources Sought Notice

Independence Day	Any other national holiday as declared by the President of the United States
------------------	--

7. **Period of Performance.** Delivery within 90 days from contract award.

8. **Invoicing.**

Contractor shall submit a consolidated invoice once all services are completed. One all-inclusive invoice shall be provided, which shall contain an invoice number, contract number, obligations order number, and an itemized list of services provided. An acceptable master (all-inclusive) itemized list of services shall be agreed to between the Contractor and the COR prior to submission of the first invoice.

All invoices from the contractor shall be submitted electronically in accordance with VAAR Clause 852.232-72 Electronic Submission of Payment Requests.

VA's Electronic Invoice Presentment and Payment System (IPPS) – The VA Financial Services Center (FSC) uses a third-party contractor, Tungsten, to transition vendors from paper to electronic invoice submission. Please go to this website to begin submitting electronic invoices, free of charge:

<http://www.tungsten-network.com/US/en/veterans-affairs/>.

More information on the VA FSC is available at: <http://www.fsc.va.gov/einvoice.asp>.

Vendor e-Invoice Set-Up Information:

To begin submitting your electronic invoices to the VA FSC for payment processing, free of charge, please contact Tungsten at the below phone number or email address:

Tungsten e-Invoice Setup Information: 1-877-489-6135

Tungsten e-Invoice Email: VA.Registration@Tungsten-Network.com

If you have a question about the e-invoicing program or Tungsten, please contact the FSC at the below phone number email address.

FSC e-Invoice Contact Information: 1-877-353-9791

FSC e-Invoice Email: vafscshd@va.gov

9. **Contract Type.**

Firm-Fixed Priced Contract

The information identified above is intended to be descriptive, not restrictive and to indicate the quality of the supplies/services that will be satisfactory. It is the

Sources Sought Notice

responsibility of the interested source to demonstrate to the government that the interested parties can provide the supplies/services that fulfill the required specifications. If you are interested and are capable of providing the sought-out supplies/services, please provide the requested information as well as the information indicated below. Response to this notice should include company name, address, point of contact, size of business pursuant to the following questions:

(1) Please indicate the size status and representations of your business, such as but not limited to: Service Disabled Veteran Owned Small Business (SDVOSB), Veteran Owned Small Business (VOSB), HUBZone, Woman Owned Small Business (WOSB), Large Business, etc.

(2) Is your company considered small under the NAICS code identified under this sources sought?

(3) Are you the manufacturer or distributor of the items being referenced above (or equivalent product/solution)? What is the manufacturing country of origin of these items?

(4) If you are a small business and an authorized distributor/reseller for the items identified above (or equivalent product/solution), do you alter; assemble; modify; the items requested in any way? If you do, state how and what is altered; assembled; modified?

(5) Does your company have an FSS contract with GSA or the NAC or are you a contract holder with NASA SEWP or any other federal contract? If so, please provide the contract type and number.

(6) If you are an FSS GSA/NAC or NASA SEWP contract holder or other federal contract holder, are the referenced items/solutions available on your schedule/contract?

(7) Please provide general pricing for your products/solutions for market research purposes.

(8) Please submit capabilities in regard to the salient characteristics being provided and any other information for planning purposes.

Submissions addressing Section (8) should show clear, compelling and convincing evidence that all equal items meet all required salient characteristics.

Telephone responses shall not be accepted. If a solicitation is issued it shall be announced on a later date, and all interested parties must respond to that solicitation announcement separately from the responses to this sources sought. Responses to this notice are not a request to be added to a prospective bidders list or to receive a copy of the solicitation.