

Vendor Question and Customer Response

RFQ 80NSSC26929758Q - Microwave Digester to Dissolve Materials in Acid Solutions

Vendor Question 1: While reviewing the specifications, we noticed that there is no indication of the vessel rack configuration or vial size to be included. To ensure the quote aligns with the intended application and requirements, could you please clarify which vessel type and capacity should be specified?

Customer Response: Typically, I only need about 5 mL to digest samples between 100 – 200 mg. I use larger vessels (50 mL polypropylene Digitubes) with a hot block, but I certainly don't need all that volume. We don't process a high volume of sample, so large capacity is not very important to us. Vessels should be compatible with acid mixtures mostly consisting of hydrochloric, nitric and hydrofluoric acids.