

STATEMENT OF WORK/PERFORMANCE WORK STATEMENT

SERVICE OF ACCELEROMETERS (CALIBRATION & EVALUATION)

1. INTRODUCTION

Naval Surface Warfare Center (NSWC), Corona Division, Measurement Science Department, Code MS35, requires the evaluation calibration services for accelerometers in support of the Navy Calibration Standards Program.

2. BACKGROUND

Measurement Science Department Corona has a requirement to provide Test and Measurement Diagnostic Equipment (TMDE) and/or associated components for the Calibration Standards Program to support the Navy Calibration Labs with calibration standard test equipment.

The accelerometers requiring service are used as calibration standards to support vibration measurement systems and related TMDE. Timely servicing is critical to ensure continued operational readiness and prevent delays in calibration support for fleet operations. The standard is listed in 9 ICPs and is listed in the NCE. Failure to service these assets will impact calibration capabilities and mission readiness.

In order to meet the requirements, the equipment listed in the specifications is required to be repaired/serviced.

3. SCOPE

Accredited calibration services for the following accelerometers:

- Model: 353B04 (Primary Accelerometer)
 - Serial Number: TBD
 - Calibration range: 5 Hz to 20 kHz (laser interferometry)
- Model: Q353B51 (Low-Frequency Transfer Standard)
 - Serial Number: TBD
 - Calibration range: 0.1 Hz to 4000 Hz

4. APPLICABLE DOCUMENTS

N/A

5. REQUIREMENTS

The contractor shall:

- Provide all labor, equipment, and materials required
- Perform calibration using laser interferometry (as applicable)
- Ensure accreditation standards are met (A2LA or equivalent)
- Provide calibration certificates and data on disk or electronic format
- Notify the Government of any out-of-tolerance conditions

6. GOVERNMENT FURNISHED PROPERTY

6.1 Government Furnished Material

N/A

6.2 Government Furnished Equipment

Accelerometers listed in Section 3

6.3 Government Furnished Information

N/A

7. DELIVERABLES

NUMBER	NAME
A001	Calibration Certificate and Data Package

ADDITIONAL INFORMATION

A. PERIOD OF PERFORMANCE / REQUIRED DELIVERY DATES

Service must be completed 60 days after contract award

B. PLACE OF PERFORMANCE

TBD

C. SHIPPING ADDRESS

TBD

D. INSPECTION AND ACCEPTANCE

TBD

E. TECHNICAL POINT OF CONTACT

TBD

DRAFT