

CONTINUATION SHEET	REFERENCE NO. OF DOCUMENT BEING CONTINUED	PAGES
	140G0225Q0115/0002	PAGE 2 OF 3
NAME OF OFFEROR OR CONTRACTOR		

ITEM NO. (A)	SUPPLIES/SERVICES (B)	QUANTITY (C)	UNIT (D)	UNIT PRICE (E)	AMOUNT (F)
	Amendment 0002 is issued to incorporate Questions and Answers. The due date for proposals is unchanged from May 7, 2026, at 5PM ET. Any additional questions are to be sent to Justice Pryor, justice_pryor@ios.doi.gov. Period of Performance: 05/01/2026 to 01/31/2031 Contract Specialist: Justice Pryor, justice_pryor@ios.doi.gov Contracting Officer: Trisha Beals, trisha_beals@ios.doi.gov				

Questions and Answers

1. Are you also able to share the anticipated annual throughput on the platform and which flow cells the end user is expecting? I ask because our platform offers substantially higher output but at substantially lower cost/Gb. I think it would be important and helpful to be able to model out what the anticipated annual op-ex expenses would be. We will almost certainly be more expensive for the equipment itself, but the delta in run costs may wash that out extremely quickly.
 - a. We have a pretty small annual throughput. In the past year we have run our current MiSeq 11 times all with a 2x300bp flow cell. We are looking for an easily maintained platform that doesn't require all the washes needed when the machine is idle (since most of the time it is idle).