

## AMENDMENT TO SOLICITATION

|                                    |                                  |
|------------------------------------|----------------------------------|
| AMENDMENT NO.<br>1                 | DATE OF AMENDMENT:<br>05/05/2026 |
| SOLICITATION NO.<br>RFP-381067-GMG | DATE ISSUED:<br>04/30/2026       |
| PROJECT NO.                        | OPENING DATE:<br>05/08/2026      |

TITLE OF PROJECT:

### Cryogenic Microbulk 3000-liter Liquid Argon Dewar

Bidders must acknowledge this Addendum prior to the hour and date set forth in the solicitation by one of the following methods:

1. By separate email

If by virtue of this Amendment you desire to change a bid already submitted, such change may be made by email, provided such email makes reference to the Solicitation and this Amendment and is received prior to the time set for opening.

The above reference Solicitation is amended as follows:

This Amendment is issued to respond to questions submitted by interested parties.

**Question:** Do we know the expected monthly Argon usage (either in ccf, cf, or liters)?

**Response:** An average of approximately 300 cubic feet (8,500 liters) per month of liquid argon is used, but we have gone as high as 745 cubic feet (21,100 liters) in a month. It varies depending on what experiments are running and during the summer months, when it's warmer, the boil off (wasted and vented gas) is a bit higher.

**All other terms and conditions remain unchanged.**

Fermi National Accelerator Laboratory  
P.O. Box 500  
Batavia, IL 60510  
M.S 210

Grace Gomez  
Procurement Specialist II  
(630) 840-2864

