

STATEMENT OF WORK

1. Objective/Requirements

The NASA Armstrong Flight Research Center (AFRC) requires the procurement of (4) Ametek Programmable DC Power Supply or Equivalent.

2. Characteristics, Scope, and Specs

LINE	PART NO.	DESCRIPTION	QTY.
1	DLM40-15M51A OR EQUIVALENT	DLM Series - 600 Watt Near Linear Benchtop and 1 U 1/2 Rack System Supply, 600 W, 0-40 V, 0-15 A, Five (5) channel isolated analog control and monitoring, 5 Year Standard Warranty Included OR EQUIVALENT <u>Specs:</u> • The voltage/current/power specification (40V/15A/600W) • possess a current limiting mode • Must have a front panel interface for controlling voltage, current, and overvoltage setpoints • Must contain a means of overriding front panel and controlling the set current limit setpoint/overvoltage limit setpoint/voltage setpoint via both 0 to 10V analog voltage signals or 0 to 5kΩ external resistors • 0 to 10V analog monitoring signals for output voltage and current • Electrical isolation from the power output terminals for all of the aforementioned analog control and monitoring signals; • Must all be in a 1U, half rack width package that can be a combined two units into a single 1U 19" rack unit	4
2	DLMRK OR EQUIVALENT	Rack Mounting Kit OR EQUIVALENT	2

3. Place of Performance

NASA Armstrong Flight Research Center (AFRC)
Building 4840, Room 212
P.O. Box 273
Edwards CA 93523-0273

4. Period of Performance

4 Weeks ARO.