

Specifications for Particle Analyzer

To meet our needs the equipment must meet these requirements:

Type(s) of material: Wheat whole meal, flour, and isolated wheat starch.

Particle size range of materials to be measured: Starch 0.1 μm to 25 μm , flour 1 μm to 500 μm .

Application type: Wet and dry.

Intended application for particle counting/characterization: Examination of wheat physiochemical properties.

Prefer Model Number C53020 made by Beckman Coulter and related items. Brand name or equal.
One of each item:

C53020 LS 13320 XR MW with ULM & DPS Modules

C20930 Workstation Assembly

7800370 LS Control G15 10pk

7800372 Pkg LS Control GB500

CO6826 HiCap HEPA vacuum 110 volt CFM 99

C58400 LS 13220 XR Installation including basic O

Include shipping charges to the location in Pullman, WA

Provide verification that items comply with the Made in America Laws.