

Test Report

Verification Website: www.gtgc.net.cn

Verification Code: QFML-8716-54

No:20R000520

Issue Date: 2020-04-17

Applicant: NANNING TECBOD BIOLOGICAL TECHNOLOGY CO.,LTD

Address: ROOM 601 FLOOR 6 , B2 BUILDING,NO 19 GUOKAI DADAO,NANING,GUANGXI,CHINA

Information confirmed by applicant:

Disposable medical protective hood

SIZE: Quantity: twenty pieces

Standard Adopted:

Date Received/Date Test Started: 2020-04-10

Conclusion:

BREAKING STRENGTH

BURSTING STRENGTH

Note: "M"-Meet the standard's requirement "F"-Fail to meet the standard's requirement "---"-No comment

Remark:

All the tested items are tested under the standard condition (except for indication).

Copies of the report are valid only re-stamped.

The experiment was carried out at No.1, Zhujiang Road, Panyu District, Guangzhou, Guangdong, P.R.China.



Approved By:

ZiShan Guo

ZiShan Guo Senior Engineer

Test Report

No: 20R000520



TEST REPORT

No: 20R000520

BREAKING STRENGTH(N)

(EN 29073-3:1992, THE DISTANCE BETWEEN THE CLAMPS:200mm, RATE:100mm/min)

LENGTH 127.2

WIDTH 65.1

BURSTING STRENGTH(kPa)

(EN ISO 13938-1:1999, TEST AREA:50cm²)

87.0



——End of Report——

PAGE 3 OF 3