

Motion Artifacts in MRI: A Complex Problem With Many Partial Solutions

Maxim Zaitsev, PhD,^{1*} Julian Maclaren, PhD,^{1,2} and Michael Herbst, PhD^{1,3}

Subject motion during magnetic resonance imaging (MRI) has been problematic since its introduction as a clinical imaging modality. While sensitivity to particle motion or blood flow can be used to provide useful image contrast, bulk motion presents a considerable problem in the majority of clinical applications. It is one of the most frequent sources of artifacts. Over 30 years of research have produced numerous methods to mitigate or correct for motion artifacts, but no single method can be applied in all imaging situations. Instead, a “toolbox” of methods exists, where each tool is suitable for some tasks, but not for others. This article reviews the origins of motion artifacts and presents current mitigation and correction methods. In some imaging situations, the currently available motion correction tools are highly effective; in other cases, appropriate tools still need to be developed. It seems likely that this multifaceted approach will be what eventually solves the motion sensitivity problem in MRI, rather than a single solution that is effective in all situations. This review places a strong emphasis on explaining the physics behind the occurrence of such artifacts, with the aim of aiding artifact detection and mitigation in particular clinical situations.

J. MAGN. RESON. IMAGING 2015;42:887–901.

Compared with other imaging modalities, such as ultrasound or computed tomography, MRI has always been particularly sensitive to subject motion. This is primarily due to the prolonged time required for most MR imaging sequences to collect sufficient data to form an image. This is far longer than the timescale of most types of physiological motion, including involuntary movements, cardiac and respiratory motion, gastrointestinal peristalsis, vessel pulsation, and blood and CSF flow. The effects of motion have been well known since the early days of MRI and include blurring and ghosting in the image.^{1,2}

Recent technological improvements have improved the situation in some cases, but have exacerbated it in others. On one hand, incremental performance gains in hardware (e.g., higher performance gradients) have enabled faster imaging, as have breakthroughs such as parallel imaging. Faster imaging means that some scans can be performed in a shorter time, leading to a smaller chance of involuntarily subject motion. On the other hand, hardware improvements have improved the achievable resolution and signal-to-noise ratio (SNR), and the sensitivity to motion has, therefore, increased. Stronger gradients also mean greater phase accu-

mulation for moving spins. Higher main field strengths and stronger gradients also mean that a typical MR scan is generally louder than it was in the early days of MRI, which reduces the chance of infants sleeping through the procedure. This is particularly true for diffusion-weighted or diffusion tensor imaging, as was reviewed recently by Le Bihan et al.³

At present, there is no sign that the problem of subject motion during MRI examinations will be resolved through hardware improvements. The potential of accelerated imaging seems to be increasingly limited by biologic constraints: peripheral nerve stimulation limits gradient switching speeds; specific adsorption rate (SAR) limits the use of RF excitation pulses; and T1 and T2 relaxation times constrain the sequence repetition and echo times, depending on the required contrast.

Furthermore, it seems increasingly likely that there is not, and will not be, a single methodological solution to the problem of motion in MRI; rather, a toolbox of solutions exists. The choice of which tool to apply depends on the task at hand, namely, the imaging situation and type of motion in question. Some prevention or correction

This website utilizes technologies such as cookies to enable essential site functionality, as well as for analytics, personalization, and targeted advertising purposes. You may change your settings at any time or accept the default settings. [Privacy Policy](#)

[Manage Preferences](#)

[Accept All](#)

techniques are highly effective in particular circumstances, but useless in others. Other techniques are relatively general, but are not completely effective. It is, therefore, extremely important to recognize and understand motion artifacts to identify the physical cause of the problem in a particular clinical situation and use the best possible tool for the problem.

In this review, we first examine the physical origins of motion artifacts and then summarize the main motion prevention and correction tools currently in use or in development. Our aim is to provide a condensed guide to the function, application and limitations of these methods.

Physical Principles Behind Motion Artifacts

The appearance of motion artifacts in an image is a result of a complicated interaction between the image structure, the type of motion, the specifics of the MR pulse sequence and the k -space acquisition strategy. In this section we briefly revisit the most important physical mechanisms affecting the appearance of motion artifacts.

***k*-Space and the Image Acquisition Process**

Spatial encoding in MRI is an intrinsically slow and sequential process. To appreciate the effect of motion on MR images it is important to realize that the data acquisition occurs not directly in image space, as is the case in photography, but rather in frequency or Fourier space, which is commonly termed " k -space".

K -space corresponds to the spectrum of the spatial frequencies of the imaged object and depending on the imaging situation can be two- or three-dimensional. Spatial frequency spectra of objects with small number of contrast features and with smooth intensity variations are predominantly defined by the components close to the k -space origin. Objects or organs with sharp high-contrast edges contain a significant spectral density at the k -space periphery. The majority of biological samples show very local spectral density in k -space, centered around $k = 0$. Some organs and tissues, such as brain cortex, have a fractal-like nature⁴ and therefore show a slower decay of spectral density in k -space.

To understand the properties of k -space and its relation to image/object space, it is important to develop an intuition regarding the local and global properties of these two spaces. This topic is reviewed in detail by Paschal and Morris.⁵ K -space describes the object using a set of global

predominantly the center of k -space is affected. The latter phenomenon is the basis of a range of data-sharing techniques, which achieve a higher apparent temporal resolution in dynamic imaging by updating the central region of k -space more frequently.^{6–9}

Different strategies to populate k -space with measured data are termed " k -space sampling trajectories" or " k -space trajectories". The most common and clinically relevant approach collects data on a rectilinear grid in k -space (so-called "Cartesian" sampling), because it allows for a computationally efficient image reconstruction using the fast Fourier transform (FFT). Even though we primarily consider the Cartesian sampling strategy, it is necessary to distinguish between the multitude of k -space trajectories, which differ from each other in the way and the order the nodes of the rectilinear 2D or 3D k -space grids are visited. Other popular data acquisition schemes include radial,^{5,10} PROPELLER,¹¹ and spirals.¹²

Typical motion-induced deterioration effects observed in MR images consist of a combination of the following basic effects: blurring of sharp contrast or object edges (intuitively similar to photography), ghosting (both coherent and incoherent) originating from moving structures, signal loss due to spin dephasing or undesired magnetization evolution, and the appearance of undesired strong signals.

The first two points are related to the signal readout process, whereas the latter two are related to the signal generation and contrast preparation within the pulse sequence. Here we refer to ghosting as a partial or complete replication of the object or structure along the phase-encoding dimension, or along multiple phase-encoding dimensions for 3D imaging. Periodic motion synchronized with the k -space acquisition results in a coherent ghosting with the number of replicas corresponding to the frequency of k -space modulation: two ghosts result if every second line is altered, four if every fourth, etc. Deviations from perfect periodicity in k -space result in incoherent ghosting, appearing as multiple overlapped replicas, and sometimes seen as stripes in the phase-encoding dimension.

The main cause of readout-related motion artifacts is the inconsistency between the various portions of the k -space data used for the image reconstruction or between the data and the signal model assumed in the reconstruction. Simple reconstruction using an inverse FFT (iFFT) assumes the object has remained stationary during the time the k -space data were sampled. A violation of this assumption results in artifacts.




FIGURE 1: (a) Original FFT reconstruction without motion, (b–d) sum of first and last source images to show the range of simulated motions for (b) continuous monotonic rotation to the total angle of 10° , (c) continuous vertical translation, and (d) continuous horizontal translation (in both cases 10 pixels for a matrix of 256×256). Periodic horizontal translation has the same amplitude and is not shown. Images (e–h) demonstrate the results of averaging the motion as would correspond to photography with an equivalently long exposure time; (e) rotation, (f) vertical translation, (g) horizontal translation, and (h) periodic horizontal translation. Note the loss of detail and edge information due to the motion blur and the enhanced blurring in (h) due to the fact that with sinusoidal motion the object spends more time close to the terminal positions. Images (i–l) show simulated MRI acquisitions with linear k -space ordering. Images (m–p) show simulated two-shot interleaved MRI acquisitions. Plots of the motion as a function of k -space position for the representative images are presented in Figure 2.

horizontal directions (Figs. 1b–d, respectively) and a periodic sinusoidal oscillation in the horizontal direction with the same amplitude (not shown). To provide a reference of image quality, a photography-like image acquisition has

ordering, Fourier acquisition is actually significantly less motion-sensitive than a corresponding “photographic” image taken with the same “exposure time” (Figs. 1i–k). Periodic motion however produces very strong ghosting (Fig. 1l). In




FIGURE 2: Effective translation (as a fraction of the FOV) as a function of the *k*-space position for (a) Figure 1j and 1k, (b) Figure 1l, (c) Figure 1n and 1o, and (d) Figure 1p. As seen, if the external disturbance oscillates as a function of *k*-space position, e.g. in cases (b), (c), and (d), ghosting will result (see corresponding images in Fig. 1). Reduced ghosting in Figure 1p can be explained by the effective reduced frequency of oscillation of the position as a function of the *k*-space coordinate (compare plots b and d). Plots of rotations are omitted due to their similar appearance.

synchronized with the underlying periodic motion, the artifacts reduce further to the level comparable with that in Figure 1k. This phenomenon is often exploited in cardiac imaging and is discussed in more detail below. Plots of the motion as a function of *k*-space position for the representative images in Figure 1 are presented in Figure 2.

Although slow continuous motions do occur in clinical settings, for example due to the gradual relaxation of the neck muscles in head imaging, they only strongly affect sequences relying on interleaved *k*-space acquisitions, e.g., T2-weighted TSE/FSE (Turbo/Fast Spin Echo) imag-

have been simulated, with the motion occurring at different time points, leading to varying proportions of *k*-space data that are inconsistent. Figures 3a-d and Figures 3e-h assume linear and interleaved *k*-space ordering, respectively. In both cases 12.5% (1/8th) of the inconsistent *k*-space data result in a negligible effect; however, image quality deteriorates much faster for the interleaved acquisition. Centric *k*-space reordering demonstrates even higher resistance against artifacts (Figs. 3i-l).

The following conclusions can be drawn from these simulations: (i) MRI with a sequential *k*-space ordering is




FIGURE 3: Simulations of a single sudden orientation change during the k -space acquisition for different k -space acquisition strategies and different amounts of inconsistent k -space data. For a linear k -space ordering, images reconstructed from datasets containing 12.5% (a), 25% (b), 37.5% (c), and 50% inconsistent data (d) have been simulated. Hardly any artifacts are visible if the corruption occurs less than half of the maximum k -space distance to the center. Similar conclusions follow for the two-shot interleaved k -space acquisitions with 12.5% (e), 25% (f), 37.5% (g), and 50% of inconsistent data (h). i-l: Images present a similar simulation for the centric reordering with (i) 50% of data inconsistent with respect to the k -space center data, (j) 62.5%, (k) 75%, and (l) 87.5%, respectively. The precise threshold at which visually detectable artifacts start to appear strongly depends on the image, primarily on the presence of small high-contrast features, but the overall tendency is that below a certain distance from the k -space center a single motion event produces only negligible artifacts.

Types of Motion

The different types of motion relevant for in vivo MRI can be loosely separated into the following categories: rigid body motion, elastic motion and flow.

Rigid body motion, or “bulk motion”, includes 1D translations (requiring a single parameter for its mathematical representation), multi-dimensional translations (requiring several parameters), and completely unconstrained rigid motion (requiring six parameters). For example, diaphragm motion is sometimes approximated using a one-dimensional translation, but involuntary head motion normally requires six degrees of

one-dimensional, for example for laminar blood flow in small or medium-sized vessels or CSF flow in the cervical spine. It then requires only a single parameter (velocity) for its complete representation. More complex flow is observed in the heart and aorta. This requires a two- or three-dimensional velocity vector field for its complete description.

The most complete description is provided by the displacement field, which consists of the trajectories of each particle within the object as a function of time. Although being complete, the displacement field is often not measura-




FIGURE 4: Velocity effects on the phase of MR signals. Stationary spins can be refocused with a pair of gradient pulses of equal area and opposite polarity, which is a basis of gradient echo. Spins moving along the gradient direction acquire additional phase due to the fact that dephasing and refocusing occur at different physical locations where the same gradient induces different frequency shifts, leading to different amounts of the phase acquired during dephasing and refocusing periods. Greyed segments on the corresponding phase circles mark the amount of phase acquired by the moving spins during the last gradient pulse.

Data Corruption Mechanisms

tionary during imaging. As seen in Figure 4, spins

This website utilizes technologies such as cookies to enable essential site functionality, as well as for analytics, personalization, and targeted advertising purposes. You may change your settings at any time or accept the default settings. [Privacy Policy](#)




FIGURE 5: Excitation history effects visualized for a three-pulse MR pulse sequence. In the case of no motion, regions excited by all three pulses overlap resulting in the desired signal evolution. The respective brain regions are marked using half-transparent bands with the color of the corresponding slice-selective RF pulse. Overlap of red and green results in light brown. When light brown is followed by an overlap with blue, a grey color results, which indicates the desired spin evolution. In the case of motion, (rotation after the first pulse, followed by a combination of rotation and translation after the second pulse, original position is shown as dotted line) the grey region corresponding to the desired spin evolution is substantially reduced, while colored regions with undesired signal evolution appear. Therefore, desired signals are decreased, and undesired signals are produced, resulting in artifacts, such as ghosting or signal disturbances. The unwanted signals may also cancel some of the desired signals, causing signal voids.

evolution of the signal and typically results in generation of signals that are too strong or too weak, or causes alternations in image contrast. For multi-shot imaging this results in magnitude inconsistencies in k -space;

- iii. Motion of the tissues and body parts often affects magnetic fields used in MRI. The B_0 field within the imaging volume changes significantly if the orientation of

the body part changes but may also be affected by motion of relatively distant body parts due to long-ranging magnetic susceptibility effects.^{13,14} Both transmit and receive B_1 fields may also change depending on the body position or due to the motion of the surface coils themselves.

- iv. Rotations during the k -space data acquisition process in multi-shot imaging produces inconsistencies between the effective readout directions for different k -space lines (see Fig. 6). Even if the true motion during the acquisition is known and compensated for, the resulting k -space is not sampled homogeneously, which may result in ghosting or streaking artifacts.
- v. Cardiac or respiratory motion leads to so-called “physiologic noise” in fMRI. Because the fMRI signal is very small,¹⁵ any other dynamic changes caused by, for example, periodic blood flow, appears as a confounding factor in the analysis. The situation is further exacerbated by the relatively low temporal resolution of typical fMRI acquisitions leading to a spectral aliasing of the physiologic signals.




FIGURE 6: The effect of rotations on the k -space sample locations in object coordinates. In the absence of rotations (a), the

TABLE 1. Common motion artifact mitigation strategies

Motion prevention	Artifact reduction	Motion correction
Training	Faster imaging	Navigators
Distraction	Insensitive sequences	Self-navigated trajectories
Feed and wrap (for babies)	Gradient moment nulling	Prospective correction
Foam restraints	Saturation bands	Retrospective correction
Sedation	Triggering and gating	
Bitebars/head holders	Phase reordering	
Breathhold		

Artifact Mitigation Strategies

As discussed above, there are many types of motion and many possible mechanisms for data corruption resulting from each motion type. It follows, therefore, that many different strategies have been developed to address these challenges. In the following, we categorize these methods into three distinct groups: motion prevention, artifact reduction and motion correction (Table 1).

Motion prevention is the most obvious method of suppressing motion artifacts. If motion can be avoided, then the effects discussed early in the review are prevented and other, more complex, strategies become redundant. Unfortunately, preventing motion is not always practical and so artifact reduction or motion correction strategies must often be used. Artifact reduction is defined in this review as acquisition strategies that reduce artifacts in the resulting image or replace them with those with a less dramatic appearance, when compared with standard Cartesian acquisition methods. Motion correction, on the other hand, normally involves the explicit estimation and compensation of motion. Combinations of all three methods are possible, just as multiple tools may be used together in a complementary manner in the toolbox analogy mentioned above.

Motion Prevention

Prevention is the most commonly attempted approach to avoid motion artifacts in clinical MRI. The main prevention tools available are briefly summarized here.

For young children, training with a mock MRI is useful to avoid bulk motion by reducing anxiety. However, this method is not widely available and is not particularly cost effective. In infants, imaging can be timed after feeding to

Moderate sedation, or in some cases, general anesthesia, is often used^{17,18} when imaging young children. Surprisingly enough, sedation is not always effective, with occurrence of motion artifacts in approximately 12% of cases¹⁷ and an increased prevalence of failures in older children. According to Malviya et al.,¹⁷ general anesthesia is more reliable, as it has a much lower incidence of motion artifacts (approximately 0.7%) but is associated with much greater costs (in part due to the requirement of a presence of an anesthetist), more severe side effects and greater health risks.

In the early days of MRI, bite bars mounted on head coils^{19,20} were deemed an effective means of avoiding head motion. However, due to their cumbersome set up and significant discomfort, these approaches have not found wide acceptance in clinical practice. The same applies to other immobilizing systems, such as a personalized plaster cast head holder.²¹ Vacuum immobilizer cushions, which are available from several suppliers, address the set-up issues, but, in the experience of the authors, are perceived as very uncomfortable after a time period of 10 to 30 min.

In the case of abdomen or cardiac imaging, breath-holding is an obvious, and effective, methods of suppressing motion artifacts due to breathing (see Chapter 12.3 in Bernstein et al.).²² Breath-holding, however, restricts scan duration severely (between 10 and 20 s in clinical settings), which in turn limits the image quality, resolution and coverage. Image quality is further restricted by the quality of breath-hold (e.g., presence of drifts or tremor). To overcome time limitations multiple breath-holds are often used despite limitations associated with breathing hysteresis and limited reproducibility of breath-hold positions.

but new ways of achieving the same goals have appeared. Therefore, for the sake of completeness, we review the current artifact reduction strategies below.

FASTER IMAGING. In analogy to photography, shortening the image acquisition time below the typical time scale of motion is the most straightforward approach to reduce motion artifacts. The wider availability of fast imaging techniques is largely responsible for the improvements seen in MRI with regard to the immunity to motion, rather than the more complicated motion correction techniques discussed below. In the case of spontaneous movements, shorter scan times mean that the patient is less likely to become uncomfortable and move. In the case of cardiac or breathing motion, fast sequences acquire more data within a single period. An early example of fast gradient echo imaging is the FLASH (fast low-angle shot) sequence,²⁴ which uses low-angle excitation and spoils transverse magnetization between TRs. Progress in gradient hardware has resulted in the wide availability of fully-refocused gradient echo sequences²⁵ (so-called balanced steady-state free precession, bSSFP). For the same reason, echo planar imaging (EPI),²⁶ which was initially deemed barely feasible, has became a workhorse in functional neuroimaging. T2-weighted imaging has became clinically feasible due to the adoption of the echo-train principles,²⁷ but still remains too slow for certain applications, such as breathhold abdominal imaging. In such cases a single-shot fast spin-echo (a.k.a. HASTE) is often used as a means of freezing motion at expense of a moderate loss in spatial resolution.

Parallel imaging^{28–30} has revolutionized many MRI applications by shortening acquisition times, typically by a factor of 2 to 3. More recently, these approaches have been extended to acceleration in two encoding dimensions^{31,32} leading to even higher gains for 3D imaging or novel approaches, such as simultaneous multi-slice imaging.³³ The introduction of compressed sensing MRI³⁴ in 2007 has provided a further tool for acceleration of MRI acquisitions and has paved a way to new motion-compensated applications.³⁵

Fast imaging, however, compared with the standard counterpart, frequently represents a compromise between the acquisition time and other confounding issues. Single-shot imaging typically has a limited spatial resolution. EPI suffers from geometric distortions and chemical-shift artifacts. Parallel imaging methods rely on the accuracy of the coil calibration data and may result in new types of artifacts if motion happens during the acquisition of the reference




FIGURE 7: The effect of linear motion on the phase acquired by the spins under the action of the encoding gradients (frequency encoding in this example). A bipolar gradient (a) refocuses the stationary spins at the moment when the net area under the gradient reaches zero. This is because the phase accrual in this case is proportional to a product of the gradient amplitude and the time duration. Contrary to that, the phase acquired by a spin moving along the gradient direction is proportional to the gradient amplitude, duration and the displacement of the spin. Because the displacement under continuous motion is proportional to the time, the phase shows a quadratic dependency on time. Therefore, a bipolar gradient fails to refocus moving spins. Consequently, such spins will be poorly located and a motion artifact will be seen in vascular structures and moving fluids. It is possible to account for the quadratic behavior of the phase by introducing a third gradient lobe, as shown in (b). Here, to keep the polarity of the frequency encoding gradient the sign of the two preceding lobes has been adjusted. The gradient area ratio of 1:2:1 does not induce additional dephasing for the stationary spins and allows for refocusing the spins moving with a constant velocity. Other gradient schemes are possible taking into account gradient amplitude and slew rate limits, but velocity compensation always increases the minimal echo time.

gradient are not. With gradient moment nulling (Fig. 7b), both stationary spins, and spins moving with a constant velocity, are refocused. Mathematically this is equivalent to minimizing both the 0th gradient moment (area under the gradient waveform) and the 1st gradient moment at the echo time. The gradient moment nulling condition is exactly or approximately fulfilled for a range of multi-echo imaging sequences⁴⁰ such as multi-gradient echo or turbo spin echo, but only for even echoes within the echo train. This phenomenon is termed “even echo refocusing” in the literature.

Incorporation of gradient moment nulling almost always slows down the sequence as it requires additional gradient lobes to be added. Currently the majority of scanner vendors provide this feature for one or several encoding dimensions as an option amongst the sequence parameters, typically referred to as “flow compensation”.

MOTION INSENSITIVE SEQUENCES. Imaging methods relying on spiral or radial encoding strategies are often more motion-tolerant than Cartesian sequences, in that the motion artifacts have a more benign appearance in the images.⁴¹ The reason for this is that these strategies typically oversample the center of k -space, which either smears out the artifacts, due to the inherent averaging^{41–44} or may be used for more advanced correction in postprocessing.^{45–49} Furthermore, if the readout starts at or close to the center of k -space, it is easier to achieve gradient moment nulling, which reduces phase errors between the shots.

PHASE REORDERING. The order of acquisition of k -space lines can make a substantial difference to the appearance and severity of motion artifacts when imaging a moving object. Changing the ordering to minimize artifacts is particularly common to prevent artifacts from respiratory motion. An early example of this is respiratory-ordered phase encoding (ROPE),⁵⁰ where neighboring lines in k -space are acquired at points in the respiratory waveform that are also close together. This minimizes side lobe peaks in the point-spread function (PSF), which reduces ghosting. Such techniques are often generally referred to as “respiratory compensation”. A major advantage of respiratory compensation over respiratory gating is a much shorter scan time.

Respiratory compensation requires information about motion caused by the patient breathing. Typically, respiratory bellows are used to obtain a waveform that gives information about the patient’s chest position by means of measurement of the movement of air in the bellows. It is

motion in abdominal imaging, etc.) can generate motion artifacts that overlay the organ of interest.^{52,53} Spatial saturation bands can be used to suppress signal from such moving tissue. To achieve this, one or more spatially selective 90° RF pulses are applied before the actual excitation pulse. The transverse magnetization is then dephased by spoiler gradients. Moving spins from the saturated region are, therefore, prevented from contributing to artifacts in the imaged volume. However, the additional RF pulses lead to an increase in SAR and measurement time – clinically, this results in a decreased number of slices for a given TR.

TRIGGERING AND GATING. Patient motion resulting from physiologic effects such as heartbeat or breathing is generally periodic. This is an advantage for motion correction, as MR data acquisition can be timed to the different points within the cardiac cycle⁵⁴ or breathing position.⁵⁵

There are two general approaches to enforce consistency of the data acquired in presence of (quasi-) periodic motion. Data can either be collected at the same point in the cycle (triggering) or acquired continuously and then reordered retrospectively (gating). Triggering is often easier to implement and it has the advantage of a more precise synchronization with the underlying motion. On the other hand triggering by necessity introduces a delay, which both disrupts the signal steady state and misses some part of the dynamic cycle in applications like cardiac cine imaging. Gating is free of these shortcomings, but it has to sacrifice some scanning efficiency and allow for a certain degree of redundancy to ensure that sufficient data are available for retrospective image reconstruction.

Both triggering and gating rely on an additional signal or several signals correlated to the physiologic motion in question. For cardiac imaging, electrodes are normally placed on the subject’s chest to record the electrocardiogram (ECG). This has the drawback of slower patient workflow and has some safety implications.⁵⁶ At higher field strengths (3 T and above) the quality of the ECG trace may be reduced substantially, so it is, therefore, important to optimize the placement of the ECG electrodes and leads.⁵⁷ Another possibility is to use a pulse oximeter, typically attached to subject’s finger. However, peripheral signals are less sharp and precise and are more suitable for gating because of the delay associated with a limited velocity of the pulse wave in the human body. Respiratory gating and triggering can rely on respiration sensors, such as belts or bel-

separate section below. Another popular approach acquires one or several nonencoded samples following the RF excitation. This is commonly referred to as FID-gating or self-gating.^{58,59}

Generic limitations to both triggering and gating arise from the violation of the two basic assumptions behind these techniques: periodicity of the disturbing motion and stationarity of the imaged object or organ. Therefore, cardiac MR imaging in patients with arrhythmia remains challenging. In pediatrics both breathing and heartbeat may be very irregular, which limits the achievable image quality severely. Also in adults free-breathing exams fail sporadically, due to breathing position drifts and changes in breathing patterns. Imaging of fast nonstationary and irreproducible events, such those observed in cardiac perfusion, is an area of active research (e.g., Schmidt et al.).⁶⁰

Motion Correction

MR NAVIGATORS. MR navigator methods acquire the required position information during an imaging sequence using the MR scanner itself.^{61–63} This is achieved by frequently playing out ultra-short (on the order of milliseconds) navigator sequences during the imaging process. These navigator sequences allow for determination of the object's position by comparing each acquired navigator signal to a reference.

Generally, navigators can be separated into two categories: those that work in k -space and those that work in image-space. They can be further classified in terms of their dimensionality: 1D, 2D, or 3D. Because any additional information encoded by the navigator comes at a cost of increased acquisition and processing times, and increased complexity, navigators are typically simplified to an extent acceptable for the target application. Therefore, 1D and 2D navigators are popular especially in abdominal and cardiac imaging to correct for breathing motion, by tracking the position of the diaphragm.^{64,65} In brain imaging, 2D navigators allow for a correction of rigid body motion in the navigator plane.^{66–69} In DWI even the smallest movements caused by brain pulsation can lead to phase changes due to the strong gradients used for diffusion weighting. Variations of the image phase leads to substantial image artifacts in multishot-DWI. MR-navigators measuring the image phase after diffusion weighting (in 2D) allow the correction of these phase instabilities.^{70,71}

Spectroscopic acquisitions offer even more “empty space” within the sequence, which is sufficient to acquire several navigator echoes, or even specialized shim navigators to correct for shim alterations due to motion.^{75,76} Recently, substantially faster 3D navigators, relying on parallel imaging or multiband excitation,³³ have been presented at various conferences and workshops.

To improve the accuracy of the navigation sequence, and to reduce interference with the imaging process, it is possible to use additional spatial-frequency-tuned markers,⁷⁷ miniature radio-frequency probes,^{78,79} or signals of the endogenous fat.^{80,81} However, many of these methods were so far restricted to 1D⁸⁰ and 2D,^{77,81} and are generally only used in research applications.

An alternative to the navigator methodology is self-navigated methods, which allow the determination of motion from the MR imaging data themselves.^{45–47,49,58} All of these methods are very specific to a particular imaging scheme and are often limited to 2D acquisitions. One popular variant of self-navigated sequences is PROPELLER, which is discussed in detail in the next section. Self-navigation methods can also be extended to 3D, for example in conjunction with a three dimensional radial imaging scheme.⁴⁸

Navigator information can be used either retrospectively to correct the acquired imaging data⁸² or prospectively to adapt the imaging scan to the patient's position. Note that for prospective motion correction, motion parameters need to be extracted from the navigator data in real time. Advantages and disadvantages of both prospective and retrospective approaches are discussed below.

A detailed overview of different navigator approaches can be found in Chapter 12.2 of Bernstein et al.²²

PROPELLER

The acronym “PROPELLER” is an abbreviation for “periodically rotated overlapping parallel lines with enhanced reconstruction”.¹¹ PROPELLER imaging, and its variants, follow a strategy similar to that of radial imaging. However, instead of acquiring the data projection-by-projection, strips of several parallel k -space lines are collected. This can be realized using different readout strategies like RARE,⁸³ EPI,^{84,85} or GRASE.⁸⁴ This strip is rotated around the k -space origin in the subsequent acquisition cycles until a disc of k -space is filled.

This sampling scheme provides substantially reduced motion sensitivity due to the strong oversampling of the k -

detect and correct in plane rotations of the single k -space strips. By cross-correlating the central disc of the corrected strips, through-plane motion can be detected. Such motion cannot be corrected for retrospectively; however, by introducing a factor that puts less weight on corrupted blades, the influence of through-plane motion can be reduced.

The PROPELLER sampling strategy can be combined with undersampling techniques.⁸⁷ It can also be used to introduce diffusion weighting to the RARE readout module,⁸⁸ which was shown to be of interest for DTI at high fields.⁸⁹

The PROPELLER methodology has been shown to be a very useful tool in clinical settings⁹⁰ and is available on the majority of imaging platforms (under different names). Drawbacks of PROPELLER are the increased image acquisition time, due to the strong oversampling around the k -space center, the limitation to 2D imaging and a limited robustness against through-plane motions.

Prospective Correction

Prospective motion correction is an intuitive approach where the relative spatial position and orientation between scanning coordinates and the object of interest is kept constant. For the correction of head motion, a rigid object is assumed, moving in 6 degrees-of-freedom. If the position and orientation of such an object can be measured in real time, the magnetic gradients, RF pulses, and receiver frequency and phase can be adjusted accordingly. Rotations of the object require a rotation of the encoding gradients; translations require a change of transmit and receive frequency and phase.⁹¹

In the case of head motion, head pose can be measured using MR navigators, working either in image space (e.g., PROMO^{92,93}, PACE⁹⁴), or k -space (e.g., cloverleaf, spherical or orbital navigators^{66,67,73}). Alternatively, an external tracking device can be used, including stereo camera systems,⁹⁵ miniature RF probes,^{78,79,96} in-bore camera systems,^{97–99} or ultrasound systems.¹⁰⁰ Navigators need to be compatible with the sequence timing and typically have a low temporal sample rate. On the other hand, external tracking typically requires an MR compatible design and accurate calibration to the scanner coordinate system. Furthermore, the majority of the external approaches require a tracking device to be mounted on the subject, which is a concern for routine applications.

The topic of prospective motion correction has gained popularity in the last few years resulting in numerous new applications including fMRI,^{101,102} DWI,^{103,104} and spectroscopy.^{105–107}

Postprocessing Techniques and Retrospective Correction

While prospective correction methods aim at maintaining data quality during the acquisition, retrospective techniques attempt to improve data consistency afterward, by modifying the collected data or the reconstruction model. This can be done by either including the exact knowledge about the motion during the scan (e.g., using navigator data) or with iterative algorithms (e.g., optimizing image entropy, gradient entropy or other measures of artifacts).^{110–112}

The basic idea behind these methods is to undo the motion-related changes that occurred to the MR data. For rigid body motion these are described according to the Fourier theorems: a translation of the object leads to a phase ramp in the acquired k -space, an object rotation corresponds to a rotation of k -space.¹¹³ While translations are relatively easy to correct by applying a phase change to the acquired data, the correction of rotations requires the use of non-Cartesian reconstruction methods^{114,115} and includes some sophisticated algorithms,^{116–119} which are computationally intensive.

In 3D imaging, arbitrary rigid body motion can be corrected, as long as the acquired signals are only damaged but not lost (e.g., there are no signal dropouts due to intra-voxel dephasing). For 2D imaging these correction methods are limited to in-plane motion, as through-plane motion during the scan results in inconsistencies, which cannot be corrected.

Elastic motion remains a major challenge for both prospective and retrospective correction approaches. Especially in abdominal and cardiac MRI, the complexity of the underlying motion restricts the current concepts to a combination of gating/triggering followed by a correction of affine motion within the gating window.^{60,120}

An alternative approach to retrospectively improve the quality of datasets which are only partially affected by motion is described by Bydder et al.¹²¹ By discarding motion-corrupted data and filling the resulting parts of k -space using parallel imaging techniques, the quality of the reconstructed images could be increased at a cost of SNR and possibly some residual blurring.

By including a general description of motion into the MR signal equation¹²² using measured or estimated displacement fields¹²³ free-breathing cardiac imaging has been demonstrated. Recently this approach has been extended to incorporate compressed sensing methodology.³⁵ Unfortunately such retrospective image reconstruction methods are extremely computationally intensive, which presently

space sampling methods in MRI causes these artifacts to appear differently from scan to scan. This, coupled with a large range of motion types that occur *in vivo*, means that there is no single motion correction tool that can be applied to every motion problem. Instead, there is a toolbox of techniques, where different tools are applicable in particular situations. In some cases, powerful tools are already available; in other cases, tools still need to be developed. Nonetheless, it is likely that developing dedicated tools for specific situations will be what eventually solves the motion problem in MRI, rather than a single approach that solves all problems.

Acknowledgments

Contract grant sponsor: NIH; Contract grant number: 1R01 DA021146; Contract grant number: 2R01 DA021146; Contract grant number: 5R01 EB011654; Contract grant number: 5R21 EB017616; Contract grant sponsor: Alexander von Humboldt Foundation.

References

- Wood ML, Henkelman RM. MR image artifacts from periodic motion. *Med Phys* 1985;12:143–151.
- Van de Walle R, Lemahieu I, Achten E. Magnetic resonance imaging and the reduction of motion artifacts: review of the principles. *Technol Health Care* 1997;5:419–435.
- Le Bihan D, Poupon C, Amadon A, Lethimonnier F. Artifacts and pitfalls in diffusion MRI. *J Magn Reson Imaging* 2006;24:478–488.
- Kiselev VG, Hahn KR, Auer DP. Is the brain cortex a fractal? *Neuroimage* 2003;20:1765–1774.
- Paschal CB, Morris HD. K-space in the clinic. *J Magn Reson Imaging* 2004;19:145–159.
- van Vaals JJ, Brummer ME, Dixon WT, et al. "Keyhole" method for accelerating imaging of contrast agent uptake. *J Magn Reson Imaging* 1993;3:671–675.
- Doyle M, Walsh EG, Blackwell GG, Pohost GM. Block regional interpolation scheme for k-space (BRISK): a rapid cardiac imaging technique. *Magn Reson Med* 1995;33:163–170.
- Korosec FR, Frayne R, Grist TM, Mistretta CA. Time-resolved contrast-enhanced 3D MR angiography. *Magn Reson Med* 1996;36:345–351.
- Lim RP, Shapiro M, Wang EY, et al. 3D time-resolved MR angiography (MRA) of the carotid arteries with time-resolved imaging with stochastic trajectories: comparison with 3D contrast-enhanced Bolus-Chase MRA and 3D time-of-flight MRA. *AJNR Am J Neuroradiol* 2008;29:1847–1854.
- Ranger AV, Block WF, Toropov Y, Grist TM, Mistretta CA. Time-
- Birn RM, Bandettini PA, Cox RW, Jesmanowicz A, Shaker R. Magnetic field changes in the human brain due to swallowing or speaking. *Magn Reson Med* 1998;40:55–60.
- Van de Moortele PF, Pfeuffer J, Glover GH, Ugurbil K, Hu X. Respiration-induced B0 fluctuations and their spatial distribution in the human brain at 7 Tesla. *Magn Reson Med* 2002;47:888–895.
- Kruger G, Glover GH. Physiological noise in oxygenation-sensitive magnetic resonance imaging. *Magn Reson Med* 2001;46:631–637.
- Windram J, Grosse-Wortmann L, Shariat M, Greer ML, Crawford MW, Yoo SJ. Cardiovascular MRI without sedation or general anesthesia using a feed-and-sleep technique in neonates and infants. *Pediatr Radiol* 2012;42:183–187.
- Malviya S, Voepel-Lewis T, Eldevik OP, Rockwell DT, Wong JH, Tait AR. Sedation and general anaesthesia in children undergoing MRI and CT: adverse events and outcomes. *Br J Anaesth* 2000;84:743–748.
- Hubbard AM, Markowitz RI, Kimmel B, Kroger M, Bartko MB. Sedation for pediatric patients undergoing CT and MRI. *J Comput Assist Tomogr* 1992;16:3–6.
- Bettinardi V, Scardaoni R, Gilardi MC, et al. Head holder for PET, CT, and MR studies. *J Comput Assist Tomogr* 1991;15:886–892.
- Menon V, Lim KO, Anderson JH, Johnson J, Pfefferbaum A. Design and efficacy of a head-coil bite bar for reducing movement-related artifacts during functional MRI scanning. *Behav Res Methods Instrum Comput* 1997;29:589–594.
- Edward V, Windischberger C, Cunnington R, et al. Quantification of fMRI artifact reduction by a novel plaster cast head holder. *Hum Brain Mapp* 2000;11:207–213.
- Bernstein MA, King KF, Zhou XJ. *Handbook of MRI Pulse Sequences*. Boston: Elsevier Academic Press; 2004. 1017 p.
- Bellon EM, Haacke EM, Coleman PE, Sacco DC, Steiger DA, Gangarosa RE. MR artifacts: a review. *AJR Am J Roentgenol* 1986;147:1271–1281.
- Haase A, Frahm J, Matthaei D, Hanicke W, Merboldt KD. FLASH imaging. Rapid NMR imaging using low flip-angle pulses. *J Magn Reson* 1986;67:258–266.
- Scheffler K, Lehnhardt S. Principles and applications of balanced SSFP techniques. *Eur Radiol* 2003;13:2409–2418.
- Mansfield P. Multi-planar image-formation using NMR spin echoes. *J Phys C Solid State Phys* 1977;10:L55–L58.
- Hennig J, Nauerth A, Friedburg H. RARE imaging: a fast imaging method for clinical MR. *Magn Reson Med* 1986;3:823–833.
- Sodickson DK, Manning WJ. Simultaneous acquisition of spatial harmonics (SMASH): fast imaging with radiofrequency coil arrays. *Magn Reson Med* 1997;38:591–603.
- Pruessmann KP, Weiger M, Scheidegger MB, Boesiger P. SENSE: sensitivity encoding for fast MRI. *Magn Reson Med* 1999;42:952–962.
- Griswold MA, Jakob PM, Heidemann RM, et al. Generalized autocalibrating partially parallel acquisitions (GRAPPA). *Magn Reson Med* 2002;47:1202–1210.
- Breuer FA, Blaimer M, Heidemann RM, Mueller MF, Griswold MA, Jakob PM. Controlled aliasing in parallel imaging results in higher acceleration (CAPIRINHA) for multi-slice imaging. *Magn Reson Med* 2005;53:684–691.
- Breuer FA, Blaimer M, Mueller MF, et al. Controlled aliasing in volumetric parallel imaging (2D CAPIRINHA). *Magn Reson Med* 2006;55:549–556.

36. Guo JY, Kholmovski EG, Zhang L, Parker DL. Evaluation of motion effects on parallel MR imaging with precalibration. *Magn Reson Imaging* 2007;25:1130–1137.
37. Larkman DJ, Nunes RG. Parallel magnetic resonance imaging. *Phys Med Biol* 2007;52:R15–R55.
38. Peeters JM, Fuderer M. SENSE with improved tolerance to inaccuracies in coil sensitivity maps. *Magn Reson Med* 2013;69:1665–1669.
39. Zhang T, Chowdhury S, Lustig M, et al. Clinical performance of contrast enhanced abdominal pediatric MRI with fast combined parallel imaging compressed sensing reconstruction. *J Magn Reson Imaging* 2014;40:13–25.
40. Katz J, Peshock RM, Malloy CR, Schaefer S, Parkey RW. Even-echo rephasing and constant velocity flow. *Magn Reson Med* 1987;4:422–430.
41. Glover GH, Pauly JM. Projection reconstruction techniques for reduction of motion effects in MRI. *Magn Reson Med* 1992;28:275–289.
42. Jackson JL, Nishimura DG, Macovski A. Twisting radial lines with application to robust magnetic resonance imaging of irregular flow. *Magn Reson Med* 1992;25:128–139.
43. Sarty GE. Single trajectory radial (STAR) imaging. *Magn Reson Med* 2004;51:445–451.
44. Pipe JG. An optimized center-out k-space trajectory for multishot MRI: comparison with spiral and projection reconstruction. *Magn Reson Med* 1999;42:714–720.
45. Glover GH, Lai S. Self-navigated spiral fMRI: interleaved versus single-shot. *Magn Reson Med* 1998;39:361–368.
46. Welch EB, Rossman PJ, Felmlee JP, Manduca A. Self-navigated motion correction using moments of spatial projections in radial MRI. *Magn Reson Med* 2004;52:337–345.
47. Liu C, Bammer R, Kim DH, Moseley ME. Self-navigated interleaved spiral (SNAILS): application to high-resolution diffusion tensor imaging. *Magn Reson Med* 2004;52:1388–1396.
48. Anderson AG III, Velikina J, Block W, Wieben O, Samsonov A. Adaptive retrospective correction of motion artifacts in cranial MRI with multicoil three-dimensional radial acquisitions. *Magn Reson Med* 2013;69:1094–1103.
49. Vaillant G, Prieto C, Kolbitsch C, Penney G, Schaeffter T. Retrospective rigid motion correction in k-space for segmented radial MRI. *IEEE Trans Med Imaging* 2013 doi: 10.1109/TMI.2013.2268898.
50. Bailes DR, Gilderdale DJ, Bydder GM, Collins AG, Firmin DN. Respiratory ordered phase encoding (ROPE): a method for reducing respiratory motion artefacts in MR imaging. *J Comput Assist Tomogr* 1985;9:835–838.
51. Markl M, Harloff A, Bley TA, et al. Time-resolved 3D MR velocity mapping at 3T: improved navigator-gated assessment of vascular anatomy and blood flow. *J Magn Reson Imaging* 2007;25:824–831.
52. Edelman RR, Buxton RB, Brady TJ. Rapid MR imaging. *Magn Reson Annu* 1988;189–216.
53. Felmlee JP, Ehman RL, Riederer SJ, Korin HW. Adaptive motion compensation in MRI: accuracy of motion measurement. *Magn Reson Med* 1991;18:207–213.
54. Lanzer P, Barta C, Botvinick EH, Wiesendanger HU, Modin G, Higgins CB. ECG-synchronized cardiac MR imaging: method and evaluation. *Radiology* 1985;155:681–686.
55. Ehman RL, McNamara MT, Pallack M, Hricak H, Higgins CB. Magnetic resonance imaging with respiratory gating: techniques and advan-
59. Dwyerfeldt P, Deshpande VS, Kober T, Krueger G, Saloner D. Reduction of motion artifacts in carotid MRI using free-induction decay navigators. *J Magn Reson Imaging* 2014;40:214–220.
60. Schmidt JF, Wissmann L, Manka R, Kozerke S. Iterative k-t principal component analysis with nonrigid motion correction for dynamic three-dimensional cardiac perfusion imaging. *Magn Reson Med* 2014;72:68–79.
61. Ehman RL, Felmlee JP. Adaptive technique for high-definition MR imaging of moving structures. *Radiology* 1989;173:255–263.
62. Korin HW, Felmlee JP, Ehman RL, Riederer SJ. Adaptive technique for three-dimensional MR imaging of moving structures. *Radiology* 1990;177:217–221.
63. Atkinson D, Hill DLG. Reconstruction after rotational motion. *Magn Reson Med* 2003;49:183–187.
64. Uribe S, Muthurangu V, Boubertakh R, et al. Whole heart cine MRI using real time respiratory self gating. *Magn Reson Med* 2007;57:606–613.
65. Wang Y, Riederer SJ, Ehman RL. Respiratory motion of the heart: kinematics and the implications for the spatial resolution in coronary imaging. *Magn Reson Med* 1995;33:713–719.
66. Fu ZW, Wang Y, Grimm RC, et al. Orbital navigator echoes for motion measurements in magnetic resonance imaging. *Magn Reson Med* 1995;34:746–753.
67. Welch EB, Manduca A, Grimm RC, Ward HA, Jack CR JR. Spherical navigator echoes for full 3D rigid body motion measurement in MRI. *Magn Reson Med* 2002;47:32–41.
68. Kadah YM, Abaza AA, Fahmy AS, Youssef ABM, Heberlein K, Hu XPP. Floating navigator echo (FNAV) for in-plane 2D translational motion estimation. *Magn Reson Med* 2004;51:403–407.
69. Bydder M, Atkinson D, Larkman DJ, Hill DLG, Hajnal JV. SMASH navigators. *Magn Reson Med* 2003;49:493–500.
70. Norris DG, Driesel W. Online motion correction for diffusion-weighted imaging using navigator echoes: application to RARE imaging without sensitivity loss. *Magn Reson Med* 2001;45:729–733.
71. Butts K, de Crespigny A, Pauly JM, Moseley M. Diffusion weighted interleaved echo planar imaging with a pair of orthogonal navigator echoes. *Magn Reson Med* 1996;35:763–770.
72. Costa AF, Petrie DW, Yen YF, Drangova M. Using the axis of rotation of polar navigator echoes to rapidly measure 3D rigid-body motion. *Magn Reson Med* 2005;53:150–158.
73. van der Kouwe AJW, Benner T, Dale AM. Real-time rigid body motion correction and shimming using cloverleaf navigators. *Magn Reson Med* 2006;56:1019–1032.
74. Tisdall MD, Hess AT, Reuter M, Meintjes EM, Fischl B, van der Kouwe AJ. Volumetric navigators for prospective motion correction and selective reacquisition in neuroanatomical MRI. *Magn Reson Med* 2012;68:389–399.
75. Hess AT, Tisdall MD, Andronesi OC, Meintjes EM, van der Kouwe AJ. Real-time motion and B0 corrected single voxel spectroscopy using volumetric navigators. *Magn Reson Med* 2011;66:314–323.
76. Keating B, Ernst T. Real-time dynamic frequency and shim correction for single-voxel magnetic resonance spectroscopy. *Magn Reson Med* 2012;68:1339–1345.

80. Nguyen TD, Nuval A, Mulukutla S, Wang Y. Direct monitoring of coronary artery motion with cardiac fat navigator echoes. *Magn Reson Med* 2003;50:235–241.
81. Kawaji K, Spincemaille P, Nguyen TD, et al. Direct coronary motion extraction from a 2D fat image navigator for prospectively gated coronary MR angiography. *Magn Reson Med* 2013 doi: 10.1002/mrm.24698.
82. Bammer R, Aksoy M, Liu C. Augmented generalized SENSE reconstruction to correct for rigid body motion. *Magn Reson Med* 2007;57:90–102.
83. Pipe JG. Multishot diffusion weighted FSE with PROPELLER. In: Proceedings of the 9th Annual Scientific Meeting of ISMRM, Glasgow, 2001. (abstract 289).
84. Pipe JG, Zwart N. Turboprop: improved PROPELLER imaging. *Magn Reson Med* 2006;55:380–385.
85. Skare S, Newbould RD, Clayton DB, Bammer R. Propeller EPI in the other direction. *Magn Reson Med* 2006;55:1298–1307.
86. Forbes KP, Pipe JG, Karis JP, Farthing V, Heiserman JE. Brain imaging in the unsedated pediatric patient: comparison of periodically rotated overlapping parallel lines with enhanced reconstruction and single-shot fast spin-echo sequences. *AJNR Am J Neuroradiol* 2003;24:796–798.
87. Arfanakis K, Tamhane AA, Pipe JG, Anastasio MA. K-space undersampling in PROPELLER imaging. *Magn Reson Med* 2005;53:675–683.
88. Pipe JG, Farthing VG, Forbes KP. Multishot diffusion-weighted FSE using PROPELLER MRI. *Magn Reson Med* 2002;47:42–52.
89. Wang FN, Huang TY, Lin FH, et al. PROPELLER EPI: an MRI technique suitable for diffusion tensor imaging at high field strength with reduced geometric distortions. *Magn Reson Med* 2005;54:1232–1240.
90. Forbes KPN, Pipe JG, Bird CR, Heiserman JE. PROPELLER MRI: clinical testing of a novel technique for quantification and compensation of head motion. *J Magn Reson Imaging* 2001;14:215–222.
91. Nehrke K, Börnert P. Prospective correction of affine motion for arbitrary MR sequences on a clinical scanner. *Magn Reson Med* 2005;54:1130–1138.
92. White N, Roddey C, Shankaranarayanan A, et al. PROMO: real-time prospective motion correction in MRI using image-based tracking. *Magn Reson Med* 2010;63:91–105.
93. Brown TT, Kuperman JM, Erhart M, et al. Prospective motion correction of high-resolution magnetic resonance imaging data in children. *Neuroimage* 2010;53:139–145.
94. Thesen S, Heid O, Mueller E, Schad LR. Prospective acquisition correction for head motion with image-based tracking for real-time fMRI. *Magn Reson Med* 2000;44:457–463.
95. Zaitsev M, Dold C, Sakas G, Hennig J, Speck O. Magnetic resonance imaging of freely moving objects: prospective real-time motion correction using an external optical motion tracking system. *Neuroimage* 2006;31:1038–1050.
96. Ooi MB, Krueger S, Thomas WJ, Swaminathan SV, Brown TR. Prospective real-time correction for arbitrary head motion using active markers. *Magn Reson Med* 2009;62:943–954.
97. Qin L, Gelderen P, Zwart J, Jin F, Tao Y, Duyn J. Head movement correction for MRI with a single camera. In: Proceedings of the 16th Annual Scientific Meeting of ISMRM, Toronto, Ontario, 2008. (abstract 1467).
98. Maclarens J, Armstrong BS, Barrows RT, et al. Measurement and correction of microscopic head motion during magnetic resonance imaging of the brain. *PLoS One* 2012;7:e48088.
99. Schulz J, Siegert T, Reimer E, et al. An embedded optical tracking sys-
102. Schulz J, Siegert T, Bazin PL, et al. Prospective slice-by-slice motion correction reduces false positive activations in fMRI with task-correlated motion. *Neuroimage* 2014;84:124–132.
103. Gumus K, Keating B, Poser BA, et al. Prevention of motion-induced signal loss in diffusion-weighted echo-planar imaging by dynamic restoration of gradient moments. *Magn Reson Med* 2014;71:2006–2013.
104. Herbst M, Maclarens J, Weigel M, Korvink J, Hennig J, Zaitsev M. Prospective motion correction with continuous gradient updates in diffusion weighted imaging. *Magn Reson Med* 2012;67:326–338.
105. Keating B, Deng W, Roddey JC, et al. Prospective motion correction for single-voxel 1H MR spectroscopy. *Magn Reson Med* 2010;64:672–679.
106. Zaitsev M, Speck O, Hennig J, Buchert M. Single-voxel MRS with prospective motion correction and retrospective frequency correction. *NMR Biomed* 2010;23:325–332.
107. Lange T, Maclarens J, Buechert M, Zaitsev M. Spectroscopic imaging with prospective motion correction and retrospective phase correction. *Magn Reson Med* 2012;67:1506–1514.
108. Boegle R, Maclarens J, Zaitsev M. Combining prospective motion correction and distortion correction for EPI: towards a comprehensive correction of motion and susceptibility-induced artifacts. *Magn Reson Mater Phys* 2010;23:263–273.
109. Maclarens J, Herbst M, Speck O, Zaitsev M. Prospective motion correction in brain imaging: a review. *Magn Reson Med* 2013;69:621–636.
110. Atkinson D, Hill DLG, Stoyle PNR, et al. Automatic compensation of motion artifacts in MRI. *Magn Reson Med* 1999;41:163–170.
111. McGee KP, Felmlee JP, Jack CR Jr, Manduca A, Riederer SJ, Ehman RL. Autocorrection of three-dimensional time-of-flight MR angiography of the Circle of Willis. *AJR Am J Roentgenol* 2001;176:513–518.
112. Loktyushin A, Nickisch H, Pohmann R, Scholkopf B. Blind retrospective motion correction of MR images. *Magn Reson Med* 2013;70:1608–1618.
113. Lauzon ML, Rutt BK. Generalized K-space analysis and correction of motion effects in MR imaging. *Magn Reson Med* 1993;30:438–446.
114. Beatty PJ, Nishimura DG, Pauly JM. Rapid gridding reconstruction with a minimal oversampling ratio. *IEEE Trans Med Imaging* 2005;24:799–808.
115. Pruessmann KP, Weiger M, Bornert P, Boesiger P. Advances in sensitivity encoding with arbitrary k-space trajectories. *Magn Reson Med* 2001;46:638–651.
116. Atkinson D, Hill DL. Reconstruction after rotational motion. *Magn Reson Med* 2003;49:183–187.
117. Marxen M, Marmurek J, Baker N, Graham SJ. Correcting magnetic resonance k-space data for in-plane motion using an optical position tracking system. *Med Phys* 2009;36:5580–5585.
118. Leung G, Plewes DB. Retrospective motion compensation using variable-density spiral trajectories. *J Magn Reson Imaging* 2005;22:373–380.
119. Aksoy M, Forman C, Straka M, Cukur T, Horngger J, Bammer R. Hybrid prospective and retrospective head motion correction to mitigate cross-calibration errors. *Magn Reson Med* 2012;67:1237–1251.
120. Schmidt JF, Buehrer M, Boesiger P, Kozerke S. Nonrigid retrospective respiratory motion correction in whole-heart coronary MRA. *Magn Reson Med* 2011;66:1541–1549.
121. Bydder M, Larkman DJ, Hajnal JV. Detection and elimination of