

# Development of Analysis Techniques for dynamic magnetic resonance imaging

– Master Thesis –

to be awarded

Master of Science in Medical Photonics

submitted by

Aayush Nepal

April 23, 2024

Aayush Nepal  
aayush.nepal@uni-jena.de  
Student ID: 198683  
Address:  
Schützengasse 2  
99423 Weimar

Supervisor

1st: Prof. Dr. rer. nat. med. habil. Jürgen Reichenbach  
2nd: Dr. rer. nat. Martin Krämer

Friedrich-Schiller-Universität Jena  
Medizinischen Fakultät  
Institut für Diagnostische und Interventionelle Radiologie  
Philosophenweg 3  
07743 Jena

## **Abstract**

Short summary of your thesis (max. 250 words) ...

## **Acknowledgements**

If you want to thank anyone (optional) . . .

# Contents

<b>List of Figures</b>	<b>v</b>
<b>List of Tables</b>	<b>vi</b>
<b>List of Acronyms</b>	<b>vii</b>
<b>1 Introduction</b>	<b>1</b>
1.1 The Knee . . . . .	1
1.2 Dynamic MRI . . . . .	4
1.2.1 CINE imaging . . . . .	5
1.2.2 Gating . . . . .	5
1.2.3 Knee loading device . . . . .	6
1.2.4 Radial golden-angle acquisition . . . . .	7
1.2.5 Gradient Echo FLASH sequence . . . . .	10
1.3 Research Necessity . . . . .	12
1.4 Thesis Structure . . . . .	12
<b>2 Methodology</b>	<b>12</b>
2.1 Data Collection Methods . . . . .	13
2.1.1 Procedure Details . . . . .	13
2.1.2 Sequence Parameters and Reconstruction . . . . .	13
2.2 Data Analysis . . . . .	13
2.2.1 Segmentation . . . . .	13
2.3 Validity and Reliability . . . . .	14
<b>3 Results</b>	<b>14</b>
3.1 Segmentation . . . . .	14
3.2 Parameter Extraction . . . . .	14
<b>4 Discussion</b>	<b>14</b>
4.1 Technique Evaluation . . . . .	14
4.2 Biomechanical Insights . . . . .	14
4.3 Comparison with Existing Methods . . . . .	14

<b>5</b>	<b>Conclusion</b>	<b>15</b>
5.1	Summary of Contributions . . . . .	15
5.2	Limitations and Challenges . . . . .	15
5.3	Future Work . . . . .	15
	<b>References</b>	<b>16</b>
<b>A</b>	<b>Appendix</b>	<b>19</b>

## List of Figures

1	Anterior (flexed,top) and posterior (extended, bottom) view of the right knee with their articular surfaces labeled. (Netter 2023, p.519) .	2
2	patellar surface . . . . .	3
3	The knee joint motion in three dimensions, described using six independent variables (degrees of freedom) (Standring and Gray 2021, p. 1412). . . . .	4
4	The MR330 controller (left) and MR338 optical fiber position sensor system (right), showcasing the connection from the controller to the sensor via duplex LC multimode fiber optic cabling. . . . .	6
5	A 3D rendering of the knee motion/loading device (left) and two photographs of the device in a Siemens Magnetom Prisma Fit MRI scanner (right). The top right image (rear view) demonstrates how the distal part of the device sits atop the rails within the scanner bore, which are normally used to guide and support the patient table. The bottom right image (front, oblique view) demonstrates an individual positioned in the device with a flexible coil around the index knee, the thigh secured between the proximal pillow blocks and the ankle fastened to the leg support, as well as the coil plug setup. . . . .	7
6	Generic radial k-space sampling (Brown et al. 2014, p.306) . . . . .	8
7	Radial k-space sampling strategies. (a) Fixed angular increment, showing potential for uneven k-space coverage. (b) Golden-angle sampling, ensuring uniform distribution across k-space (Winkelmann et al. 2007). . . . .	9
8	<b>Example of k-space and reconstructed image.</b> Left: K-space data from a specific angle during the knee flexion cycle. Right: Corresponding reconstructed MRI image, illustrating the detailed visualization achievable through golden-angle sampling. . . . .	10
9	Simplified diagram of the GRE sequence that depicts the refocusing gradient applied in the frequency encoding direction (Berry and Bulpitt 2009, p.233). . . . .	11

## List of Tables

# List of Acronyms

FSU Jena    Friedrich-Schiller-Universität Jena



# 1 Introduction

## 1.1 The Knee

The knee, a pivotal structure in the human body, plays an essential role in locomotion. Additionally, it functions to control the center of body mass and posture in the activities of daily living by facilitating a wide range of movements. As such, the knee's health and integrity are vital for overall quality of life, influencing everything from basic mobility to participation in complex sports. In a comprehensive survey across 15 European countries and Israel, it was found that knee pain was the third most commonly reported location of chronic pain, highlighting the significant concern it poses to public health (Breivik et al. 2006). Furthermore, arthritis/osteoarthritis (OA) was identified as the most common cause of this pain. And this situation has not improved with time. In Germany for instance, a recent retrospective study has found that the number of patients with OA is steadily rising (Obermüller et al. 2024). Part of the reason knee-related issues are so prevalent and impactful is the inherent complexity of the knee joint itself. As a hub of various anatomical structures working in unison, the knee supports a range of movements and bears significant loads, making it susceptible to a variety of injuries and conditions. Understanding the anatomy of the knee is the first step in tackling this problem.

### Anatomy and Function

The knee joint is the largest and the most superficial synovial joint in the body. (Dalley, Agur, and Moore 2023). It comprises three primary articulations: two femorotibial articulations (see Figure 1) and one femoropatellar articulation. These articulations, illustrated in figures 1 and 2, are defined by their complexity and incongruence, features that are crucial for the joint's biomechanical performance.




Figure 1: Anterior (flexed,top) and posterior (extended, bottom) view of the right knee with their articular surfaces labeled. (Netter 2023, p.519)




Figure 2: Right joint opened, knee slightly in flexion, with the patellar articulation labelled. (Netter 2023, p.517)

This design allows the knee to manage a wide array of movements and bear significant loads. Figure 3 illustrates the knee's capacity for multidimensional movement, highlighting the joint's sophisticated structural design that enables this versatility. However, this inherent design also renders the knee vulnerable to a range of forces (Standring and Gray 2021).

Such mechanical forces, influenced by activities and body mass index (BMI), are significant causes of OA and represent one of the most modifiable risk factors (Heidari 2011). Furthermore, abnormal joint loading is considered a key mechanical driver of osteochondral changes thought to contribute to the initiation and progression of knee OA (Coburn et al. 2023).




Figure 3: The knee joint motion in three dimensions, described using six independent variables (degrees of freedom) (Standring and Gray 2021, p. 1412).

The relationship between biomechanical stresses and knee health highlights the need for diagnostic tools that fulfill two key criteria: they must capture the knee’s dynamic behavior during motion and assess the joint under varying loads. Dynamic magnetic resonance imaging (MRI) excels in meeting these requirements as one of the most advanced imaging modalities.

## 1.2 Dynamic MRI

In a broad sense, dynamic MRI is an umbrella term encompassing various MRI techniques designed to capture and visualize physiological processes and motion over time. This dynamic aspect of MRI is crucial when studying systems or structures

where motion is inherent, such as blood flow, tissue perfusion, or cardiac movement. For the purposes of this project, the focus of "dynamic MRI" narrows down to capturing the bulk movement of the knee joint undergoing active flexion extension cycles inside the scanner. Compared to traditional static MRI scans, which are highly susceptible to motion artifacts compared to other imaging modalities (Zaitsev, Maclaren, and Herbst 2015), dynamic MRI techniques not only accommodates motion, but can even leverage it to offer comprehensive insights into the functional and biomechanical properties of the concerned structures. Among these techniques, CINE imaging, particularly renowned in cardiac MRI, is considered the gold standard for evaluating cardiac function (Menchón-Lara et al. 2019). This high regard in cardiac MRI illustrates the method's precision and adaptability, traits we leverage for knee imaging.

### 1.2.1 CINE imaging

CINE is derived from the word cinematography, to create a movie. In the context of dynamic knee MRI, this technique can be used to reconstruct a "movie" of a single flexion-extension cycle. This visualization is achieved by aggregating multiple partial datasets acquired over various cycles. The knee's movement cycle is segmented into distinct stages, each corresponding to a specific angle of flexion or extension. For each stage, k-space — the Fourier transform space from which MR images are reconstructed — is incrementally sampled. This data is gathered across multiple repetitions of the movement cycle, ensuring comprehensive coverage of k-space and thus, high-resolution imaging of each movement phase. This technique works robustly only if the cycles are sufficiently similar (Curtis and Cheng 2022). Achieving this consistency in dynamic MRI of the knee, where natural variations in movement can be significant, requires precise synchronization of the imaging process.

### 1.2.2 Gating

This synchronization, known as gating, aligns image acquisition with specific, repeatable points in the knee's movement cycle. By ensuring each captured image corresponds accurately to a consistent phase of motion, it minimizes the discrepancies that can arise from cycle-to-cycle variations, enhancing the reliability of biomechanical analyses. Although there are two types of gating, namely prospective and retrospective, it is the latter that is the standard for CINE imaging as it provides a more consistent image quality (Edelman 2005, p. 102). In retrospective

gating, the data is acquired continuously throughout the motion cycle. The data is then reordered post-acquisition into a coherent sequence that shows the knee at different positions along the motion cycle. This necessitates external information that captures the precise moment and position of the knee throughout its range of movement beyond just the raw data acquired during the MRI scan. A previously reported novel knee loading device is employed to address this, which comes with an optical sensor attachment that provides this vital information (Brisson et al. 2022).

### 1.2.3 Knee loading device

Figure 5 depicts this MR safe device that was used. This device allowed for a range of motion of approximately 25 to 45 degrees (subject-dependent), enabling subjects to perform knee flexion and extension cycles under both loaded and unloaded conditions. For loading, the device was equipped with compartments for weight plates and sandbags, providing a physiological load of 10 to 12 kilograms. Central to this device’s functionality is an optical fiber position sensor (MR338-Y10C10, Micronor, 155 Camarillo, CA, USA), which precisely measures the absolute angle from  $0^\circ$  to  $360^\circ$  with a resolution of  $0.025^\circ$  (Rickenbach 2013). This measurement capability is critical for synchronizing the knee’s movement with MRI data reconstruction.




Figure 4: The MR330 controller (left) and MR338 optical fiber position sensor system (right), showcasing the connection from the controller to the sensor via duplex LC multimode fiber optic cabling.

As shown in Figure 4, the MR330 controller interfaces with the MRI-safe MR338 position sensor to facilitate precise measurement within the MRI environment, accommodating up to 300 meters of fiber optic cabling. The controller is connected to a laptop placed outside the scanner while the sensor is attached to the device within. To enhance signal acquisition and the clarity of imaging, two flexible coils (cite the coils here, perhaps also show a picture) were positioned at key anatomical



locations: one at the distal femur and another at the proximal tibia, as specified in the MRI protocol.




Figure 5: A 3D rendering of the knee motion/loading device (left) and two photographs of the device in a Siemens Magnetom Prisma Fit MRI scanner (right). The top right image (rear view) demonstrates how the distal part of the device sits atop the rails within the scanner bore, which are normally used to guide and support the patient table. The bottom right image (front, oblique view) demonstrates an individual positioned in the device with a flexible coil around the index knee, the thigh secured between the proximal pillow blocks and the ankle fastened to the leg support, as well as the coil plug setup.

To further enhance the quality and precision of our imaging, and to maximize the potential for precise image reconstruction, a previously reported radial golden-angle gradient-echo FLASH sequence was used, which is more robust against motion artifacts than Cartesian sequences (Aleksiev et al. 2022). The following sections will detail the key components of this technique — radial golden-angle acquisition and the gradient-echo FLASH sequence.

#### 1.2.4 Radial golden-angle acquisition

As the name suggests, radial golden-angle acquisition specifically modifies how k-space is sampled. Unlike traditional static MRI, which typically employs Cartesian sampling, this method utilizes radial sampling. Here, the data points are collected along radial lines spreading out from the center of k-space, resembling spokes on a

wheel. Figure 6 depicts a generic radial sampling scheme. The separation between any two adjacent circles defines the separation  $\Delta k_r$  in the radial sampling direction, whereas, the angular separation between any two successive angular lines in k-space (such as between the two example lines shown in the figure) defines  $\Delta\theta$ . The two quantities  $\Delta k_r$  and  $\Delta\theta$  are constrained by the Nyquist sampling criterion. (Brown et al. 2014)




Figure 6: Generic radial k-space sampling (Brown et al. 2014, p.306)

The 'golden-angle' strategy, which we employ, optimizes this approach by spacing the radial lines at an angle of approximately 111.25 degrees. This specific angle helps in evenly covering the k-space without overlapping, ensuring that each new image frame provides unique information, thus enhancing image quality and temporal resolution.






Figure 7: Radial k-space sampling strategies. (a) Fixed angular increment, showing potential for uneven k-space coverage. (b) Golden-angle sampling, ensuring uniform distribution across k-space (Winkelmann et al. 2007).

Figure 7 provides a visual comparison between traditional fixed increment and the golden-angle methods in k-space sampling. It highlights how each approach affects the distribution of sampling lines across k-space. In the figure, part (a) shows radial sampling using a fixed angular increment. This traditional method can lead to gaps or overlaps in data collection, depending on the number of radial profiles used. Part (b) demonstrates the golden-angle method, where each new radial line is placed at an increment of approximately 111.25 degrees. This approach allows for a more uniform distribution of sampling lines across the k-space, enhancing image quality by preventing gaps and reducing redundancy in data collection. Figure blank shows the k-space of a specific frame during the knee flexion cycle and the corresponding reconstructed MRI image.




Figure 8: **Example of k-space and reconstructed image.** Left: K-space data from a specific angle during the knee flexion cycle. Right: Corresponding reconstructed MRI image, illustrating the detailed visualization achievable through golden-angle sampling.

### 1.2.5 Gradient Echo FLASH sequence

Gradient echo (GRE) is a prime candidate for dynamic MRI because it is a fast scanning process. There are two major factors that contribute to its speed. First, unlike the spin echo, we do not need to wait for a long time for the longitudinal magnetization component  $M_z$  to sufficiently recover for another repetition. This is because in a GRE sequence, a smaller flip angle is used. Secondly, we do not need to wait for the spins to refocus after the application of a  $180^\circ$  rf pulse, because there is no such pulse applied. Instead, the refocusing is done here by using magnetic field gradients of opposite polarity. The scheme of GRE sequence is depicted in Figure 9. The notation is as follows:  $\alpha^\circ$ : flip angle (less than  $90^\circ$ ), RF: radio frequency pulse, SS: slice selecting gradient, PE: phase encoding, FE: frequency encoding, TE: echo time and TR: repetition time. Note that in the frequency encoding direction, a negative dephasing lobe is followed by a positive gradient that brings the spins back into phase to generate the signal.




Figure 9: Simplified diagram of the GRE sequence that depicts the refocusing gradient applied in the frequency encoding direction (Berry and Bulpitt 2009, p.233).

After a series of rf pulses are delivered, we reach a condition in gradient echo sequences called the **steady state**. In this state, the magnitude of the longitudinal magnetization component  $M_z$  is the same at the end of every TR. Similarly, we can also have a residual transverse magnetization  $M_{xy}$  at the start of every repetition cycle. If we deliberately dephase this residual transverse magnetization, the sequence is described as incoherent or spoiled gradient echo sequences.

FLASH is a type of spoiled GRE. It is a naming scheme adopted by Siemens that spells Fast Low-Angle Shot. Other vendors have their own naming schemes for such a sequence. General Electric calls it Spoiled Gradient Echo (SPGR) and Philips calls it by the name  $T_1$  fast field echo (T1-FFE). (Bernstein, King, and Zhou 2004, p.583). The introduction of the FLASH technique in the 1980s, known for its rapid acquisition capabilities, marked a significant advancement in MRI technology. It allowed for clearer imaging of dynamic processes and reduced artifacts associated with movement. The original paper on FLASH highlights the technique’s potential for ‘recording NMR movies’ to visualize physiological changes, a capability now fully integrated into contemporary MRI practices.” (Haase et al. 1986)

### 1.3 Research Necessity

With the advanced capabilities to capture high-resolution CINE MRI images of the knee during flexion-extension cycles under both loaded and unloaded conditions, we are now faced with the challenge of extracting meaningful biomechanical insights from this data. The first critical step in any such analysis typically involves segmenting the key anatomical structures within the images—namely, the femur and tibia. This process is crucial because accurate segmentation directly influences the reliability of subsequent biomechanical measurements.

Manual segmentation, while simple, is exceedingly time-consuming and susceptible to human error, highlighting the need for more efficient methodologies. In response, this thesis proposes the development of a semi-automated segmentation pipeline. This approach aims to reduce the manual effort required while maintaining high accuracy, thereby accelerating the analysis process and reducing potential biases introduced by manual methods.

Once segmentation is achieved, the next step involves measuring a biomechanical parameter—specifically, the distance between chosen anatomical landmarks on the femur and tibia. This measurement serves as an indirect indicator of joint space, which is crucial for assessing knee health. Derived from 2D images captured in the sagittal plane throughout the knee’s flexion-extension cycle, this parameter allows for an analysis of how relative positions between the femur and tibia vary under different conditions of movement and load.

### 1.4 Thesis Structure

Structure of the thesis is as follows:

## 2 Methodology

### 2.1 Data Collection Methods

#### 2.1.1 Procedure Details

\*\*\*MRI measurements were performed on four healthy volunteers (aged between 28 and 37 years, body mass between 55 and 90 kg) using a clinical 3 T Siemens Prisma fit scanner. Volunteers had no known musculoskeletal conditions and gave written

informed consent in accordance with the guidelines set out by the institutional ethics committee.<sup>\*\*\*</sup> from device paper For all of these subjects, the left leg was used.

The thigh is secured on a wedge positioner, and the lower leg is attached to an ankle rest, just above the malleoli, using Velcro straps to minimize lateral movement. Once positioned at the scanner’s isocenter, the volunteer then engages in a controlled exercise, following a metronome set at 60 beats per minute. This pace dictates a four-beat flexion to extension cycle, where the leg is flexed at the first beat and fully extended by the fourth. This exercise is performed for approximately 2 minutes throughout the duration of the scan, totaling between 100 to 120 repetitions. Initially conducted under a loaded condition with weights, the process is repeated without the added resistance to compare both states.

### 2.1.2 Sequence Parameters and Reconstruction

## 2.2 Data Analysis

All the analysis and data visualization were done using the python programming language (v3.11.5). To begin, the data in nifti (Neuroimaging Informatics Technology Initiative) format is loaded using nibabel (v5.1.0) library. It is then visualized using napari(v0.4.18), a multi-dimensional interactive image viewer in Python.

### 2.2.1 Segmentation

Step 1: Edge Detection Using the Canny Algorithm

The Canny filter (Canny 1986), as implemented in the scikit-image’s feature library (v0.21.0), was employed to apply an edge filter to the images. Various parameters of the Canny algorithm were adjusted, including edge thresholds and Gaussian blur, to optimize edge detection. Subsequently, the scikit-image’s morphology library was utilized to remove small elements from the binary image. The image was then skeletonized to a one-pixel width, retaining only long and consistent edges. The final selection of the desired edge was accomplished using scipy’s label algorithm (v1.11.3).

## **2.3 Validity and Reliability**

Validation Methods: Detail the steps taken to validate the segmentation techniques and the biomechanical parameters you derived. Reliability Measures: Describe any repeat analyses or cross-verifications done to ensure the consistency and reliability of your results.

# **3 Results**

## **3.1 Segmentation**

some text

## **3.2 Parameter Extraction**

some text

# **4 Discussion**

## **4.1 Technique Evaluation**

Assess the effectiveness and accuracy of your segmentation techniques.

## **4.2 Biomechanical Insights**

Discuss the biomechanical parameters obtained and their implications for understanding knee movement.

## **4.3 Comparison with Existing Methods**

Compare your results with current analysis techniques.

## **5 Conclusion**

### **5.1 Summary of Contributions**

Recap the novel analysis techniques developed and their significance.

### **5.2 Limitations and Challenges**

Discuss any limitations encountered and the challenges in the analysis process

### **5.3 Future Work**

Suggest potential improvements and future directions for research

## References

- Aleksiev, Martin et al. (Oct. 2022). “High-resolution CINE imaging of active guided knee motion using continuously acquired golden-angle radial MRI and rotary sensor information”. In: *Magnetic Resonance Imaging* 92, pp. 161–168. ISSN: 0730725X. DOI: [10.1016/j.mri.2022.06.015](https://doi.org/10.1016/j.mri.2022.06.015). URL: <https://linkinghub.elsevier.com/retrieve/pii/S0730725X22001096> (visited on 02/12/2024).
- Bernstein, Matt A., Kevin Franklin King, and Ziaohong Joe Zhou (2004). *Handbook of MRI pulse sequences*. Amsterdam ; Boston: Academic Press. ISBN: 978-0-12-092861-3.
- Berry, Elizabeth and Andrew J. Bulpitt (2009). *Fundamentals of MRI: an interactive learning approach*. eng. OCLC: 872673253. Boca Raton, Florida: CRC Press. ISBN: 978-1-58488-902-1.
- Breivik, Harald et al. (May 2006). “Survey of chronic pain in Europe: Prevalence, impact on daily life, and treatment”. In: *European Journal of Pain* 10.4, pp. 287–287. ISSN: 1090-3801, 1532-2149. DOI: [10.1016/j.ejpain.2005.06.009](https://doi.org/10.1016/j.ejpain.2005.06.009). URL: <https://onlinelibrary.wiley.com/doi/10.1016/j.ejpain.2005.06.009> (visited on 03/27/2024).
- Brisson, Nicholas M. et al. (Nov. 2022). “A novel multipurpose device for guided knee motion and loading during dynamic magnetic resonance imaging”. en. In: *Zeitschrift für Medizinische Physik* 32.4, pp. 500–513. ISSN: 09393889. DOI: [10.1016/j.zemedi.2021.12.002](https://doi.org/10.1016/j.zemedi.2021.12.002). URL: <https://linkinghub.elsevier.com/retrieve/pii/S093938892100115X> (visited on 02/12/2024).
- Brown, Robert W. et al. (2014). *Magnetic resonance imaging: physical principles and sequence design*. Second edition. Hoboken, New Jersey: John Wiley & Sons, Inc. 944 pp. ISBN: 978-0-471-72085-0.
- Canny, John (Nov. 1986). “A Computational Approach to Edge Detection”. In: *IEEE Transactions on Pattern Analysis and Machine Intelligence* PAMI-8.6, pp. 679–698. ISSN: 0162-8828. DOI: [10.1109/TPAMI.1986.4767851](https://doi.org/10.1109/TPAMI.1986.4767851). URL: <https://ieeexplore.ieee.org/document/4767851> (visited on 02/18/2024).
- Coburn, Sally L. et al. (July 14, 2023). “Immediate and Delayed Effects of Joint Loading Activities on Knee and Hip Cartilage: A Systematic Review and Meta-analysis”. In: *Sports Medicine - Open* 9.1, p. 56. ISSN: 2198-9761. DOI: [10.1186/s40798-023-00602-7](https://doi.org/10.1186/s40798-023-00602-7). URL: <https://sportsmedicine-open.springeropen.com/articles/10.1186/s40798-023-00602-7> (visited on 04/07/2024).



- Curtis, Aaron D. and Hai-Ling M. Cheng (Feb. 2022). “Primer and Historical Review on Rapid Cardiac *¿span style=*” font-variant:small-caps;” ¿CINE MRI¿/span¿”. en. In: *Journal of Magnetic Resonance Imaging* 55.2, pp. 373–388. ISSN: 1053-1807, 1522-2586. DOI: [10.1002/jmri.27436](https://doi.org/10.1002/jmri.27436). URL: <https://onlinelibrary.wiley.com/doi/10.1002/jmri.27436> (visited on 04/09/2024).
- Dalley, Arthur F., A. M. R. Agur, and Keith L. Moore (2023). *Moore’s clinically oriented anatomy*. Ninth edition. Philadelphia: Wolters Kluwer. 1 p. ISBN: 978-1-975154-08-0.
- Edelman, Robert R., ed. (2005). *Clinical magnetic resonance imaging*. Philadelphia, Pa.: Saunders. ISBN: 978-0-7216-0306-3.
- Haase, A et al. (Apr. 1986). “FLASH imaging. Rapid NMR imaging using low flip-angle pulses”. en. In: *Journal of Magnetic Resonance (1969)* 67.2, pp. 258–266. ISSN: 00222364. DOI: [10.1016/0022-2364\(86\)90433-6](https://doi.org/10.1016/0022-2364(86)90433-6). URL: <https://linkinghub.elsevier.com/retrieve/pii/0022236486904336> (visited on 04/19/2024).
- Heidari, Behzad (2011). “Knee osteoarthritis prevalence, risk factors, pathogenesis and features: Part I”. In: *Caspian Journal of Internal Medicine* 2.2, pp. 205–212. ISSN: 2008-6164.
- Menchón-Lara, Rosa-María et al. (Sept. 23, 2019). “Reconstruction techniques for cardiac cine MRI”. In: *Insights into Imaging* 10.1, p. 100. ISSN: 1869-4101. DOI: [10.1186/s13244-019-0754-2](https://doi.org/10.1186/s13244-019-0754-2).
- Netter, Frank H. (2023). *Netter atlas of human anatomy: classic regional approach*. Eighth edition. Philadelphia, PA: Elsevier. ISBN: 978-0-323-68042-4.
- Obermüller, Dominik et al. (Mar. 2024). “Epidemiology and treatment of pain associated with osteoarthritis of the knee in Germany: A retrospective health claims data analysis”. In: *Osteoarthritis and Cartilage Open* 6.1, p. 100430. ISSN: 26659131. DOI: [10.1016/j.ocarto.2023.100430](https://doi.org/10.1016/j.ocarto.2023.100430). URL: <https://linkinghub.elsevier.com/retrieve/pii/S2665913123000973> (visited on 03/27/2024).
- Rickenbach, Robert (June 11, 2013). “Optical spectrum modulated position sensor having a controller with at least one fiber optic line”. U.S. pat. 8461514B1. Individual. URL: <https://patents.google.com/patent/US8461514B1/en?q=US-8461514-B1> (visited on 04/23/2024).
- Standring, Susan and Henry Gray, eds. (2021). *Gray’s anatomy: the anatomical basis of clinical practice*. Forty-second edition. OCLC: on1202943188. Amsterdam: Elsevier. 1588 pp. ISBN: 978-0-7020-7705-0.

- Winkelmann, Stefanie et al. (Jan. 2007). “An Optimal Radial Profile Order Based on the Golden Ratio for Time-Resolved MRI”. en. In: *IEEE Transactions on Medical Imaging* 26.1, pp. 68–76. ISSN: 0278-0062. DOI: [10.1109/TMI.2006.885337](https://doi.org/10.1109/TMI.2006.885337). URL: <http://ieeexplore.ieee.org/document/4039540/> (visited on 04/18/2024).
- Zaitsev, Maxim, Julian Maclaren, and Michael Herbst (Oct. 2015). “Motion artifacts in MRI: A complex problem with many partial solutions”. eng. In: *Journal of magnetic resonance imaging: JMRI* 42.4, pp. 887–901. ISSN: 1522-2586. DOI: [10.1002/jmri.24850](https://doi.org/10.1002/jmri.24850).

## **A Appendix**

If needed for supplementary material, such as detailed description of data collection, tables, or figures.

### **Statutory Declaration:**

I declare that I have developed and written the enclosed Master Thesis completely by myself, and have not used sources or means without declaration in the text. Any thoughts from others or literal quotations are clearly marked. The Master Thesis was not used in the same or in a similar version to achieve an academic grading or is being published elsewhere.

---

Place, Date

---

Signature