\\USER\lxblab\mackey\blackblood\perf-asl_acq-loft5pld

TA: 7:18 min Coil Selection: Auto Voxel Size: 3.0×3.0×3.0 mm³ Acc:: 2 Rel. SNR: 1.00

Properties

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slab Group	1
Slabs	1
Distance Factor	50 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Slices per Slab	40
Phase Oversampling	0 %
Slice Oversampling	10.0 %
FOV Read	240 mm
FOV Phase	100.0 %
Slice Thickness	3.0 mm
TR	4100.0 ms
TE	26.70 ms
Averages	1
Concatenations	1
AutoAlign	---

Contrast - Common

TR	4100.0 ms
TE	26.70 ms
Flip Angle	120 deg
Fat-Water Contrast	Fat Saturation
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	26
Multiple Series	Off
Delay in TR	0.00 ms
Reordering	Centric

Contrast - ASL

Perfusion Mode	None

Resolution - Common

FOV Read	240 mm
FOV Phase	100.0 %
Slice Thickness	3.0 mm
Base Resolution	80
Phase Resolution	98 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	GRAPPA
Reference Scans	GRE/Separate
Acceleration Factor PE	2
Reference Lines PE	80
Acceleration Factor 3D	1
Reference Lines 3D	44
Phase Partial Fourier	Off
Slice Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Normalize	Off

Geometry - Common

Slab Group	1
Slabs	1
Distance Factor	50 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Slices per Slab	40
Phase Oversampling	0 %
Slice Oversampling	10.0 %
FOV Read	240 mm
FOV Phase	100.0 %
Slice Thickness	3.0 mm
TR	4100.0 ms
Concatenations	1

Geometry - AutoAlign

Slab Group	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm

Geometry - AutoAlign

Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	Auto Coil Select
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Adaptive Combine
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
CoilShim	Off
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	240 mm
R >> L	240 mm
F >> H	120 mm
Reset	Off

System - pTx

B1 Shim	TrueForm

System - Tx/Rx

Frequency 1H	123.252072 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	4100.0 ms

Physio - Signal

Segments	4
Concatenations	1

Sequence - Part 1

Sequence Name	tgse
Dimension	3D
RF Pulse Type	Normal
Gradient Mode	Performance
Reordering	Centric
Bandwidth	2604 Hz/Px
Echo Spacing	0.54 ms
Turbo Factor	11
Segments	4
EPI Factor	39

Sequence - Part 2

Introduction	Off

Sequence - Special

CASL Method	Multi-slice
label Offset	90.00 mm
M0-PLD	2000000.00 us
Num RF Blocks	82.00
RF GAP	360.00 us
label FA	25.0 deg
Multi-PLD 0.	500000 us
Multi-PLD 1.	1000000 us
Multi-PLD 2.	1500000 us
Multi-PLD 3.	2000000 us
Multi-PLD 4.	2500000 us
PLD reps 0.	2
PLD reps 1.	2
PLD reps 2.	2
PLD reps 3.	3
PLD reps 4.	3
Crusher Gradient	0.00 s/mm2
mean Gz x10	6.00 mT/m
phi adjust	100.00 percent
BS	On

Sequence - Assistant

SAR Assistant	Off

\\USER\lxblab\mackey\blackblood\fmap-epi_acq-asl_dir-pa

TA: 44 sec Coil Selection: Auto Voxel Size: 3.0×3.0×3.0 mm³ Acc:: 2 Rel. SNR: 1.00

Properties

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slab Group	1
Slabs	1
Distance Factor	50 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	P >> A
Slices per Slab	40
Phase Oversampling	0 %
Slice Oversampling	10.0 %
FOV Read	240 mm
FOV Phase	100.0 %
Slice Thickness	3.0 mm
TR	4100.0 ms
TE	26.70 ms
Averages	1
Concatenations	1
AutoAlign	---

Contrast - Common

TR	4100.0 ms
TE	26.70 ms
Flip Angle	120 deg
Fat-Water Contrast	Fat Saturation
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	2
Multiple Series	Off
Delay in TR	0.00 ms
Reordering	Centric

Contrast - ASL

Perfusion Mode	None

Resolution - Common

FOV Read	240 mm
FOV Phase	100.0 %
Slice Thickness	3.0 mm
Base Resolution	80
Phase Resolution	98 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	GRAPPA
Reference Scans	GRE/Separate
Acceleration Factor PE	2
Reference Lines PE	80
Acceleration Factor 3D	1
Reference Lines 3D	44
Phase Partial Fourier	Off
Slice Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Normalize	Off

Geometry - Common

Slab Group	1
Slabs	1
Distance Factor	50 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	P >> A
Slices per Slab	40
Phase Oversampling	0 %
Slice Oversampling	10.0 %
FOV Read	240 mm
FOV Phase	100.0 %
Slice Thickness	3.0 mm
TR	4100.0 ms
Concatenations	1

Geometry - AutoAlign

Slab Group	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	P >> A
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm

Geometry - AutoAlign

Initial Orientation	Transversal
Initial Rotation	180.00 deg

Geometry - Saturation

Special Saturation	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	Auto Coil Select
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Adaptive Combine
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
CoilShim	Off
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	180.00 deg
A >> P	240 mm
R >> L	240 mm
F >> H	120 mm
Reset	Off

System - pTx

B1 Shim	TrueForm

System - Tx/Rx

Frequency 1H	123.252072 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	4100.0 ms

Physio - Signal

Segments	4
Concatenations	1

Sequence - Part 1

Sequence Name	tgse
Dimension	3D
RF Pulse Type	Normal
Gradient Mode	Performance
Reordering	Centric
Bandwidth	2604 Hz/Px
Echo Spacing	0.54 ms
Turbo Factor	11
Segments	4
EPI Factor	39

Sequence - Part 2

Introduction	Off

Sequence - Special

CASL Method	Multi-slice
label Offset	90.00 mm
M0-PLD	2000000.00 us
Num RF Blocks	82.00
RF GAP	360.00 us
label FA	25.0 deg
Multi-PLD 0.	500000 us
Multi-PLD 1.	1000000 us
Multi-PLD 2.	1500000 us
Multi-PLD 3.	2000000 us
Multi-PLD 4.	2500000 us
PLD reps 0.	0
PLD reps 1.	0
PLD reps 2.	0
PLD reps 3.	0
PLD reps 4.	0
Crusher Gradient	0.00 s/mm2
mean Gz x10	6.00 mT/m
phi adjust	100.00 percent
BS	On

Sequence - Assistant

SAR Assistant	Off