\\USER\crlab\generic\ep3d\anat-scout_ses-1 *

TA: 14 sec Coil Selection: Auto Voxel Size: 1.6×1.6×1.6 mm³ Acc:: 3 Rel. SNR: 1.00

Properties

Start measurement without further preparation	Off
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	On
Graphic segment	Default
Inline Movie	Off

Routine

Slab Group	1
Slabs	1
Distance Factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slices per Slab	128
Phase Oversampling	0 %
Slice Oversampling	0.0 %
FOV Read	260 mm
FOV Phase	100.0 %
Slice Thickness	1.6 mm
TR	3.15 ms
TE	1.37 ms
Averages	1
Concatenations	1
AutoAlign	Head

Contrast - Common

TR	3.15 ms
TE	1.37 ms
Flip Angle	8 deg
Fat-Water Contrast	Standard
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Time to Center	6.2 s

Resolution - Common

FOV Read	260 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	1.6 mm
Base Resolution	160
Phase Resolution	100 %
Slice Resolution	69 %
Trajectory	Cartesian

Resolution - Acceleration

Acceleration Mode	GRAPPA
Reference Scans	Integrated
Acceleration Factor PE	3
Reference Lines PE	24
Acceleration Factor 3D	1
Phase Partial Fourier	6/8
Slice Partial Fourier	6/8
Asymmetric Echo	Weak

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Distortion Correction	3D
Normalize	Prescan
Noise Masking	Off
Image Filter	Off

Geometry - Common

Slab Group	1
Slabs	1
Distance Factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slices per Slab	128
Phase Oversampling	0 %
Slice Oversampling	0.0 %
FOV Read	260 mm
FOV Phase	100.0 %
Slice Thickness	1.6 mm
TR	3.15 ms
Multi-Slice Mode	Sequential
Series	Ascending
Concatenations	1

Geometry - AutoAlign

Slab Group	1
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
AutoAlign	Head
Initial Position	Isocenter
L	0.0 mm

Geometry - AutoAlign

P	0.0 mm
H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	Auto Coil Select
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Adaptive Combine
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Tune up
B1 Shim	TrueForm
CoilShim	Off
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	123.252048 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - PACE

Resp. Control	Off
Concatenations	1

Inline - Subtraction

Subtract	Off
Measurements	1
StdDev	Off
Save Original Images	On

Inline - MIP

MIP Sag	Off
MIP Cor	Off
MIP Tra	Off
MIP Time	Off
Radial MIP	Off
Save Original Images	On
MPR Sag	Off
MPR Cor	Off
MPR Tra	Off

Inline - Composing

Inline Composing	Off

Sequence - Part 1

Sequence Name	fl
Dimension	3D
Excitation	Non-sel.
RF Pulse Type	Fast
Gradient Mode	Normal
Bandwidth	540 Hz/Px
Asymmetric Echo	Weak

Sequence - Part 2

Introduction	On
RF Spoiling	On
Breast Application	Off
Phase Enc. Order	Automatic

Sequence - Assistant

SAR Assistant	Off

\\USER\crlab\generic\ep3d\func-bold_task-rest_acq-3d_run-1 *

TA: 28 sec Coil Selection: Auto Voxel Size: 2.8×2.8×2.8 mm³ Acc:: 4 Rel. SNR: 1.00

Properties

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slab Group	1
Slabs	1
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slices per Slab	44
Phase Oversampling	0.0 %
Slice Oversampling	0.0 %
FOV Read	211 mm
FOV Phase	100.0 %
Slice Thickness	2.80 mm
TR	46.1 ms
Vol. TR	1014.2 ms
TE 1	30.00 ms
Averages	1
Multi-echo Shots	1
AutoAlign	Head > Brain

Contrast - Common

TR	46.1 ms
Vol. TR	1014.2 ms
TE 1	30.00 ms
Multi-echo spacing	30.20 ms
MTC	Off
Flip Angle	45 deg
Fat-Water Contrast	Standard
Contrasts	1
Reconstruction	Magn./Phase

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	12
Reordering	Linear

Resolution - Common

FOV Read	211 mm
FOV Phase	100.0 %
Slice Thickness	2.80 mm
Base Resolution	76
Phase Resolution	100 %
Slice Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	CAIPIRINHA
Reference Scans	GRE/Separate
Acceleration Factor PE	2
Reference Lines PE	68
Acceleration Factor 3D	2
Reference Lines 3D	44
Reordering Shift 3D	0
Phase Partial Fourier	Off
Slice Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Distortion Correction	Off
Normalize	Off
Image Filter	Off

Geometry - Common

Slab Group	1
Slabs	1
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slices per Slab	44
Phase Oversampling	0.0 %
Slice Oversampling	0.0 %
FOV Read	211 mm
FOV Phase	100.0 %
Slice Thickness	2.80 mm
TR	46.1 ms
Vol. TR	1014.2 ms
Multi-echo Shots	1

Geometry - AutoAlign

Slab Group	1
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm

Geometry - AutoAlign

P	0.0 mm
H	0.0 mm
Initial Orientation	Sagittal
Initial Rotation	0.00 deg

Geometry - Saturation

Saturation Mode	Standard

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	Auto Coil Select
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Adaptive Combine
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Brain
B1 Shim	TrueForm
CoilShim	Off
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Sagittal
Rotation	0.00 deg
A >> P	211 mm
F >> H	211 mm
R >> L	124 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	123.252048 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Sequence - Part 1

Sequence Name	ep v3.1
Dimension	3D
Excitation	Non-sel.
RF Pulse Type	Normal
Gradient Mode	Fast
Reordering	Linear
Bandwidth	1430 Hz/Px
Echo Spacing	0.78 ms
Segmentation	1
EPI Factor	38

Sequence - Part 2

Introduction	On
RF Spoiling	On

Sequence - Special

RF duration	1000 us
RF time x BW	20
PAT ref. FA	3 deg
Fat sat. FA	110 deg
T1 (Ernst FA)	1200 ms
Invert PE	Off
Min. TE w/ PF	On
Ramp Sampling	On
Trigger per shot	Off
Noise image	Off
Relax spoilers	Off
Round up Vol. TR	Off
MT flip angle	500 deg
MT off-res.	2000 Hz
MT RF duration	10240 us
Custom Water Exc.	-none-
Phase Correction	per Series
Saturation RF	per Shot
EPI rise time factor	1.10
G. spoil dephasing[1]	0.0 pi
G. spoil dephasing[2]	4.0 pi
G. spoil dephasing[3]	2.0 pi
Modify Ice Config	On
G-factor map	Off
GRAPPA Regularization	5000 /10^6
Slab Scale	0 %
RF spoil scheme	Conventional
Read polarity	Regular RO

Sequence - Assistant

SAR Assistant	Off

\\USER\crlab\generic\ep3d\func-bold_task-rest_acq-3d3echo_run-2 *

TA: 30 sec Coil Selection: Auto Voxel Size: 2.8×2.8×2.8 mm³ Acc:: 4 Rel. SNR: 1.00

Properties

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slab Group	1
Slabs	1
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slices per Slab	44
Phase Oversampling	0.0 %
Slice Oversampling	0.0 %
FOV Read	211 mm
FOV Phase	100.0 %
Slice Thickness	2.80 mm
TR	56.2 ms
Vol. TR	1236.4 ms
TE 1	13.00 ms
TE 2	40.10 ms
Averages	1
Multi-echo Shots	1
AutoAlign	Head > Brain

Contrast - Common

TR	56.2 ms
Vol. TR	1236.4 ms
TE 1	13.00 ms
TE 2	40.10 ms
Multi-echo spacing	27.10 ms
MTC	Off
Flip Angle	45 deg
Fat-Water Contrast	Standard
Contrasts	2
Reconstruction	Magn./Phase

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	12
Reordering	Linear

Resolution - Common

FOV Read	211 mm
FOV Phase	100.0 %
Slice Thickness	2.80 mm
Base Resolution	76
Phase Resolution	100 %
Slice Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	CAIPIRINHA
Reference Scans	GRE/Separate
Acceleration Factor PE	2
Reference Lines PE	68
Acceleration Factor 3D	2
Reference Lines 3D	44
Reordering Shift 3D	0
Phase Partial Fourier	7/8
Slice Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Distortion Correction	Off
Normalize	Off
Image Filter	Off

Geometry - Common

Slab Group	1
Slabs	1
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slices per Slab	44
Phase Oversampling	0.0 %
Slice Oversampling	0.0 %
FOV Read	211 mm
FOV Phase	100.0 %
Slice Thickness	2.80 mm
TR	56.2 ms
Vol. TR	1236.4 ms
Multi-echo Shots	1

Geometry - AutoAlign

Slab Group	1
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm

Geometry - AutoAlign

P	0.0 mm
H	0.0 mm
Initial Orientation	Sagittal
Initial Rotation	0.00 deg

Geometry - Saturation

Saturation Mode	Standard

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	Auto Coil Select
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Adaptive Combine
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Brain
B1 Shim	TrueForm
CoilShim	Off
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Sagittal
Rotation	0.00 deg
A >> P	211 mm
F >> H	211 mm
R >> L	124 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	123.252048 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Sequence - Part 1

Sequence Name	ep v3.1
Dimension	3D
Excitation	Non-sel.
RF Pulse Type	Normal
Gradient Mode	Fast
Reordering	Linear
Bandwidth	1430 Hz/Px
Echo Spacing	0.78 ms
Segmentation	1
EPI Factor	34

Sequence - Part 2

Introduction	On
RF Spoiling	On

Sequence - Special

RF duration	1000 us
RF time x BW	20
PAT ref. FA	3 deg
Fat sat. FA	110 deg
T1 (Ernst FA)	1200 ms
Invert PE	Off
Min. TE w/ PF	On
Ramp Sampling	On
Trigger per shot	Off
Noise image	Off
Relax spoilers	Off
Round up Vol. TR	Off
MT flip angle	500 deg
MT off-res.	2000 Hz
MT RF duration	10240 us
Custom Water Exc.	-none-
Phase Correction	per Series
Saturation RF	per Shot
EPI rise time factor	1.10
G. spoil dephasing[1]	0.0 pi
G. spoil dephasing[2]	4.0 pi
G. spoil dephasing[3]	2.0 pi
Modify Ice Config	On
G-factor map	Off
GRAPPA Regularization	5000 /10^6
Slab Scale	0 %
RF spoil scheme	Conventional
Read polarity	Regular RO

Sequence - Assistant

SAR Assistant	Off

\\USER\crlab\generic\ep3d\func-bold_task-rest_acq-3d2shots_run-3 *

TA: 40 sec Coil Selection: Auto Voxel Size: 2.8×2.8×2.8 mm³ Acc:: 4 Rel. SNR: 1.00

Properties

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slab Group	1
Slabs	1
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slices per Slab	44
Phase Oversampling	0.0 %
Slice Oversampling	0.0 %
FOV Read	211 mm
FOV Phase	100.0 %
Slice Thickness	2.80 mm
TR	46.1 ms
Vol. TR	1014.2 ms
TE 1	16.10 ms
TE 2	30.00 ms
Averages	1
Multi-echo Shots	2
AutoAlign	Head > Brain

Contrast - Common

TR	46.1 ms
Vol. TR	1014.2 ms
TE 1	16.10 ms
TE 2	30.00 ms
Multi-echo spacing	30.20 ms
MTC	Off
Flip Angle	45 deg
Fat-Water Contrast	Standard
Contrasts	2
Reconstruction	Magn./Phase

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	12
Reordering	Linear

Resolution - Common

FOV Read	211 mm
FOV Phase	100.0 %
Slice Thickness	2.80 mm
Base Resolution	76
Phase Resolution	100 %
Slice Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	CAIPIRINHA
Reference Scans	GRE/Separate
Acceleration Factor PE	2
Reference Lines PE	68
Acceleration Factor 3D	2
Reference Lines 3D	44
Reordering Shift 3D	0
Phase Partial Fourier	Off
Slice Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Distortion Correction	Off
Normalize	Off
Image Filter	Off

Geometry - Common

Slab Group	1
Slabs	1
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slices per Slab	44
Phase Oversampling	0.0 %
Slice Oversampling	0.0 %
FOV Read	211 mm
FOV Phase	100.0 %
Slice Thickness	2.80 mm
TR	46.1 ms
Vol. TR	1014.2 ms
Multi-echo Shots	2

Geometry - AutoAlign

Slab Group	1
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm

Geometry - AutoAlign

P	0.0 mm
H	0.0 mm
Initial Orientation	Sagittal
Initial Rotation	0.00 deg

Geometry - Saturation

Saturation Mode	Standard

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	Auto Coil Select
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Adaptive Combine
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Brain
B1 Shim	TrueForm
CoilShim	Off
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Sagittal
Rotation	0.00 deg
A >> P	211 mm
F >> H	211 mm
R >> L	124 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	123.252048 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Sequence - Part 1

Sequence Name	ep v3.1
Dimension	3D
Excitation	Non-sel.
RF Pulse Type	Normal
Gradient Mode	Fast
Reordering	Linear
Bandwidth	1430 Hz/Px
Echo Spacing	0.78 ms
Segmentation	1
EPI Factor	38

Sequence - Part 2

Introduction	On
RF Spoiling	On

Sequence - Special

RF duration	1000 us
RF time x BW	20
PAT ref. FA	3 deg
Fat sat. FA	110 deg
T1 (Ernst FA)	1200 ms
Invert PE	Off
Min. TE w/ PF	On
Ramp Sampling	On
Trigger per shot	Off
Noise image	Off
Relax spoilers	Off
Round up Vol. TR	Off
MT flip angle	500 deg
MT off-res.	2000 Hz
MT RF duration	10240 us
Custom Water Exc.	-none-
Phase Correction	per Series
Saturation RF	per Shot
EPI rise time factor	1.10
G. spoil dephasing[1]	0.0 pi
G. spoil dephasing[2]	4.0 pi
G. spoil dephasing[3]	2.0 pi
Modify Ice Config	On
G-factor map	Off
GRAPPA Regularization	5000 /10^6
Slab Scale	0 %
RF spoil scheme	Conventional
Read polarity	Regular RO

Sequence - Assistant

SAR Assistant	Off