**TAM Assistant Analysis Report**
## *Arrhenius*

## Regression Input

| Temperature | Po |
|---|---|
| 75°C | 201.21μW/g |
| 85°C | 1.015mW/g |

## Regression Results

| | |
|---|---|
| Ea: | 167.77kJ/mol [Error could not be calculated] |
| dH A: | 2.9834ZW/g [Error could not be calculated] |
| Standard deviation: | InfinityW/g |
| NDF: | 0 |

o **Measured** —— **Calculated**



## Rate Constants Calculation Input
| | |
|---|---|
| dH: | 800J/g |

## Rate Constants Calculation Results

| Temperature | k |
|---|---|
| 25°C | 0.0000000000151131/s |