Resolution Calibration

fitting parameters:

a = 2.86e-06±1.13e-06
b = -0.00266±0.00135
c = 1.97±0.185

$R = \frac{\sqrt{a*E^2 + b*E + c}}{E}$

fitted curve
data

Resolution

Reference Photopeak Energy/keV