CTI

ilac-MRA CNAS 中国认可 国际互认 校准 CALIBRATION CNAS L5102

# 华测计量检测有限公司

CTI MEASUREMENT AND TESTING CO., LTD.

# 校准证书

Calibration Certificate

| | | | |
|---|---|---|---|
| 证书编号 Certificate No. | C2506112780299 | | 第 1 页 共 3 页 Page of |
| 委托单位 Customer | 上海格琳电子有限公司 | | |
| 委托单位地址 Address | 上海市嘉定区安亭镇先锋村先锋路200号 | | |
| 器具名称 Name of instrument | 槽深仪（百分表） | | |
| 型号规格 Model | （0~10）mm | | |
| 制造商 Manufacturer | UNIVERSAL PUNCH CORP | | |
| 出厂编号 Serial No. | HAE197 | 管理编号 Management No. | --------- |
| 接收日期 Received date | 2025/07/21 | 校准日期 Calibration date | 2025/07/22 |
| 发布日期 Issue date | 2025/07/24 | 建议下次校准日期 Next calibration date | 2026/07/21 |

证书/报告章 Stamp

| | |
|---|---|
| 批准 Approved by | 李泽文 |
| 审核 Inspected by | 李泽文 |
| 校准 Calibrated by | 洪通 |


总部地址：广东省深圳市宝安区西乡街道铁岗社区桃花源科技创新园B、C栋

Headquarter address: Building B and C, Taohuayuan Sci-Tech Innovation Park, Tiegang Community, Xixiang Sub-district, Bao'an District, Shenzhen, Guangdong, China

实验室地址：上海市闵行区万芳路1351号

Laboratory address: No.1351,Wanfang Road, Minhang District, Shanghai, China

邮编 Post code：201112 电话 Tel.：021-3107 3300 电子邮箱 E-mail：calibration@cti-cert.com

# 说 明

## Directions

证书编号 C2506112780299
Certificate No.



1. 本证书校准结果均可溯源至国际单位制（SI）单位。
The results are traceable to International System of Units(SI).

2. 证书未盖本公司证书/报告章及骑缝章无效。未经本公司书面批准，不得部分复制此证书。
Any certificate is deemed to be invalid without both the certificate/report seal and its across-page seal. This certificate shall not be copied partly without the written approval.

3. 本证书校准结果只与受校准仪器有关。如证书中的英文内容与中文内容有差异，以中文为准。
The results relate only to the items calibrated. In case of any discrepancy between the English version and Chinese version of the certificate(if generated), the Chinese version shall prevail.

4. 本次校准的技术依据：
Reference documents for the calibration
JJG 830-2007 深度指示表检定规程

5. 校准地点、环境条件：
Place and environment condition during calibration
地点：上海市嘉定区安亭镇先锋村先锋路200号二楼实验室
Place
温度：21.0℃~ 21.5℃ 相对湿度：52%~ 54%
Temperature R.H.

计量溯源性声明

6. 本次校准所使用的主要计量标准器具：
Main mearsurement standards used in the calibration

| 名称/型号规格 Name/Model | 编号 Serial No. | 测量范围 Measurement range | 准确度等级/最大允许误差/不确定度 Accuracy class/Maximum permissible error/Uncertainty | 证书号/溯源机构 Certificate No./Traceability to | 有效期 Due date |
|---|---|---|---|---|---|
| 量块 (0.5~100)mm (122) | 84817 | (0.5~100)mm | 3等 | 2024D12-20-5603305001 上海市计量测试技术研究院 | 2025/11/21 |

# 校准结果

## Results of calibration

证书编号 C2506112780299
Certificate No.



1. 外观及工作正常性检查
   Appearance and function check
   正常 Normal

2. 重复性 0.001mm 允差 0.003mm 结论 Pass
   Repeatability MPE Conclusion

3. 示值误差

| 校准点(mm) Calibrated point | 示值误差(mm) Error of indication | 允差（mm） MPE | 结论 Conclusion |
|---|---|---|---|
| 0.5 | −0.001 | ± 0.02 | Pass |
| 1 | −0.002 | ± 0.02 | Pass |
| 2 | −0.002 | ± 0.02 | Pass |
| 3 | −0.002 | ± 0.02 | Pass |

4. 扩展不确定度 $U$= 0.003mm ，$k$=2。
   Expanded uncertainty

5. 分度值/分辨力 0.01mm
   Division/resolution

备注：
Notes

1. 依据JJF1059.1-2012测量不确定度评定与表示。
   According to JJF1059.1-2012 Evaluation and Expression of Uncertainty in Measurement.
2. 校准项目符合技术要求。
   The calibrated measurand are accord with the technical specifications.

以下空白
Blank below