



TRAINING EXAMPLE — NOT A REAL IEC 62304 DELIVERABLE

This project is a **training site, not a real medical device** under any regulatory definition. The device described in this document — “SentinelFlow 500” — is entirely **fictional**, invented for this course: nothing here is a genuine IEC 62304 deliverable, a real regulatory submission, or evidence of any real organisation’s compliance with anything. It illustrates the *kind* of document a real SaMD/SiMD project would produce. See the [document register](#) for the full explanation.

Clause: 5.8 — Software Release · **Register:** [Device Example Register](#)

Document ID: SOP-09 · **Device:** SentinelFlow 500 (fictional) · **Safety class:** C

Status: SOP — template (Procedure not yet written for SentinelFlow 500)

Fictional worked example. See [SOP-01](#) for the disclaimer that applies to this whole document set, and the [Device Example Register](#) for the SentinelFlow 500 device description and what “template” status means here.

1. Purpose

This SOP defines the release gate SentinelFlow 500's software must pass before it can ship: verification complete, anomalies evaluated, the version archived, and reliable delivery assured, per Clause 5.8.

2. Scope

Applies to release of SentinelFlow 500's control software (Class C — see [SOP-01 §7.3](#)). Covers Clause 5.8 in full — Amendment 1 moved §5.8.1, 5.8.2, 5.8.7 and 5.8.8 from Class B/C to all classes, so at Class C the whole clause applies.

3. Responsibilities

ROLE	RESPONSIBILITY UNDER THIS SOP
Software Development Lead	Confirms all planned development activities are complete before release
QA/RA	Authorises release under the QMS; IEC 62304 does not itself say who authorises release — that is an ISO 13485 quality management system matter, not a requirement of this standard
Risk Manager	Confirms residual risk from any known anomaly is evaluated as acceptable
Configuration Manager	Archives the released software, its configuration items, and its documentation per SOP-12

4. Definitions & Abbreviations

TERM	MEANING
Residual anomaly	A known defect present in the software at release, evaluated and either accepted or scheduled for correction
Archive	The retained copy of source, build instructions, tools, and SOUP versions sufficient to reproduce a released build

5. References

- IEC 62304:2006+AMD1:2015, §5.8
- **SOP-11 — Software Risk Management File:** source of the residual-risk evaluation §5.8.3 requires
- **SOP-12 — Software Configuration Management Plan:** owns the archive §5.8.7 requires

6. What the standard requires

REF	TITLE	APPLIES AT CLASS C	WHAT MUST BE PRODUCED
5.8.1	Ensure software verification is complete	Yes	(No standard-named output.)
5.8.2	Document known residual anomalies	Yes	All known residual anomalies, documented
5.8.3	Evaluate known residual anomalies	Yes	Evaluation confirming that known residual anomalies do not contribute to an unacceptable risk
5.8.4	Document released versions	Yes	The version of the medical device software being released, documented
5.8.5	Document how released software was created	Yes	The procedure and environment used to create the released software, documented
5.8.6	Ensure activities and tasks are complete	Yes	Evidence that all development or maintenance plan activities and tasks are complete, with their associated documentation
5.8.7	Archive software	Yes	Archive of the medical device software, its configuration items and the documentation, retained for the longer of the software lifetime or the applicable regulatory retention period
5.8.8	Assure reliable delivery of released software	Yes	Procedures assuring reliable delivery of the released software, addressing replication, media labelling, packaging, protection, storage and delivery

(Quoted directly from `applicability.json` , identical in content to the self-referential set's [09](#), whose Status is itself marked **Partial** — that project has no

formal release-versioning scheme. This device-example set's version is a template for the same clause, not a completed alternative.)

7. Procedure

Not yet written for SentinelFlow 500 — see the [Device Example Register](#) for what "template" status means. Once complete, this section would give SentinelFlow 500's actual release checklist, its versioning scheme, and its archive contents.

5.8.1–5.8.3 — Verification complete; residual anomalies documented and evaluated

[Template.]

5.8.4–5.8.6 — Version, build procedure, and activity completeness

[Template.]

5.8.7–5.8.8 — Archive and reliable delivery

[Template — see SOP-12 for the archive this section would rely on.]

8. Records

RECORD	PRODUCED BY	RETAINED IN
Release checklist / gate review record	Software Development Lead	Project technical file
Residual anomaly evaluation	Risk Manager	Risk management file (SOP-11)
Software archive	Configuration Manager	Configuration management system (SOP-12)

9. Revision History

Illustrative only — SentinelFlow 500 has no real revision history.

VERSION	STATUS	DESCRIPTION
1.0	Template	Structure and requirements table complete; Procedure section pending