



TRAINING EXAMPLE — NOT A REAL IEC 62304 DELIVERABLE

This project is a **training site, not a real medical device** under any regulatory definition. The device described in this document — “SentinelFlow 500” — is entirely **fictional**, invented for this course: nothing here is a genuine IEC 62304 deliverable, a real regulatory submission, or evidence of any real organisation’s compliance with anything. It illustrates the *kind* of document a real SaMD/SiMD project would produce. See the [document register](#) for the full explanation.

Clause: 6 — Software Maintenance Process · **Register:** [Device Example Register](#)

Document ID: SOP-10 · **Device:** SentinelFlow 500 (fictional) · **Safety class:** C

Status: SOP — template (Procedure not yet written for SentinelFlow 500)

Fictional worked example. See [SOP-01](#) for the disclaimer that applies to this whole document set, and the [Device Example Register](#) for the SentinelFlow 500 device description and what “template” status means here.

1. Purpose

This SOP defines how SentinelFlow 500 monitors feedback on released software, evaluates and implements changes, and re-applies development activities and, where necessary, re-classifies affected software items, per Clause 6.

2. Scope

Applies to all released versions of SentinelFlow 500's control software (Class C — see [SOP-01 §7.3](#)). Covers Clause 6 in full — every sub-clause applies to all three safety classes; Amendment 1's move of §6.2.3 from Class B/C to all classes made Clause 6 uniformly all-classes.

3. Responsibilities

ROLE	RESPONSIBILITY UNDER THIS SOP
Software Development Lead	Monitors and evaluates feedback; analyses change requests for effect on the organisation, software, and interfacing systems
Risk Manager	Evaluates every problem report's effect on safety; re-evaluates classification if scope changes
QA/RA	Approves change requests; confirms users and regulators are informed as required

4. Definitions & Abbreviations

TERM	MEANING
Feedback	Post-market information about released software — complaints, field reports, vigilance reports
Change request	A proposed modification to released software, evaluated before implementation

5. References

- IEC 62304:2006+AMD1:2015, Clause 6
- **SOP-13 — Software Problem Resolution Process:** every problem identified through maintenance feedback is managed through this process
- **SOP-01 — General Requirements & Classification:** re-applied if a change materially expands intended use or introduces significant new functionality

6. What the standard requires

REF	TITLE	APPLIES AT CLASS C	WHAT MUST BE PRODUCED
6.1	Establish software maintenance plan	Yes	Software maintenance plan

REF	TITLE	APPLIES AT CLASS C	WHAT MUST BE PRODUCED
6.2.1.1	Monitor feedback	Yes	(No standard-named output.)
6.2.1.2	Document and evaluate feedback	Yes	Documented and evaluated feedback, with any problem recorded as a problem report
6.2.1.3	Evaluate problem report's effects on safety	Yes	Evaluation of each problem report for its effect on the safety of released software, and whether a change is needed
6.2.2	Use software problem resolution process	Yes	(No standard-named output.)
6.2.3	Analyse change requests	Yes	Analysis of each change request for its effect on the organisation, released software and interfacing systems
6.2.4	Change request approval	Yes	Evaluated and approved change requests
6.2.5	Communicate to users and regulators	Yes	Identification of approved change requests affecting released software, and the information provided to users and regulators as local regulation requires
6.3.1	Use established process to implement modification	Yes	(No standard-named output.)
6.3.2	Re-release modified software system	Yes	(No standard-named output.)

(Quoted directly from `applicability.json` , identical in content to the self-referential set's [10](#).)

7. Procedure

Not yet written for SentinelFlow 500 — see the [Device Example Register](#) for what "template" status means. Once complete, this section would give SentinelFlow 500's actual feedback channels (e.g. field service reports, complaint handling), its change-evaluation criteria, and its re-release process.

6.1 — Maintenance plan

[Template.]

6.2.1.1–6.2.1.3 — Monitor, document, and evaluate feedback for safety effect

[Template.]

6.2.2–6.2.5 — Problem resolution, change analysis, approval, communication

[Template — see SOP-13 for the problem resolution process this section relies on.]

6.3.1–6.3.2 — Implement and re-release

[Template.]

8. Records

RECORD	PRODUCED BY	RETAINED IN
Feedback log	Software Development Lead	Project technical file
Change request records, evaluated and approved	Software Development Lead / QA/RA	Project technical file

9. Revision History

Illustrative only — SentinelFlow 500 has no real revision history.

VERSION	STATUS	DESCRIPTION
1.0	Template	Structure and requirements table complete; Procedure section pending