



TRAINING EXAMPLE — NOT A REAL IEC 62304 DELIVERABLE

This project is a **training site, not a real medical device** under any regulatory definition. The device described in this document — “SentinelFlow 500” — is entirely **fictional**, invented for this course: nothing here is a genuine IEC 62304 deliverable, a real regulatory submission, or evidence of any real organisation’s compliance with anything. It illustrates the *kind* of document a real SaMD/SiMD project would produce. See the [document register](#) for the full explanation.

Clause: 7 — Software Risk Management · **Register:** [Device Example Register](#)

Document ID: SOP-11 · **Device:** SentinelFlow 500 (fictional) · **Safety class:** C

Status: SOP — partial worked example (hazard analysis drafted; risk control verification and traceability not yet written)

Fictional worked example. See [SOP-01](#) for the disclaimer that applies to this whole document set, and the [Device Example Register](#) for the SentinelFlow 500 device description and what “template” status means here.

1. Purpose

This SOP defines how software items that could contribute to a hazardous situation are identified, how their risk is evaluated and controlled, and how those activities integrate with the ISO 14971 device risk management process, per Clause 7. It is the SOP the classification rationale in [SOP-01 §7.3](#) is drawn from.

2. Scope

Applies to all software items in SentinelFlow 500's control software (Class C — see SOP-01 §7.3). Covers Clause 7 in full. §7.4.1 (analyse changes to the software with respect to safety) is the only requirement in Clause 7 that applies at Class A; every

other sub-clause applies at Class B and C only, since a Class A software item cannot by definition contribute to a hazardous situation.

3. Responsibilities

ROLE	RESPONSIBILITY UNDER THIS SOP
Risk Manager	Owns the risk management file; integrates software risk activities with the device-level ISO 14971 process
Software Development Lead	Identifies software items and their potential causes of contribution to a hazardous situation
QA/RA	Confirms verification of risk control measures is documented before release

4. Definitions & Abbreviations

TERM	MEANING
Hazard	A potential source of harm (e.g. incorrect dose delivered)
Hazardous situation	The circumstance in which a person is exposed to a hazard (e.g. patient receives the incorrect dose) — the Amendment 1 terminology this SOP uses throughout
Risk control measure	An action implemented in software to reduce the probability or severity of a hazardous situation

5. References

- IEC 62304:2006+AMD1:2015, Clause 7
- ISO 14971 — the device-level risk management process this file feeds
- **SOP-01 — General Requirements & Classification:** the classification this file's hazard analysis supports
- **SOP-03 — Software Requirements Specification:** where risk control measures identified here are implemented as requirements (§5.2.3)

6. What the standard requires

REF	TITLE	APPLIES AT CLASS C	WHAT MUST BE PRODUCED
7.1.1	Identify software items that could contribute to a hazardous situation	Yes	(No standard-named output.)
7.1.2	Identify potential causes of contribution to a hazardous situation	Yes	(No standard-named output.)
7.1.3	Evaluate published SOUP anomaly lists	Yes	(No standard-named output.)
7.1.4	Document potential causes	Yes	Potential causes of the software item contributing to a hazardous situation, documented in the risk management file
7.1.5	Deleted by Amendment 1	No — deleted	—
7.2.1	Define risk control measures	Yes	Documented risk control measures for each case where a software item could contribute to a hazardous situation
7.2.2	Risk control measures implemented in software	Yes	Risk control measures included in the software requirements, with a safety class assigned to each implementing software item
7.3.1	Verify risk control measures	Yes	Documented verification of each risk control measure, and a review of whether it could itself result in a new hazardous situation

REF	TITLE	APPLIES AT CLASS C	WHAT MUST BE PRODUCED
7.3.2	Not used (Amendment 1)	No — deleted	—
7.3.3	Document traceability	Yes	Documented traceability from hazardous situation to software item, to software cause, to risk control measure, to verification of that measure
7.4.1	Analyse changes to medical device software with respect to safety	Yes (only Clause 7 requirement reaching Class A)	Analysis of changes to the software, including SOUP, for additional potential causes and additional risk control measures needed
7.4.2	Analyse impact of software changes on existing risk control measures	Yes	Analysis of whether software changes, including SOUP changes, could interfere with existing risk control measures
7.4.3	Perform risk management activities based on analyses	Yes	(No standard-named output.)

(Quoted directly from `applicability.json` , identical in content to the self-referential set's [11](#), whose Status is itself marked "Documented — reflexive": that document uses risk management to describe the risk of teaching a reader something wrong, rather than patient harm — the gap this device-example version exists to fill.)

7. Procedure

7.1 — Hazard analysis (§7.1.1–7.1.4)

The three failure modes already reasoned through in [SOP-01 §7.3](#) to justify Class C form the starting hazard analysis for the delivery-control software item:

#	SOFTWARE CAUSE	HAZARDOUS SITUATION	POTENTIAL HARM	RISK CONTROL MEASURE (§7.2.1)
H-1	Occlusion-sensor software path fails to raise an alarm on a blocked line	Time-critical drug (e.g. vasoactive infusion) silently stops reaching the patient	Death or serious injury, depending on drug	Independent software watchdog on the occlusion-sensor read path; audible + visual alarm within a bounded response time
H-2	Software fails to keep the motor stopped when it should be closed (e.g. during a pause, or set change)	Free-flow over-infusion — the reservoir empties into the patient far faster than programmed	Death or serious injury	Software-enforced motor lock during any non-delivering state; state confirmed by sensor feedback, not command alone
H-3	Parameter validation fails to catch a rate outside the configured safe range for the selected drug profile	Order-of-magnitude dosing error delivered as if correct	Death or serious injury	Dose-rate validation against a configurable per-profile safe range before delivery starts, rejecting out-of-range programming

Each of these is the sole software-level control between a plausible failure and patient harm, which is the Class C condition applied in SOP-01. §7.1.3 (evaluate published SOUP anomaly lists) has no entries yet: no SOUP component has been identified in the architecture (SOP-04 is itself still a template), so there is nothing to evaluate against published anomaly lists until that section is written.

7.2–7.3 — Risk control measures, their implementation, and verification

[Template — the risk control measures in the table above are named but not yet carried through to specific software requirements (SOP-03) or verified (§7.3.1). Traceability (§7.3.3) from hazardous situation → software item →

cause → control → verification cannot be completed until SOP-03 and SOP-04 are written.]

7.4 — Change analysis

[Template — no software change has occurred yet for SentinelFlow 500 to analyse.]

8. Records

RECORD	PRODUCED BY	RETAINED IN
Hazard analysis (§7.1)	Risk Manager	This risk management file
Risk control measure verification (§7.3.1), once written	Risk Manager	This risk management file
Traceability matrix (hazard → cause → control → verification), once written	Risk Manager	This risk management file

9. Revision History

Illustrative only — SentinelFlow 500 has no real revision history.

VERSION	STATUS	DESCRIPTION
1.0	Partial	Hazard analysis (§7.1) drafted for the three failure modes underlying the Class C classification; risk control verification and traceability pending