



## TRAINING EXAMPLE — NOT A REAL IEC 62304 DELIVERABLE

This project is a **training site, not a real medical device** under any regulatory definition. The device described in this document — “SentinelFlow 500” — is entirely **fictional**, invented for this course: nothing here is a genuine IEC 62304 deliverable, a real regulatory submission, or evidence of any real organisation’s compliance with anything. It illustrates the *kind* of document a real SaMD/SiMD project would produce. See the [document register](#) for the full explanation.

**Clause:** 5.8 — Software Release · **Register:** [Device Example Register](#)

**Document ID:** SOP-09 · **Device:** SentinelFlow 500 (fictional) · **Safety class:** C

**Status:** SOP — template (Procedure not yet written for SentinelFlow 500)

**Fictional worked example.** See [SOP-01](#) for the disclaimer that applies to this whole document set, and the [Device Example Register](#) for the SentinelFlow 500 device description and what “template” status means here.

## 1. Purpose

This SOP defines the release gate SentinelFlow 500's software must pass before it can ship: verification complete, anomalies evaluated, the version archived, and reliable delivery assured, per Clause 5.8.

## 2. Scope

Applies to release of SentinelFlow 500's control software (Class C — see [SOP-01 §7.3](#)). Covers Clause 5.8 in full — Amendment 1 moved §5.8.1, 5.8.2, 5.8.7 and 5.8.8 from Class B/C to all classes, so at Class C the whole clause applies.

## 3. Responsibilities

| ROLE                      | RESPONSIBILITY UNDER THIS SOP                                                                                                                                                      |
|---------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Software Development Lead | Confirms all planned development activities are complete before release                                                                                                            |
| QA/RA                     | Authorises release under the QMS; IEC 62304 does not itself say who authorises release — that is an ISO 13485 quality management system matter, not a requirement of this standard |
| Risk Manager              | Confirms residual risk from any known anomaly is evaluated as acceptable                                                                                                           |
| Configuration Manager     | Archives the released software, its configuration items, and its documentation per SOP-12                                                                                          |

## 4. Definitions & Abbreviations

| TERM             | MEANING                                                                                                            |
|------------------|--------------------------------------------------------------------------------------------------------------------|
| Residual anomaly | A known defect present in the software at release, evaluated and either accepted or scheduled for correction       |
| Archive          | The retained copy of source, build instructions, tools, and SOUP versions sufficient to reproduce a released build |

## 5. References

- IEC 62304:2006+AMD1:2015, §5.8
- **SOP-11 — Software Risk Management File:** source of the residual-risk evaluation §5.8.3 requires
- **SOP-12 — Software Configuration Management Plan:** owns the archive §5.8.7 requires

## 6. What the standard requires

| REF   | TITLE                                         | APPLIES<br>AT CLASS<br>C | WHAT MUST BE PRODUCED                                                                                                                                                                 |
|-------|-----------------------------------------------|--------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| 5.8.1 | Ensure software verification is complete      | Yes                      | (No standard-named output.)                                                                                                                                                           |
| 5.8.2 | Document known residual anomalies             | Yes                      | All known residual anomalies, documented                                                                                                                                              |
| 5.8.3 | Evaluate known residual anomalies             | Yes                      | Evaluation confirming that known residual anomalies do not contribute to an unacceptable risk                                                                                         |
| 5.8.4 | Document released versions                    | Yes                      | The version of the medical device software being released, documented                                                                                                                 |
| 5.8.5 | Document how released software was created    | Yes                      | The procedure and environment used to create the released software, documented                                                                                                        |
| 5.8.6 | Ensure activities and tasks are complete      | Yes                      | Evidence that all development or maintenance plan activities and tasks are complete, with their associated documentation                                                              |
| 5.8.7 | Archive software                              | Yes                      | Archive of the medical device software, its configuration items and the documentation, retained for the longer of the software lifetime or the applicable regulatory retention period |
| 5.8.8 | Assure reliable delivery of released software | Yes                      | Procedures assuring reliable delivery of the released software, addressing replication, media labelling, packaging, protection, storage and delivery                                  |

(Quoted directly from `applicability.json` , identical in content to the self-referential set's [09](#), whose Status is itself marked **Partial** — that project has no

formal release-versioning scheme. This device-example set's version is a template for the same clause, not a completed alternative.)

## 7. Procedure

*Not yet written for SentinelFlow 500 — see the [Device Example Register](#) for what "template" status means. Once complete, this section would give SentinelFlow 500's actual release checklist, its versioning scheme, and its archive contents.*

### 5.8.1–5.8.3 — Verification complete; residual anomalies documented and evaluated

**[Template.]**

### 5.8.4–5.8.6 — Version, build procedure, and activity completeness

**[Template.]**

### 5.8.7–5.8.8 — Archive and reliable delivery

**[Template — see SOP-12 for the archive this section would rely on.]**

## 8. Records

| RECORD                                 | PRODUCED BY               | RETAINED IN                              |
|----------------------------------------|---------------------------|------------------------------------------|
| Release checklist / gate review record | Software Development Lead | Project technical file                   |
| Residual anomaly evaluation            | Risk Manager              | Risk management file (SOP-11)            |
| Software archive                       | Configuration Manager     | Configuration management system (SOP-12) |

## 9. Revision History

*Illustrative only — SentinelFlow 500 has no real revision history.*

| VERSION | STATUS   | DESCRIPTION                                                          |
|---------|----------|----------------------------------------------------------------------|
| 1.0     | Template | Structure and requirements table complete; Procedure section pending |