



TRAINING EXAMPLE — NOT A REAL IEC 62304 DELIVERABLE

This project is a **training site, not a real medical device** under any regulatory definition. The device described in this document — “SentinelFlow 500” — is entirely **fictional**, invented for this course: nothing here is a genuine IEC 62304 deliverable, a real regulatory submission, or evidence of any real organisation’s compliance with anything. It illustrates the *kind* of document a real SaMD/SiMD project would produce. See the [document register](#) for the full explanation.

Clause: 9 — Problem Resolution Process · **Register:** [Device Example Register](#)

Document ID: SOP-13 · **Device:** SentinelFlow 500 (fictional) · **Safety class:** C

Status: SOP — template (Procedure not yet written for SentinelFlow 500)

Fictional worked example. See [SOP-01](#) for the disclaimer that applies to this whole document set, and the [Device Example Register](#) for the SentinelFlow 500 device description and what “template” status means here.

1. Purpose

This SOP defines how every problem found in SentinelFlow 500's software — during development, testing, or after release — is reported, investigated, resolved, verified, and analysed for trends, per Clause 9. Every other SOP in this set that finds an anomaly (SOP-06 through SOP-10) hands it to this process rather than resolving it informally.

2. Scope

Applies to every problem detected in SentinelFlow 500's control software at any lifecycle stage (Class C — see [SOP-01 §7.3](#)). Covers Clause 9 in full — every sub-clause applies to all three safety classes.

3. Responsibilities

ROLE	RESPONSIBILITY UNDER THIS SOP
Reporter (any role)	Prepares a problem report for each problem detected
Software Development Lead	Investigates the problem; determines a change request or documents a rationale for no action
Risk Manager	Evaluates each problem report's effect on safety
QA/RA	Verifies resolution closed the problem, introduced no new problems, and analyses trends across problem reports

4. Definitions & Abbreviations

TERM	MEANING
Problem report	The record of a detected problem, including its criticality and information aiding resolution
Trend analysis	Reviewing problem data over time to detect a recurring, systemic design or process weakness

5. References

- IEC 62304:2006+AMD1:2015, Clause 9
- **SOP-11 — Software Risk Management File:** updated whenever a resolved problem affects the risk picture (§9.5)
- **SOP-10 — Software Maintenance Plan:** the feedback channel that raises most post-release problem reports

6. What the standard requires

REF	TITLE	APPLIES AT CLASS C	WHAT MUST BE PRODUCED
9.1	Prepare problem reports	Yes	A problem report for each problem detected, including a statement of criticality and information aiding resolution
9.2	Investigate the problem	Yes	Documented outcome of the investigation and evaluation, and either a change request or a documented rationale for taking no action
9.3	Advise relevant parties	Yes	(No standard-named output.)
9.4	Use change control process	Yes	(No standard-named output.)
9.5	Maintain records	Yes	Records of problem reports and their resolution including verification, and an updated risk management file
9.6	Analyse problems for trends	Yes	(No standard-named output.)
9.7	Verify software problem resolution	Yes	Verification that each resolution closed the problem, reversed adverse trends, was implemented in the right software and activities, and introduced no new problems
9.8	Test documentation contents	Yes	Test documentation for testing, retesting or regression testing after a change: results, anomalies, software version, hardware and software test configuration, test tools, date and tester

(Quoted directly from `applicability.json` , identical in content to the self-referential set's [13](#), whose Status is itself marked "Documented — strong evidence".)

7. Procedure

Not yet written for SentinelFlow 500 — see the [Device Example Register](#) for what "template" status means. Once complete, this section would give SentinelFlow 500's actual problem-report format, its criticality scale, and worked examples of problems raised against the hazards identified in SOP-11.

9.1–9.2 — Prepare and investigate problem reports

[Template.]

9.3–9.4 — Advise relevant parties; use change control

[Template — see SOP-12 for the change control process this section relies on.]

9.5 — Maintain records

[Template — see SOP-11, which this section keeps updated.]

9.6–9.8 — Trend analysis, resolution verification, test documentation

[Template.]

8. Records

RECORD	PRODUCED BY	RETAINED IN
Problem reports	Reporter	Problem resolution log
Investigation and resolution records	Software Development Lead	Problem resolution log
Trend analysis	QA/RA	Problem resolution log

9. Revision History

Illustrative only — SentinelFlow 500 has no real revision history.

VERSION	STATUS	DESCRIPTION
1.0	Template	Structure and requirements table complete; Procedure section pending