Review Production Records for Compliance

Nicolás Figueroa – Senior Quality Engineer
Hospira Costa Rica

Hospira

About Hospira Costa Rica



About Hospira Costa Rica

- > 2800 Employees
- Operated 24/7
- 3 Products:
  - Gravity Sets
  - Injectable Syringes
  - Infusion Pumps







“If we deliver on time, but the product has defects, we have not delivered on time."

— Philip Crosby

Identifying Production Records

Hospira

What is a production record?

- Documentation of:
  - the specifications of the product to be manufactured with the critical quality attributes of the product
  - the procedure to perform the processes required to manufacture the product including the critical process parameters that need to be controlled to guarantee successful product
  - the procedures to perform the tests required to control the product adherence to the specifications
  - the documented resulting values and data captured in data forms that supports adherence to all specifications, procedures and testing performed over the product

Hospira

What is a production record?

- There is no difference between pharma and device worlds:
  - Master Production and Control Records = Device Master Records
  - Production and Control Records = Device History Records
- Also the requirements for review are related since both pharma and devices require the review of these production records prior to the distribution of the product to the market.
- The difference between them exists only in the difference that exists between products and the corresponding processes to manufacture them, differences between the master production records.

Hospira

How are production records generated?

- Production records are generated through
  - the filling of data forms
  - electronic data recording
  - documented data analysis and conversion
  - automated data recording
- Each procedure should have a way of providing documented evidence that the process established by the procedure was followed and generate records as evidence of the process execution.

Hospira

Why are production records required?

- Production records are required because:
  - there are requirements in the regulation to have them (if you're only looking to be compliant)
- But, if you're looking to make business sense:
  - they are the documented evidence that the product is built according to expected quality attributes = supports that the product is good!
  - they guarantee the detectability of flaws in the manufacturing procedure that could be transported into the product = triggers the quality system before the issue translates into a complaint!
  - they standardize the process and makes the process and product measurable, trendable and capable of being analyzed.

Hospira

Regulatory requirements for production records review

Hospira

Regulatory Requirement for Medical Devices

- (d) Final acceptance activities. Each manufacturer shall establish and maintain procedures for finished device acceptance to ensure that each production run, lot, or batch of finished devices meets acceptance criteria. Finished devices shall be held in quarantine or otherwise adequately controlled until released. Finished devices shall not be released for distribution until:
  - (1) The activities required in the DMR are completed;
  - (2) the associated data and documentation is reviewed;
  - (3) the release is authorized by the signature of a designated individual(s); and
  - (4) the authorization is dated.

Hospira

Regulatory Requirement for Medical Devices

- (e) Acceptance records. Each manufacturer shall document acceptance activities required by this part. These records shall include:
  - (1) The acceptance activities performed;
  - (2) the dates acceptance activities are performed;
  - (3) the results;
  - (4) the signature of the individual(s) conducting the acceptance activities; and
  - (5) where appropriate the equipment used. These records shall be part of the DHR.

Hospira

Regulatory Requirements for Pharma

- Sec. 211.192 Production record review.
  - All drug product production and control records, including those for packaging and labeling, shall be reviewed and approved by the quality control unit to determine compliance with all established, approved written procedures before a batch is released or distributed.

Hospira

Regulatory Guidance for Pharma

- Sec. 211.192 Production record review (cont…)
  - Any unexplained discrepancy (including a percentage of theoretical yield exceeding the maximum or minimum percentages established in master production and control records) or the failure of a batch or any of its components to meet any of its specifications shall be thoroughly investigated, whether or not the batch has already been distributed. The investigation shall extend to other batches of the same drug product and other drug products that may have been associated with the specific failure or discrepancy. A written record of the investigation shall be made and shall include the conclusions and followup.

Hospira

ISO 13485 (Devices)

- **4.2.4 Control of records**

  Records shall be established and maintained to provide evidence of conformity to requirements and of the effective operation of the quality management system. Records shall **remain legible, readily identifiable and retrievable**. A documented procedure shall be established to define the controls needed for the **identification, storage, protection, retrieval, retention time and disposition of records**.

Hospira

Ensure Continued Compliance of Production Records

Hospira

Master Records: Specifications

- Examples of specification documents include:
  - drawings, composition, formulation, component specifications, and software specifications
- Requirements for specifications include:
  - Unique identification of the item being specified
  - Physical characteristics of the item being specified
  - Quantities involved in the item specified
  - Acceptable levels of control or quality attributes for the item specified
  - Description of packaging, labeling as applicable
  - Verification methods of the quality attributes

Hospira

Procedures

- Examples of procedures include:
  - equipment specifications,
  - production methods,
  - production procedures,
  - Packaging and labeling procedures
  - Installation, maintenance, and
  - servicing procedures and methods.
- Procedures shall be managed through Document Control requirements (change control)

Hospira

Test Methods / Processes – Quality Controls

- Test Method requirements include:
  - Acceptance criteria
  - quality assurance equipment to be used
- Methods must be traceable to known standards when feasible.
- The output of the methods must be capable of being measured and documented in the respective records

Hospira

Storage and Availability of production records

- Records must be readily available during inspection or request from regulatory authorities
  - This is a general requirement in all standards and regulation.
  - Electronic records are more easily retrievable than paper documents, specially for older records.
  - Recommended to be kept at manufacturing location as much as possible for fast retrieval
- Storage requirements
  - Records must be maintained as original records or as true copies such as photocopies, microfilm, microfiche, or other accurate reproductions of the original records.
  - Electronic records subject to storage conditions determined by Part 11.
  - Environmental protection requirements

Hospira

Auditing Production Records

- Production records must be visually inspected to confirm that the following elements are included (Drugs):
  - (1) Dates;
  - (2) Identity of individual major equipment and lines used;
  - (3) Specific identification of each batch of component or in-process material used;
  - (4) Weights and measures of components used in the course of processing;
  - (5) In-process and laboratory control results;
  - (6) Inspection of the packaging and labeling area before and after use;
  - (7) A statement of the actual yield and a statement of the percentage of the theoretical yield at appropriate phases of processing;
  - (8) Complete labeling control records, including specimens or copies of all labeling used;
  - (9) Description of drug product containers and closures;
  - (10) Any sampling performed;
  - (11) Identification of the persons performing and directly supervising or checking each significant step in the operation;
  - (12) Any investigation made according to § 211.192.
  - (13) Results of examination made in accordance with § 211.134.

Hospira

Auditing Production Records

- Production records must be visually inspected to confirm that the following elements are included (Devices):
  - (a) The dates of manufacture;
  - (b) The quantity manufactured;
  - (c) The quantity released for distribution;
  - (d) The acceptance records which demonstrate the device is manufactured in accordance with the DMR;
  - (e) The primary identification label and labeling used for each production unit; and
  - (f) Any device identification(s) and control number(s) used.

Hospira

Retention periods of production records

- Drugs:
  - 1 year after the expiration date of the batch
  - 3 years after distribution of the batch for drugs that have no expiration date.
- Devices:
  - For the lifetime of the device or at least 2 years after the device has been distributed.
- Keeping the records of the product aids in handling identified non-conforming product and complaint analysis.
- Final disposition or discard of records must be documented in a permanent record for evidence of compliance with the requirement.

Hospira

QA
Experience Sharing

Hospira

Thank You

Hospira