

KB — Cameras

The detector is half the instrument. This note holds what each camera on the roster actually is, what its optical path does to the spectrum, and what changing camera would cost and buy.

Written 2026-09-04, after the halogen measurement in `KB_lamps.md` §4 turned "the camera's red response" from an assumption into a number.

Companions: `KB_lamps.md` (the light source, and the measurement this note builds on), `KB_spectroscopy_physics.md` §7 (the physical instrument), `SPEC_real_camera_capture.md` (the capture wiring), `SPEC_capture_quality.md` §16 (the error budget), `SPEC_dev_capture_view.md` (the view every measurement here was taken in).

CONTENTS

- 0 · ★★★★★ The answer up front: a better sensor does NOT fix the pumpkin measurement
- 1 · The roster, at a glance
- 2 · ELP 32e4:8830 — the bench unit, and what the archive was captured through
- 3 · Microdia/Sonix 0c45:6366 — the production camera, and the surprise
- 4 · ★ The candidate: ToupTek GPCMOS02000KMA (IMX290 mono)
 - 4.1 What it would buy — and what it would NOT
 - ★★★★ 4.1a How much is bit depth actually worth — measured on 225 archived runs
 - 4.1b △ And gamma is not merely a nuisance — it spends codes where they are needed
 - 4.2 Coverage arithmetic — how much spectrum fits on 5.57 mm
 - 4.3 — The blockers, in the order they would bite
 - 4.4 So — can a NIR spectrometer to 1000 nm be built on this camera?
 - 4.5 △ And the question behind the question — what is at 700–1000 nm worth having?
 - ★★★★ 4.5a The AOCS colour and chlorophyll methods — what each one actually needs (2026-09-04)
 - — 4.5b The obvious reading of that is WRONG — PCI is the one we CANNOT execute
 - 4.5c What the archive says about the three terms we CAN see
 - ★★★★ 4.5d Could the device report LOVIBOND or AOCS colour? — the optics say yes at 690 nm (2026-09-04)
 - But three gates remain, and two are worse than the optics
 - ★ PCI is the exception, and it is the one within reach
- 5 · What would have to change in the code
- 6 · How to characterise any camera on this bench
- 7 · What would change these conclusions

0 · ★★★★★ The answer up front: a better sensor does NOT fix the pumpkin measurement

(Edwin, 2026-09-04: "isn't it that the biggest problem we have is sample preparation and a noise floor?" — it is, and this section is the arithmetic that settles it. Recorded here so the camera question is never re-opened on the wrong grounds.)

The whole instrument — camera, lamp, optics, quantisation, every electronic term — contributes 0.063 Q% units. That is `DevSpectralPlugin.SINGLE_WINDOW_SIGMA`: the standard deviation of Q% over ten

repeats with the jar untouched (SPEC_capture_quality.md §16.36.6). Against it:

what is being measured	Q% units	× the instrument
the entire instrument, jar untouched	0.063	1×
second pour of the same dilution (§36)	0.076	1.2×
clean set, aliquots kept dark (§40)	0.198	3.1×
σ_{fill} — five separate preparations (§28)	0.276	4.4×
archive within-fill scatter (§28)	1.255	19.9×

⇒ ★★★★★ A PERFECT detector — zero read noise, infinite bit depth, no Bayer array, no IR-cut — would improve σ_{fill} by 2.6 % and the archive scatter by 0.1 %. Not "12-bit buys 2.6 %": *anything* confined to the instrument floor buys at most that, because $\sqrt{(0.276^2 - 0.063^2)} = 0.269$ and no detector change can subtract more than the whole floor. **This is an upper bound on every camera upgrade ever proposed, and it needs no model of the new camera at all.**

⊖ So the monochrome/12-bit case must not be made on measurement quality. §4.1a re-quantised 225 archived runs to check the specific claim: quantisation contributes at most 0.063 and 12 bits shrinks it 3.1×. Real, and irrelevant. Mono's ~3× photon gain is a better argument than the bits — photon noise is genuinely in the floor — but it is subject to the identical 2.6 % ceiling.

★ What a camera change IS for, then: buying spectral range that does not exist today (§4.2–4.5), and unlocking two things the clamp forbids — **normed chlorophyll methods** (§4.5b narrows those to exactly one, needing 710 nm) and ★★ any COLOUR number at all (§4.5d: $\bar{x}$ is 11 % truncated at 632.6 nm and 0.2 % at 690.8, so a tristimulus integral goes from impossible to essentially complete). Those are capability arguments, not precision arguments. △ Judge any camera proposal by which of the two it is making — and note that both of the capability arguments are about the RED END, not the sensor.

⇒ The measurement is gated on the fill, not the detector. SPEC_settled_measurement.md is where the 4.4× lives, and until its one-fill/capillary work lands, no detector purchase moves the verdict.

1 · The roster, at a glance

	ELP 32e4:8830	Microdia/Sonix 0c45:6366	ToupTek GPCMOS02000KMA
role	bench / dev — the archive camera	intended production camera	△ candidate only, not owned
sensor	Sony IMX179 (inferred)	unrecorded	Sony IMX290LLR mono
colour	Bayer RGB	Bayer RGB	★ monochrome
native	3264 × 2448 (8 MP)	unrecorded	1945 × 1097 (2.13 MP)
captured at	2592 × 1944 (pinned)	⊖ not wired	1920 × 1080
pixel pitch	1.4 µm native	unrecorded	2.9 µm
imaging width	3.63–4.57 mm (§4.1)	unrecorded	5.57 mm
bit depth	8	8	★ 12

	ELP 32e4:8830	Microdia/Sonix 0c45:6366	ToupTek GPCMOS02000KMA
transfer curve	△ gamma-encoded (pow2.2)	△ gamma-encoded	★ linear raw
IR-cut filter	🚫 YES — measured, λ_{50} = 641.8 nm	★ NO (remote test)	★ NO — AR-coated clear window, IR-transmitting
red reach	62 DN @ 650, 17 DN @ 660 nm	△ unmeasured	△ unmeasured
interface	UVC / V4L2 → cv2	UVC / V4L2 → cv2	🚫 proprietary libtoupcam SDK
mount	M12 (S-mount)	M12	1.25" barrel + C-mount adapter
price class	~€60	★ cheap — the reason it is the production part	~€180–230

★ The single most important row is the IR-cut row, and it is the one that was assumed rather than known until 2026-09-04.




Figure 1 — what each camera can reach, and why the three limits are different kinds of limit. Solid = measured. Hatched = projected, never measured. The faded tail is signal present but below the 16 DN working floor. Only the silicon bandgap at ~1100 nm is a law of physics; the other two boundaries are purchasing decisions.

2 · ELP 32e4:8830 — the bench unit, and what the archive was captured through

Identity. CaptureBackend records the sensor maximum as 3264 × 2448 and pins capture to 2592 × 1944. 8 MP at 4:3 in that size is the ELP-USB8MP02G class, i.e. Sony IMX179, 1.4 μm pixels. △ Inferred from the resolution — there is no datasheet or invoice in the repo, and KNOWLEDGE_BASE.md §8 calls it "ELP 4K", which does not fit 3264 × 2448. Worth confirming from the purchase record.

Why 2592 × 1944 is pinned, and must stay pinned. CaptureBackend.py carries the full argument: the ROI corners and the px → nm cubic on SpectrometerCalibrationProfile were authored at that size, so

any other capture size mis-maps every wavelength. Not the sensor maximum either — that would need a recalibration and is slower still (~1.5 fps at 2592 over USB 2). ⚠ The file's own `TOD0: make this per-sensor ... when a second camera lands` is now due (§5).

🚫 **It has an IR-cut filter, and the edge is measured.** `KB_lamps.md` §4: dividing a 60 W halogen frame by Planck leaves the instrument response, which carries a dielectric edge at $\lambda_{50} = 641.8 \text{ nm}$, **10 → 90 % over 16.5 nm** — $641 \pm 1 \text{ nm}$ under every decode assumption and filament temperature. At 650 nm ~82 % and at 660 nm ~87 % of the attenuation is that edge; everything else costs under 2× across 620–660 nm.

⚠ **Two artefacts this camera imprints on every archived spectrum.** Both are the Bayer array, not the sample:

max-channel notches	The reduction takes the maximum of R, G, B, so where two channels hand over the maximum dips. Measured at B=G 486.2 nm and G=R 581.0/581.5 nm on two different lamps. The 581 one is why <code>KB_spectroscopy_physics.md</code> §4.1a reads the Q band at 568 nm rather than 574
★ and they are useful	Being a property of the colour filters, they are a free wavelength-scale check — no calibration lamp needed. Two lamp families agreeing on both to 0.5 nm is what pins the scale in <code>KB_lamps.md</code>

🚫 **Do not modify this unit.** Removing an IR-cut filter shifts the focal plane by $\approx t(1-1/n)$ — about 0.3 mm for 1 mm of glass — so the lens must be refocused, refocusing an M12 lens changes magnification, and changed magnification changes dispersion ⇒ **a full recalibration**, which breaks registration with the ~98-run archive that every fitted metric constant rests on.

3 · Microdia/Sonix 0c45:6366 — the production camera, and the surprise

`KB_spectroscopy_physics.md` §7 records this as *"the cheap Chinese cam intended for the production batch"*. Almost nothing else about it is written down: no resolution, no calibration profile (the second `spectrometer_calibration_profile` row is all NULL), and `CaptureBackend` has no branch for it.

★★ **But it has no IR-cut filter.** Edwin ran the remote test on 2026-09-04: in the dark it shows **a white dot with a purple/violet halo**.

The halo is the diagnostic, not the dot. A silicon sensor's three Bayer dyes all become transparent in the near infrared, so an unfiltered sensor renders an 850/940 nm source as white with a colour fringe — every channel responding at once. A camera *with* an IR-cut returns nothing, or a dim dot with no colour fringe.

⇒ **The production camera does not share the ELP's 642 nm wall.** That re-opens the deep-red half of `DOC_lamp_410_680.md` §5.4/§7.3 for the shipped product, and it means the de-filtered camera that `KB_lamps.md` §7 wanted to buy is **already on the bench**.

⚠ **What is not established:** "no IR-cut" ⇒ "good 660–690 nm response". Removing the *edge* leaves silicon QE and the red dye, which are near plateau there, so a good response is *expected* — expected is not measured. **The measurement is `KB_lamps.md` §7.1 and it is the cheapest open experiment in the project:** halogen → this camera → look at the raw red end for a cliff. A 16 nm collapse is scale-free, so step 1 needs no calibration at all.

⚠ **And two risks appear only once the filter is gone**, neither of which the ELP data can speak to: **stray NIR scatter** raising the floor across the whole band, and **second-order diffraction** (§4.3).

4 • ★ The candidate: ToupTek GPCMOS02000KMA (IMX290 mono)

Sold as an astronomical **guiding** camera — which is exactly why it is interesting here, because guide cameras are built to keep near-infrared rather than throw it away.

item	value	source
sensor	Sony IMX290LLR , 1/2.8", back-illuminated (STARVIS), monochrome	Sony / vendor
effective pixels	1945 × 1097 (~2.13 MP); camera outputs 1920 × 1080	Sony datasheet
unit cell	2.9 μm square	Sony datasheet
active area	5.57 × 3.13 mm , diagonal 6.39–6.46 mm	vendor / Sony
ADC	12 bit	vendor
peak QE	~81 %	vendor
read noise	0.53–0.84 e ⁻	vendor
full well	~11 200 e ⁻	vendor
exposure	0.105 ms – 1000 s	vendor
frame rate	~16–18 fps at full resolution (USB 2.0)	vendor
window	★★ AR-coated clear glass, "also transparent in the infrared"	vendor
mount	1.25" barrel; C-mount and CS-mount adapters	vendor
back focus	8.5 mm (1.25"), 17.5 mm (C), 12.5 mm (CS)	vendor
interface	USB 2.0 + ST-4; 🚫 libtoupcam SDK, not UVC	vendor / SDK docs
mass	70 g	vendor

4.1 What it would buy — and what it would NOT

⚠ **12-bit linear raw removes a whole class of complexity — but NOT a whole class of error.** The current chain is built around an 8-bit gamma-encoded frame: the quantisation window in `ImageSpectrumAcquisitionLogicModule`, the `__TIE_WINDOW_CODES` tie-breaker, the low-DN guard at 16 DN (§16.23.10f), the `pow2.2` decode and §17's "decode before you average" rule. A linear 12-bit sensor makes all of that unnecessary, and the `MAD==0` collapse (`spectracs-mad-zero-collapse`) could not occur.

🚫🚫 **What it would NOT do is make the verdict meaningfully more repeatable, and an earlier draft of this section wrongly implied it would.** `diagnostics/bit_depth_gain.py` re-quantises **225 archived runs** onto both grids and recomputes `Q%`. See §4.1a. The short version: **quantisation contributes at most 0.063 Q% units — the entire jar-untouched floor — and removing it improves σ_{fill} by 2.6 %.**

★★ 4.1a How much is bit depth actually worth — measured on 225 archived runs

(Added 2026-09-04 after Edwin's challenge: "isn't it that the biggest problem we have is sample preparation and a noise floor?" — it is, and here is the arithmetic.)

`diagnostics/bit_depth_gain.py` takes each archived report's stored **linear** reference and sample, re-quantises **both** onto today's 8-bit gamma grid and onto a 12-bit linear grid, and recomputes `Q% = -100 · (A_valley - A_Q) / A_Soret`. 225 runs, `Q%` spanning -14.9 to 39.9.

grid	mean \	shift\		sd	95th pct	max
8-bit gamma	0.0517	0.0668	0.1270	0.2360		
12-bit linear	0.0055	0.0215	0.0094	0.2853		

⇒ 12 bits shrinks the quantisation term **3.1×**. ⚠ These are a **no-dither worst case**: archived values are the mean of 60 frames, so real capture dithers and the true term is smaller.

★ **The simulation's own sanity check gives a better bound than the simulation does.** 0.0668 exceeds the measured jar-untouched floor of **0.063** (`DevSpectralPlugin.SINGLE_WINDOW_SIGMA` , §16.36.6) — and it cannot, because quantisation is one *component* of that floor. That over-shoot is positive evidence that capture dithers, and it hands us a model-free ceiling: **quantisation contributes at most 0.063 Q% units, because that is the entire spread of ten repeats with the jar untouched.**

Granting 12-bit that entire ceiling — the most generous case it could ever claim:

what is being measured	today	with 12-bit	gain
second pour of the same dilution (§36)	0.076	0.043	44 %
clean set, aliquots kept dark (§40)	0.198	0.188	5.2 %
σ_fill — five separate preparations (§28)	0.276	0.269	2.6 %
archive within-fill scatter (§28)	1.255	1.253	0.1 %

⇒ ★★ **Edwin is right. The bit depth is not where the error is.** Preparation and re-seating dominate by 4–20×, and `SPEC_capture_quality.md` §16.26 already measured that directly: instrument floor **0.42 %** against a jar re-seat of **median 1.7–3.0 %, max 14.4 %**. Buying 12 bits to improve the verdict would be **spending on the smallest term in the budget.**

4.1b ⚠ And gamma is not merely a nuisance — it spends codes where they are needed

One code, as a percentage of the level, with the reference parked at 90 % of full scale:

A	sample level	8-bit gamma	12-bit linear	12-bit is
0.5	64.26	1.62 %	0.10 %	16.7× better
1.0	20.32	2.74 %	0.31 %	9.0× better
1.5	6.43	4.66 %	0.97 %	4.8× better
2.0	2.03	7.92 %	3.06 %	2.6× better
2.5	0.64	13.56 %	9.69 %	1.4× better
3.0	0.20	23.44 %	30.64 %	🚫 0.8× worse

★ **This is why 8 bits has held up as well as it has:** `pow2.2` concentrates codes at the dark end, which is exactly where an absorbance measurement lives. A *linear* 12-bit grid is uniform, so it wins by 3–10× through the working range but actually **loses past A ≈ 3.3** — where the bin is dead on any grid.

⇒ **Where 12-bit genuinely earns its place is the STARVED regime**, not the verdict: bins at 2–7 DN, where one code is 8–23 %. That is the regime §7.13 measured as producing a *concentration-dependent compression* that corrupted `r_Q` , the regime that forced the dilution-protocol change, and the regime the **red**

extension would put us back into (the halogen gives 62 DN at 650 nm but 17 DN at 660). ★ So bit depth is an enabler for **new range**, not an improvement to the **existing verdict**.

★ **Monochrome removes the Bayer array**, and with it: the two max-channel notches (§2), the ~3× light loss to the colour filters, the demosaic, and the red dye's own roll-off. △ It also removes the free wavelength-scale check the crossovers provide — the CFL becomes the only scale reference. △ ★ The ~3× light gain is worth more than the bit depth: it is a **photon** gain, and photon noise is a real term in the floor, whereas quantisation demonstrably is not.

★ **The 1.25" barrel is a filter thread**. An order-sorting long-pass filter (§4.3) screws straight in. That is a genuinely lucky mechanical fit for the one hard requirement of NIR work.

4.2 Coverage arithmetic — how much spectrum fits on 5.57 mm

With the optics unchanged, the span a sensor holds is its **imaging width × the dispersion in nm/mm**. The ELP ROI holds 290.8 nm across 2055 of 2592 columns, so the full ELP frame holds **~367 nm**:

if the 2592 mode is...	ELP frame width	dispersion	IMX290 span	starting at 400 nm
a crop (1.40 µm pitch)	3.63 mm	101.1 nm/mm	563 nm	400–963 nm
a scale (1.76 µm pitch)	4.57 mm	80.3 nm/mm	447 nm	400–847 nm

△ **Which one is unresolved** — nobody has recorded whether the ELP's 2592 × 1944 mode crops or downscales the 3264 × 2448 array. It is one `v4l2-ctl --list-formats-ext` away.

⇒ **One IMX290 frame covers roughly 400–850 ... 400–960 nm at today's dispersion**. To *guarantee* 400–1000 (a 600 nm span) needs 107.8 nm/mm — spreading the spectrum **7–34 % less tightly**, via a coarser grating or a shorter focal length.

★ **That is affordable, because the instrument massively oversamples**. The optical resolution is ~2 nm (the Hg 576.96 / 579.07 doublet is *marginally* resolved at ~14 px), while sampling is 0.14 nm/px on the ELP and would be 0.23 nm/px on the IMX290 — roughly **10× finer than the optics deliver** in both cases. Fewer, larger pixels cost nothing here; the slit and the grating are the limit, not the detector.

4.3 — The blockers, in the order they would bite

1. — Order overlap — **this is the real one, and it is not optional**. A grating sends 2nd order of 400 nm to the same place as 1st order of 800 nm. Any instrument spanning 400 → 1000 nm therefore has the blue end **folded on top of** the infrared end. It must be fixed with an **order-sorting long-pass filter** (an OG/RG-type glass, e.g. ~695 nm) covering the red half, or by taking two exposures — one plain, one long-passed — and splicing. ★ The 1.25" filter thread makes this trivial mechanically. — Ignore it and the NIR readings are pure artefact, silently.

2. — **It is not a UVC camera**. Touptek uses a proprietary `libtoupcam` SDK on Linux; `cv2.VideoCapture` will not see it. `CaptureBackend` is V4L2-only, so this needs a **second backend implementation**. Not fatal — the SDK ships Linux `.so` builds and there is a third-party GStreamer element — but it is real work, and △ it is hostile to the **Android port** (`SPEC_android_port.md`), where a vendor `.so` plus USB permissions is a much worse story than a UVC device.

3. △ **NIR focus shift**. Camera lenses are corrected for the visible. At 900 nm the focal plane moves noticeably, so the infrared end of the spectrum would sit out of focus — which broadens lines exactly where the new range is being added. Either accept the blur (it degrades resolution, not position) or budget for an apochromatic or reflective collimator.

4. **△ A full mechanical and calibration rebuild.** New mount (C or 1.25", not M12), new back focus, a new grating holder, and a fresh ROI + px → nm calibration. **🚫 Archive registration breaks** — every metric constant is fitted on 98 runs taken through the ELP.

5. **△ Silicon runs out at ~1100 nm** and is failing well before that (§4.4).

4.4 So — can a NIR spectrometer to 1000 nm be built on this camera?

To ~850–900 nm: yes, and comfortably. The window transmits IR, the sensor is back-illuminated STARVIS that Sony explicitly designed for 850 nm, it is mono, and 12-bit. Nothing in the way but the order-sorting filter and the rebuild.

To 1000 nm: physically possible, but do not promise it. Silicon's bandgap is 1.12 eV ⇒ a hard cutoff near **1100 nm**, so 1000 nm is inside the sensor's range — but only just. The absorption depth in silicon grows from ~1 μm at 550 nm to well over 100 μm approaching 1000 nm, while a back-illuminated photodiode is only a few μm thick. ⇒ **QE at 1000 nm is single-digit percent** on any sensor of this class. Workable against a bright halogen with long exposures and frame averaging; hopeless for a dim source.

🚫 And Sony publishes no numeric QE beyond "improved sensitivity at 850 nm" and a graph. No figure for 900, 940 or 1000 nm exists in any source checked. **△ Treat 900–1000 nm as unquantified until measured.**

★ **The good news: we now own the method to measure it.** `KB_lamps.md` §4's halogen ÷ Planck division returns the instrument response of *whatever* camera it is pointed through. Applied to a new camera it answers "what is the QE at 950 nm on my actual optical stack" directly — no datasheet needed. That is the acceptance test to run on day one, before any redesign is committed.

△ A halogen is also the right source for it, and increasingly so: a 2900 K blackbody's radiance keeps climbing to ~1000 nm, so the lamp is *strongest* exactly where the sensor is weakest. The two curves work against each other in the helpful direction.

4.5 △ And the question behind the question — what is at 700–1000 nm worth having?

target	wavelength	verdict
AOCS Cc 13i-96 chlorophyll baseline	710 nm	★ comfortably in range ⇒ a normed method becomes executable
literature-comparable Kreft DI	> 700 nm	★ in range
chlorophyll a	430 / 662 nm	★ in range (hemp oil, §95.2 of the lamp memory)
the 660–680 quiet window	660–680 nm	★ in range — the pigment-free baseline anchor the metric has never had
C–H 3rd overtone (fat)	~900–930 nm	△ weak overtone; needs long path + chemometrics
O–H 2nd overtone (water)	~970 nm	△ the classic NIR moisture band — see below

★ **The 970 nm water band is the one with a business story attached.** `spectracs-alwera-group` records that ALWERA/Estyria already use a **humimeter FSA** to check contract farmers' drying, and *return goods on it*. Moisture is measured in the NIR, and 970 nm is where. **🚫 But do not read that as a product.** Moisture in seed is a **diffuse-reflectance** measurement with a chemometric calibration against oven reference values — a different instrument geometry from a transmission cuvette, a different corpus, and a

validated incumbent already in the customer's hand. It is a reason to keep the door open at 1000 nm, not a reason to claim the door leads anywhere yet.

★★ 4.5a The AOCS colour and chlorophyll methods — what each one actually needs
(2026-09-04)

The **AOCS Official Methods and Recommended Practices** carries a whole Cc 13* family for the colour of fats and oils. Read off the official 2003 method index (saved to spectracs-references/standards/AOCS_2003_method_index.pdf):

method	subject
Cc 13a-43	Color — FAC method
Cc 13b-45	Color — Wesson method (AOCS Lovibond)
★★ Cc 13c-50	Color — SPECTROPHOTOMETRIC method
Cc 13d-55	Chlorophyll pigments (refined and bleached oils)
Cc 13e-92	Lovibond (per ISO standard)
★ Cc 13i-96 (01)	Chlorophyll pigments (crude vegetable oils)
Cc 13j-97	Automated method
Ak 2-92	Chlorophyll pigments (rapeseed)

Cc 13i-96 — chlorophyll pigments in crude vegetable oils. Absorbance of the neat oil against air at three wavelengths, reported as pheophytin a:

$$c = \frac{345.3 (A_{670} - 0.5 A_{630} - 0.5 A_{710})}{L}$$

read: c is mg pheophytin a per kg of oil; L is the cell thickness in mm.

The 630 and 710 points are a two-point BASELINE under the 670 nm Qy peak — that is all they are for.

⚠ This corrects the form carried in the project record, which had the 345.3 as a divisor. It is a multiplier, and there is a path-length term. Stated detection limit: > 1 mg/kg.

Cc 13c-50 — the photometric colour index, and ★★ the more interesting one for Spectracs:

$$PCI = 1.29 A_{460} + 69.7 A_{550} + 41.2 A_{620} - 56.4 A_{670}$$

Three of its four wavelengths — 460, 550, 620 nm — are inside today's 400–630 nm window already; only **670 nm** is missing, which is precisely what the ELP's 641.8 nm IR-cut destroys.

⊖⊖ 4.5b The obvious reading of that is WRONG — PCI is the one we CANNOT execute

(diagnostics/aocs_pci_feasibility.py , 225 archived runs. A first draft of §4.5a called PCI "the nearest normed method, one wavelength away". It is the opposite, and the reason is scale, not wavelength.)

Cc 13c-50 specifies a 1-inch (25.4 mm) cell — the method was built to grade light refined oils against a Lovibond red. Scaling the archive's median A460 = 0.411 (diluted, 1.3 cm jar) back to neat oil at that cell:

dilution factor	A_{460} neat @ 25.4 mm	A_{460} neat @ 0.7 mm	path for $A_{460} = 1.0$
60	48	1.3	527 μm
100	80	2.2	316 μm
120	96	2.7	264 μm

➡ Neat pumpkin oil in the AOCS cell reads $A \approx 48\text{--}96$ — transmittance 10^{-48} . The method is **inapplicable by about fifty orders of magnitude**. PCI has no path-length term (unlike Cc 13i-96), so it is *defined* at its cell; measuring at a thinner path and scaling yields a number that is **not** the norm's PCI. You could publish a PCI-like index; you could not claim Cc 13c-50.

★★★ Cc 13i-96 is the one that survives, and for a structural reason: it carries $/L$ explicitly. Being path-length-general by construction, its *arithmetic* applies at any cell — including a thin one. $\Rightarrow$ **the wavelength that unlocks a normed method is 710 nm, not 670**, and the bar for a camera change is correspondingly higher.

△△ But only the arithmetic travels. Cc 13i-96 specifies a **5 mm or 10 mm cell**, and neat pumpkin oil reads $A_{630} \approx 3\text{--}12$ there — out of a spectrophotometer's range by 3–6 $\times$. A standard is a *procedure*, not only a formula, so whether a read at an unspecified path may be reported *per Cc 13i-96* is a question for the method text or a certifying body. ➡ **Do not promise "we execute the AOCS method" on a portable formula alone.** ★ The geometry news is nonetheless good: at this project's existing **~1 mm** cell the same oil reads $A_{630} \approx 0.6\text{--}1.2$, dead centre of the usable window — it is the *lab's* cells that are wrong for a dark oil, not ours. Full working: SPEC_metric_research.md §16.20.5d.

★ And the thin cell is not exotic — it is the geometry the trade already uses. 260–530 μm puts A_{460} at 1.0, and the **Kernöltestgerät is a backlit 0.7 mm viewer** (spectracs-oelmuehlen-verzeichnis) where the same oil reads $A \approx 1.3\text{--}2.7$. The gatekeeper's own instrument is already in the right regime.

4.5c What the archive says about the three terms we CAN see

question	answer
★ dynamic range — can one path serve all four bands?	Yes, easily. At a path giving $A_{460} = 1.0$ the bands span $A = 0.24 \dots 1.73$ ($T = 58 \% \dots 1.8 \%$) — a factor of 7.3. ➡ This corrects an earlier guess in this note that neat oil would put PCI back in the starved regime: it does not, because PCI reads 460 nm on the Soret flank , never the 432 nm peak
➡ is $1.29 \cdot A_{460} + 69.7 \cdot A_{550} + 41.2 \cdot A_{620}$ a new axis?	It is 77 % concentration ($r = +0.77$ with A_{Soret}). Divided by A_{Soret} it is uncorrelated with $Q\%$ ($r = +0.07$) — △ but that is <i>not</i> evidence of a new axis
➡ is the residual signal or noise?	Noise. Grouped by source (15 oils, ≥ 4 runs each) its between/within ratio is 0.74 , against $Q\%$'s 0.93 on the identical grouping. The residual discriminates oils <i>worse</i> than the shipped metric

$\Rightarrow$ **Everything rides on the $-56.4 \cdot A_{670}$ term** — 40 % of PCI's weight by coefficient, and the one wavelength the instrument cannot see. △ A prediction was recorded before the test (that the three visible terms would land on the metric family's single axis, per spectracs-metric-family-2026-08-21); it **failed** — they land nowhere, which is a weaker result than either alternative.

△△ But both methods measure **NEAT OIL against air, not a solvent dilution** — and that is a bigger change than the wavelength. Consequences, none of them small:

🚫 no dilution	the whole <code>SPEC_settled_measurement.md</code> protocol, σ_{fill} , the clearing gate and every fitted constant assume oil-in-isopropanol. A neat-oil method shares none of that corpus
🚫 dynamic range	neat pumpkin oil is near-opaque below ~550 nm at any normal path length. Cc 13c needs <code>A460</code> — $\triangle$ this is the starved regime again, and the one place §4.1b says bit depth genuinely helps
★ path length is the lever	AOCS specifies <code>L</code> explicitly. The Kernöltestgerät is a backlit 0.7 mm viewer (<code>spectracs-oelmuehlen-verzeichnis</code>) — that is the right order of magnitude for neat dark oil, and it is the geometry the gatekeeper already accepts
★ simpler for the operator	no capillaries, no solvent, no waiting for a fill to settle. $\triangle$ Which also removes the 4.4× error term §0 says dominates everything

⇒ ★ **A neat-oil, short-path mode is a genuinely different product shape from the diluted 0% workflow**, and §4.5b names the one that could carry a norm: **Cc 13i-96 at a thin cell, needing 710 nm**. It would answer §89's fourth condition ("*the absence of a normed alternative*") in the opposite direction — by *executing* the norm instead of avoiding it.

$\triangle\triangle$ **But count what that trades away.** The moat on record is *the validated corpus, not the formula* (`spectracs-alwera-group`). A normed number has **no corpus moat by construction** — anyone with a spectrophotometer computes the same figure. What would remain is form factor and price: ~€900 at the press against a lab instrument plus sample transport, which is `spectracs-international-market` §91's "only non-destructive / at-the-mill survives" argument, now with a normed number attached instead of a proprietary one. 🚫 Whether that trade is good is a business question this note cannot answer.

🚫 **Nothing here is a plan.** It is the first time the norms have been read closely enough to see what they cost, and the headline is that one of the two is arithmetically impossible on this oil.

$\triangle$ **On the source.** AOCS Official Methods is a **copyrighted commercial publication** — there is no legitimate free full text, and none was obtained. What is on disk is the **2003 method INDEX** (22 pp, numbers and titles only). The two formulas above are quoted from the open literature, not from the standard. 🚫 **Before anything is certified against either method, buy the actual method text** (AOCS sells them individually) — an index cannot tell you the cell, the blanking or the tolerances.

★★ 4.5d Could the device report LOVIBOND or AOCS colour? — the optics say yes at 690 nm (2026-09-04)

A **band metric** (`Q%` , `Rv` , `PCI`) needs a handful of wavelengths. A **colour** number — Lovibond, AOCS-Tintometer/Wesson, a literature Kreft dichromaticity index — is a **tristimulus integral** and needs the whole visible band. `spectracs-colorimeter-idea` records the objection in one line: "*with the 440–630 nm window a literature-comparable DI is NOT computable ($\bar{x}$ runs past 700).*"

★★ That is true of the CLAMP and false of the INSTRUMENT.

`diagnostics/cie_truncation_cost.py` measures the fraction of each D65-weighted CIE 1931 2° colour-matching function lying above each cutoff — once for a white source, once weighted by what an archived oil actually transmits:

cutoff	$\bar{x}$ (red)	$\bar{y}$ (luminance)	$\bar{z}$ (blue)
632.6 nm — the pipeline clamp	9.9 % / 11.1 %	3.6 % / 3.7 %	0.0 %
690.8 nm — the extended ROI	0.22 % / 0.25 %	0.08 %	0.0 %
780.0 nm — the full visible	0.00 %	0.00 %	0.0 %

(illuminant-weighted / sample-weighted. Δ The sample figures hold the measured absorbance FLAT past 632.6 nm at its 620–630 value; real transmittance rises into the red, so they are **lower bounds**.)

⇒ ★★ $\bar{x}$ — the red primary — is the only one truncated, and the extended ROI closes it. 780 nm is not needed. $\bar{z}$ is finished long before either cutoff and $\bar{y}$ almost. A colour number is not computable at 632.6 nm and is computable at 690.8 nm — on the filterless camera already on the bench, through the ROI the capture view already draws. ★ This is the sharpest single argument for the red extension yet produced: it turns a *categorical* "not computable" into a **0.2 %** residual.

⊖ But three gates remain, and two are worse than the optics

gate	
⊖⊖ the Lovibond scale is PROPRIETARY	converting a spectrum to R/Y/B needs Tintometer's glass transmission data and their matching algorithm — licensed, not published. It is why a PFX990 can do it and a generic spectrophotometer cannot. An approximation is possible; "approximate Lovibond" is not Lovibond. Δ This is the binding constraint, and it is contractual, not technical
⊖ scope	Cc 13j-97 is "refined oils only ... providing no turbidity is present" (§4.5a) ⇒ cold-pressed is out whatever we compute
Δ path length	Lovibond is defined at 133 mm . We would measure at ~1 mm and scale by Beer–Lambert — valid arithmetic, but a 133× extrapolation, and neat pumpkin at 133 mm is $A_{460} \approx 250\text{--}500$, off-scale in principle. Plausible for a light oil
Δ photometric accuracy	a ratio metric tolerates shape distortion that a tristimulus integral does not — stray light especially. ★ Not a problem: the 486/581 nm max-channel notches (§2) are common to reference and sample and cancel in $T = S/R$

★ PCI is the exception, and it is the one within reach

	Lovibond / AOCS-Tintometer	PCI (Cc 13c-50)
needs the full visible band	✓ yes — ★ solved at 690 nm	⊖ no — four band means
needs proprietary data	⊖ yes	✓ no
computable by us	⊖ only as an approximation	★ yes
reportable as the norm	⊖ scope + cell	⊖ cell — §4.5b's fifty decades on pumpkin

⇒ The device could compute PCI on a light oil at the right cell, and a Lovibond-like number on anything — but could not legitimately report either as the AOCS value. ⊖ The remaining obstacles are procedural and contractual; the optical one closes at 690 nm.

★★★ And there is a better target than approximating a licensed scale: CIELAB is ISO/CIE 11664 — open, and exactly what modern food colorimetry uses. ⇒ emit a *fully standard, non-indicative* coordinate any lab reproduces exactly, and keep a declared-indicative Lovibond bridge as a separate courtesy. Two jobs; only one carries liability. ★★ ★ The number that actually matters here is Kreft's dichromaticity index, which §4.5d's 0.2 % makes *literature-comparable rather than a proxy* — and whose whole value is legitimacy, which a proxy cannot supply. ⇒ the red extension is a precondition for that argument, not an enhancement of it. The commercial reading is spectracs-references/business/SPEC_oelmuehlen_verzeichnis.md §142 (with §92/§93, which still conclude: feature yes, business no).

5 · What would have to change in the code

△ None of this is proposed work — it is the cost side of §4, so the trade is visible.

change	where	why
capture resolution per sensor	<code>CaptureBackend.py</code> — its own <code>TODO: make this per-sensor when a second camera lands</code>	2592 × 1944 is hardcoded and is the ELP's calibration size. Blocks <i>any</i> second camera, including the Microdia
a non-V4L2 backend	new sibling of <code>CaptureBackend</code>	<code>libtoupcam</code> is not UVC (§4.3)
bit depth through the chain	<code>ImageSpectrumAcquisitionLogicModule</code> , <code>SpectralColorUtil</code>	everything assumes 8-bit gamma-encoded: the decode, the tie window, the 16 DN guard
mono path	the same	max-channel over three channels is meaningless with one channel
a second calibration profile	already supported	per- <code>SpectrometerProfile</code> , so this is data, not code

★ The first row is required by the Microdia experiment too, so it is not specific to a camera purchase.

6 · How to characterise any camera on this bench

1. **Remote test in the dark** — white dot with a **violet halo** ⇒ no IR-cut; nothing or a dim dot ⇒ filter.
2. **Halogen frame, then divide by Planck** (`KB_lamps.md` §4) ⇒ the instrument response of that whole optical stack, with no datasheet and no reference detector.
3. **Look for a cliff before doing any arithmetic** — a dielectric edge is a 15–25 nm collapse and is scale-free, so it is visible before the camera is calibrated at all.
4. **CFL frame for the scale** — Hg 435.83 / 546.07 for a two-point linear px → nm (`KB_spectroscopy_physics.md` §7.2 did this at 0.5057 nm/px), then the full cubic once the ROI is authored.
5. △ **Never compare two cameras' absolute levels** — only their shapes. Exposure, aperture, gain and transfer curve all differ.

7 · What would change these conclusions

- ★★ **The Microdia halogen frame (§3)** ⇒ decides whether the production camera already has the red range. Cheapest open experiment in the project; needs no purchase and no calibration.
- A `v4l2-ctl --list-formats-ext` on the ELP ⇒ settles crop-vs-scale and halves §4.2's uncertainty.
- **The ELP purchase record** ⇒ confirms or refutes the IMX179 identification in §2.
- **A halogen ÷ Planck run on an IMX290** ⇒ the only thing that would turn §4.4's 900–1000 nm from unquantified into a number.
- 🚫 **A measured NIR QE in the low single digits at 900 nm** ⇒ the 1000 nm ambition is dead and the honest ceiling is ~850 nm.

- ★★ **A repeat of §4.1a's arithmetic against a re-seat-free protocol** ⇒ the only thing that would make bit depth matter is removing the term that dwarfs it. `SPEC_settled_measurement.md` 's capillary/one-fill work is that; until it lands, no detector change moves the verdict.