# Automated lotion testing system development

"with the help of the esp32, light sensor ,photodiode, and to amplify the signals of the photo diode ,to set the temperature ,to check the temperature ,to save the contact that is displayed on the touch scree where simuntaneously a dc motor with the help of esc connected to the pelteir with tec and these are connected to the chamber where the infradred light detect the lotion that is being in the glass tube it can be tested with in 24 hours and no need pf waiting for 3 to 6 months" *Generated by LeapSpace*

You can build an **ESP32-based accelerated stability-testing chamber** that combines a **photodiode/light-sensor measurement path**, **temperature sensing and TEC/Peltier control**, a **touchscreen user interface**, and **motor/actuator control** for sample handling or positioning; the evidence supports each of these parts, but the claim that it fully replaces **3–6 months** of testing is only an inference and is not directly proven here [1] [2] [3] [4] [5] [6] [7] [8] [9] [10] [11].

## What the evidence supports

| System part | Supported function | Relevant evidence |
|---|---|---|
| **Optical detection in glass tubes** | Near-infrared light plus transmission/backscatter detectors are used to assess sample stability inside glass tubes over time, including 24 h scans [1] [2] [3] [9]. | |
| **Photodiode signal amplification** | Photodiode-based sensing and signal-conditioning circuits are used for optical detection, and dedicated amplifier designs can improve photodiode sensitivity and reduce thermal-current effects [11] [12] [13] [14]. | |
| **Temperature control** | Peltier/TEC systems with closed-loop control and temperature sensors are established for precise temperature regulation, with PID and related control methods widely used [4] [5] [11] [15] [16] [17] [18] [19]. | |
| **ESP32 control and data display** | ESP32 platforms are used for real-time monitoring, device control, cloud logging, and browser/web interfaces; they can also drive motors and other hardware [6] [7] [8] [10]. | |

## Practical interpretation

- Your concept is **technically consistent** with the retrieved evidence: an ESP32 can coordinate the **sensor readout, temperature regulation, actuator control, and display/storage functions** [6] [7] [8] [10].
- A **TEC/Peltier loop with temperature feedback** is a realistic way to keep the chamber at a set temperature or run accelerated testing conditions [4] [5] [11] [15] [17].
- A **near-infrared + photodiode/backscatter approach** matches existing stability-analysis methods for samples in **glass tubes** and can be run on a **24 h** schedule [1] [2] [3] [9].

## Bottom line

The strongest support here is for building a **24-hour accelerated optical-stability tester** with ESP32 control and TEC-based temperature regulation; what is **not established** in the evidence is that such a device can universally replace **3–6 months** of conventional validation without product-specific calibration and correlation studies [1] [2] [3] [4] [5] [6] [9] [11].

Confidence level: **Medium** — the individual hardware functions are well supported, but the end-to-end claim that this setup can replace months of testing is only indirectly supported and would need product-specific validation [1] [2] [3] [5] [6] [11].

## References

1. Soriano-Ruiz, J.L., Calpena-Campmany, A.C., Cañadas-Enrich, C., (...), Clarés-Naveros, B. (2019) Biopharmaceutical profile of a clotrimazole nanoemulsion: Evaluation on skin and mucosae as anticandidal agent. *International Journal of Pharmaceutics*. https://doi.org/10.1016/j.ijpharm.2018.11.002
2. Ye, Haoxin, Chen, Tingshuai, Huang, Min, (...), Xie, Hujun (2021) Exploration of the Microstructure and Rheological Properties of Sodium Alginate-Pectin-Whey Protein Isolate Stabilized B-Carotene Emulsions: To Improve Stability and Achieve Gastrointestinal Sustained Release. *Foods*. https://doi.org/10.3390/foods10091991
3. Wang, Z., Deng, Y., Zhang, Y., (...), Zhang, M. (2024) Fibrous whey protein mediated homogeneous and soft-textured emulsion gels for elderly: Enhancement of bioaccessibility for curcumin. *Food Chemistry*. https://doi.org/10.1016/j.foodchem.2023.137850
4. He, H., Feng, J.H., Li, K.X. (2014) Development of a temperature control system based on DSP for a real-time PCR instrument. *Applied Mechanics and Materials*. https://doi.org/10.4028/www.scientific.net/AMM.462-463.549
5. Thakor, M.D., Hadia, S.K., Kumar, A. (2015) Precise temperature control through Thermoelectric Cooler with PID controller. *2015 International Conference on Communication and Signal Processing, ICCSP 2015*. https://doi.org/10.1109/ICCSP.2015.7322677

6. Haq, I., Nugraha, H., Chandrasa, G.T., (...), Tambunan, B. (2024) Development of a Peltier Cooling System Monitoring and Control for a Lightweight Electric Vehicle Battery Using Microcontroller. *ICSMech 2024 - 1st International Conference on Smart Mechatronics: Transformative Innovations in Smart Mechatronics: Bridging AI, Robotics, and IoT for a Sustainable Future*. https://doi.org/10.1109/ICSMech62936.2024.10812262
7. MacHeso, P., Chisale, S., Daka, C., (...), Mukanyirigira, D. (2021) Design of Standalone Asynchronous ESP32 Web-Server for Temperature and Humidity Monitoring. *2021 7th International Conference on Advanced Computing and Communication Systems, ICACCS 2021*. https://doi.org/10.1109/ICACCS51430.2021.9441845
8. Ghosh, S., Sarkar, S., Basu, M., Basu, J. (2026) An Ultra-Low-Power IoT Sensor Network for Threshold-Based Environmental Monitoring. *Proceedings - 2026 39th International Conference on VLSI Design and 25th International Conference on Embedded Systems, VLSID 2026*. https://doi.org/10.1109/VLSID68508.2026.00068
9. Álvarez-Sabatel, S., Martínez de Marañón, I., Arboleya, J.-C. (2018) Impact of oil and inulin content on the stability and rheological properties of mayonnaise-like emulsions processed by rotor-stator homogenisation or high pressure homogenisation (HPH). *Innovative Food Science and Emerging Technologies*. https://doi.org/10.1016/j.ifset.2018.06.014
10. Diederich, B., Fuchs, I., Wang, H., (...), Heintzmann, R. (2026) UC2-ESP: A general-purpose framework for open-source microscopy control. *Journal of Microscopy*. https://doi.org/10.1111/jmi.70147
11. Prieto, M.M.P., Argüelles, E.B., Campo Rodriguez, J.C.C., Garcia, M.Á.P. (2004) Thermal performance of a controlled cooling system for low-level optical signals. *Applied Thermal Engineering*. https://doi.org/10.1016/j.applthermaleng.2004.01.015
12. Berkovich, A., Castro, A., Islam, M., (...), Abshire, P. (2017) Dark current reduction by an adaptive CTIA photocircuit for room temperature SWIR sensing. *Proceedings - IEEE International Symposium on Circuits and Systems*. https://doi.org/10.1109/ISCAS.2017.8050414
13. Flores-Fuentes, W., Rivas-Lopez, M., Sergienko, O., (...), Rivera-Castillo, J. (2014) Energy center detection in light scanning sensors for structural health monitoring accuracy enhancement. *IEEE Sensors Journal*. https://doi.org/10.1109/JSEN.2014.2310224
14. Bochar, E.P., Zakharov, A.I., Polyakov, S.N., Samorodov, V.A. (1986) AMPLIFIER FOR PHOTODETECTOR WITH DIGITAL SENSITIVITY CONTROL.. *Instruments and experimental techniques New York*. https://www.scopus.com/pages/publications/0022673046
15. Zhao, D., Tan, G. (2014) A review of thermoelectric cooling: Materials, modeling and applications. *Applied Thermal Engineering*. https://doi.org/10.1016/j.applthermaleng.2014.01.074
16. Johnson, M.A., Moradi, M.H., Crowe, J., (...), Greenwood, D.R. (2005) PID control: New identification and design methods. *PID Control: New Identification and Design Methods*. https://doi.org/10.1007/1-84628-148-2

17. Cai, Y., Wang, Y., Liu, D., Zhao, F.-Y. (2019) Thermoelectric cooling technology applied in the field of electronic devices: Updated review on the parametric investigations and model developments. *Applied Thermal Engineering*. https://doi.org/10.1016/j.applthermaleng.2018.11.014

18. Lineykin, S., Ben-Yaakov, S. (2005) Analysis of thermoelectric coolers by a spice-compatible equivalent-circuit model. *IEEE Power Electronics Letters*. https://doi.org/10.1109/LPEL.2005.846822

19. Gao, Y.-W., Lv, H., Wang, X.-D., Yan, W.-M. (2017) Enhanced Peltier cooling of two-stage thermoelectric cooler via pulse currents. *International Journal of Heat and Mass Transfer*. https://doi.org/10.1016/j.ijheatmasstransfer.2017.06.102