



TrioDocs

Version: 0.4.0

Date: May 21, 2025

Download the latest version on:
<https://triodocs.org>

Close the Loop

Device Configuration

This section covers the first two sections of Trio settings: [Closed Loop](#) and [Devices](#)

Closed Loop

Closed loop is turned off by default. This means Trio cannot make adjustments autonomously. Until you enable closed loop, the system relies on your approval of its actions.

When Should I Close the Loop?

Closed loop allows the system to make insulin adjustments automatically. Your goal for looping is to enable this setting to reduce the time spent managing your diabetes. However, entering closed loop too early can make it difficult for new loopers wanting to optimize their application settings. If this is your first time looping, this setting should remain disabled until your profile settings (CR, ISF, basal rates) are correctly set. You should monitor Trio's predictions and suggestions to see if they are appropriate for you. Please look at the corresponding CR, ISF, and basal rates sections for more information on their adjustment.

Settings

I've been using Loop. Can I copy my settings?

No. Users coming from Loop usually adjust their ISF and CR settings to compensate for fluctuations in insulin needs throughout the day. When using those settings for Trio, the algorithm also adjusts to compensate for the fluctuations, which may result in over- or under-dosing. Please review the steps of monitoring Trio predictions and change your settings accordingly. It would help if you also read the section on [Coming from Other AID Systems](#).



Tip

Remember that everyone's diabetes is different, and there is no one-size-fits-all, so take the time to optimize your settings for the Oref algorithm.