



TrioDocs

Version: 0.4.0

Date: July 03, 2026

Download the latest version on:  
<https://triodocs.org>

# Meal Settings

# Meal Settings

## Limits

### Max Carbs

**Default:** 250 g

**Setting Limits:** 0-300 g

The maximum value of carbs in grams allowed per meal entry in the bolus calculator.

### Max Fat

**Default:** 250 g

**Setting Limits:** 0-300 g

The maximum value of fat in grams allowed per meal entry in the bolus calculator.



#### Tip

This will only be visible if [Enable Fat and Protein Entries](#) is turned on.

### Max Protein

**Default:** 250 g

**Setting Limits:** 0-300 g

The maximum value of protein in grams allowed per meal entry in the bolus calculator.



#### Tip

This will only be visible if [Enable Fat and Protein Entries](#) is turned on.

---

## Max Meal Absorption Time

**Default:** 6 hours

**Setting Limits:** 4-10 hours

Carb entries will be fully decayed by the number of hours specified as Max Meal Absorption Time. Meals that are high in fat and/or protein can have long lasting effects on BG levels. To allow such late meal effects to be considered by the carb decay model, a longer Max Meal Absorption Time than the default 6 hours can be set.

If carb entries decay too slowly, it is possible to set a lower than default setting. But this should typically be addressed by tuning ISF and CR settings instead, which in combination determines the rate of carb decay.

## Gastroparesis

If you have gastroparesis, increasing this setting may help keep carbs around for delayed digestion. Please also adjust [Recommended Bolus Percentage](#) to a lower percent to avoid lows due to delayed digestion.

## Enable Fat and Protein Entries

**Default:** *OFF*

Enabling this setting allows you to log fat and protein, which are then converted into future carb equivalents (FPU) using the Warsaw Method.

### Warsaw Method

The Warsaw Method helps account for the delayed glucose spikes caused by fat and protein in meals. It uses Fat-Protein Units (FPU) to calculate the carb effect from fat and protein. The system spreads insulin delivery over several hours to mimic natural insulin release, helping to manage post-meal glucose spikes.

For more information, click to see the [Omnicalculator](#).

**\*\*Fat Conversion\*\***

$$F = fat(g) \times 90\%$$

**\*\*Protein conversion\*\***

$$P = protein(g) \times 40\%$$

**\*\*FPU Conversion\*\***

$$F + P = FPU(g)$$

or

$$(fat(g) \times 90\%) + (protein(g) \times 40\%) = FPU(g)$$

You can personalize the conversion calculation by adjusting the following settings that will appear when this option is enabled:

- [Fat and Protein Delay](#)
- [Maximum duration](#)
- [Spread Interval](#)
- [Fat and protein percentage](#)

## Fat and Protein Delay

**Default:** 60 min

**Setting Limits:** 15-120 min

This setting defines the time between when you log fat & protein and when the system start delivering insulin for the Fat-Protein Unit Carb Equivalents (FPU's).

This delay accounts for the slower absorption of fat and protein, as calculated by the Warsaw Method, ensuring insulin delivery is properly times to manage glucose spikes caused by high-fat, high-protein meals.

**Increasing** this setting results in FPU entries that begin later.

**Decreasing** this setting results in FPU entries that begin earlier.

---

## Maximum Duration

**Default:** 8 hours

**Setting Limits:** 5-12 hours

This sets the maximum length of time that Fat and Protein Carb Equivalents (FPU's) will be extended over from a single Fat and/or Protein bolus calculator entry.

It is one factor used in combination with the Fat and Protein Delay, Spread Interval, and Fat and Protein Factor to create the FPU entries.

**Increasing** this setting may result in **more** FPU entries with **smaller** carb values.

**Decreasing** this setting may result in **fewer** FPU entries with **larger** carb values.

---

## Spread Interval

**Default:** 30 min

**Setting Limits:** 10-60 min

This determines how many minutes will be between individual Fat-Protein Unit Carb Equivalent (FPU) entries from a single Fat and/or Protein bolus calculator entry.

The shorter the interval, the smoother the correlating dosing result.

**Increasing** this setting may result in **fewer** FPU entries with **larger** carb values.

**Decreasing** this setting may result in **more** FPU entries with **smaller** carb values.

---

## Fat and Protein Percentage

**Default:** 50%

**Setting Limits:** 10-120%

This setting changes how much effect the fat and protein entry has on the FPU calculations.

At the default setting of 50%, this reduces the converted amount by half.

Use the grid below to see how adjusting this percentage changes the fat and protein conversion rates:

10%

### Minimum Adjustment

This is the smallest allowed FPU conversion in Trio

$$(F \times 10\%) + (P \times 10\%) =$$

$$(fat(g) \times 90\% \times 10\%) + (protein(g) \times 40\% \times 10\%) =$$

**10% Adjusted FPU Calculation:**

$$(fat(g) \times 9\%) + (protein(g) \times 4\%) = FPU(g)$$

30%

### Reduced Adjustment

$$(F \times 30\%) + (P \times 30\%) =$$

$$(fat(g) \times 90\% \times 30\%) + (protein(g) \times 40\% \times 30\%) =$$

**30% Adjusted FPU Calculation:**

$$(fat(g) \times 27\%) + (protein(g) \times 12\%) = FPU(g)$$

50%

### Default

$$(F \times 50\%) + (P \times 50\%) =$$

$$(fat(g) \times 90\% \times 50\%) + (protein(g) \times 40\% \times 50\%) =$$

**50% Adjusted FPU Calculation:**



#### Tip

You may find your normal carb ratio needs to increase to a larger number (weaker) when you begin adding fat and protein entries. For this reason, it is best to start with a factor of 50% and adjust as you find the right balance between carb ratio and FPU conversion.

---

## How Fat and Protein Entry Settings Work Together

***Coming Soon!***

---