



TrioDocs

Version: 0.4.0

Date: April 15, 2026

Download the latest version on:
<https://triodocs.org>

Migrating from AndroidAPS

Coming from AAPS

What to Expect After Install

- **Onboarding Wizard**

Trio will walk you through the initial setup of the app. Learn more in the [New User Setup Guide](#).

- **User Interface**

You can find more about the Trio User Interface in [this walk-through](#) of the app.

- **Remote Commands & iOS Shortcuts**

Trigger actions like carb entry, bolus, overrides, or temporary targets [remotely](#) or through [shortcut automation](#).

- **Use Loop Follow v4.0 or Later for remote commands**

More info at [LoopFollowDocs](#)

- **Bolus Calculator**

The [Bolus Calculator](#) in Trio offers simple dosing with clear breakdowns and safety logic.

- **In-App Statistics**

There are multiple statistics and graphs available in the Trio app. Click the link to learn more about the [Statistics](#) available in the app.

- **Important Re: Autotune**

Autotune has not performed as it was intended for years with the addition of Dynamic ISF and the change from a single ISF and CR in Therapy settings. For this reason, we have removed it until it can be rewritten to work with Trio or a new Autotune-like feature can be built.

What to Expect in Your First 3 Hours

Here are some things to keep in mind after moving from AndroidAPS to Trio 0.5+ in the first 3 hours:

- **Onboarding Wizard**

Take this time to go through the Onboarding Wizard. Reference the [New User Setup Guide](#) or ask questions on [Facebook](#) or [Discord](#) if you have any questions.

- **Test Settings**

This is a great time to test your settings so you can start fresh on solid footing.

- **User Interface**

Use this time to learn the [User Interface](#).

- **Dynamic ISF**

- Dynamic ISF is disabled for the first **7 days**.
 - If you were using Dynamic ISF in AAPS, you will have to wait a week before enabling dynamic ISF in Trio.
 - Data import from AAPS/Nightscout is not available.

- **Settings** When migrating from AAPS to Trio, many users will achieve similar results by using similar settings. The core algorithm is the same between AndroidAPS and Trio.



Dynamic ISF & Closed Loop

Trio must be in **Closed Loop** for 7 days in order to enable Dynamic ISF.

- **Autosens**

Autosens will require **8–24 hours** of data before it starts making adjustments. *You may need to enter carbs and bolus during this time if you aren't already doing so.*



Initial Limitation with Autosens

There's a long standing issue with Autosens where it will likely be stuck at 1 for the first 24 hours when there are carb entries or SMBs.

- When carb entries or SMBs **are not** present, Autosens will kick in after 8 hours.
- When carb entries or SMBs **are** present, Autosens will kick in after 24 hours.

Whether or not to enable SMBs or enter carbs is a personal choice based on what you prioritize in your first 24 hours.



Do not use a pump or cgm simulator

- Doing so will result in false data being stored and you will have to delete the app with all the data and reinstall before you can use it on a live pump.
- This will also reset your 7-day waiting period for Dynamic ISF.
- [More information on simulator use](#)

What to Expect After 24 Hours

A full day on Trio! Congratulations! Here's what to look for:

- **Autosens**

Autosens now has enough data to make adjustments.

- **Review & Test Settings**

This is a good time to review your last 24 hours of results to see if your core settings performed as expected. If you had unexpected challenges with your settings, look at how to test them over the next 7 days:

- [Basal Testing](#)
- [ISF Testing](#)
- [Carb Ratio Testing](#)

- **Review Dynamic ISF Documentation**

Review the documentation on [Using Dynamic ISF](#) in preparation for [Day 7](#).

What to Expect After 7 Days

You've completed a week on Trio! Here's what you should see:

- **Dynamic ISF**

- Dynamic ISF now has enough data. You now have the option to enable it.
- The Logarithmic formula in Trio is the most similar to AAPS dynamic ISF. The Sigmoid formula behaves very differently (informative links below). Which one you use is up to you.
- This is a great time to refresh [Using Dynamic ISF](#) and check your settings for [Logarithmic Dynamic ISF](#) or [Sigmoid Dynamic ISF](#).



Can't Enable Dynamic ISF?

If Dynamic ISF does not give you the option to enable, you may have experienced one or more of the following:

- You were not in closed loop for the full 7 days.
 - *Trio needs 7 days of closed loop to safely enable Dynamic ISF*
- You had significant loss of connection with your CGM and/or pump.
 - *Check your last week of Looping Statistics in the Stats tab in the Trio app. You need both **7 days** and an **85% success rate** to enable Dynamic ISF.*
- You enabled a significant number of manual Temp Basals with long run times.
 - *When a manual Temp Basal is set, Trio is unable to complete a loop cycle for the duration of the temp basal. This will cause a reduction in your success rate. If you do not have **85% success rate** for **7 days**, you cannot enable Dynamic ISF.*

Convert AAPS Settings to Trio Settings

When you move your settings from AAPS to Trio, there are multiple settings that have similar names. We've created this chart to help you identify which settings are equivalent between the two apps and where to find them.



Some Settings May Need Conversion

- To convert a value from a decimal to a percentage, multiply it by 100.
- To convert from a percentage to a decimal, divide by 100.

Therapy Settings

Trio Name	Setting Format (example)	Location in Trio	AAPS Name	Setting Format (example)	Location in AAPS
Glucose Targets	decimal (100 mg/dL 5.5 mmol/L)	Settings → Therapy → Glucose Targets	TARG	decimal range (100-120 mg/dL)	LP → TARG
Basal Rates	decimal (1.0 U/hr)	Settings → Therapy → Basal Rates	BAS	decimal (1.0 U/hr)	LP → BAS
Carb Ratios	decimal (10 g/U)	Settings → Therapy → Carb Ratios	IC	decimal (10 g/U)	LP → IC
Insulin Sensitivities	decimal (54 mg/dL/U 3.0 mmol/L/U)	Settings → Therapy → Insulin Sensitivities	ISF	decimal (54)	LP → ISF

Units and Limits

Trio Name	Setting Format (example)	Location in Trio	AAPS Name	Setting Format (example)	Location in AAPS
Glucose Units	selection (mg/dL or mmol/L)	Settings → Therapy → Units and Limits	Units	selection (mg/dL or mmol/L)	Setup Wizard → Units
Maximum Insulin On Board (IOB)	decimal (2 U)	Settings → Therapy → Units and Limits → Maximum Insulin On Board (IOB)	Maximum total IOB OpenAPS can't go over	decimal (2)	Preferences → OpenAPS SMB → Maximum total IOB OpenAPS can't go over
Maximum Bolus	decimal (10 U)	Settings → Therapy → Units and Limits → Maximum Bolus	Max allowed bolus	decimal (10.0)	Preferences → Treatments safety → Max allowed bolus

Maximum Basal Rate	decimal (2 U/hr)	Settings → Therapy → Units and Limits → Maximum Basal Rate	Max U/h a Temp Basal can be set to	decimal (4.0)	Preferences → OpenAPS SMB
Maximum Carbs on Board (COB)	decimal (120 g)	Settings → Therapy → Units and Limits → Maximum Carbs on Board (COB)	Max allowed carbs	decimal (120)	Preferences → Treatments safety → Max allowed carbs
Minimum Safety Threshold	decimal (60 mg/dL)	Settings → Therapy → Units and Limits → Minimum Safety Threshold	BG level below which low glucose suspend occurs	decimal (65)	Preferences → OpenAPS SMB → Enable dynamic sensitivity (ON) → BG level below which low glucose suspend occurs

Algorithm Settings

Autosens

Trio Name	Setting Format (example)	Location in Trio	AAPS Name	Setting Format (example)	Location in AAPS
Autosens Min	percentage (70%)	Settings → Algorithm → Autosens → Autosens Min	Min autosens ratio	decimal (0.7)	Preferences → Absorption Settings → Advanced Settings Preferences → Min autosens ratio
Autosens Max	percentage (120%)	Settings → Algorithm → Autosens → Autosens Max	Max autosens ratio	decimal (1.2)	Preferences → Absorption Settings → Advanced Settings → Max autosens ratio
Rewind Resets Autosens	toggle (On/Off)	Settings → Algorithm → Autosens → Autosens Max	Log a Pump Site Change	<i>Always On</i>	<i>Always On</i>

SMB (Super Micro Bolus)

Trio Name	Setting Format (example)	Location in Trio	AAPS Name	Setting Format (example)	Location in AAPS
Enable SMB Always	toggle (On/Off)	Settings → Algorithm → Super Micro Bolus (SMB) → Enable SMB Always	Enable SMB always	toggle (On/Off)	Preferences → Enable SMB → Enable SMB always
Enable SMB With COB	toggle (On/Off)	Settings → Algorithm → Super Micro Bolus (SMB) → Enable SMB Always(OFF) → Enable SMB With COB	Enable SMB with COB	toggle (On/Off)	Preferences → Enable SMB → Enable SMB with COB
Enable SMB with TempTarget	toggle (On/Off)	Settings → Algorithm → Super Micro Bolus (SMB) → Enable SMB Always(OFF) → Enable SMB With TempTarget	Enable SMB with temp targets	toggle (On/Off)	Preferences → Enable SMB → Enable SMB with temp targets
Enable SMB After Carbs	toggle (On/Off)	Settings → Algorithm → Super Micro Bolus (SMB) → Enable SMB Always(OFF) → Enable SMB After Carbs	Enable SMB after carbs	toggle (On/Off)	Preferences → Enable SMB → Enable SMB after carbs
Enable SMB With High Glucose	toggle (On/Off)	Settings → Algorithm → Super Micro Bolus (SMB) → Enable SMB Always(OFF) → Enable SMB With High Glucose	No Equivalent Setting	N/A	N/A
High Glucose Target	decimal (110 mg/dL / 6.1 mmol/L)	Settings → Algorithm → Super Micro Bolus (SMB) → Enable SMB With High Glucose			

		→ High Glucose Target	No Equivalent Setting	N/A	N/A
Allow SMB with High Temp Target	toggle (On/Off)	Settings → Algorithm → Super Micro Bolus (SMB) → Allow SMB With High Temp target	Enable SMB with high temp targets	toggle (On/Off)	Preferences → OpenAPS SMB → Enable SMB → Enable SMB with high temp targets
Enable UAM	toggle (On/Off)	Settings → Algorithm → Super Micro Bolus (SMB) → Enable UAM	Enable UAM	toggle (On/Off)	Preferences → OpenAPS SMB → Enable UAM
Max SMB Basal Minutes	decimal (30 min)	Settings → Algorithm → Super Micro Bolus (SMB) → Max SMB Basal Minutes	Max minutes of basal to limit SMB to	decimal (30)	Preferences → OpenAPS SMB → Enable SMB → Max minutes of basal to limit SMB to
Max UAM Basal Minutes	decimal (30 min)	Settings → Algorithm → Super Micro Bolus (SMB) → Max UAM Basal Minutes	Max minutes of basal to limit SMB to for UAM	decimal (30)	Preferences → OpenAPS SMB → Enable UAM → Max minutes of basal to limit SMB to for UAM
Max Allowed Glucose Rise for SMB	percentage (20%)	Settings → Algorithm → Super Micro Bolus (SMB) → Max Allowed Glucose Rise for SMB	No Equivalent Setting	N/A	N/A

Dynamic Settings

Trio Name	Setting Format (example)	Location in Trio	AAPS Name	Setting Format (example)	Location in AAPS
Dynamic ISF	selection (Disabled / Logarithmic / Sigmoid)	Settings → Algorithm → Dynamic Settings → Dynamic ISF	Enable dynamic sensitivity	toggle (On/Off)	Preferences → OpenAPS SMB → Enable dynamic sensitivity

Adjustment Factor	percentage (80%)	Settings → Algorithm → Dynamic Settings → Dynamic ISF (Logarithmic) → Adjustment Factor	DynamicISF Adjustment Factor %	percentage (100)	Preferences → OpenAPS SMB → Enable dynamic sensitivity (ON) → DynamicISF Adjustment Factor %
Sigmoid Adjustment Factor	percentage (50%)	Settings → Algorithm → Dynamic Settings → Dynamic ISF (Sigmoid) → Sigmoid Adjustment Factor	No Equivalent Setting	N/A	N/A
Weighted Average of TDD	percentage (35%)	Settings → Algorithm → Dynamic Settings → Dynamic ISF (Logarithmic/Sigmoid) → Weighted Average of TDD	No Equivalent Setting	N/A	N/A
Adjust Basal	toggle (On/Off)	Settings → Algorithm → Dynamic Settings → Dynamic ISF (Logarithmic/Sigmoid) → Adjust Basal	Enable TDD based sensitivity ratio for basal and glucose target modification	toggle (On/Off)	Preferences → OpenAPS SMB → Enable dynamic sensitivity (ON) → Enable TDD based sensitivity ratio for basal and glucose target modification

Target Behavior

Trio Name	Setting Format (example)	Location in Trio	AAPS Name	Setting Format (example)	Location in AAPS
High Temp Target Raises Sensitivity	toggle (On/Off)	Settings → Algorithm → Target Behavior → High Temp Target Raises Sensitivity	High temp-target raises sensitivity	toggle (On/Off)	Preferences → OpenAPS SMB → High temp-target raises sensitivity

Low Temp Target Lowers Sensitivity	toggle (On/Off)	Settings → Algorithm → Target Behavior → Low Temp Target Lowers Sensitivity	Low temp-target lowers sensitivity	toggle (On/Off)	Preferences → OpenAPS SMB → Low temp-target lowers sensitivity
Sensitivity Raises Target	toggle (On/Off)	Settings → Algorithm → Target Behavior → Sensitivity Raises Target	Sensitivity raises target	toggle (On/Off)	Preferences → OpenAPS SMB → Sensitivity raises target
Resistance Lowers Target	toggle (On/Off)	Settings → Algorithm → Target Behavior → Resistance Lowers Target	Resistance lowers target	toggle (On/Off)	Preferences → OpenAPS SMB → Resistance lowers target
Half Basal Exercise Target	decimal (160 mg/dL 8.9 mmol/L)	Settings → Algorithm → Target Behavior → Half Basal Exercise Target	Activity Target Value	decimal (160 mg/dL 8.9 mmol/L)	Preferences → Overview Default → Temp-targets → Activity target value

Additional

 **Warning**

The settings in this section typically do not require any modifications.
Do not alter them without a solid understanding of what you are changing and the full impact it will have on the algorithm.

Trio Name	Setting Format (example)	Location in Trio	AAPS Name	Setting Format (example)	Location in AAPS
Max Daily Safety Multiplier	percentage (300%)	Settings → Algorithm → Additional → Max Daily Safety Multiplier	Max daily safety multiplier	decimal (3.0)	Preferences → OpenAPS SMB → Advanced Settings → Max daily safety multiplier

Current Basal Safety Multiplier	percentage (400%)	Settings → Algorithm → Additional → Current Basal Safety Multiplier	Current basal safety multiplier	decimal (4.0)	Preferences → OpenAPS SMB → Advanced Settings → Current basal safety multiplier
Duration of Insulin Action	decimal (10 hr)	Settings → Algorithm → Additional → Duration of Insulin Action	DIA	decimal (10)	LP → DIA
Use Custom Peak Time	toggle (On/Off)	Settings → Algorithm → Additional → Use Custom Peak Time	Free-Peak Oref	selection (Free- Peak Oref)	Config Builder → Profile → Insulin → Free-Peak Oref
Insulin Peak Time	decimal (65 min)	Settings → Algorithm → Additional → Use Custom Peak Time (ON) → Insulin Peak Time	Free-Peak Oref	decimal (45)	Config Builder → Profile → Insulin → Free-Peak Oref → Peak Time
Skip Neutral Temps	toggle (On/Off)	Settings → Algorithm → Additional → Skip Neutral Temps	No Equivalent Setting	N/A	N/A
Unsuspend If No Temp	toggle (On/Off)	Settings → Algorithm → Additional → Unsuspend If No Temp	No Equivalent Setting	N/A	N/A
SMB Delivery Ratio	percentage (50%)	Settings → Algorithm → Additional → SMB			

Delivery Ratio			No Equivalent Setting	Always 50%	Always 50%
SMB Interval	decimal (3 min)	Settings → Algorithm → Additional → SMB Interval	How frequently SMBs will be given in min	decimal (3)	Preferences → Enable SMB → How frequently SMBs will be given in min
Min 5m Carb Impact	decimal (8 mg/dL)	Settings → Algorithm → Additional → Min 5m Carb Impact	min_5m_carbimpact	decimal (8)	Preferences → Absorption settings → Advanced → min_5m_carbimpact
Remaining Carbs Percentage	percentage (100%)	Settings → Algorithm → Additional → Remaining Carbs Percentage	No Equivalent Setting	N/A	N/A
Remaining Carbs Cap	decimal (90g)	Settings → Algorithm → Additional → Remaining Carbs Cap	No Equivalent Setting	N/A	N/A
Noisy CGM Target Increase	percentage (130%)	Settings → Algorithm → Additional → Noisy CGM Target Increase	No Equivalent Setting	N/A	N/A

Features

Bolus Calculator

Trio Name	Setting Format (example)	Location in Trio	AAPS Name	Setting Format (example)	Location in AAPS
Display Meal Presets	toggle (On/Off)	Settings → Features → Bolus Calculator →			

		Display Meal Presets	Quick Wizard	N/A	Preferences → Overview → QuickWizard Settings → Add / edit Quick Wizard
Recommended Bolus Percentage	percentage (80%)	Settings → Features → Bolus Calculator → Recommended Bolus Percentage	Deliver this part of bolus wizard result	percentage (100)	Preferences → Overview → Deliver this part of bolus wizard result
Enable Reduced Bolus	toggle (On/Off)	Settings → Features → Bolus Calculator → Enable Reduced Bolus	No Equivalent Setting	N/A	N/A
Reduced Bolus Percentage	percentage (30%)	Settings → Features → Bolus Calculator → Enable Reduced Bolus (ON) → Reduced Bolus Percentage	No Equivalent Setting	N/A	N/A
Enable Super Bolus	toggle (On/Off)	Settings → Features → Bolus Calculator → Enable Super Bolus	Enable super bolus functionality in wizard	toggle (On/Off)	Preferences → Overview → Advanced → Enable super bolus functionality in wizard
Super Bolus Percentage	percentage (100%)	Settings → Features → Bolus Calculator → Enable Super Bolus (ON) → Super Bolus Percentage	No Equivalent Setting	N/A	N/A
Very Low Glucose Warning	toggle (On/Off)	Settings → Features → Bolus Calculator → Very Low Glucose Warning	No Equivalent Setting	N/A	N/A

Meal Settings

Trio Name	Setting Format (example)	Location in Trio	AAPS Name	Setting Format (example)	Location in AAPS
Max Carbs	decimal (250 g)	Settings → Meal Settings → Max Carbs	Max allowed carbs	decimal (100)	Preferences → Treatments safety → Max allowed carbs
Max Protein	decimal (250 g)	Settings → Meal Settings → Enable Fat and Protein Entries (ON) → Max Protein	No Equivalent Setting	N/A	N/A
Max Fat	decimal (250 g)	Settings → Meal Settings → Enable Fat and Protein Entries (ON) → Max Fat	No Equivalent Setting	N/A	N/A
Max Meal Absorption Time	decimal (6 hr)	Settings → Meal Settings → Max Meal Absorption Time	Meal max absorption time	decimal (6)	Preferences → Absorption Settings → Advanced → Meal max absorption time
Enable Fat and Protein Entries	toggle (On/Off)	Settings → Meal Settings → Enable Fat and Protein Entries	No Equivalent Setting	N/A	N/A
Fat and Protein Delay	decimal (60 min)	Settings → Meal Settings → Enable Fat and Protein Entries (ON) → Fat and Protein Delay	No Equivalent Setting	N/A	N/A
Maximum Duration	decimal (8 hr)	Settings → Meal Settings → Enable Fat and Protein Entries (ON) → Maximum Duration	No Equivalent Setting	N/A	N/A

Spread Interval	decimal (30 min)	Settings → Meal Settings → Enable Fat and Protein Entries (ON) → Spread Interval	No Equivalent Setting	N/A	N/A
Fat and Protein Percentage	percentage (50%)	Settings → Meal Settings → Enable Fat and Protein Entries (ON) → Fat and Protein Percentage	No Equivalent Setting	N/A	N/A

Other Notable Settings

Trio Name	Setting Format (example)	Location in Trio	AAPS Name	Setting Format (example)	Location in AAPS
Smooth Glucose Value	toggle (On/Off)	Settings → Devices → CGM → Smooth Glucose Value	Smoothing	selection (No smoothing / Exponential smoothing / Average smoothing)	Config Builder → Smoothing

 Custom Automations

Trio does not offer certain custom automations that are available in AAPS, such as automating a hypo temp target when glucose drops below a certain value. This can, however, be achieved one of two ways:

- Enable "Sensitivity Raises Target" which will automatically increase your target glucose when sensitivity ratio is below 100%.
- Enable "High Temp Target Raises Sensitivity" and save a custom temp target of <100 mg/dL (5.5 mmol/L) under adjustments that you can manually enable when recovering from a low glucose.