

#### Available online at www.sciencedirect.com

# **ScienceDirect**



# Review Article Recent advances in electrochemiluminescence devices for point-of-care testing



Wenyue Gao<sup>1,2</sup>, Muhammad Saqib<sup>1,2</sup>, Liming Qi<sup>1,2</sup>, Wei Zhang<sup>1,\*</sup> and Guobao Xu<sup>1,\*</sup>

A dynamic progress of methodologies has increased demand for high performance detection technologies for point-of-care testing (POCT). Electrochemiluminescence (ECL) is now established as an important, highly sensitive detection strategy for the development of POCT devices. In this short review, we summarize the recent advances of portable ECL devices, such as portable power sources, bipolar ECL devices, wireless ECL devices, ECL detectors, and microfluidic chips. Moreover, we address the remaining challenges and future perspectives to integrate ECL sensing devices into point-of-care solutions.

#### **Addresses**

- <sup>1</sup> State Key Laboratory of Electroanalytical Chemistry, Changchun Institute of Applied Chemistry, Chinese Academy of Sciences, Changchun, Jilin 130022, PR China
- <sup>2</sup> University of Chinese Academy of Sciences, Beijing 100049, PR China

#### Current Opinion in Electrochemistry 2017, 3:4-10

This review comes from a themed issue on **Sensors and Biosensors 2017** 

#### Edited by Robert Forster

For a complete overview see the Issue and the Editorial

Available online 11 March 2017

http://dx.doi.org/10.1016/j.coelec.2017.03.003

2451-9103/© 2017 Elsevier B.V. All rights reserved.

# Introduction

Due to the urgent need of fast diagnosis at or near patients, point-of-care testing (POCT) was proposed and have attracted immense attention in the past few years [1–5]. It aims to provide rapid analysis and show health-care trends using portable, low-cost and user-friendly devices [4]. POCT devices are highly desirable in global market especially in the developing countries for hospital/nonhospital settings, and for patients at home. Many sensors for POCT have been designed based on different analytical methods including electrochemical and optical methods [5,6], and some of them have been successfully commercialized [1,7]. The successful implementation and application of these sensors still has several challenges and largely depends on the development of analytical and engineering techniques.


Electrochemiluminescence (ECL), also known as electrogenerated chemiluminescence, is a powerful analytical technique and has been extensively studied [8–27]. As the combination of chemiluminescence and electrochemistry, ECL possesses several advantages, such as simplicity, rapidity, high sensitivity and easy to control. Due to its excellent features, it has been widely used in many fields including bioanalysis, immunoassays, clinical diagnostics, food and environmental monitoring, biowarfare agent detection, etc. [28–30]. It receives increasing attention for POCT devices. Figure 1 shows the key components of a point-of-care device including the methodologies used to drive ECL reactions, the luminophores and detectors. This review summarizes the recent development of power supply methods, ECL detectors, and ECL microchips.

### Electrical energy supply devices Wired electrical energy supply devices

In traditional ECL detection methods, the luminescence is initiated and controlled by an electrochemical workstation. Due to high cost of electrochemical workstation, great efforts have been made to develop alternative power sources for ECL POCT. Portable and low-cost rechargeable battery is one example [31–33]. The advantages including high efficiency, output voltage stability, large capacity and low self-discharge rate, make rechargeable batteries appropriate for the fabrication of portable devices for POCT and on-site test [34°]. An inexpensive lightrechargeable supercapacitor was also used to supply voltage to screen-printed carbon electrodes for ECL light generation by Rusling and co-workers [35\*\*]. Moreover, a mobile phone that can serve the basic functions of a potentiostat in controlling an applied potential was also demonstrated by Hogan and co-workers to initiate ECL emission [36°°]. The voltage was initiated from the headphone jack of the device, and modulated by an audio output. Later, they further devised a novel, universal, standards-based approach exploiting the USB On-The-Go (USB-OTG) specification [37°]. In this approach, current can be drawn from any USB-OTG certified mini-USB port and modulated via the audio jack using a tonedetection circuit. A self-powered system was developed to further simplify and miniaturize the design of ECL sensors. A portable thermal-powered ECL visual sensor composed of a tiny power supply and a facile prepared electrode array was successfully developed by Chen's group (Figure 2a) [38]. The unique power supply is valuable for the miniaturization of ECL devices for field operation and

<sup>\*</sup>Corresponding authors: Xu, Guobao (guobaoxu@ciac.ac.cn)

Figure 1.



The elements of ECL sensing process and the recent developments of ECL sensing devices access to POCT.

POCT. A self-powered 3D microfluidic ECL biosensing platform based on the origami technique was developed by Zhang et al. and used to detect glucose to demonstrate the sensing function of the design [39°]. By assembling the energy part and the sensing part together in a 3D paper chip, this proposed microfluidic origami ECL device exhibits broad application prospect in point-of-care diagnostics.

## Bipolar electrochemistry devices

In contrast to conventional three-electrode ECL, bipolar electrochemistry-based ECL methods only needs one conductor which serves as both cathode and anode without any direct electrical connection [40°°]. Bipolar ECL systems (BPE) are usually simple, low-cost, highthroughput, convenient to operate, easy to be miniaturized and thus ideally suited for application in POCT. Recently, bipolar electrochemistry has attracted much attention and has been widely used in constructing ECL sensors for chemical and biological analysis [41–44]. A number of attempts have been made to develop signal amplification approaches to improve the sensitivity and broaden its analytical applications [45-48]. Also, ECL sensing platforms based on multichannel bipolar systems have been proposed to demonstrate its potential in multianalysis [44,49,50]. Wang and co-workers constructed a novel microfluidic-based bipolar system with two-directional driving electrodes and dual-channel mode [51\*\*]. As shown in Figure 2b, this novel BPE system achieved 100% current efficiency, and completely eliminated the background ECL signals from the driving electrodes. Conventional bipolar ECL is a two-dimensional process where light emission is strictly confined to the geometric surface area of the electrode. To overcome this intrinsic limitation, the bulk generation of ECL in a threedimensional configuration using BPE as the driving force was demonstrated by Sojic et al. [52\*\*] As shown in Figure 2c, a suspension of multi-walled carbon nanotubes where each one acts as an individual ECL nano-emitter was polarized in a bipolar electrochemical manner and ECL light was generated simultaneously from the entire solution. The higher sensitivity of this approach provides attractive potential in visual detection.


#### Wireless energy transmission devices

BPE-ECL can avoid the direct contact of electrochemical instruments with sensing electrodes, but they still require a direct contact of electrochemical instruments with driving electrodes in solutions. Besides, signal-on sensing electrodes in BPE-ECL sensors usually suffer from the background noise from driving electrodes. Thus, new ECL methods are still required to overcome these serious drawbacks. In 2014, our group proposed a wireless ECL minidevice based on the wireless energy transmission technique [53\*\*]. Wireless energy transmission (wireless power transmission) is the transmission of electrical energy from a power source to an electrical load without using conductors [54]. As shown in Figure 2d, it consists of a disposable transmitter and a coiled energy receptor that used as the electrode. It can avoid the direct electrical contact of electrochemical instruments with any electrodes. It is promising for the development of portable ECL devices for various applications including POCT and field analysis. Later, a mini-diode that can rectify alternating current into direct current was embedded into a wireless ECL electrode microarray chip and thus strikingly enhanced the ECL intensity [55\*\*]. This enables the visual detection using ordinary cameras or smartphones as facile and inexpensive detectors. We believe that this technique has great potential in construction of ECL minidevices integrating power supplier, and signal detector for POCT.

# Light detection technologies

Signal collection is one of the most important parts in ECL sensing systems. It is often accomplished using a photomultiplier tube (PMT) because of its higher sensitivity. The demand for portable, cost-effective and userfriendly ECL detectors for POCT devices with acceptable sensitivity has given rise to use of alternative light sensors, such as charge-coupled devices (CCD), complementary metal oxide semiconductors (CMOS) devices, and silicon and organic photodiodes [56]. The use of a digital camera as an ECL detector was also reported to reduce the cost of the measurement [33,38,57 •• -59]. Smartphones, thanks to their high imaging and computing capabilities, open-source operation systems, increasingly play

Figure 2.



(a) Design of the thermo-powered high-throughput visual ECL sensor. Reprinted with permission from [38]. Copyright 2013 American Chemical Society. (b) Fundamental principle of the dual-channel bipolar ECL sensor. Reprinted with permission from [51\*\*]. Copyright 2013 American Chemical Society. (c) Principle of bulk ECL produced by a suspension of multi-walled carbon nanotubes. They are polarized by the electric field generated between the feeder electrodes in a bipolar electrochemical manner. Reproduced from Ref. [52\*\*] with permission from the Royal Society of Chemistry. (d) Schematic description of the wireless ECL system (A), the photo (B) and ECL image (C) of the wireless ECL minidevice using two short gold wires connected with small lacquered copper coils. Reprinted with permission from [53\*\*]. Copyright 2014 American Chemical Society.


a pivotal role in recreational activities and healthcare delivery [7,60]. Possessing various advanced features, smartphone has emerged as a promising digital platform for the development of various bioanalytical devices for rapid, real-time, and point-of-care monitoring, which can significantly simplify design and reduce cost of the detection systems [56,60–64]. The portability and ubiquitous availability of smartphone offer great convenience for the development of portable ECL devices. The integration of ECL sensing elements into "all-in-one device" is one attractive direction. It is forecasted that the development of portable ECL sensors will greatly depend on the advance of light detection and analysis technologies. Many existing techniques may pave the way for this goal, such as wireless power technology, three-dimensional printing technology, Wireless Fidelity and Bluetooth transmission protocols.

#### The integrated portable ECL devices

The integration of ECL sensing processes including sample collection, power supply and signal detection into

one minidevice is an exciting research area and will play an important role in POCT. Microfluidic platforms can be easily integrated with various analytical techniques to fabricate miniaturized systems. Their various useful capabilities, such as low cost, short assay time, low sample and reagents consumption, and the possibility for automation and integration of several analytical steps into one device, enable them to be widely used in the diagnostic and life science fields [65]. Different materials have been reported for preparing the microfluidics ranging from silicon and glass to soft materials, such as polymer (e.g., poly(dimethylsiloxane), PDMS) and paper [63]. Since Whitesides and co-workers proposed the first patterned paper as a microfluidic platform for bioassays [66], the paper-based devices have attracted rapidly increasing interest. It holds great potential for applications in POCT due to low-cost, portability, flexibility, and the porous structure of paper enables it to easily immobilize sensing materials with reduced sample volumes [67]. The combination of paper microfluidics with ECL was first demonstrated by Hogan and co-workers in the year

Figure 3.



(a) Schematic diagram of a paper-based microfluidic ECL sensor. Using inkjet-printed paper fluidic substrates and screen-printed electrodes, this ECL sensor can be read with mobile phone cameras. Reprinted with permission from [57\*\*]. Copyright 2011 American Chemical Society. (b) A mobile phone used for ECL sensing. The audio jack supplies the potential to the paper microfluidic sensor, while the resultant ECL emission is detected by the camera in video mode. Both the excitation and detection processes can be controlled by a software application which can also transmit the results via e-mail. The black plastic sleeve surrounding the top of the phone holds the sensor adjacent to the camera and blocks ambient light [36\*\*]. Copyright 2013, reprinted with permission from Elsevier.

of 2011 [57 ••]. By utilizing inkjet-printed paper fluidic substrates and screen-printed electrodes, the authors constructed low-cost and disposable ECL sensors which can be read-out with conventional photodetectors or mobile phone cameras (Figure 3a). Since then, several paperbased ECL sensors have been constructed [67,68]. The integration of disposable paper-based platform with bipolar electrode was first demonstrated by Chen's group for the sensitive ECL determination of prostate-specific antigen [40°°]. Later, many paper-based ECL sensors utilizing bipolar electrodes have been proposed for the analysis of biomolecules [33,69] and DNAs [70]. Besides, an electrophoretic separation technique was successfully coupled with microfluidic paper-based analytical device with an on-column wireless ECL detector for the first time by Ge et al. [71]. Yu's group carried out extensive works on paper-based ECL origami devices for sensitive detection of DNAs [72], proteins [73–77], cancer cells [78–80], etc. As shown in Figure 3b, the integration of power resource and signal detector into smartphone was developed by Hogan et al. The power was initiated from the headphone jack of the device, and modulated by an audio output. The light emitted from the reaction was easily detected using the in-built phone camera. It significantly reduced the analysis cost due to generation, detection and analysis of the ECL signal in a single step.

# **Conclusions and future perspectives**

ECL methods have been widely used in constructing sensors for bioanalysis, immunoassays, clinical diagnostics, food and environmental monitoring, and an increasing attention has been made to develop ECL sensors for POCT due to its higher sensitivity, easy to control, realtime analysis and high throughput readout. Recent developments in bipolar electrochemistry, microfluidic chips, wireless power and data transmission, and light detection technologies have unprecedentedly improved the performance of ECL sensors toward POCT. There is no doubt that the significant advances and major breakthroughs in POCT in the next few years will greatly enhance the development of these emerging technologies. Research efforts toward integrating a variety of processes (including sample collection, separation and detection) into a miniaturized analytical system will continue. The smart combination of different techniques and elements to fabricate integrated ECL sensors, and ensuring high specificity, sensitivity and reproducibility of the sensors in POCT are still challenges. Finally, these devices will play much important roles in pharmaceutical assay, food safety, environmental monitoring, homeland security, and other pointof-care applications.

# **Acknowledgments**

Funding: This work was supported by the National Key Research and Development Program of China (Grant number 2016YFA0201300), the National Natural Science Foundation of China (grant number 21475123), and the Chinese Academy of Sciences (CAS)-the Academy of Sciences for the Developing World (TWAS) President's Fellowship Programme (Grant number 2013-053).

# References and recommended reading

Papers of particular interest, published within the period of review, have been highlighted as:

- · of special interest
- • of outstanding interest.
- Luppa PB, Müller C, Schlichtiger A, Schlebusch H: Point-of-care testing (POCT): current techniques and future perspectives. TRAC Trends Anal Chem 2011, 30:887-898.
- Gubala V, Harris LF, Ricco AJ, Tan MX, Williams DE: Point of care diagnostics: status and future. Anal Chem 2012, 84:487-515.

- Kaushik A, Vasudev A, Arya SK, Pasha SK, Bhansali S: Recent advances in cortisol sensing technologies for point-of-care application. Biosens Bioelectron 2014, 53:499-512.
- Vashist SK, Luppa PB, Yeo LY, Ozcan A, Luong JHT: Emerging technologies for next-generation point-of-care testing. Trends Biotechnol 2015, 33:692-705.
- Luppa PB, Bietenbeck A, Beaudoin C, Giannetti A: Clinically relevant analytical techniques, organizational concepts for application and future perspectives of point-of-care testing. *Biotechnol Adv* 2016, **34**:139–160.
- Vashist SK, Mudanyali O, Schneider EM, Zengerle R, Ozcan A: Cellphone-based devices for bioanalytical sciences. Anal Bioanal Chem 2014, 406:3263-3277.
- Ozcan A: Mobile phones democratize and cultivate next-generation imaging, diagnostics and measurement tools. Lab Chip 2014, 14:3187–3194.
- Hu L, Xu G: Applications and trends in electrochemiluminescence. Chem Soc Rev 2010, 39:3275-3304.
- Liu Z, Qi W, Xu G: Recent advances in electrochemiluminescence. Chem Soc Rev 2015, 44:3117-3142.
- Richter MM: Electrochemiluminescence (ECL). Chem Rev 2004, 104:3003-3036.
- 11. Bertoncello P, Forster RJ: Nanostructured materials for electrochemiluminescence (ECL)-based detection methods: recent advances and future perspectives. Biosens Bioelectron 2009, 24:3191-3200.
- 12. Forster RJ, Bertoncello P, Keyes TE: Electrogenerated chemiluminescence. Annu Rev Anal Chem 2009, 2:359-385.
- 13. Chen X-M, Su B-Y, Song X-H, Chen Q-A, Chen X, Wang X-R: Recent advances in electrochemiluminescent enzyme biosensors. TRAC Trends Anal Chem 2011, 30:665-676.
- 14. Li J, Guo S, Wang E: Recent advances in new luminescent nanomaterials for electrochemiluminescence sensors. RSC Adv 2012, 2:3579.
- 15. Nepomnyashchii AB, Bard AJ: Electrochemistry and electrogenerated chemiluminescence of BODIPY Dyes. Acc Chem Res 2012. 45:1844-1853.
- 16. Wu P, Hou X, Xu J-J, Chen H-Y: Electrochemically generated versus photoexcited luminescence from semiconductor nanomaterials: bridging the valley between two worlds. *Chem Rev* 2014, **114**:11027–11059.
- 17. Kirschbaum SEK, Baeumner AJ: A review of electrochemiluminescence (ECL) in and for microfluidic analytical devices. Anal Bioanal Chem 2015, 407:3911-3926.
- 18. Li S-P, Guan H-M, Xu G-B, Tong Y-J: Progress in molecular imprinting electrochemiluminescence analysis. Chin J Anal Chem 2015, 43:294-299.
- 19. Bano K, Rusling JF: Electrochemiluminescence arrays for studies of metabolite-related toxicity. Electroanalysis 2016, **28**:2636–2643.
- 20. Bouffier L, Arbault S, Kuhn A, Sojic N: Generation of electrochemiluminescence at bipolar electrodes: concepts and applications. Anal Bioanal Chem 2016, 408:7003-7011.
- 21. Cui H, Paolucci F, Sojic N, Xu G: Analytical electrochemiluminescence. Anal Bioanal Chem 2016, 408:7001-7002.
- 22. Hai H, Huang W-G, Liang S-C, Li J-P: Ultrasensitive electrochemiluminescence biosensor for DNA determination based on nicking endonuclease assisted signal amplification. Chin J Anal Chem 2016, 44:779-786.
- 23. Han S, Zhang Z, Li S, Qi L, Xu G: Chemiluminescence and electrochemiluminescence applications of metal nanoclusters. Sci China Chem 2016, 59:794-801.
- 24. Hesari M, Ding Z: Review electrogenerated chemiluminescence: light years ahead. J Electrochem Soc 2016, 163:H3116-H3131

- 25. Majeed S, Gao W, Zholudov Y, Muzyka K, Xu G: Electrochemiluminescence of acridines. Electroanalysis 2016, 28:2672-2679.
- Valenti G, Fiorani A, Li H, Sojic N, Paolucci F: Essential role of electrode materials in electrochemiluminescence applications. Chemelectrochem 2016, 3:1990-1997.
- 27. Muzyka K, Saqib M, Liu Z, Zhang W, Xu G: Progress and challenges in electrochemiluminescent aptasensors. Biosens Bioelectron 2017. http://dx.doi.org/10.1016/j.bios.2017.01.015.
- 28. Miao W: Electrogenerated chemiluminescence and its biorelated applications. Chem Rev 2008, 108:2506-2553.
- Deng S, Ju H: Electrogenerated chemiluminescence of nanomaterials for bioanalysis. Analyst 2013, 138:43-61.
- Bertoncello P, Stewart AJ, Dennany L: **Analytical applications of nanomaterials in electrogenerated chemiluminescence**. *Anal* Bioanal Chem 2014, 406:5573-5587,
- 31. Wang S, Ge L, Zhang Y, Song X, Li N, Ge S, Yu J:
  Battery-triggered microfluidic paper-based multiplex
  electrochemiluminescence immunodevice based on potential-resolution strategy. Lab Chip 2012, 12:4489-4498.
- Yang H, Kong Q, Wang S, Xu J, Bian Z, Zheng X, Ma C, Ge S, Yu J: Hand-drawn&written pen-on-paper electrochemiluminescence immunodevice powered by rechargeable battery for low-cost point-of-care testing. Biosens Bioelectron 2014, 61:21-27.
- Chen L, Zhang C, Xing D: Paper-based bipolar electrode-electrochemiluminescence (BPE-ECL) device with battery energy supply and smartphone read-out: a handheld ECL system for biochemical analysis at the point-of-care level. Sensor Actuat B, Chem 2016, 237:308-317.
- 34. Wang S, Dai W, Ge L, Yan M, Yu J, Song X, Ge S, Huang J:

   Rechargeable battery-triggered electrochemiluminescence detection on microfluidic origami immunodevice based on two electrodes. Chem Commun 2012, 48:9971–9973.

   A rechargeable battery was used to fabricate a screen-printed two-electrode electrochemiluminescent immunoassay on a 3D

microfluidic origami electronic device.

 35. Kadimisetty K, Mosa IM, Malla S, Satterwhite-Warden JE,
 Kuhns TM, Faria RC, Lee NH, Rusling JF: 3D-printed supercapacitor-powered electrochemiluminescent protein immunoarray. Biosens Bioelectron 2016, 77:188-193

The authors reported a 3D-printed microfluidic immunoarray for multiple protein detection, which was powered by an inexpensive light-rechargeable supercapacitor to screen-printed carbon electrodes for ECL light generation.

36. Delaney JL, Doeven EH, Harsant AJ, Hogan CF: Reprint of: Use of a mobile phone for potentiostatic control with low cost paper-based microfluidic sensors. Anal Chim Acta 2013,

The authors demonstrate that a mobile phone with suitable software installation can serve the basic functions of a potentiostat in controlling an applied potential to oxidise ECL-active molecules. Using the in-built phone camera, the ECL emission from the reaction can be easily detected. Combined with paper microfluidic sensors, it opens new possibilities for low-cost and instrument-free sensing

- 37. Doeven EH, Barbante GJ, Harsant AJ, Donnelly PS, Connell TU,
  Hogan CF, Francis PS: Mobile phone-based electrochemiluminescence sensing exploiting the 'USB On-The-Go' protocol. Sensor Actuat B, Chem 2015, 216:608–613.

A universal low-cost system to generate, control and detect ECL signals using a mobile phone and related USB On-The-Go ready devices is described.

- 38. Hao N, Xiong M, Zhang J-D, Xu J-J, Chen H-Y: Portable thermo-powered high-throughput visual electrochemiluminescence sensor. Anal Chem 2013, 85:11715-11719.
- 39. Zhang X, Li J, Chen C, Lou B, Zhang L, Wang E: A self-powered microfluidic origami electrochemiluminescence biosensing platform. Chem Commun 2013, 49:3866-3868

The authors developed a self-powered 3D microfluidic ECL biosensing platform, which integrated with a stable, environmentally-friendly and noble metal-free primary battery (C|FeCl3|NaCl|AlCl3|Al). Combined

with the origami technique, this device is portable, low-cost, disposable and convenient to use in POCT.

- 40. Feng Q-M, Pan J-B, Zhang H-R, Xu J-J, Chen H-Y: Disposable
- paper-based bipolar electrode for sensitive electrochemiluminescence detection of a cancer biomarker. Chem Commun 2014, 50:10949-10951

The integration of disposable paper-based platform with bipolar electrode was demonstrated for the first time for the sensitive ECL detection of a prostate specific antigen. This paper-based bipolar sensing platform is simple, low-cost, portable, and may provide more opportunities for the development of analytical methods and diagnostic platforms.

- Sentic M, Arbault S, Goudeau B, Manojlovic D, Kuhn A, Bouffier L, Sojic N: **Electrochemiluminescent swimmers for dynamic** enzymatic sensing. Chem Commun 2014, 50:10202-10205.
- 42. Wu S, Zhou Z, Xu L, Su B, Fang Q: Integrating bipolar electrochemistry and electrochemiluminescence imaging with microdroplets for chemical analysis. Biosens Bioelectron 2014,
- 43. Essmann V, Jambrec D, Kuhn A, Schuhmann W: Linking glucose oxidation to luminol-based electrochemiluminescence using bipolar electrochemistry. Electrochem Commun 2015, 50:77-80.
- 44. Zhang H-R, Wang Y-Z, Zhao W, Xu J-J, Chen H-Y: **Visual** color-switch electrochemiluminescence biosensing of cancer cell based on multichannel bipolar electrode chip. Anal Chem 2016. 88:2884-2890.
- 45. Wu M-S, Qian G-S, Xu J-J, Chen H-Y: Sensitive electrochemiluminescence detection of c-Myc mRNA in breast cancer cells on a wireless bipolar electrode. Anal Chem 2012, 84:5407-5414.
- Shi H-W, Wu M-S, Du Y, Xu J-J, Chen H-Y: Electrochemiluminescence aptasensor based on bipolar electrode for detection of adenosine in cancer cells. Biosens Bioelectron 2014, 55:459-463.
- 47. Zhang J-D, Yu T, Li J-Y, Xu J-J, Chen H-Y: An ITO bipolar array for electrochemiluminescence imaging of H<sub>2</sub>O<sub>2</sub>. Electrochem Commun 2014, 49:75-78.
- 48. Wu M-S, Liu Z, Shi H-W, Chen H-Y, Xu J-J: Visual electrochemiluminescence detection of cancer biomarkers on a closed bipolar electrode array chip. Anal Chem 2015,
- 49. Zhang X, Li J, Jia X, Li D, Wang E: Full-featured electrochemiluminescence sensing platform based on the multichannel closed bipolar system. Anal Chem 2014, 86:5595-5599.
- 50. Zhai Q, Zhang X, Han Y, Zhai J, Li J, Wang E: A nanoscale multichannel closed bipolar electrode array for electrochemiluminescence sensing platform. Anal Chem 2016, 88:945-951.
- 51. Zhang X, Chen C, Li J, Zhang L, Wang E: New insight into a 
  o microfluidic-based bipolar system for an electrochemiluminescence sensing platform. Anal Chem 2013, 85:5335-5339

The novel microfluidic-based bipolar system with two-directional driving electrodes and dual-channel mode achieved 100% current efficiency and completely eliminated the background ECL signals from the driving electrodes.

52. Sentic M, Arbault S, Bouffier L, Manojlovic D, Kuhn A, Sojic N: 3D electrogenerated chemiluminescence: from surface-confined reactions to bulk emission. Chem Sci 2015, 6:4433-4437.

The bulk generation of ECL in a 3D configuration using bipolar electrode as the driving force was demonstrated for the first time. Every object acts as an individual ECL emitter and they generate collectively the strong luminescence in the whole volume of the solution. The high sensitivity of this approach provides attractive potential in visual detection.

- 53. Qi W, Lai J, Gao W, Li S, Hanif S, Xu G: Wireless
- electrochemiluminescence with disposable minidevice. Anal Chem 2014, 86:8927-8931.

A wireless energy transmission technique has been explored for ECL for the first time. The adoption of wireless energy transmission avoid the

need of direct electrical connections and the use of large and expensive electrochemical instruments. Since the handy and cheap wireless electrochemical minidevice can use laptops or batteries as a power supply, it is promising for the development of portable ECL devices for various applications including POCT and field analysis.

- 54. Sekitani T, Takamiya M, Noguchi Y, Nakano S, Kato Y, Sakurai T, Someya T: A large-area wireless power-transmission sheet using printed organic transistors and plastic MEMS switches. Nat Mater 2007, 6:413-417.
- 55. Qi L, Xia Y, Qi W, Gao W, Wu F, Xu G: Increasing
- electrochemiluminescence intensity of a wireless electrode array chip by thousands of times using a diode for sensitive visual detection by a digital camera. *Anal Chem* 2016, **88**:1123–1127.

A diode was embedded in a wireless ECL electrode microarray chip to rectify the alternating current efficiently to dramatically enhance ECL intensities. The high sensitivity of this wireless mini-device with the diode enables the sensitive visual detection using a common portable digital camera.

- Roda A, Mirasoli M, Michelini E, Di Fusco M, Zangheri M, Cevenini L, Roda B, Simoni P: Progress in chemical luminescence-based biosensors: A critical review. Biosens Bioelectron 2016, 76:164-179.
- 57. Delaney JL, Hogan CF, Tian J, Shen W: Electrogenerated
- chemiluminescence detection in paper-based microfluidic sensors. Anal Chem 2011, 83:1300-1306.

By utilizing inkjet-printed paper fluidic substrates and screen-printed electrodes, the authors constructed low-cost and disposable ECL sensors which can be read-out with conventional photodetectors or mobile phone cameras.

- 58. Doeven EH, Barbante GJ, Kerr E, Hogan CF, Endler JA, Francis PS: Red-green-blue electrogenerated chemiluminescence utilizing a digital camera as detector. Anal Chem 2014, 86:2727-2732.
- 59. Khoshfetrat SM, Ranjbari M, Shayan M, Mehrgardi MA, Kiani A: Wireless electrochemiluminescence bipolar electrode array for visualized genotyping of single nucleotide polymorphism. Anal Chem 2015, 87:8123-8131.
- 60. Roda A, Michelini E, Cevenini L, Calabria D, Calabretta MM, Simoni P: Integrating biochemiluminescence detection on smartphones: mobile chemistry platform for point-of-need analysis. Anal Chem 2014, 86:7299-7304.
- 61. Martinez AW, Phillips ST, Carrilho E, Thomas SW, Sindi H, Whitesides GM: Simple telemedicine for developing regions: camera phones and paper-based microfluidic devices for real-time, off-site diagnosis. Anal Chem 2008, 80:3699-3707.
- 62. Zangheri M, Cevenini L, Anfossi L, Baggiani C, Simoni P, Di Nardo F, Roda A: A simple and compact smartphone accessory for quantitative chemiluminescence-based lateral flow immunoassay for salivary cortisol detection. Biosens Bioelectron 2015, 64:63-68.
- 63. Hu J, Cui X, Gong Y, Xu X, Gao B, Wen T, Lu TJ, Xu F: Portable microfluidic and smartphone-based devices for monitoring of cardiovascular diseases at the point of care. Biotechnol Adv 2016, 34:305-320.
- 64. Zhang D, Liu Q: Biosensors and bioelectronics on smartphone for portable biochemical detection. Biosens Bioelectron 2016, **75**:273–284.
- 65. Valenti G, Zangheri M, Sansaloni SE, Mirasoli M, Penicaud A, Roda A, Paolucci F: Transparent carbon nanotube network for efficient electrochemiluminescence devices. Chem Eur J 2015, 21:12640-12645
- 66. Martinez AW, Phillips ST, Butte MJ, Whitesides GM: Patterned paper as a platform for inexpensive, low-volume, portable bioassays. Angew Chem Int Ed 2007, 46:1318-1320
- 67. Ge L, Yu J, Ge S, Yan M: Lab-on-paper-based devices using chemiluminescence and electrogenerated chemiluminescence detection. Anal Bioanal Chem 2014, 406:5613-5630.
- 68. Xu Y, Liu M, Kong N, Liu J: Lab-on-paper micro- and nano-analytical devices: Fabrication, modification, detection and emerging applications. Microchim Acta 2016, **183**:1521-1542

- Liu R, Zhang C, Liu M: Open bipolar electrode-electrochemiluminescence imaging sensing using paper-based microfluidics. Sensor Actuat B, Chem 2015, 216:255–262.
- Feng Q, Chen H, Xu J: Disposable paper-based bipolar electrode array for multiplexed electrochemiluminescence detection of pathogenic DNAs. Sci China Chem 2015, 58:810–818.
- Ge L, Wang S, Ge S, Yu J, Yan M, Li N, Huang J: Electrophoretic separation in a microfluidic paper-based analytical device with an on-column wireless electrogenerated chemiluminescence detector. Chem Commun 2014, 50:5699–5702.
- Li M, Wang Y, Zhang Y, Yu J, Ge S, Yan M: Graphene functionalized porous Au-paper based electrochemiluminescence device for detection of DNA using luminescent silver nanoparticles coated calcium carbonate/carboxymethyl chitosan hybrid microspheres as labels. Biosens Bioelectron 2014, 59:307–313.
- Li L, Li W, Ma C, Yang H, Ge S, Yu J: Paper-based electrochemiluminescence immunodevice for carcinoembryonic antigen using nanoporous gold-chitosan hybrids and graphene quantum dots functionalized Au@Pt. Sensor Actuat B-Chem 2014, 202:314–322.
- 74. Li W, Yang H, Ma C, Li L, Ge S, Song X, Yu J, Yan M: Electrochemiluminescence of peroxydisulfate using flower-like Ag@Au-paper electrode and Pd@Au-assisted multiple enzymatic labels. Electrochim Acta 2014, 141:391–397.

- Yan J, Yan M, Ge L, Ge S, Yu J: An origami electrochemiluminescence immunosensor based on gold/graphene for specific, sensitive point-of-care testing of carcinoembryonic antigen. Sensor Actuat B, Chem 2014, 193:247–254.
- Gao C, Su M, Wang Y, Ge S, Yu J: A disposable paper-based electrochemiluminescence device for ultrasensitive monitoring of CEA based on Ru(bpy)<sub>3</sub><sup>2+</sup>@Au nanocages. RSC Adv 2015, 5:28324–28331.
- 77. Wu L, Ma C, Zheng X, Liu H, Yu J: Paper-based electrochemiluminescence origami device for protein detection using assembled cascade DNA-carbon dots nanotags based on rolling circle amplification. Biosens Bioelectron 2015, 68:413–420.
- Liu F, Ge S, Su M, Song X, Yan M, Yu J: Electrochemiluminescence device for in-situ and accurate determination of CA153 at the MCF-7 cell surface based on graphene quantum dots loaded surface villous Au nanocage. Biosens Bioelectron 2015, 71:286–293.
- Wu L, Ma C, Ge L, Kong Q, Yan M, Ge S, Yu J: Paper-based electrochemiluminescence origami cyto-device for multiple cancer cells detection using porous AuPd alloy as catalytically promoted nanolabels. Biosens Bioelectron 2015, 63:450–457.
- 80. Wu L, Zhang Y, Wang Y, Ge S, Liu H, Yan M, Yu J: A paper-based electrochemiluminescence electrode as an aptamer-based cytosensor using PtNi@carbon dots as nanolabels for detection of cancer cells and for in-situ screening of anticancer drugs. Microchim Acta 2016, 183:1873–1880.