Albumin (BCG)

Analyze	Calib	Range	Others

Assay/Time/Point	1 Point		5		15	0	0	0	
Wave(2nd/Primary)	700		600						
S.Vol(Normal)	2.0	0.0	0						
S.Vol(Decrease)	2.0	0.0	0						
S.Vol(Increaase)	4.0	0.0	0						
Diluent	Water	0							
Reagent (R1)	200	0	00350	0	350 000 935 015 188 844 018				
Reagent (R2)	0	0	00350	0					
Reagent (R3)	0	0	00350	0					
Reagent (R4)	0	0	00350	0					
Abs.Limit	0	Increase		2Tests					
Prozone Limit	0	34	Lower						
Cell Detergent	Detergent 1,2								

Analyze	Calib	Range	Others

Calibration Type	Linear			
Point	2	Span Point	2	
SD Limit	0.1			
Duplicate Limit	10	%	500	Abs
Sensitivity Limit	-99999		999999	
S1Abs Limit	-32000		32000	

Analyze	Calib	Range	Others

Aplication Code	350		Unit	g/dl
Report Name	Albumin (BCP)			
Technical Limit	1	6		
Expected Value	3.5	5		

Analyze	Calib	Range	Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calib.Code	*	*				
Concentration	0.0	#				
Position						
Sample Volume	2.0	2.0	2.0	2.0	2.0	2.0
Diluted S.Vol	0.0	0.0	0.0	0.0	0.0	0.0
Diluent Volume	0	0	0	0	0	0

Alkaline phosphatase

Analyze	Calib	Range	Others

Assay/Time/Point	Rate-A		10		19	28	0	0	
Wave(2nd/Primary)	660		415						
S.Vol(Normal)	4.0	0.0	0						
S.Vol(Decrease)	2.0	0.0	0						
S.Vol(Increaase)	8.0	0.0	0						
Diluent	Water	0							
Reagent (R1)	160	0	00351	0	351 000 935 077 518 844 023				
Reagent (R2)	0	0	00351	0					
Reagent (R3)	40	0	00351	0	351 000 935 032 948 904 011				
Reagent (R4)	0	0	00351	0					
Abs.Limit	14000	Increase			2Tests				
Prozone Limit	0	34	Lower						
Cell Detergent	Detergent 1,2								

Analyze	Calib	Range	Others

Calibration Type	Linear			
Point	2	Span Point	2	
SD Limit	0.1			
Duplicate Limit	10	%	500	Abs
Sensitivity Limit	-99999		999999	
S1Abs Limit	-32000		32000	

Analyze	Calib	Range	Others

Aplication Code	351	Unit	U/L	
Report Name	Alkaline phosphatase			
Technical Limit	1	1000		
Expected Value	98	279		

Analyze	Calib	Range	Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calib.Code	*	*				
Concentration	0	#				
Position						
Sample Volume	4.0	4.0	4.0	4.0	4.0	4.0
Diluted S.Vol	0.0	0.0	0.0	0.0	0.0	0.0
Diluent Volume	0	0	0	0	0	0

ALT/GPT

Analyze	Calib	Range	Others

Assay/Time/Point	Rate-A		10		19	28	0	0	
Wave(2nd/Primary)	700		340						
S.Vol(Normal)	20.0	0.0	0						
S.Vol(Decrease)	2.0	0.0	0						
S.Vol(Increaase)	30.0	0.0	0						
Diluent	Water	0							
Reagent (R1)	160	0	00352	0	352 000 935 018 268 844 029				
Reagent (R2)	0	0	00352	0					
Reagent (R3)	40	0	00352	0	352 000 935 013 218 904 024				
Reagent (R4)	0	0	00352	0					
Abs.Limit	5000	Decrease		2Tests					
Prozone Limit	0	34	Lower						
Cell Detergent	Detergent 1,2								

Analyze	Calib	Range	Others

Calibration Type	Linear			
Point	2	Span Point	2	
SD Limit	0.1			
Duplicate Limit	10	%	500	Abs
Sensitivity Limit	-99999		999999	
S1Abs Limit	-32000		32000	

Analyze	Calib	Range	Others

Aplication Code	352	Unit	U/L	
Report Name	ALT/GPT			
Technical Limit	0	300		
Expected Value	0	40		

Analyze	Calib	Range	Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calib.Code	*	*				
Concentration	0	#				
Position						
Sample Volume	20.0	20.0	20.0	20.0	20.0	20.0
Diluted S.Vol	0.0	0.0	0.0	0.0	0.0	0.0
Diluent Volume	0	0	0	0	0	0

Amylase - total

Analyze		Calib		Range		Others				
Assay/Time/Point	Rate-A		4		3	12	0	0		
Wave(2nd/Primary)	660		415							
S.Vol(Normal)	4.0	0.0	0							
S.Vol(Decrease)	2.0	0.0	0							
S.Vol(Increaase)	8.0	0.0	0							
Diluent	Water	0								
Reagent (R1)	200	0	00354	0	354 000 935 095 158 844 042					
Reagent (R2)	0	0	00354	0						
Reagent (R3)	0	0	00354	0						
Reagent (R4)	0	0	00354	0						
Abs.Limit	20000	Increase			2Tests					
Prozone Limit	0	34	Lower							
Cell Detergent	Detergent 1,2									

Analyze		Calib		Range		Others
Calibration Type	Linear					
Point	2	Span Point	2			
SD Limit	0.1					
Duplicate Limit	10	%	500	Abs		
Sensitivity Limit	-99999		999999			
S1Abs Limit	-32000		32000			

Analyze		Calib		Range		Others
Aplication Code	354		Unit	U/L		
Report Name	Amylase - total					
Technical Limit	0	2000				
Expected Value	0	90				

Analyze		Calib		Range		Others	
<Standard>	(1)	(2)	(3)	(4)	(5)	(6)	
Calib.Code	*	*					
Concentration	0	#					
Position							
Sample Volume	4.0	4.0	4.0	4.0	4.0	4.0	
Diluted S.Vol	0.0	0.0	0.0	0.0	0.0	0.0	
Diluent Volume	0	0	0	0	0	0	

AST/GOT

Analyze	Calib	Range	Others

Assay/Time/Point	Rate-A		10		19	28	0	0	
Wave(2nd/Primary)	700		340						
S.Vol(Normal)	20.0	0.0	0						
S.Vol(Decrease)	2.0	0.0	0						
S.Vol(Increaase)	30.0	0.0	0						
Diluent	Water	0							
Reagent (R1)	160	0	00358	0	358 000 935 043 588 844 008				
Reagent (R2)	0	0	00358	0					
Reagent (R3)	40	0	00358	0	358 000 935 072 238 904 100				
Reagent (R4)	0	0	00358	0					
Abs.Limit	5000	Decrease		2Tests					
Prozone Limit	0	34	Lower						
Cell Detergent	Detergent 1,2								

Analyze	Calib	Range	Others

Calibration Type	Linear			
Point	2	Span Point	2	
SD Limit	0.1			
Duplicate Limit	10	%	500	Abs
Sensitivity Limit	-99999		999999	
S1Abs Limit	-32000		32000	

Analyze	Calib	Range	Others

Aplication Code	358	Unit	U/L	
Report Name	AST/GOT			
Technical Limit	0	300		
Expected Value	0	38		

Analyze	Calib	Range	Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calib.Code	*	*				
Concentration	0	#				
Position						
Sample Volume	20.0	20.0	20.0	20.0	20.0	20.0
Diluted S.Vol	0.0	0.0	0.0	0.0	0.0	0.0
Diluent Volume	0	0	0	0	0	0

Complement 3

Analyze	Calib	Range	Others

Assay/Time/Point	2 Point End		10		15	34	0	0	
Wave(2nd/Primary)	700		340						
S.Vol(Normal)	2.0	0.0	0						
S.Vol(Decrease)	2.0	0.0	0						
S.Vol(Increaase)	4.0	0.0	0						
Diluent	Water	0							
Reagent (R1)	180	0	361	0	361 000 935 055 438 844 042				
Reagent (R2)	0	0	361	0					
Reagent (R3)	35	0	361	0	361 000 935 078 218 904 086				
Reagent (R4)	0	0	361	0					
Abs.Limit	0	Increase		2Tests					
Prozone Limit	0	34	Lower						
Cell Detergent	Detergent 1,2								

Analyze	Calib	Range	Others

Calibration Type	Logit-Log (4P)			
Point	6	Span Point	6	
SD Limit	999.9			
Duplicate Limit	99	%	2000	Abs
Sensitivity Limit	-99999		999999	
S1Abs Limit	-32000		32000	

Analyze	Calib	Range	Others

Aplication Code	361		Unit	mg/dl
Report Name	Complement 3			
Technical Limit	2	500		
Expected Value	80	180		

Analyze	Calib	Range	Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calib.Code	*	*	*	*	*	*
Concentration	0	#	#	#	#	#
Position						
Sample Volume	2.0	2.0	2.0	2.0	2.0	2.0
Diluted S.Vol	0.0	0.0	0.0	0.0	0.0	0.0
Diluent Volume	0	0	0	0	0	0

Complement 4

Analyze	Calib	Range	Others

Assay/Time/Point	2 Point End		10		15	34	0	0	
Wave(2nd/Primary)	700		340						
S.Vol(Normal)	4.0	0.0	0						
S.Vol(Decrease)	2.0	0.0	0						
S.Vol(Increaase)	8.0	0.0	0						
Diluent	Water	0							
Reagent (R1)	180	0	362	0	362 000 935 028 508 844 052				
Reagent (R2)	0	0	362	0					
Reagent (R3)	35	0	362	0	362 000 935 024 118 904 021				
Reagent (R4)	0	0	362	0					
Abs.Limit	0	Increase		2Tests					
Prozone Limit	0	34	Lower						
Cell Detergent	Detergent 1,2								

Analyze	Calib	Range	Others

Calibration Type	Logit-Log (4P)			
Point	6	Span Point	6	
SD Limit	999.9			
Duplicate Limit	99	%	2000	Abs
Sensitivity Limit	-99999		999999	
S1Abs Limit	-32000		32000	

Analyze	Calib	Range	Others

Aplication Code	362		Unit	mg/dl
Report Name	Complement 4			
Technical Limit	1	80		
Expected Value	10	40		

Analyze	Calib	Range	Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calib.Code	*	*	*	*	*	*
Concentration	0.0	#	#	#	#	#
Position						
Sample Volume	4.0	4.0	4.0	4.0	4.0	4.0
Diluted S.Vol	0.0	0.0	0.0	0.0	0.0	0.0
Diluent Volume	0	0	0	0	0	0

Calcium Arsenazo III

Analyze	Calib	Range	Others

Assay/Time/Point	1 Point		3		10	0	0	0
Wave(2nd/Primary)	700		660					
S.Vol(Normal)	3.0	0.0	0					
S.Vol(Decrease)	2.0	0.0	0					
S.Vol(Increaase)	6.0	0.0	0					
Diluent	Water	0						
Reagent (R1)	200	0	363	0	363 000 935 017 418 844 046			
Reagent (R2)	0	0	363	0				
Reagent (R3)	0	0	363	0				
Reagent (R4)	0	0	363	0				
Abs.Limit	0	Increase		2Tests				
Prozone Limit	0	34	Lower					
Cell Detergent	Detergent 1,2							

Analyze	Calib	Range	Others

Calibration Type	Linear			
Point	2	Span Point	2	
SD Limit	0.1			
Duplicate Limit	10	%	500	Abs
Sensitivity Limit	-99999		999999	
S1Abs Limit	-32000		32000	

Analyze	Calib	Range	Others

Aplication Code	363	Unit	mg/dl	
Report Name	Calcium Arsenazo III			
Technical Limit	0	0		
Expected Value	8.4	10.4		

Analyze	Calib	Range	Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calib.Code	*	*				
Concentration	0	#				
Position						
Sample Volume	3.0	3.0	3.0	3.0	3.0	3.0
Diluted S.Vol	0.0	0.0	0.0	0.0	0.0	0.0
Diluent Volume	0	0	0	0	0	0

Cholesterol

Analyze	Calib	Range	Others

Assay/Time/Point	1 Point		10		34	0	0	0
Wave(2nd/Primary)	700		505					
S.Vol(Normal)	2.0	0.0	0					
S.Vol(Decrease)	2.0	0.0	0					
S.Vol(Increaase)	4.0	0.0	0					
Diluent	Water	0						
Reagent (R1)	200	0	00366	0	366 000 935 080 928 844 017			
Reagent (R2)	0	0	00366	0				
Reagent (R3)	0	0	00366	0				
Reagent (R4)	0	0	00366	0				
Abs.Limit	0	Increase		2Tests				
Prozone Limit	0	34	Lower					
Cell Detergent	Detergent 1,2							

Analyze	Calib	Range	Others

Calibration Type	Linear			
Point	2	Span Point	2	
SD Limit	0.1			
Duplicate Limit	10	%	500	Abs
Sensitivity Limit	-99999		999999	
S1Abs Limit	-32000		32000	

Analyze	Calib	Range	Others

Aplication Code	366	Unit	mg/dl	
Report Name	Cholesterol			
Technical Limit	5	600		
Expected Value	0	200		

Analyze	Calib	Range	Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calib.Code	*	*				
Concentration	0	#				
Position						
Sample Volume	2.0	2.0	2.0	2.0	2.0	2.0
Diluted S.Vol	0.0	0.0	0.0	0.0	0.0	0.0
Diluent Volume	0	0	0	0	0	0

CK-NAC

Analyze		Calib		Range		Others			
Assay/Time/Point	Rate-A		10		24	34	0	0	
Wave(2nd/Primary)	546		340						
S.Vol(Normal)	10.0	0.0	0						
S.Vol(Decrease)	2.0	0.0	0						
S.Vol(Increaase)	20.0	0.0	0						
Diluent	Water	0							
Reagent (R1)	180	0	00368	0	368 000 935 093 638 844 035				
Reagent (R2)	0	0	00368	0					
Reagent (R3)	35	0	00368	0	368 000 935 025 298 904 067				
Reagent (R4)	0	0	00368	0					
Abs.Limit	10000	Increase			2Tests				
Prozone Limit	0	34	Lower						
Cell Detergent	Detergent 1,2								

Analyze		Calib		Range		Others
Calibration Type	Linear					
Point	2	Span Point	2			
SD Limit	0.1					
Duplicate Limit	10	%	500	Abs		
Sensitivity Limit	-99999		999999			
S1Abs Limit	-32000		32000			

Analyze		Calib		Range		Others
Aplication Code	368		Unit	U/L		
Report Name	CK-NAC					
Technical Limit	1	1500				
Expected Value	24	190				

Analyze	Calib		Range		Others	
<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calib.Code	*	*				
Concentration	0	#				
Position						
Sample Volume	10.0	10.0	10.0	10.0	10.0	10.0
Diluted S.Vol	0.0	0.0	0.0	0.0	0.0	0.0
Diluent Volume	0	0	0	0	0	0

Creatinine (Jaffe)

Analyze	Calib	Range	Others

Assay/Time/Point	2 Point Rate		4		8	12	0	0	
Wave(2nd/Primary)	700		505						
S.Vol(Normal)	20.0	0.0	0						
S.Vol(Decrease)	2.0	0.0	0						
S.Vol(Increaase)	30.0	0.0	0						
Diluent	Water	0							
Reagent (R1)	100	0	00370	0	370 000 935 066 008 844 097				
Reagent (R2)	100	0	00370	0	370 000 935 004 178 874 011				
Reagent (R3)	0	0	00370	0					
Reagent (R4)	0	0	00370	0					
Abs.Limit	32000	Increase		2Tests					
Prozone Limit	0	34	Lower						
Cell Detergent	Detergent 1,2								

Analyze	Calib	Range	Others

Calibration Type	Linear			
Point	2	Span Point	2	
SD Limit	0.1			
Duplicate Limit	10	%	500	Abs
Sensitivity Limit	-99999		999999	
S1Abs Limit	-32000		32000	

Analyze	Calib	Range	Others

Aplication Code	370		Unit	mg/dl
Report Name	Creatinine (Jaffe)			
Technical Limit	0.1	20		
Expected Value	0.7	1.4		

Analyze	Calib	Range	Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calib.Code	*	*				
Concentration	0.00	#				
Position						
Sample Volume	20.0	20.0	20.0	20.0	20.0	20.0
Diluted S.Vol	0.0	0.0	0.0	0.0	0.0	0.0
Diluent Volume	0	0	0	0	0	0

C-REACTIVE PROTEIN

Analyze	Calib	Range	Others

Assay/Time/Point	2 Point End		10		15	34	0	0	
Wave(2nd/Primary)	700		340						
S.Vol(Normal)	8.0	0.0	0						
S.Vol(Decrease)	4.0	0.0	0						
S.Vol(Increaase)	16.0	0.0	0						
Diluent	Water	0							
Reagent (R1)	180	0	00371	0	371 000 935 055 178 844 067				
Reagent (R2)	0	0	00371	0					
Reagent (R3)	35	0	00371	0	371 000 935 063 128 904 037				
Reagent (R4)	0	0	00371	0					
Abs.Limit	0	Increase		2Tests					
Prozone Limit	0	34	Lower						
Cell Detergent	Detergent 1,2								

Analyze	Calib	Range	Others

Calibration Type	Logit-Log (4P)			
Point	6	Span Point	6	
SD Limit	999.9			
Duplicate Limit	99	%	2000	Abs
Sensitivity Limit	-99999		999999	
S1Abs Limit	-32000		32000	

Analyze	Calib	Range	Others

Aplication Code	371		Unit	mg/L
Report Name	C-REACTIVE PROTEIN			
Technical Limit	1	200		
Expected Value	0	9		

Analyze	Calib	Range	Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calib.Code	*	*	*	*	*	*
Concentration	0.0	#	#	#	#	#
Position						
Sample Volume	8.0	8.0	8.0	8.0	8.0	8.0
Diluted S.Vol	0.0	0.0	0.0	0.0	0.0	0.0
Diluent Volume	0	0	0	0	0	0

COPPER

Analyze	Calib	Range	Others

Assay/Time/Point	1 Point		5		15	0	0	0
Wave(2nd/Primary)	700		570					
S.Vol(Normal)	10.0	0.0	0					
S.Vol(Decrease)	2.0	0.0	0					
S.Vol(Increaase)	20.0	0.0	0					
Diluent	Water	0						
Reagent (R1)	200	0	00373	0	373 000 935 048 468 844 062			
Reagent (R2)	0	0	00373	0				
Reagent (R3)	0	0	00373	0				
Reagent (R4)	0	0	00373	0				
Abs.Limit	0	Increase		2Tests				
Prozone Limit	0	34	Lower					
Cell Detergent	Detergent 1,2							

Analyze	Calib	Range	Others

Calibration Type	Linear			
Point	2	Span Point	2	
SD Limit	0.1			
Duplicate Limit	10	%	500	Abs
Sensitivity Limit	-99999		999999	
S1Abs Limit	-32000		32000	

Analyze	Calib	Range	Others

Aplication Code	373	Unit	µg/dl	
Report Name	COPPER			
Technical Limit	20	400		
Expected Value	70	140		

Analyze	Calib	Range	Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calib.Code	*	*				
Concentration	0	#				
Position						
Sample Volume	10.0	10.0	10.0	10.0	10.0	10.0
Diluted S.Vol	0.0	0.0	0.0	0.0	0.0	0.0
Diluent Volume	0	0	0	0	0	0

Glucose

Analyze	Calib	Range	Others

Assay/Time/Point	1 Point		10		34	0	0	0	
Wave(2nd/Primary)	700		505						
S.Vol(Normal)	2.0	0.0	0						
S.Vol(Decrease)	2.0	0.0	0						
S.Vol(Increaase)	4.0	0.0	0						
Diluent	Water	0							
Reagent (R1)	200	0	379	0	379 000 935 084 778 844 100				
Reagent (R2)	0	0	379	0					
Reagent (R3)	0	0	379	0					
Reagent (R4)	0	0	379	0					
Abs.Limit	0	Increase		2Tests					
Prozone Limit	0	34	Lower						
Cell Detergent	Detergent 1,2								

Analyze	Calib	Range	Others

Calibration Type	Linear			
Point	2	Span Point	2	
SD Limit	0.1			
Duplicate Limit	10	%	500	Abs
Sensitivity Limit	-99999		999999	
S1Abs Limit	-32000		32000	

Analyze	Calib	Range	Others

Aplication Code	379	Unit	mg/dl	
Report Name	Glucose			
Technical Limit	0	400		
Expected Value	70	110		

Analyze	Calib	Range	Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calib.Code	*	*				
Concentration	0	#				
Position						
Sample Volume	2.0	2.0	2.0	2.0	2.0	2.0
Diluted S.Vol	0.0	0.0	0.0	0.0	0.0	0.0
Diluent Volume	0	0	0	0	0	0

HDL Cholesterol Direct

Analyze	Calib	Range	Others

Assay/Time/Point	2 Point End		10		15	34	0	0	
Wave(2nd/Primary)	700		600						
S.Vol(Normal)	2.0	0.0	0						
S.Vol(Decrease)	2.0	0.0	0						
S.Vol(Increaase)	4.0	0.0	0						
Diluent	Water	0							
Reagent (R1)	180	0	381	0	381 000 935 079 938 844 023				
Reagent (R2)	0	0	381	0					
Reagent (R3)	60	0	381	0	381 000 935 063 738 904 053				
Reagent (R4)	0	0	381	0					
Abs.Limit	0	Increase		2Tests					
Prozone Limit	0	34	Lower						
Cell Detergent	Detergent 1,2								

Analyze	Calib	Range	Others

Calibration Type	Linear			
Point	2	Span Point	2	
SD Limit	0.1			
Duplicate Limit	10	%	500	Abs
Sensitivity Limit	-99999		999999	
S1Abs Limit	-32000		32000	

Analyze	Calib	Range	Others

Aplication Code	381	Unit	mg/dl	
Report Name	HDL Cholesterol Direct			
Technical Limit	5	120		
Expected Value	30	90		

Analyze	Calib	Range	Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calib.Code	*	*				
Concentration	0	#				
Position						
Sample Volume	2.0	2.0	2.0	2.0	2.0	2.0
Diluted S.Vol	0.0	0.0	0.0	0.0	0.0	0.0
Diluent Volume	0	0	0	0	0	0

Immunoglobulin A

Analyze	Calib	Range	Others

Assay/Time/Point	2 Point End		10		15	34	0	0	
Wave(2nd/Primary)	700		340						
S.Vol(Normal)	2.0	0.0	0						
S.Vol(Decrease)	2.0	0.0	0						
S.Vol(Increaase)	4.0	0.0	0						
Diluent	Water	0							
Reagent (R1)	180	0	383	0	383 000 935 080 098 844 099				
Reagent (R2)	0	0	383	0					
Reagent (R3)	35	0	383	0	383 000 935 015 688 904 079				
Reagent (R4)	0	0	383	0					
Abs.Limit	0	Increase		2Tests					
Prozone Limit	0	34	Lower						
Cell Detergent	Detergent 1,2								

Analyze	Calib	Range	Others

Calibration Type	Logit-Log (4P)			
Point	6	Span Point	6	
SD Limit	999.9			
Duplicate Limit	99	%	2000	Abs
Sensitivity Limit	-99999		999999	
S1Abs Limit	-32000		32000	

Analyze	Calib	Range	Others

Aplication Code	383		Unit	mg/dl
Report Name	Immunoglobulin A			
Technical Limit	1	700		
Expected Value	70	400		

Analyze	Calib	Range	Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calib.Code	*	*	*	*	*	*
Concentration	0	#	#	#	#	#
Position						
Sample Volume	2.0	2.0	2.0	2.0	2.0	2.0
Diluted S.Vol	0.0	0.0	0.0	0.0	0.0	0.0
Diluent Volume	0	0	0	0	0	0

Immunoglobulin G

Analyze	Calib	Range	Others

Assay/Time/Point	2 Point End		10		15	34	0	0	
Wave(2nd/Primary)	Cancel		600						
S.Vol(Normal)	2.0	0.0	0						
S.Vol(Decrease)	2.0	0.0	0						
S.Vol(Increaase)	4.0	0.0	0						
Diluent	Water	0							
Reagent (R1)	180	0	384	0	384 000 935 050 248 844 044				
Reagent (R2)	0	0	384	0					
Reagent (R3)	35	0	384	0	384 000 935 009 708 904 077				
Reagent (R4)	0	0	384	0					
Abs.Limit	0	Increase		2Tests					
Prozone Limit	0	34	Lower						
Cell Detergent	Detergent 1,2								

Analyze	Calib	Range	Others

Calibration Type	Logit-Log (4P)			
Point	6	Span Point	6	
SD Limit	999.9			
Duplicate Limit	99	%	2000	Abs
Sensitivity Limit	-99999		999999	
S1Abs Limit	-32000		32000	

Analyze	Calib	Range	Others

Aplication Code	384		Unit	mg/dl
Report Name	Immunoglobulin G			
Technical Limit	1	3500		
Expected Value	700	1600		

Analyze	Calib	Range	Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calib.Code	*	*	*	*	*	*
Concentration	0	#	#	#	#	#
Position						
Sample Volume	2.0	2.0	2.0	2.0	2.0	2.0
Diluted S.Vol	0.0	0.0	0.0	0.0	0.0	0.0
Diluent Volume	0	0	0	0	0	0

Immunoglobulin M

Analyze	Calib	Range	Others

Assay/Time/Point	2 Point End		10		15	34	0	0	
Wave(2nd/Primary)	700		340						
S.Vol(Normal)	2.0	0.0	0						
S.Vol(Decrease)	2.0	0.0	0						
S.Vol(Increaase)	4.0	0.0	0						
Diluent	Water	0							
Reagent (R1)	180	0	385	0	385 000 935 014 908 844 001				
Reagent (R2)	0	0	385	0					
Reagent (R3)	35	0	385	0	385 000 935 048 418 904 025				
Reagent (R4)	0	0	385	0					
Abs.Limit	0	Increase		2Tests					
Prozone Limit	0	34	Lower						
Cell Detergent	Detergent 1,2								

Analyze	Calib	Range	Others

Calibration Type	Logit-Log (4P)			
Point	6	Span Point	6	
SD Limit	999.9			
Duplicate Limit	99	%	2000	Abs
Sensitivity Limit	-99999		999999	
S1Abs Limit	-32000		32000	

Analyze	Calib	Range	Others

Aplication Code	385		Unit	mg/dl
Report Name	Immunoglobulin M			
Technical Limit	1	250		
Expected Value	40	230		

Analyze	Calib	Range	Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calib.Code	*	*	*	*	*	*
Concentration	0	#	#	#	#	#
Position						
Sample Volume	2.0	2.0	2.0	2.0	2.0	2.0
Diluted S.Vol	0.0	0.0	0.0	0.0	0.0	0.0
Diluent Volume	0	0	0	0	0	0

Lactate Dehydrogenase

Analyze	Calib	Range	Others

Assay/Time/Point	Rate-A		10		19	28	0	0	
Wave(2nd/Primary)	415		340						
S.Vol(Normal)	4.0	0.0	0						
S.Vol(Decrease)	2.0	0.0	0						
S.Vol(Increaase)	8.0	0.0	0						
Diluent	Water	0							
Reagent (R1)	180	0	00388	0	388 000 935 022 068 844 036				
Reagent (R2)	0	0	00388	0					
Reagent (R3)	35	0	00388	0	388 000 935 050 788 904 047				
Reagent (R4)	0	0	00388	0					
Abs.Limit	3000	Decrease			2Tests				
Prozone Limit	0	34	Lower						
Cell Detergent	Detergent 1,2								

Analyze	Calib	Range	Others

Calibration Type	Linear			
Point	2	Span Point	2	
SD Limit	0.1			
Duplicate Limit	10	%	500	Abs
Sensitivity Limit	-99999		999999	
S1Abs Limit	-32000		32000	

Analyze	Calib	Range	Others

Aplication Code	388		Unit	U/L
Report Name	Lactate Dehydrogenase			
Technical Limit	1	1200		
Expected Value	200	400		

Analyze	Calib	Range	Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calib.Code	*	*				
Concentration	0	#				
Position						
Sample Volume	4.0	4.0	4.0	4.0	4.0	4.0
Diluted S.Vol	0.0	0.0	0.0	0.0	0.0	0.0
Diluent Volume	0	0	0	0	0	0

LDL Cholesterol Direct

Analyze	Calib	Range	Others

Assay/Time/Point	2 Point End		10		15	34	0	0	
Wave(2nd/Primary)	700		546						
S.Vol(Normal)	3.0	0.0	0						
S.Vol(Decrease)	2.0	0.0	0						
S.Vol(Increaase)	6.0	0.0	0						
Diluent	Water	0							
Reagent (R1)	180	0	389	0	389 000 935 070 218 844 037				
Reagent (R2)	0	0	389	0					
Reagent (R3)	60	0	389	0	389 000 935 046 468 904 025				
Reagent (R4)	0	0	389	0					
Abs.Limit	0	Increase		2Tests					
Prozone Limit	0	34	Lower						
Cell Detergent	Detergent 1,2								

Analyze	Calib	Range	Others

Calibration Type	Linear			
Point	2	Span Point	2	
SD Limit	0.1			
Duplicate Limit	10	%	500	Abs
Sensitivity Limit	-99999		999999	
S1Abs Limit	-32000		32000	

Analyze	Calib	Range	Others

Aplication Code	389	Unit	mg/dl	
Report Name	LDL Cholesterol Direct			
Technical Limit	5	700		
Expected Value	0	130		

Analyze	Calib	Range	Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calib.Code	*	*				
Concentration	0	#				
Position						
Sample Volume	3.0	3.0	3.0	3.0	3.0	3.0
Diluted S.Vol	0.0	0.0	0.0	0.0	0.0	0.0
Diluent Volume	0	0	0	0	0	0

Lipase

Analyze	Calib	Range	Others

Assay/Time/Point	Rate-A		10		22	28	0	0	
Wave(2nd/Primary)	700		570						
S.Vol(Normal)	4.0	0.0	0						
S.Vol(Decrease)	2.0	0.0	0						
S.Vol(Increaase)	8.0	0.0	0						
Diluent	Water	0							
Reagent (R1)	160	0	00390	0	390 000 935 034 448 844 009				
Reagent (R2)	0	0	00390	0					
Reagent (R3)	40	0	00390	0	390 000 935 076 928 904 039				
Reagent (R4)	0	0	00390	0					
Abs.Limit	10000	Increase			2Tests				
Prozone Limit	0	34	Lower						
Cell Detergent	Detergent 1,2								

Analyze	Calib	Range	Others

Calibration Type	Linear			
Point	2	Span Point	2	
SD Limit	0.1			
Duplicate Limit	10	%	500	Abs
Sensitivity Limit	-99999		999999	
S1Abs Limit	-32000		32000	

Analyze	Calib	Range	Others

Aplication Code	390	Unit	U/L	
Report Name	Lipase			
Technical Limit	1	200		
Expected Value	0	60		

Analyze	Calib	Range	Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calib.Code	*	*				
Concentration	0	#				
Position						
Sample Volume	4.0	4.0	4.0	4.0	4.0	4.0
Diluted S.Vol	0.0	0.0	0.0	0.0	0.0	0.0
Diluent Volume	0	0	0	0	0	0

Micro Albumin

Analyze	Calib	Range	Others

Assay/Time/Point	2 Point End		10		15	34	0	0	
Wave(2nd/Primary)	700		340						
S.Vol(Normal)	14.0	0.0	0						
S.Vol(Decrease)	4.0	0.0	0						
S.Vol(Increaase)	20.0	0.0	0						
Diluent	Water	0							
Reagent (R1)	180	0	00392	0	392 000 935 064 908 844 036				
Reagent (R2)	0	0	00392	0					
Reagent (R3)	35	0	00392	0	392 000 935 098 408 904 053				
Reagent (R4)	0	0	00392	0					
Abs.Limit	0	Increase		2Tests					
Prozone Limit	0	34	Lower						
Cell Detergent	Detergent 1,2								

Analyze	Calib	Range	Others

Calibration Type	Logit-Log (4P)			
Point	6	Span Point	6	
SD Limit	999.9			
Duplicate Limit	99	%	2000	Abs
Sensitivity Limit	-99999		999999	
S1Abs Limit	-32000		32000	

Analyze	Calib	Range	Others

Aplication Code	392		Unit	mg/L
Report Name	Micro Albumin			
Technical Limit	1	240		
Expected Value	0	30		

Analyze	Calib	Range	Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calib.Code	*	*	*	*	*	*
Concentration	0.0	#	#	#	#	#
Position						
Sample Volume	14.0	14.0	14.0	14.0	14.0	14.0
Diluted S.Vol	0.0	0.0	0.0	0.0	0.0	0.0
Diluent Volume	0	0	0	0	0	0

CK-MB

Analyze	Calib	Range	Others

Assay/Time/Point	Rate-A		10		24	34	0	0	
Wave(2nd/Primary)	546		340						
S.Vol(Normal)	10.0	0.0	0						
S.Vol(Decrease)	2.0	0.0	0						
S.Vol(Increaase)	20.0	0.0	0						
Diluent	Water	0							
Reagent (R1)	180	0	00367	0	367 000 935 093 828 844 045				
Reagent (R2)	0	0	00367	0					
Reagent (R3)	35	0	00367	0	367 000 935 088 798 904 054				
Reagent (R4)	0	0	00367	0					
Abs.Limit	10000	Increase			2Tests				
Prozone Limit	0	34	Lower						
Cell Detergent	Detergent 1,2								

Analyze	Calib	Range	Others

Calibration Type	Linear			
Point	2	Span Point	2	
SD Limit	0.1			
Duplicate Limit	10	%	500	Abs
Sensitivity Limit	-99999		999999	
S1Abs Limit	-32000		32000	

Analyze	Calib	Range	Others

Aplication Code	367	Unit	U/L	
Report Name	CK-MB			
Technical Limit	1	1000		
Expected Value	0	24		

Analyze	Calib	Range	Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calib.Code	*	*				
Concentration	0.0	#				
Position						
Sample Volume	10.0	10.0	10.0	10.0	10.0	10.0
Diluted S.Vol	0.0	0.0	0.0	0.0	0.0	0.0
Diluent Volume	0	0	0	0	0	0

Magnesium

Analyze	Calib	Range	Others

Assay/Time/Point	1 Point		3		10	0	0	0
Wave(2nd/Primary)	700		505					
S.Vol(Normal)	2.0	0.0	0					
S.Vol(Decrease)	2.0	0.0	0					
S.Vol(Increaase)	4.0	0.0	0					
Diluent	Water	0						
Reagent (R1)	200	0	393	0	393 000 935 074 158 844 003			
Reagent (R2)	0	0	393	0				
Reagent (R3)	0	0	393	0				
Reagent (R4)	0	0	393	0				
Abs.Limit	0	Increase		2Tests				
Prozone Limit	0	34	Lower					
Cell Detergent	Detergent 1,2							

Analyze	Calib	Range	Others

Calibration Type	Linear			
Point	2	Span Point	2	
SD Limit	0.1			
Duplicate Limit	10	%	500	Abs
Sensitivity Limit	-99999		999999	
S1Abs Limit	-32000		32000	

Analyze	Calib	Range	Others

Aplication Code	393		Unit	0
Report Name	Magnesium			
Technical Limit	0	0		
Expected Value	0	0		

Analyze	Calib	Range	Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calib.Code	*	*				
Concentration	0	#				
Position						
Sample Volume	2.0	2.0	2.0	2.0	2.0	2.0
Diluted S.Vol	0.0	0.0	0.0	0.0	0.0	0.0
Diluent Volume	0	0	0	0	0	0

Phosphorus

Analyze	Calib	Range	Others

Assay/Time/Point	1 Point		5		15	0	0	0
Wave(2nd/Primary)	700		340					
S.Vol(Normal)	2.0	0.0	0					
S.Vol(Decrease)	2.0	0.0	0					
S.Vol(Increaase)	4.0	0.0	0					
Diluent	Water	0						
Reagent (R1)	200	0	00395	0	395 000 935 042 108 844 022			
Reagent (R2)	0	0	00395	0				
Reagent (R3)	0	0	00395	0				
Reagent (R4)	0	0	00395	0				
Abs.Limit	0	Increase		2Tests				
Prozone Limit	0	34	Lower					
Cell Detergent	Detergent 1,2							

Analyze	Calib	Range	Others

Calibration Type	Linear			
Point	2	Span Point	2	
SD Limit	0.1			
Duplicate Limit	10	%	500	Abs
Sensitivity Limit	-99999		999999	
S1Abs Limit	-32000		32000	

Analyze	Calib	Range	Others

Aplication Code	395	Unit	mg/dl	
Report Name	Phosphorus			
Technical Limit	0	25		
Expected Value	2.5	5		

Analyze	Calib	Range	Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calib.Code	*	*				
Concentration	0.0	#				
Position						
Sample Volume	2.0	2.0	2.0	2.0	2.0	2.0
Diluted S.Vol	0.0	0.0	0.0	0.0	0.0	0.0
Diluent Volume	0	0	0	0	0	0

Triglycerides

Analyze	Calib	Range	Others

Assay/Time/Point	1 Point		10		34	0	0	0	
Wave(2nd/Primary)	700		505						
S.Vol(Normal)	2.0	0.0	0						
S.Vol(Decrease)	2.0	0.0	0						
S.Vol(Increaase)	4.0	0.0	0						
Diluent	Water	0							
Reagent (R1)	200	0	00398	0	398 000 935 041 038 844 027				
Reagent (R2)	0	0	00398	0					
Reagent (R3)	0	0	00398	0					
Reagent (R4)	0	0	00398	0					
Abs.Limit	0	Increase		2Tests					
Prozone Limit	0	34	Lower						
Cell Detergent	Detergent 1,2								

Analyze	Calib	Range	Others

Calibration Type	Linear			
Point	2	Span Point	2	
SD Limit	0.1			
Duplicate Limit	10	%	500	Abs
Sensitivity Limit	-99999		999999	
S1Abs Limit	-32000		32000	

Analyze	Calib	Range	Others

Aplication Code	398	Unit	mg/dl	
Report Name	Triglycerides			
Technical Limit	5	700		
Expected Value	0	160		

Analyze	Calib	Range	Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calib.Code	*	*				
Concentration	0	#				
Position						
Sample Volume	2.0	2.0	2.0	2.0	2.0	2.0
Diluted S.Vol	0.0	0.0	0.0	0.0	0.0	0.0
Diluent Volume	0	0	0	0	0	0

Total Protein

Analyze	Calib	Range	Others

Assay/Time/Point	1 Point		5		15	0	0	0
Wave(2nd/Primary)	700		546					
S.Vol(Normal)	5.0	0.0	0					
S.Vol(Decrease)	2.0	0.0	0					
S.Vol(Increaase)	10.0	0.0	0					
Diluent	Water	0						
Reagent (R1)	200	0	00397	0	397 000 935 086 128 844 031			
Reagent (R2)	0	0	00397	0				
Reagent (R3)	0	0	00397	0				
Reagent (R4)	0	0	00397	0				
Abs.Limit	0	Increase		2Tests				
Prozone Limit	0	34	Lower					
Cell Detergent	Detergent 1,2							

Analyze	Calib	Range	Others

Calibration Type	Linear			
Point	2	Span Point	2	
SD Limit	0.1			
Duplicate Limit	10	%	500	Abs
Sensitivity Limit	-99999		999999	
S1Abs Limit	-32000		32000	

Analyze	Calib	Range	Others

Aplication Code	397	Unit	g/dl	
Report Name	Total Protein			
Technical Limit	1	12		
Expected Value	6.6	8.3		

Analyze	Calib	Range	Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calib.Code	*	*				
Concentration	0.0	#				
Position						
Sample Volume	5.0	5.0	5.0	5.0	5.0	5.0
Diluted S.Vol	0.0	0.0	0.0	0.0	0.0	0.0
Diluent Volume	0	0	0	0	0	0

Uric acid

Analyze	Calib	Range	Others

Assay/Time/Point	1 Point		0		15	0	0	0
Wave(2nd/Primary)	700		546					
S.Vol(Normal)	5.0	0.0	0					
S.Vol(Decrease)	2.0	0.0	0					
S.Vol(Increaase)	10.0	0.0	0					
Diluent	Water	0						
Reagent (R1)	100	0	00402	0	402 000 935 040 508 844 025			
Reagent (R2)	100	0	00402	0	402 000 935 049 598 874 000			
Reagent (R3)	0	0	00402	0				
Reagent (R4)	0	0	00402	0				
Abs.Limit	0	Increase		2Tests				
Prozone Limit	0	34	Lower					
Cell Detergent	Detergent 1,2							

Analyze	Calib	Range	Others

Calibration Type	Linear				
Point	2	Span Point	2		
SD Limit	0.1				
Duplicate Limit	10	%	500	Abs	
Sensitivity Limit	-99999		999999		
S1Abs Limit	-32000		32000		

Analyze	Calib	Range	Others

Aplication Code	402	Unit	mg/dl		
Report Name	Uric acid				
Technical Limit	1	20			
Expected Value	3.6	7.7			

Analyze	Calib	Range	Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calib.Code	*	*				
Concentration	0.0	#				
Position						
Sample Volume	5.0	5.0	5.0	5.0	5.0	5.0
Diluted S.Vol	0.0	0.0	0.0	0.0	0.0	0.0
Diluent Volume	0	0	0	0	0	0

Urea/BUN

Analyze	Calib	Range	Others

Assay/Time/Point	2 Point Rate		4		8	12	0	0	
Wave(2nd/Primary)	415		340						
S.Vol(Normal)	2.0	0.0	0						
S.Vol(Decrease)	2.0	0.0	0						
S.Vol(Increaase)	4.0	0.0	0						
Diluent	Water	0							
Reagent (R1)	160	0	00403	0	403 000 935 000 848 844 043				
Reagent (R2)	40	0	00403	0	403 000 935 007 608 874 084				
Reagent (R3)	0	0	00403	0					
Reagent (R4)	0	0	00403	0					
Abs.Limit	5000	Decrease		2Tests					
Prozone Limit	0	34	Lower						
Cell Detergent	Detergent 1,2								

Analyze	Calib	Range	Others

Calibration Type	Linear			
Point	2	Span Point	2	
SD Limit	0.1			
Duplicate Limit	10	%	500	Abs
Sensitivity Limit	-99999		999999	
S1Abs Limit	-32000		32000	

Analyze	Calib	Range	Others

Aplication Code	403	Unit	mg/dl	
Report Name	Urea/BUN			
Technical Limit	2	350		
Expected Value	15	45		

Analyze	Calib	Range	Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calib.Code	*	*				
Concentration	0.0	#				
Position						
Sample Volume	2.0	2.0	2.0	2.0	2.0	2.0
Diluted S.Vol	0.0	0.0	0.0	0.0	0.0	0.0
Diluent Volume	0	0	0	0	0	0

ZINC

Analyze	Calib	Range	Others

Assay/Time/Point	1 Point		5		15	0	0	0
Wave(2nd/Primary)	700		570					
S.Vol(Normal)	10.0	0.0	0					
S.Vol(Decrease)	2.0	0.0	0					
S.Vol(Increaase)	20.0	0.0	0					
Diluent	Water	0						
Reagent (R1)	200	0	00404	0	404 000 935 043 698 844 010			
Reagent (R2)	0	0	00404	0				
Reagent (R3)	0	0	00404	0				
Reagent (R4)	0	0	00404	0				
Abs.Limit	0	Increase		2Tests				
Prozone Limit	0	34	Lower					
Cell Detergent	Detergent 1,2							

Analyze	Calib	Range	Others

Calibration Type	Linear			
Point	2	Span Point	2	
SD Limit	0.1			
Duplicate Limit	10	%	500	Abs
Sensitivity Limit	-99999		999999	
S1Abs Limit	-32000		32000	

Analyze	Calib	Range	Others

Aplication Code	404		Unit	µg/dl
Report Name	ZINC			
Technical Limit	20	400		
Expected Value	67	127		

Analyze	Calib	Range	Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calib.Code	*	*				
Concentration	0	#				
Position						
Sample Volume	10.0	10.0	10.0	10.0	10.0	10.0
Diluted S.Vol	0.0	0.0	0.0	0.0	0.0	0.0
Diluent Volume	0	0	0	0	0	0