BIOMEDIC

Mindray BS 800

Albumin (BCG)

Chem	ALB	No.		Sample Type	Serum
Chemistry	Albumin (BCP)			Print Name	ALB
Reaction Type	Endpoint			Reaction Direction	Positive
Pri Wave	605nm			Sec Wave	700nm
Unit	g/dl			Decimal	0.01
Blank Time	3	4		Reaction Time	14	15

	Sample Vol		Aspirated		Diluent			Reagent Vol		Diluent	
Standard	2	uL		uL		uL	R1	200	uL		uL
Decreased		uL		uL		uL	R2		uL		uL
Increased		uL		uL		uL	R3		uL		uL
							R4		uL		uL

Calibration Settings

Math Model	Two-point linear		
Factor		Replicates	2

TKS TOSEE KIMIA SAADAT
توسعه کیمیای سعادت

BIOMEDIC

Mindray BS 800

Alkaline phosphatase

Chem	ALP	No.		Sample Type	Serum
Chemistry	Alkaline phosphatase			Print Name	ALP
Reaction Type	Kinetic			Reaction Direction	Positive
Pri Wave	412nm			Sec Wave	660nm
Unit	U/L			Decimal	0
Blank Time	0	0		Reaction Time	20	30

	Sample Vol		Aspirated		Diluent			Reagent Vol		Diluent	
Standard	4	uL		uL		uL	R1	160	uL		uL
Decreased		uL		uL		uL	R2	40	uL		uL
Increased		uL		uL		uL	R3		uL		uL
							R4		uL		uL

Calibration Settings

Math Model	Two-point linear		
Factor		Replicates	2

TKS TOSEE KIMIA SAADAT
توسعه کیمیای سعادت

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Mindray BS 800

ALT/GPT

Chem	ALT	No.		Sample Type	Serum
Chemistry	ALT/GPT			Print Name	ALT
Reaction Type	Kinetic			Reaction Direction	negative
Pri Wave	340nm			Sec Wave	412nm
Unit	U/L			Decimal	0
Blank Time	0	0		Reaction Time	20	30

	Sample Vol		Aspirated		Diluent			Reagent Vol		Diluent	
Standard	20	uL		uL		uL	R1	160	uL		uL
Decreased		uL		uL		uL	R2	40	uL		uL
Increased		uL		uL		uL	R3		uL		uL
							R4		uL		uL

Calibration Settings

Math Model	Two-point linear		
Factor		Replicates	2

TKS TOSEE KIMIA SAADAT
توسعه کیمیای سعادت

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Mindray BS 800

Amylase - total

Chem	AMY	No.		Sample Type	Serum
Chemistry	Amylase - total			Print Name	AMY
Reaction Type	Kinetic			Reaction Direction	Positive
Pri Wave	412nm			Sec Wave	700nm
Unit	U/L			Decimal	0
Blank Time	0	0		Reaction Time	9	19

	Sample Vol		Aspirated		Diluent			Reagent Vol		Diluent	
Standard	4	uL		uL		uL	R1	200	uL		uL
Decreased		uL		uL		uL	R2		uL		uL
Increased	0	uL		uL		uL	R3		uL		uL
							R4		uL		uL

Calibration Settings

Math Model	Two-point linear		
Factor		Replicates	2

TKS TOSEE KIMIA SAADAT
توسعه کیمیای سعادت

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Mindray BS 800

AST/GOT

Chem	AST	No.		Sample Type	Serum ↓
Chemistry	AST/GOT			Print Name	AST
Reaction Type	Kinetic ↓			Reaction Direction	negative ↓
Pri Wave	340nm ↓			Sec Wave	412nm ↓
Unit	U/L ↓			Decimal	0 ↓
Blank Time	0	0		Reaction Time	20	30

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	20 uL	uL	uL	R1	160 uL	uL
Decreased	uL	uL	uL	R2	40 uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

TKS TOSEE KIMIA SAADAT
توسعه کیمیای سعادت

BIOMEDIC

Mindray BS 800

Complement C3

Chem	C3	No.		Sample Type	Serum
Chemistry	Complement C3			Print Name	C3
Reaction Type	Endpoint			Reaction Direction	Positive
Pri Wave	340nm			Sec Wave	700nm
Unit	mg/dl			Decimal	0
Blank Time	15	16		Reaction Time	32	33

	Sample Vol		Aspirated		Diluent			Reagent Vol		Diluent	
Standard	2	uL		uL		uL	R1	180	uL		uL
Decreased		uL		uL		uL	R2	35	uL		uL
Increased		uL		uL		uL	R3		uL		uL
							R4		uL		uL

Calibration Settings

Math Model	Logit-Log 4P		
Factor		Replicates	1

TKS TOSEE KIMIA SAADAT
توسعه کیمیای سعادت

BIOMEDIC

Mindray BS 800

Complement C4

Chem	C4	No.		Sample Type	Serum
Chemistry	Complement C4			Print Name	C4
Reaction Type	Endpoint			Reaction Direction	Positive
Pri Wave	340nm			Sec Wave	700nm
Unit	mg/dl			Decimal	0.1
Blank Time	15	16		Reaction Time	32	33

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	4 uL	uL	uL	R1	180 uL	uL
Decreased	uL	uL	uL	R2	35 uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Logit-Log 4P		
Factor		Replicates	1

TKS TOSEE KIMIA SAADAT
توسعه کیمیای سعادت

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Mindray BS 800

Calcium Arsenazo III

Chem	CA	No.		Sample Type	Serum
Chemistry	Calcium Arsenazo III			Print Name	CA
Reaction Type	Endpoint			Reaction Direction	Positive
Pri Wave	660nm			Sec Wave	800nm
Unit	mg/dl			Decimal	0.1
Blank Time	3	4		Reaction Time	14	15

	Sample Vol		Aspirated		Diluent			Reagent Vol		Diluent	
Standard	4	uL		uL		uL	R1	200	uL		uL
Decreased		uL		uL		uL	R2		uL		uL
Increased		uL		uL		uL	R3		uL		uL
							R4		uL		uL

Calibration Settings

Math Model	Two-point linear		
Factor		Replicates	2

TKS TOSEE KIMIA SAADAT
توسعه کیمیای سعادت

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Mindray BS 800

Cholesterol

Chem	CHOL	No.		Sample Type	Serum
Chemistry	Cholesterol			Print Name	CHOL
Reaction Type	Endpoint			Reaction Direction	Positive
Pri Wave	505			Sec Wave	700nm
Unit	mg/dl			Decimal	0
Blank Time	3	4		Reaction Time	32	33

	Sample Vol		Aspirated		Diluent			Reagent Vol		Diluent	
Standard	2	uL		uL		uL	R1	200	uL		uL
Decreased		uL		uL		uL	R2		uL		uL
Increased		uL		uL		uL	R3		uL		uL
							R4		uL		uL

Calibration Settings

Math Model	Two-point linear		
Factor		Replicates	2

TKS TOSEE KIMIA SAADAT
توسعه کیمیای سعادت

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Mindray BS 800

CPK

Chem	CPK	No.		Sample Type	Serum ↓
Chemistry	CPK			Print Name	CPK
Reaction Type	Kinetic ↓			Reaction Direction	Positive ↓
Pri Wave	340nm ↓			Sec Wave	546nm ↓
Unit	U/L ↓			Decimal	0 ↓
Blank Time	0	0		Reaction Time	24	33

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	8 uL	uL	uL	R1	150 uL	uL
Decreased	uL	uL	uL	R2	30 uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

TKS TOSEE KIMIA SAADAT توسعه کیمیای سعادت

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Mindray BS 800

Creatinine (Jaffe)

Chem	CREAT	No.		Sample Type	Serum
Chemistry	Creatinine (Jaffe)			Print Name	CREAT
Reaction Type	Fixed-time			Reaction Direction	Positive
Pri Wave	505nm			Sec Wave	700nm
Unit	mg/dl			Decimal	0.01
Blank Time	0	0		Reaction Time	19	24

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	24 uL	uL	uL	R1	120 uL	uL
Decreased	uL	uL	uL	R2	120 uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear		
Factor		Replicates	2

TKS TOSEE KIMIA SAADAT
توسعه کیمیای سعادت

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Mindray BS 800

C-Reactive Protein (CRP)

Chem	CRP	No.		Sample Type	Serum
Chemistry	C-Reactive Protein (CRP)			Print Name	CRP
Reaction Type	Endpoint			Reaction Direction	Positive
Pri Wave	340nm			Sec Wave	700nm
Unit	mg/L			Decimal	0.1
Blank Time	15	16		Reaction Time	32	33

	Sample Vol		Aspirated		Diluent		Reagent Vol		Diluent	
Standard	9	uL		uL		uL	R1	160	uL		uL
Decreased		uL		uL		uL	R2	30	uL		uL
Increased		uL		uL		uL	R3		uL		uL
							R4		uL		uL

Calibration Settings

Math Model	Spline		
Factor		Replicates	1

TKS TOSEE KIMIA SAADAT
توسعه کیمیای سعادت

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Mindray BS 800

Glucose

Chem	GLU	No.		Sample Type	Serum ↓
Chemistry	Glucose			Print Name	GLU
Reaction Type	Endpoint ↓			Reaction Direction	Positive ↓
Pri Wave	505nm ↓			Sec Wave	700nm ↓
Unit	mg/dl ↓			Decimal	0 ↓
Blank Time	3	4		Reaction Time	32	33

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	2 uL	uL	uL	R1	200 uL	uL
Decreased	uL	uL	uL	R2	0 uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

TKS TOSEE KIMIA SAADAT
توسعه کیمیای سعادت

BIOMEDIC

Mindray BS 800

HDL Cholesterol

Chem	HDL	No.		Sample Type	Serum
Chemistry	HDL Cholesterol			Print Name	HDL
Reaction Type	Endpoint			Reaction Direction	Positive
Pri Wave	605nm			Sec Wave	700nm
Unit	mg/dl			Decimal	0
Blank Time	15	16		Reaction Time	32	33

	Sample Vol		Aspirated		Diluent			Reagent Vol		Diluent	
Standard	2	uL		uL		uL	R1	150	uL		uL
Decreased		uL		uL		uL	R2	50	uL		uL
Increased		uL		uL		uL	R3		uL		uL
							R4		uL		uL

Calibration Settings

Math Model	Two-point linear		
Factor		Replicates	2

TKS TOSEE KIMIA SAADAT
توسعه کیمیای سعادت

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Mindray BS 800

Immunoglobulin A

Chem	IgA	No.		Sample Type	Serum
Chemistry	Immunoglobulin A			Print Name	IgA
Reaction Type	Endpoint			Reaction Direction	Positive
Pri Wave	340nm			Sec Wave	700nm
Unit	mg/dl			Decimal	0
Blank Time	15	16		Reaction Time	32	33

	Sample Vol		Aspirated		Diluent			Reagent Vol		Diluent	
Standard	2	uL		uL		uL	R1	180	uL		uL
Decreased		uL		uL		uL	R2	35	uL		uL
Increased		uL		uL		uL	R3		uL		uL
							R4		uL		uL

Calibration Settings

Math Model	Logit-Log 4P		
Factor		Replicates	1

TKS TOSEE KIMIA SAADAT
توسعه کیمیای سعادت

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Mindray BS 800

Immunoglobulin G

Chem	IgG	No.		Sample Type	Serum
Chemistry	Immunoglobulin G			Print Name	IgG
Reaction Type	Endpoint			Reaction Direction	Positive
Pri Wave	605nm			Sec Wave	0
Unit	mg/dl			Decimal	0
Blank Time	15	16		Reaction Time	32	33

	Sample Vol	Aspirated	Diluent
Standard	2 uL	uL	uL
Decreased	uL	uL	uL
Increased	uL	uL	uL

	Reagent Vol	Diluent
R1	180 uL	uL
R2	35 uL	uL
R3	uL	uL
R4	uL	uL

Calibration Settings

Math Model	Logit-Log 4P		
Factor		Replicates	1

TKS TOSEE KIMIA SAADAT
توسعه کیمیای سعادت

BIOMEDIC

Mindray BS 800

Immunoglobulin M

Chem	IgM	No.		Sample Type	Serum ↓
Chemistry	Immunoglobulin M			Print Name	IgM
Reaction Type	Endpoint ↓			Reaction Direction	Positive ↓
Pri Wave	340nm ↓			Sec Wave	700nm ↓
Unit	mg/dl ↓			Decimal	0 ↓
Blank Time	15	16		Reaction Time	32	33

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	2 uL	uL	uL	R1	180 uL	uL
Decreased	uL	uL	uL	R2	35 uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Logit-Log 4P ↓		
Factor		Replicates	1 ↓

TKS TOSEE KIMIA SAADAT توسعه کیمیای سعادت

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Mindray BS 800

Lactate Dehydrogenase (LDH)

Chem	LDH	No.		Sample Type	Serum
Chemistry	Lactate Dehydrogenase (LDH)			Print Name	LDH
Reaction Type	Kinetic			Reaction Direction	negative
Pri Wave	340nm			Sec Wave	412nm
Unit	U/L			Decimal	0
Blank Time	0	0		Reaction Time	20	30

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	3 uL	uL	uL	R1	150 uL	uL
Decreased	uL	uL	uL	R2	30 uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear		
Factor		Replicates	2

TKS TOSEE KIMIA SAADAT
توسعه کیمیای سعادت

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Mindray BS 800

LDL Cholesterol

Chem	LDL	No.		Sample Type	Serum
Chemistry	LDL Cholesterol			Print Name	LDL
Reaction Type	Endpoint			Reaction Direction	Positive
Pri Wave	546nm			Sec Wave	700nm
Unit	mg/dl			Decimal	0
Blank Time	15	16		Reaction Time	32	33

	Sample Vol		Aspirated		Diluent		Reagent Vol		Diluent	
Standard	2	uL		uL		uL	R1	150	uL		uL
Decreased		uL		uL		uL	R2	50	uL		uL
Increased		uL		uL		uL	R3		uL		uL
							R4		uL		uL

Calibration Settings

Math Model	Two-point linear		
Factor		Replicates	2

TKS TOSEE KIMIA SAADAT
توسعه کیمیای سعادت

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Mindray BS 800

Lipase

Chem	LIP	No.		Sample Type	Serum ↓
Chemistry	Lipase			Print Name	LIP
Reaction Type	Kinetic ↓			Reaction Direction	Positive ↓
Pri Wave	570nm ↓			Sec Wave	700nm ↓
Unit	U/L ↓			Decimal	0 ↓
Blank Time	0	0		Reaction Time	22	28

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	4 uL	uL	uL	R1	160 uL	uL
Decreased	uL	uL	uL	R2	40 uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

TKS TOSEE KIMIA SAADAT
توسعه کیمیای سعادت

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Mindray BS 800

Micro Albumin

Chem	MALB	No.		Sample Type	Serum ↓
Chemistry	Micro Albumin			Print Name	MALB
Reaction Type	Endpoint ↓			Reaction Direction	Positive ↓
Pri Wave	340nm ↓			Sec Wave	700nm ↓
Unit	mg/L ↓			Decimal	0.1 ↓
Blank Time	15	16		Reaction Time	32	33

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	12 uL	uL	uL	R1	150 uL	uL
Decreased	uL	uL	uL	R2	30 uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Logit-Log 4P ↓		
Factor		Replicates	1 ↓

TKS TOSEE KIMIA SAADAT
توسعه کیمیای سعادت

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Mindray BS 800

CKMB

Chem	CKMB	No.		Sample Type	Serum ↓
Chemistry	CKMB			Print Name	CKMB
Reaction Type	Kinetic ↓			Reaction Direction	Positive ↓
Pri Wave	340nm ↓			Sec Wave	546nm ↓
Unit	U/L ↓			Decimal	0 ↓
Blank Time	0	0		Reaction Time	24	33

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	8 uL	uL	uL	R1	150 uL	uL
Decreased	uL	uL	uL	R2	30 uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

TKS TOSEE KIMIA SAADAT
توسعه کیمیای سعادت

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Mindray BS 800

Triglycerides

Chem	TG	No.		Sample Type	Serum
Chemistry	Triglycerides			Print Name	TG
Reaction Type	Endpoint			Reaction Direction	Positive
Pri Wave	505			Sec Wave	700nm
Unit	mg/dl			Decimal	0
Blank Time	3	4		Reaction Time	32	33

	Sample Vol		Aspirated		Diluent			Reagent Vol		Diluent	
Standard	2	uL		uL		uL	R1	200	uL		uL
Decreased		uL		uL		uL	R2		uL		uL
Increased		uL		uL		uL	R3		uL		uL
							R4		uL		uL

Calibration Settings

Math Model	Two-point linear		
Factor		Replicates	2

TKS TOSEE KIMIA SAADAT
توسعه کیمیای سعادت

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Mindray BS 800

Total Protein

Chem	T.PRO	No.		Sample Type	Serum
Chemistry	Total Protein			Print Name	T.PRO
Reaction Type	Endpoint			Reaction Direction	Positive
Pri Wave	546nm			Sec Wave	700nm
Unit	g/dl			Decimal	0.01
Blank Time	3	4		Reaction Time	19	20

	Sample Vol		Aspirated		Diluent			Reagent Vol		Diluent	
Standard	5	uL		uL		uL	R1	200	uL		uL
Decreased		uL		uL		uL	R2		uL		uL
Increased		uL		uL		uL	R3		uL		uL
							R4		uL		uL

Calibration Settings

Math Model	Two-point linear		
Factor		Replicates	2

TKS TOSEE KIMIA SAADAT
توسعه کیمیای سعادت

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Mindray BS 800

Uric acid

Chem	UA	No.		Sample Type	Serum ↓
Chemistry	Uric acid			Print Name	UA
Reaction Type	Endpoint ↓			Reaction Direction	Positive ↓
Pri Wave	546nm ↓			Sec Wave	700nm ↓
Unit	mg/dl ↓			Decimal	0.1 ↓
Blank Time	0	0		Reaction Time	32	33

	Sample Vol		Aspirated		Diluent			Reagent Vol		Diluent	
Standard	6	uL		uL		uL	R1	120	uL		uL
Decreased		uL		uL		uL	R2	120	uL		uL
Increased		uL		uL		uL	R3		uL		uL
							R4		uL		uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

TKS TOSEE KIMIA SAADAT
توسعه کیمیای سعادت

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Mindray BS 800

Urea/BUN

Chem	UREA	No.		Sample Type	Serum
Chemistry	Urea/BUN			Print Name	UREA
Reaction Type	Fixed-time			Reaction Direction	negative
Pri Wave	340nm			Sec Wave	412nm
Unit	mg/dl			Decimal	0.1
Blank Time	0	0		Reaction Time	19	24

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	2 uL	uL	uL	R1	160 uL	uL
Decreased	uL	uL	uL	R2	40 uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear		
Factor		Replicates	2

TKS TOSEE KIMIA SAADAT
توسعه کیمیای سعادت

BIOMEDIC

Mindray BS 800

ZINC

Chem	ZINC	No.		Sample Type	Serum ↓
Chemistry	ZINC			Print Name	ZINC
Reaction Type	Endpoint ↓			Reaction Direction	Positive ↓
Pri Wave	570nm ↓			Sec Wave	700nm ↓
Unit	µg/dl ↓			Decimal	0 ↓
Blank Time	3	4		Reaction Time	14	15

	Sample Vol		Aspirated		Diluent			Reagent Vol		Diluent	
Standard	8	uL		uL		uL	R1	160	uL		uL
Decreased		uL		uL		uL	R2		uL		uL
Increased		uL		uL		uL	R3		uL		uL
							R4		uL		uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

TKS TOSEE KIMIA SAADAT
توسعه کیمیای سعادت