BIOMEDIC

biolis 50i
prestige 50i

Albumin (BCG)

Item No	*	Item Name	ALB	FULLNAME	ALB	Optical

Data Information

Units	g/dl
Decimals	2

Analysis

Type	End point method
Main Wave Length	600
Sub Wave Length	700
Method	Bromocresol Green

Correlation

Slope		Intercept
Y =	1	* X + 0

Aspiration

Kind	Kind	Vol.	Units
Vol.	Sample	2	ul
	Reagent 1	200	ul

Calibration

Type: Linear

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability			
Factor MIN		Factor MAX	
CV MAX			

Data Process

Read

	Start	End
Main	28	30
Sub		

	low	High
Abs. Limit	-3	3

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

biolis 50i
prestige 50i

Alkaline phosphatase

Item No: * ↓ | Item Name: ALP | FULLNAME: ALP | Optical

Data Information

Units	U/L
Decimals	0

Analysis

Type	Rate method
Main Wave Length	405
Sub Wave Length	700
Method	DGKC

Correlation

Slope		Intercept
Y =	1	* X + 0

Aspiration

Kind	Kind	Vol.	Units
Vol.	Sample	3	ul
	Reagent 1	120	ul
	Reagent 2	30	ul

Calibration

Type: Linear ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability:

Factor MIN: | Factor MAX:

CV MAX:

Data Process

Read

	Start	End
Main	45	70
Sub		

Abs. Limit	low	High
	-3	3

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

biolis 50i
prestige 50i

ALT/GPT

Item No * ↓ | Item Name ALT | FULLNAME ALT | Optical

Data Information

Units	U/L
Decimals	0

Analysis

Type	Rate method
Main Wave Length	340
Sub Wave Length	405
Method	IFCC

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind	Vol.	Units
Sample	15	ul
Reagent 1	120	ul
Reagent 2	30	ul

Calibration

Type: Linear ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability:

Factor MIN: | Factor MAX:

CV MAX:

Data Process

Read

	Start	End
Main	45	70
Sub		

	low	High
Abs. Limit	-3	3

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

biolis 50i
prestige 50i

Amylase - total

Item No * ↓ | Item Name AMY | FULLNAME AMY | Optical

Data Information

Units	U/L
Decimals	0

Analysis

Type	Rate method
Main Wave Length	405
Sub Wave Length	700
Method	CNPG3

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind	Vol.	Units
Sample	3	ul
Reagent 1	150	ul

Calibration

Type: Linear ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability			
Factor MIN		Factor MAX	
CV MAX			

Data Process

Read	Start	End
Main	14	34
Sub		

Abs. Limit	low	High
	-3	3

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

biolis 50i
prestige 50i

AST/GOT

Item No: * ↓ | Item Name: AST | FULLNAME: AST | Optical

Data Information

Units	U/L
Decimals	0

Analysis

Type	Rate method
Main Wave Length	340
Sub Wave Length	405
Method	IFCC

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind	Kind	Vol.	Units
Vol.	Sample	15	ul
	Reagent 1	120	ul
	Reagent 2	30	ul

Calibration

Type: Linear ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability:

Factor MIN: | Factor MAX:

CV MAX:

Data Process

Read

	Start	End
Main	45	70
Sub		

Abs. Limit	low	High
	-3	3

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

biolis 50i
prestige 50i

Complement 3

Item No * ↓ | Item Name C3 | FULLNAME C3 | Optical

Data Information

Units	mg/dl
Decimals	0

Analysis

Type	End point method
Main Wave Length	340
Sub Wave Length	700
Method	Immunoturbidimetric

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind	Vol.	Units
Sample	2	ul
Reagent 1	180	ul
Reagent 2	35	ul

Calibration

Type: Logit.log ↓

Blank	0	#1	#	#2	#
#3	#	#4	#	#5	#
#6					

Calib Stability | Factor MIN | Factor MAX | CV MAX

Data Process

Read

	Start	End
Main	72	74
Sub	32	34

Abs. Limit	low	High
	-3	3

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

biolis 50i
prestige 50i

Complement 4

Item No * ↓ | Item Name C4 | FULLNAME C4 | Optical

Data Information

Units	mg/dl
Decimals	0

Analysis

Type	End point method
Main Wave Length	340
Sub Wave Length	700
Method	Immunoturbidimetric

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind	Vol.	Units
Sample	4	ul
Reagent 1	180	ul
Reagent 2	35	ul

Calibration

Type: Logit.log ↓

Blank	0	#1	#	#2	#
#3	#	#4	#	#5	#
#6					

Calib Stability | Factor MIN | Factor MAX | CV MAX

Data Process

Read

	Start	End
Main	72	74
Sub	32	34

Abs. Limit	low	High
	-3	3

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

biolis 50i
prestige 50i

Calcium

Item No * ↓ | Item Name CA | FULLNAME CA | Optical

Data Information

Units	mg/dl
Decimals	2

Analysis

Type	End point method
Main Wave Length	660
Sub Wave Length	800
Method	Arsenazo III

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind	Vol.	Units
Sample	3	ul
Reagent 1	200	ul

Calibration

Type: Linear ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability | Factor MIN | Factor MAX | CV MAX

Data Process

Read

	Start	End
Main	28	30
Sub		

Abs. Limit	low	High
	-3	3

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

biolis 50i
prestige 50i

Cholesterol

Item No * ↓ | Item Name CHOL | FULLNAME CHOL | Optical

Data Information

Units	mg/dl
Decimals	0

Analysis

Type	End point method
Main Wave Length	505
Sub Wave Length	700
Method	Cholesterol Oxidase

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind	Vol.	Units
Sample	2	ul
Reagent 1	200	ul

Calibration

Type: Linear ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability | Factor MIN | Factor MAX | CV MAX

Data Process

Read

	Start	End
Main	72	74
Sub		

	low	High
Abs. Limit	-3	3

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

biolis 50i
prestige 50i

CK-NAC

Item No * ↓ | Item Name CPK | FULLNAME CPK | Optical

Data Information

Units	U/L
Decimals	0

Analysis

Type	Rate method
Main Wave Length	340
Sub Wave Length	546
Method	IFCC

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind	Vol.	Units
Sample	8	ul
Reagent 1	150	ul
Reagent 2	30	ul

Calibration

Type: Linear ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability | Factor MIN | Factor MAX | CV MAX

Data Process

Read

	Start	End
Main	53	74
Sub		

	low	High
Abs. Limit	-3	3

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

biolis 50i
prestige 50i

Creatinine (Jaffe)

Item No: * ↓ | Item Name: CREAT | FULLNAME: CREAT | Optical

Data Information

Units	mg/dl
Decimals	2

Analysis

Type	Rate method
Main Wave Length	505
Sub Wave Length	700
Method	Jaffe

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind	Vol.	Units
Sample	20	ul
Reagent 1	100	ul
Reagent 2	100	ul

Calibration

Type: Linear ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability:

Factor MIN: | Factor MAX:

CV MAX:

Data Process

Read	Start	End
Main	43	53
Sub		

Abs. Limit	low	High
	-3	3

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

biolis 50i
prestige 50i

C-REACTIVE PROTEIN

Item No * ↓ | Item Name CRP | FULLNAME CRP | Optical

Data Information

Units	mg/L
Decimals	1

Analysis

Type	End point method
Main Wave Length	340
Sub Wave Length	700
Method	Immunoturbidimetric

Correlation

Slope		Intercept
Y =	1	* X + 0

Aspiration

Kind	Kind	Vol.	Units
Vol.	Sample	8	ul
	Reagent 1	150	ul
	Reagent 2	30	ul

Calibration

Type: Logit.log ↓

Blank	0	#1	#	#2	#
#3	#	#4	#	#5	#
#6					

Calib Stability

Factor MIN | Factor MAX

CV MAX

Data Process

Read

	Start	End
Main	72	74
Sub	32	34

Abs. Limit	low	High
	-3	3

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

biolis 50i
prestige 50i

Glucose

Item No	*	Item Name	GLU	FULLNAME	GLU	Optical

Data Information

Units	mg/dl
Decimals	0

Analysis

Type	End point method
Main Wave Length	505
Sub Wave Length	700
Method	Glucose oxidase

Correlation

Slope		Intercept
Y =	1	* X + 0

Aspiration

Kind	Kind	Vol.	Units
Vol.	Sample	2	ul
	Reagent 1	200	ul

Calibration

Type	Linear

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability			
Factor MIN		Factor MAX	
CV MAX			

Data Process

Read

	Start	End
Main	72	74
Sub		

Abs. Limit	low	High
	-3	3

\# غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

biolis 50i
prestige 50i

HDL Cholesterol direct

Item No: * ↓ | Item Name: HDL-C | FULLNAME: HDL-C | Optical

Data Information

Units	mg/dl
Decimals	0

Analysis

Type	End point method
Main Wave Length	600
Sub Wave Length	700
Method	DIRECT

Correlation

Slope		Intercept
Y =	1	* X + 0

Aspiration

Kind	Kind	Vol.	Units
Vol.	Sample	2	ul
	Reagent 1	150	ul
	Reagent 2	50	ul

Calibration

Type: Linear ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability:

Factor MIN: | Factor MAX:

CV MAX:

Data Process

Read

	Start	End
Main	72	74
Sub	32	34

Abs. Limit	low	High
	-3	3

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

biolis 50i
prestige 50i

Immunoglobulin A

Item No * ↓ | Item Name IgA | FULLNAME IgA | Optical

Data Information

Units	mg/dl
Decimals	0

Analysis

Type	End point method
Main Wave Length	340
Sub Wave Length	700
Method	Immunoturbidimetric

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind	Vol.	Units
Sample	2	ul
Reagent 1	180	ul
Reagent 2	35	ul

Calibration

Type: Logit.log ↓

Blank	0	#1	#	#2	#
#3	#	#4	#	#5	#
#6					

Calib Stability | Factor MIN | Factor MAX | CV MAX

Data Process

Read

	Start	End
Main	72	74
Sub	32	34

Abs. Limit	low	High
	-3	3

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

biolis 50i
prestige 50i

Immunoglobulin G

Item No * ↓ | Item Name IgG | FULLNAME IgG | Optical

Data Information

Units	mg/dl
Decimals	0

Analysis

Type	End point method
Main Wave Length	600
Sub Wave Length	700
Method	Immunoturbidimetric

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind	Vol.	Units
Sample	2	ul
Reagent 1	180	ul
Reagent 2	35	ul

Calibration

Type: Logit.log ↓

Blank	0	#1	#	#2	#
#3	#	#4	#	#5	#
#6					

Calib Stability | Factor MIN | Factor MAX | CV MAX

Data Process

Read

	Start	End
Main	72	74
Sub	32	34

	low	High
Abs. Limit	-3	3

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

biolis 50i
prestige 50i

Immunoglobulin M

Item No * ↓ | Item Name IgM | FULLNAME IgM | Optical

Data Information

Units	mg/dl
Decimals	0

Analysis

Type	End point method
Main Wave Length	340
Sub Wave Length	700
Method	Immunoturbidimetric

Correlation

Slope | Intercept

Y = 1 * X + 0

Aspiration

Kind	Kind	Vol.	Units
Vol.	Sample	2	ul
	Reagent 1	180	ul
	Reagent 2	35	ul

Calibration

Type: Logit.log ↓

Blank	0	#1	#	#2	#
#3	#	#4	#	#5	#
#6					

Calib Stability

Factor MIN | Factor MAX

CV MAX

Data Process

Read

	Start	End
Main	72	74
Sub	32	34

Abs. Limit	low	High
	-3	3

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

biolis 50i
prestige 50i

Lactate Dehydrogenase

Item No * ↓ | Item Name LDH | FULLNAME LDH | Optical

Data Information

Units	U/L
Decimals	0

Analysis

Type	Rate method
Main Wave Length	340
Sub Wave Length	700
Method	(P-L) SCE

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind	Vol.	Units
Sample	3	ul
Reagent 1	150	ul
Reagent 2	30	ul

Calibration

Type: Linear ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability:

Factor MIN: | Factor MAX:

CV MAX:

Data Process

Read

	Start	End
Main	45	70
Sub		

Abs. Limit	low	High
	-3	3

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

biolis 50i
prestige 50i

LDL Cholesterol direct

Item No: * ↓ | Item Name: LDL-C | FULLNAME: LDL-C | Optical

Data Information

Units	mg/dl
Decimals	0

Analysis

Type	End point method
Main Wave Length	546
Sub Wave Length	700
Method	DIRECT

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind	Kind	Vol.	Units
Vol.	Sample	2	ul
	Reagent 1	150	ul
	Reagent 2	50	ul

Calibration

Type: Linear ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability:

Factor MIN: | Factor MAX:

CV MAX:

Data Process

Read

	Start	End
Main	72	74
Sub	32	34

Abs. Limit	low	High
	-3	3

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

biolis 50i
prestige 50i

Lipase

Item No * ↓ | Item Name LIP | FULLNAME LIP | Optical

Data Information

Units	U/L
Decimals	1

Analysis

Type	Rate method
Main Wave Length	570
Sub Wave Length	700
Method	Colorimetric

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind	Vol.	Units
Sample	4	ul
Reagent 1	160	ul
Reagent 2	40	ul

Calibration

Type: Linear ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability:
Factor MIN: | Factor MAX:
CV MAX:

Data Process

Read

	Start	End
Main	53	70
Sub		

Abs. Limit	low	High
	-3	3

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

biolis 50i
prestige 50i

Micro Albumin

Item No * ↓ | Item Name MALB | FULLNAME MALB | Optical

Data Information

Units	mg/L
Decimals	1

Analysis

Type	End point method
Main Wave Length	340
Sub Wave Length	700
Method	Immunoturbidimetric

Correlation

Slope		Intercept
Y =	1	* X + 0

Aspiration

Kind	Kind	Vol.	Units
Vol.	Sample	12	ul
	Reagent 1	150	ul
	Reagent 2	30	ul

Calibration

Type: Logit.log ↓

Blank	0	#1	#	#2	#
#3	#	#4	#	#5	#
#6					

Calib Stability

Factor MIN | Factor MAX

CV MAX

Data Process

Read

	Start	End
Main	72	74
Sub	32	34

Abs. Limit	low	High
	-3	3

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

biolis 50i
prestige 50i

CK-MB

Item No: * ↓ | Item Name: CK-MB | FULLNAME: CK-MB | Optical

Data Information

Units	U/L
Decimals	1

Analysis

Type	Rate method
Main Wave Length	340
Sub Wave Length	546
Method	Immunoinhibition

Correlation

Y = 1 * X + 0 (Slope = 1, Intercept = 0)

Aspiration

Kind	Vol.	Units
Sample	8	ul
Reagent 1	150	ul
Reagent 2	30	ul

Calibration

Type: Linear ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability			
Factor MIN		Factor MAX	
CV MAX			

Data Process

Read

	Start	End
Main	53	74
Sub		

	low	High
Abs. Limit	-3	3

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

biolis 50i
prestige 50i

Magnesium

Item No * ↓ | Item Name MG | FULLNAME MG | Optical

Data Information

Units	mg/dl
Decimals	1

Analysis

Type	End point method
Main Wave Length	505
Sub Wave Length	700
Method	Xylidyl Blue

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind	Vol.	Units
Sample	2	ul
Reagent 1	200	ul

Calibration

Type: Linear ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability | Factor MIN | Factor MAX | CV MAX

Data Process

Read

	Start	End
Main	28	30
Sub		

	low	High
Abs. Limit	-3	3

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

biolis 50i
prestige 50i

Phosphorus

Item No * ↓ | Item Name PHOS | FULLNAME PHOS | Optical

Data Information

Units	mg/dl
Decimals	1

Analysis

Type	End point method
Main Wave Length	340
Sub Wave Length	700
Method	Phosphomolybdate. UV

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind	Vol.	Units
Sample	2	ul
Reagent 1	200	ul

Calibration

Type: Linear ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability | Factor MIN | Factor MAX | CV MAX

Data Process

Read

	Start	End
Main	28	30
Sub		

	low	High
Abs. Limit	-3	3

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

biolis 50i
prestige 50i

Triglycerides

Item No * ↓ | Item Name TG | FULLNAME TG | Optical

Data Information

Units	mg/dl
Decimals	0

Analysis

Type	End point method
Main Wave Length	505
Sub Wave Length	700
Method	GPO-POD

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind	Vol.	Units
Sample	2	ul
Reagent 1	200	ul

Calibration

Type: Linear ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability			
Factor MIN		Factor MAX	
CV MAX			

Data Process

Read	Start	End
Main	72	74
Sub		

Abs. Limit	low	High
	-3	3

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

biolis 50i
prestige 50i

Total Protein

Item No * ↓ | Item Name T.PRO | FULLNAME T.PRO | Optical

Data Information

Units	g/dl
Decimals	2

Analysis

Type	End point method
Main Wave Length	546
Sub Wave Length	700
Method	Biuret

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind	Vol.	Units
Sample	5	ul
Reagent 1	200	ul

Calibration

Type: Linear ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability			
Factor MIN		Factor MAX	
CV MAX			

Data Process

Read

	Start	End
Main	28	30
Sub		

	low	High
Abs. Limit	-3	3

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

biolis 50i
prestige 50i

Uric acid

Item No * ↓ | Item Name UA | FULLNAME UA | Optical

Data Information

Units	mg/dl
Decimals	1

Analysis

Type	End point method
Main Wave Length	546
Sub Wave Length	700
Method	Uricase Peroxidase

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind	Vol.	Units
Sample	5	ul
Reagent 1	100	ul
Reagent 2	100	ul

Calibration

Type: Linear ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability | Factor MIN | Factor MAX | CV MAX

Data Process

Read

	Start	End
Main	58	60
Sub		

	low	High
Abs. Limit	-3	3

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

biolis 50i
prestige 50i

Urea/BUN

Item No * ↓ | Item Name UREA | FULLNAME UREA | Optical

Data Information

Units	mg/dl
Decimals	1

Analysis

Type	Rate method
Main Wave Length	340
Sub Wave Length	700
Method	Urease kinetic

Correlation

Slope		Intercept
Y =	1	* X + 0

Aspiration

Kind	Kind	Vol.	Units
Vol.	Sample	2	ul
	Reagent 1	160	ul
	Reagent 2	40	ul

Calibration

Type: Linear ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability

Factor MIN | Factor MAX

CV MAX

Data Process

Read

	Start	End
Main	43	53
Sub		

Abs. Limit	low	High
	-3	3

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

biolis 50i
prestige 50i

ZINC

Item No * ↓ | Item Name ZN | FULLNAME ZN | Optical

Data Information

Units	µg/dl
Decimals	0

Analysis

Type	End point method
Main Wave Length	570
Sub Wave Length	700
Method	Colorimetric

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind	Kind	Vol.	Units
Vol.	Sample	8	ul
	Reagent 1	160	ul

Calibration

Type: Linear ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability | Factor MIN | Factor MAX | CV MAX

Data Process

Read

	Start	End
Main	28	30
Sub		

Abs. Limit: low -3 | High 3

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001