BIOMEDIC

DIRUI
CS-T240 / CS-T300

Albumin BCG

Analyze Parameters

Test Item:	ALB	Test Full Name:	ALB	Decimal Digit:	2	Unit:	g/dl		Sample Blank		
Assay:	1-point end	Test Time:	13	Point:	20	0	0	0	Control Interval:	0	
Main Wave:	600	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin			
Normal:	3	0	0	3	0	0	
Decrement:	3	0	0	3	0	0	
Increment:	6	0	0	6	0	0	
Abs. Limit:	3.3			3.3	Positive Reaction		
Prozone Limit:	-3.3			-3.3	Lower Limit		

Reagent

	Vol.	Pos.
R1:	300	*
R2:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: ALB

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	3	0	0
(2)	#	S2	3	0	0
(3)			3	0	0
(4)			3	0	0
(5)			3	0	0
(6)			3	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

DIRUI
CS-T240 / CS-T300

Alkaline Phosphatase

Analyze Parameters

Test Item:	ALP	Test Full Name:	ALP	Decimal Digit:	0	Unit:	U/L		Sample Blank				
Assay:	Rate-A	Test Time:	13	Point:	24	36	0	0	Control Interval:	0			
Main Wave:	405	Second Wave:	660	Instrument Factor (Y=aX+b)	a=	1	b=	0					

Sample Vol.

	Serum			Urin		
Normal:	5	0	0	5	0	0
Decrement:	3	0	0	3	0	0
Increment:	10	0	0	10	0	0
Abs. Limit:	1.5			1.5	Positive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	200	*
R2:	50	*

Second half item of Two test analyze

Calibration Parameters

Test Item: ALP

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	5	0	0
(2)	#	S2	5	0	0
(3)			5	0	0
(4)			5	0	0
(5)			5	0	0
(6)			5	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

DIRUI
CS-T240 / CS-T300

Alanine Aminotransferase (ALT)

Analyze Parameters

Test Item:	ALT	Test Full Name:	ALT	Decimal Digit:	0	Unit:	U/L		Sample Blank				
Assay:	Rate-A	Test Time:	13	Point:	24	36	0	0	Control Interval:	0			
Main Wave:	340	Second Wave:	405	Instrument Factor (Y=aX+b)	a=	1	b=	0					

Sample Vol.

	Serum			Urin		
Normal:	25	0	0	25	0	0
Decrement:	3	0	0	3	0	0
Increment:	35	0	0	35	0	0
Abs. Limit:	0.5			0.5	Negetive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	200	*
R2:	50	*

Second half item of Two test analyze

Calibration Parameters

Test Item: ALT

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	25	0	0
(2)	#	S2	25	0	0
(3)			25	0	0
(4)			25	0	0
(5)			25	0	0
(6)			25	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

DIRUI
CS-T240 / CS-T300

Amylase CNPG3

Analyze Parameters

Test Item:	AMY	Test Full Name:	AMY	Decimal Digit:	0	Unit:	U/L		Sample Blank		
Assay:	Rate-A	Test Time:	13	Point:	4	16	0	0	Control Interval:	0	
Main Wave:	405	Second Wave:	660	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin		
Normal:	5	0	0	5	0	0
Decrement:	3	0	0	3	0	0
Increment:	10	0	0	10	0	0
Abs. Limit:	2			2	Positive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	250	*
R2:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: AMY

Caliration Methods: Linear (2-point) | Point: 2 | Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	5	0	0
(2)	#	S2	5	0	0
(3)			5	0	0
(4)			5	0	0
(5)			5	0	0
(6)			5	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

DIRUI
CS-T240 / CS-T300

Aspartate Aminotransferase (AST)

Analyze Parameters

Test Item:	AST	Test Full Name:	AST	Decimal Digit:	0	Unit:	U/L		Sample Blank		
Assay:	Rate-A	Test Time:	13	Point:	24	36	0	0	Control Interval:	0	
Main Wave:	340	Second Wave:	405	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin		
Normal:	25	0	0	25	0	0
Decrement:	3	0	0	3	0	0
Increment:	35	0	0	35	0	0
Abs. Limit:	0.5			0.5	Negetive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	200	*
R2:	50	*

Second half item of Two test analyze

Calibration Parameters

Test Item: AST

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	25	0	0
(2)	#	S2	25	0	0
(3)			25	0	0
(4)			25	0	0
(5)			25	0	0
(6)			25	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

DIRUI
CS-T240 / CS-T300

Complement 3 (C3)

Analyze Parameters

Test Item:	C3 V	Test Full Name:	C3	Decimal Digit:	0 V	Unit:	mg/dl V		Sample Blank			
Assay:	2-point end V	Test Time:	13 V	Point:	20	42	0	0	Control Interval:	0		
Main Wave:	340 V	Second Wave:	700 V	Instrument Factor (Y=aX+b)	a=	1	b=	0				

Sample Vol.

	Serum			Urin V		
Normal:	3	0	0	3	0	0
Decrement:	3	0	0	3	0	0
Increment:	4	0	0	4	0	0
Abs. Limit:	3.3			3.3	Positive Reaction V	
Prozone Limit:	-3.3			-3.3	Lower Limit V	

Reagent

	Vol.	Pos.
R1:	250	* V
R2:	50	* V

Second half item of Two test analyze V

Calibration Parameters

Test Item: C3 V

Caliration Methods: Logit-log 4P V V Point: 6 Span: 6

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1 V	3	0	0
(2)	#	S2 V	3	0	0
(3)	#	S3 V	3	0	0
(4)	#	S4 V	3	0	0
(5)	#	S5 V	3	0	0
(6)	#	S6 V	3	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

DIRUI
CS-T240 / CS-T300

Complement 4 (C4)

Analyze Parameters

Test Item:	C4	Test Full Name:	C4	Decimal Digit:	1	Unit:	mg/dl		Sample Blank			
Assay:	2-point end	Test Time:	13	Point:	20	42	0	0	Control Interval:	0		
Main Wave:	340	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0				

Sample Vol.

	Serum			Urin			
Normal:	6	0	0	6	0	0	
Decrement:	3	0	0	3	0	0	
Increment:	8	0	0	8	0	0	
Abs. Limit:	3.3			3.3	Positive Reaction		
Prozone Limit:	-3.3			-3.3	Lower Limit		

Reagent

	Vol.	Pos.
R1:	250	*
R2:	50	*

Second half item of Two test analyze

Calibration Parameters

Test Item: C4

Caliration Methods: Logit-log 4P | Point: 6 | Span: 6

Drift Rate Checkup:	3.3
Discreteness Checkup:	3.3 Abs.
Sensitivity Checkup:	0
Blank Horizontal Checkuo:	-3.3 - -3.3

Caliration Solution

(1)	0	S1	6	0	0
(2)	#	S2	6	0	0
(3)	#	S3	6	0	0
(4)	#	S4	6	0	0
(5)	#	S5	6	0	0
(6)	#	S6	6	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

DIRUI
CS-T240 / CS-T300

Calcium Arsenazo III

Analyze Parameters

Test Item:	CA	Test Full Name:	CA	Decimal Digit:	1	Unit:	mg/dl		Sample Blank		
Assay:	1-point end	Test Time:	13	Point:	20	0	0	0	Control Interval:	0	
Main Wave:	660	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin			
Normal:	3	0	0	3	0	0	
Decrement:	3	0	0	3	0	0	
Increment:	6	0	0	6	0	0	
Abs. Limit:	3.3			3.3	Positive Reaction		
Prozone Limit:	-3.3			-3.3	Lower Limit		

Reagent

	Vol.	Pos.
R1:	300	*
R2:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: CA

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	3	0	0
(2)	#	S2	3	0	0
(3)			3	0	0
(4)			3	0	0
(5)			3	0	0
(6)			3	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

DIRUI
CS-T240 / CS-T300

Cholesterol

Analyze Parameters

Test Item:	CHOL	Test Full Name:	CHOL	Decimal Digit:	0	Unit:	mg/dl		Sample Blank		
Assay:	1-point end	Test Time:	13	Point:	49	0	0	0	Control Interval:	0	
Main Wave:	505	Second Wave:	700	Instrument Factor (Y=aX+b)		a=	1	b=	0		

Sample Vol.

	Serum			Urin			
Normal:	3	0	0	3	0	0	
Decrement:	3	0	0	3	0	0	
Increment:	6	0	0	6	0	0	
Abs. Limit:	3.3			3.3	Positive Reaction		
Prozone Limit:	-3.3			-3.3	Lower Limit		

Reagent

	Vol.	Pos.
R1:	300	*
R2:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: CHOL

Caliration Methods: Linear (2-point) | Point: 2 | Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	3	0	0
(2)	#	S2	3	0	0
(3)			3	0	0
(4)			3	0	0
(5)			3	0	0
(6)			3	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

DIRUI
CS-T240 / CS-T300

Creatine Kinase (CK-NAC)

Analyze Parameters

Test Item:	CPK	Test Full Name:	CPK	Decimal Digit:	0	Unit:	U/L		Sample Blank				
Assay:	Rate-A	Test Time:	13	Point:	28	40	0	0	Control Interval:	0			
Main Wave:	340	Second Wave:	546	Instrument Factor (Y=aX+b)	a=	1	b=	0					

Sample Vol.

	Serum			Urin		
Normal:	10	0	0	10	0	0
Decrement:	3	0	0	3	0	0
Increment:	20	0	0	20	0	0
Abs. Limit:	1.5			1.5	Positive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	200	*
R2:	40	*

Second half item of Two test analyze

Calibration Parameters

Test Item: CPK

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	10	0	0
(2)	#	S2	10	0	0
(3)			10	0	0
(4)			10	0	0
(5)			10	0	0
(6)			10	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

DIRUI
CS-T240 / CS-T300

Creatinine Jaffe

Analyze Parameters

Test Item:	CREAT	Test Full Name:	CREAT	Decimal Digit:	2	Unit:	mg/dl		Sample Blank		
Assay:	2-point rate	Test Time:	13	Point:	23	29	0	0	Control Interval:	0	
Main Wave:	505	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin			
Normal:	24	0	0	24	0	0	
Decrement:	3	0	0	3	0	0	
Increment:	35	0	0	35	0	0	
Abs. Limit:	3.3			3.3	Positive Reaction		
Prozone Limit:	-3.3			-3.3	Lower Limit		

Reagent

	Vol.	Pos.
R1:	120	*
R2:	120	*

Second half item of Two test analyze

Calibration Parameters

Test Item: CREAT

Caliration Methods: Linear (2-point) | Point: 2 | Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	24	0	0
(2)	#	S2	24	0	0
(3)			24	0	0
(4)			24	0	0
(5)			24	0	0
(6)			24	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

DIRUI
CS-T240 / CS-T300

C - reactive protein (CRP)

Analyze Parameters

Test Item:	CRP	Test Full Name:	CRP	Decimal Digit:	1	Unit:	mg/L		Sample Blank		
Assay:	2-point end	Test Time:	13	Point:	20	42	0	0	Control Interval:	0	
Main Wave:	340	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin		
Normal:	15	0	0	15	0	0
Decrement:	3	0	0	3	0	0
Increment:	30	0	0	30	0	0
Abs. Limit:	3.3			3.3	Positive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	250	*
R2:	50	*

Second half item of Two test analyze

Calibration Parameters

Test Item: CRP

Caliration Methods: Logit-log 4P | Point: 6 | Span: 6

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	15	0	0
(2)	#	S2	15	0	0
(3)	#	S3	15	0	0
(4)	#	S4	15	0	0
(5)	#	S5	15	0	0
(6)	#	S6	15	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

DIRUI
CS-T240 / CS-T300

Glucose

Analyze Parameters

Test Item:	GLU	Test Full Name:	GLU	Decimal Digit:	0	Unit:	mg/dl		Sample Blank		
Assay:	1-point end	Test Time:	13	Point:	49	0	0	0	Control Interval:	0	
Main Wave:	505	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin			
Normal:	3	0	0	3	0	0	
Decrement:	3	0	0	3	0	0	
Increment:	6	0	0	6	0	0	
Abs. Limit:	3.3			3.3	Positive Reaction		
Prozone Limit:	-3.3			-3.3	Lower Limit		

Reagent

	Vol.	Pos.
R1:	300	*
R2:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: GLU

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	3	0	0
(2)	#	S2	3	0	0
(3)			3	0	0
(4)			3	0	0
(5)			3	0	0
(6)			3	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

DIRUI
CS-T240 / CS-T300

HDL Cholesterol Direct

Analyze Parameters

Test Item:	HDL	Test Full Name:	HDL	Decimal Digit:	0	Unit:	mg/dl		Sample Blank		
Assay:	2-point end	Test Time:	13	Point:	20	42	0	0	Control Interval:	0	
Main Wave:	600	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin			
Normal:	3	0	0	3	0	0	
Decrement:	3	0	0	3	0	0	
Increment:	6	0	0	6	0	0	
Abs. Limit:	3.3			3.3	Positive Reaction		
Prozone Limit:	-3.3			-3.3	Lower Limit		

Reagent

	Vol.	Pos.
R1:	240	*
R2:	80	*

Second half item of Two test analyze

Calibration Parameters

Test Item: HDL

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	3	0	0
(2)	#	S2	3	0	0
(3)			3	0	0
(4)			3	0	0
(5)			3	0	0
(6)			3	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

DIRUI
CS-T240 / CS-T300

Immunoglobulin A (IgA)

Analyze Parameters

Test Item:	IgA	Test Full Name:	IgA	Decimal Digit:	0	Unit:	mg/dl		Sample Blank		
Assay:	2-point end	Test Time:	13	Point:	20	42	0	0	Control Interval:	0	
Main Wave:	340	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin			
Normal:	3	0	0	3	0	0	
Decrement:	3	0	0	3	0	0	
Increment:	6	0	0	6	0	0	
Abs. Limit:	3.3			3.3	Positive Reaction		
Prozone Limit:	-3.3			-3.3	Lower Limit		

Reagent

	Vol.	Pos.
R1:	250	*
R2:	50	*

Second half item of Two test analyze

Calibration Parameters

Test Item: IgA

Caliration Methods: Logit-log 4P | Point: 6 | Span: 6

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	3	0	0
(2)	#	S2	3	0	0
(3)	#	S3	3	0	0
(4)	#	S4	3	0	0
(5)	#	S5	3	0	0
(6)	#	S6	3	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

DIRUI
CS-T240 / CS-T300

Immunoglobulin G (IgG)

Analyze Parameters

Test Item:	IgG	Test Full Name:	IgG	Decimal Digit:	0	Unit:	mg/dl		Sample Blank		
Assay:	2-point end	Test Time:	13	Point:	20	42	0	0	Control Interval:	0	
Main Wave:	600	Second Wave:	700	Instrument Factor (Y=aX+b)		a=	1	b=	0		

Sample Vol.

	Serum			Urin			
Normal:	3	0	0	3	0	0	
Decrement:	3	0	0	3	0	0	
Increment:	6	0	0	6	0	0	
Abs. Limit:	3.3			3.3	Positive Reaction		
Prozone Limit:	-3.3			-3.3	Lower Limit		

Reagent

	Vol.	Pos.
R1:	250	*
R2:	50	*

Second half item of Two test analyze

Calibration Parameters

Test Item: IgG

Caliration Methods: Logit-log 4P | Point: 6 | Span: 6

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	3	0	0
(2)	#	S2	3	0	0
(3)	#	S3	3	0	0
(4)	#	S4	3	0	0
(5)	#	S5	3	0	0
(6)	#	S6	3	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

DIRUI
CS-T240 / CS-T300

Immunoglobulin M (IgM)

Analyze Parameters

Test Item:	IgM	Test Full Name:	IgM	Decimal Digit:	0	Unit:	mg/dl		Sample Blank		
Assay:	2-point end	Test Time:	13	Point:	20	42	0	0	Control Interval:	0	
Main Wave:	340	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin			
Normal:	3	0	0	3	0	0	
Decrement:	3	0	0	3	0	0	
Increment:	6	0	0	6	0	0	
Abs. Limit:	3.3			3.3	Positive Reaction		
Prozone Limit:	-3.3			-3.3	Lower Limit		

Reagent

	Vol.	Pos.
R1:	250	*
R2:	50	*

Second half item of Two test analyze

Calibration Parameters

Test Item: IgM

Caliration Methods: Logit-log 4P | Point: 6 | Span: 6

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	3	0	0
(2)	#	S2	3	0	0
(3)	#	S3	3	0	0
(4)	#	S4	3	0	0
(5)	#	S5	3	0	0
(6)	#	S6	3	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

DIRUI
CS-T240 / CS-T300

Lactate Dehydrogenase (LDH)

Analyze Parameters

Test Item:	LDH	Test Full Name:	LDH	Decimal Digit:	0	Unit:	U/L		Sample Blank		
Assay:	Rate-A	Test Time:	13	Point:	24	36	0	0	Control Interval:	0	
Main Wave:	340	Second Wave:	405	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin		
Normal:	4	0	0	4	0	0
Decrement:	3	0	0	3	0	0
Increment:	8	0	0	8	0	0
Abs. Limit:	0.5			0.5	Negetive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	200	*
R2:	40	*

Second half item of Two test analyze

Calibration Parameters

Test Item: LDH

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	4	0	0
(2)	#	S2	4	0	0
(3)			4	0	0
(4)			4	0	0
(5)			4	0	0
(6)			4	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

DIRUI
CS-T240 / CS-T300

LDL Cholesterol Direct

Analyze Parameters

Test Item:	LDL	Test Full Name:	LDL	Decimal Digit:	0	Unit:	mg/dl		Sample Blank		
Assay:	2-point end	Test Time:	13	Point:	20	42	0	0	Control Interval:	0	
Main Wave:	546	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin			
Normal:	4	0	0	4	0	0	
Decrement:	3	0	0	3	0	0	
Increment:	6	0	0	6	0	0	
Abs. Limit:	3.3			3.3	Positive Reaction		
Prozone Limit:	-3.3			-3.3	Lower Limit		

Reagent

	Vol.	Pos.
R1:	240	*
R2:	80	*

Second half item of Two test analyze

Calibration Parameters

Test Item: LDL

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	4	0	0
(2)	#	S2	4	0	0
(3)			4	0	0
(4)			4	0	0
(5)			4	0	0
(6)			4	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

DIRUI
CS-T240 / CS-T300

Lipase

Analyze Parameters

Test Item:	LIP	Test Full Name:	LIP	Decimal Digit:	0	Unit:	U/L		Sample Blank			
Assay:	Rate-A	Test Time:	13	Point:	28	36	0	0	Control Interval:	0		
Main Wave:	570	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0				

Sample Vol.

	Serum			Urin		
Normal:	5	0	0	5	0	0
Decrement:	3	0	0	3	0	0
Increment:	10	0	0	10	0	0
Abs. Limit:	2			2	Positive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	200	*
R2:	50	*

Second half item of Two test analyze

Calibration Parameters

Test Item: LIP

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	5	0	0
(2)	#	S2	5	0	0
(3)			5	0	0
(4)			5	0	0
(5)			5	0	0
(6)			5	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

DIRUI
CS-T240 / CS-T300

Microalbumin

Analyze Parameters

Test Item:	MALB	Test Full Name:	MALB	Decimal Digit:	1	Unit:	mg/L		Sample Blank		
Assay:	2-point end	Test Time:	13	Point:	20	42	0	0	Control Interval:	0	
Main Wave:	340	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin		
Normal:	18	0	0	18	0	0
Decrement:	3	0	0	3	0	0
Increment:	35	0	0	35	0	0
Abs. Limit:	3.3			3.3	Positive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	250	*
R2:	50	*

Second half item of Two test analyze

Calibration Parameters

Test Item: MALB

Caliration Methods: Logit-log 4P | Point: 6 | Span: 6

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	18	0	0
(2)	#	S2	18	0	0
(3)	#	S3	18	0	0
(4)	#	S4	18	0	0
(5)	#	S5	18	0	0
(6)	#	S6	18	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

DIRUI
CS-T240 / CS-T300

CK - MB

Analyze Parameters

Test Item:	CKMB	Test Full Name:	CKMB	Decimal Digit:	1	Unit:	U/L		Sample Blank		
Assay:	Rate-A	Test Time:	13	Point:	28	40	0	0	Control Interval:	0	
Main Wave:	340	Second Wave:	546	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin		
Normal:	10	0	0	10	0	0
Decrement:	3	0	0	3	0	0
Increment:	20	0	0	20	0	0
Abs. Limit:	1.5			1.5	Positive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	200	*
R2:	40	*

Second half item of Two test analyze

Calibration Parameters

Test Item: CKMB

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	10	0	0
(2)	#	S2	10	0	0
(3)			10	0	0
(4)			10	0	0
(5)			10	0	0
(6)			10	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

DIRUI
CS-T240 / CS-T300

Magnesium Xylidyl Blue

Analyze Parameters

Test Item:	MG	Test Full Name:	MG	Decimal Digit:	1	Unit:	mg/dl		Sample Blank		
Assay:	1-point end	Test Time:	13	Point:	20	0	0	0	Control Interval:	0	
Main Wave:	505	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin			
Normal:	3	0	0	3	0	0	
Decrement:	3	0	0	3	0	0	
Increment:	6	0	0	6	0	0	
Abs. Limit:	3.3			3.3	Positive Reaction		
Prozone Limit:	-3.3			-3.3	Lower Limit		

Reagent

	Vol.	Pos.
R1:	300	*
R2:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: MG

Caliration Methods: Linear (2-point) | Point: 2 | Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	3	0	0
(2)	#	S2	3	0	0
(3)			3	0	0
(4)			3	0	0
(5)			3	0	0
(6)			3	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

DIRUI
CS-T240 / CS-T300

Phosphorous UV

Analyze Parameters

Test Item:	PHOS	Test Full Name:	PHOS	Decimal Digit:	1	Unit:	mg/dl		Sample Blank		
Assay:	1-point end	Test Time:	13	Point:	20	0	0	0	Control Interval:	0	
Main Wave:	340	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin			
Normal:	3	0	0	3	0	0	
Decrement:	3	0	0	3	0	0	
Increment:	6	0	0	6	0	0	
Abs. Limit:	3.3			3.3	Positive Reaction		
Prozone Limit:	-3.3			-3.3	Lower Limit		

Reagent

	Vol.	Pos.
R1:	300	*
R2:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: PHOS

Caliration Methods: Linear (2-point) | Point: 2 | Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	3	0	0
(2)	#	S2	3	0	0
(3)			3	0	0
(4)			3	0	0
(5)			3	0	0
(6)			3	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

DIRUI
CS-T240 / CS-T300

Triglycerides

Analyze Parameters

Test Item:	TG	Test Full Name:	TG	Decimal Digit:	0	Unit:	mg/dl		Sample Blank		
Assay:	1-point end	Test Time:	13	Point:	49	0	0	0	Control Interval:	0	
Main Wave:	505	Second Wave:	700	Instrument Factor (Y=aX+b)		a=	1	b=	0		

Sample Vol.

	Serum			Urin			
Normal:	3	0	0	3	0	0	
Decrement:	3	0	0	3	0	0	
Increment:	6	0	0	6	0	0	
Abs. Limit:	3.3			3.3	Positive Reaction		
Prozone Limit:	-3.3			-3.3	Lower Limit		

Reagent

	Vol.	Pos.
R1:	300	*
R2:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: TG

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	3	0	0
(2)	#	S2	3	0	0
(3)			3	0	0
(4)			3	0	0
(5)			3	0	0
(6)			3	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

DIRUI
CS-T240 / CS-T300

Total Protein

Analyze Parameters

Test Item:	TPRO	Test Full Name:	TPRO	Decimal Digit:	2	Unit:	g/dl		Sample Blank		
Assay:	1-point end	Test Time:	13	Point:	20	0	0	0	Control Interval:	0	
Main Wave:	546	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin			
Normal:	6	0	0	6	0	0	
Decrement:	3	0	0	3	0	0	
Increment:	12	0	0	12	0	0	
Abs. Limit:	3.3			3.3	Positive Reaction		
Prozone Limit:	-3.3			-3.3	Lower Limit		

Reagent

	Vol.	Pos.
R1:	240	*
R2:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: TPRO

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	6	0	0
(2)	#	S2	6	0	0
(3)			6	0	0
(4)			6	0	0
(5)			6	0	0
(6)			6	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

DIRUI
CS-T240 / CS-T300

Uric Acid

Analyze Parameters

Test Item:	UA	Test Full Name:	UA	Decimal Digit:	1	Unit:	mg/dl		Sample Blank		
Assay:	1-point End	Test Time:	13	Point:	35	0	0	0	Control Interval:	0	
Main Wave:	546	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin			
Normal:	6	0	0	6	0	0	
Decrement:	3	0	0	3	0	0	
Increment:	12	0	0	12	0	0	
Abs. Limit:	3.3			3.3	Positive Reaction		
Prozone Limit:	-3.3			-3.3	Lower Limit		

Reagent

	Vol.	Pos.
R1:	120	*
R2:	120	*

Second half item of Two test analyze

Calibration Parameters

Test Item: UA

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	6	0	0
(2)	#	S2	6	0	0
(3)			6	0	0
(4)			6	0	0
(5)			6	0	0
(6)			6	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

DIRUI
CS-T240 / CS-T300

UREA

Analyze Parameters

Test Item:	UREA	Test Full Name:	UREA	Decimal Digit:	0	Unit:	mg/dl		Sample Blank		
Assay:	2-point rate	Test Time:	13	Point:	23	29	0	0	Control Interval:	0	
Main Wave:	340	Second Wave:	405	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin		
Normal:	3	0	0	3	0	0
Decrement:	3	0	0	3	0	0
Increment:	6	0	0	6	0	0
Abs. Limit:	0.1			0.1	Negetive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	240	*
R2:	60	*

Second half item of Two test analyze

Calibration Parameters

Test Item: UREA

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	3	0	0
(2)	#	S2	3	0	0
(3)			3	0	0
(4)			3	0	0
(5)			3	0	0
(6)			3	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

DIRUI
CS-T240 / CS-T300

ZINC

Analyze Parameters

Test Item:	ZINC	Test Full Name:	ZINC	Decimal Digit:	0	Unit:	µg/dl		Sample Blank		
Assay:	1-point end	Test Time:	13	Point:	20	0	0	0	Control Interval:	0	
Main Wave:	570	Second Wave:	700	Instrument Factor (Y=aX+b)		a=	1	b=	0		

Sample Vol.

	Serum			Urin		
Normal:	15	0	0	15	0	0
Decrement:	3	0	0	3	0	0
Increment:	30	0	0	30	0	0
Abs. Limit:	3.3			3.3	Positive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	300	*
R2:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: ZINC

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	15	0	0
(2)	#	S2	15	0	0
(3)			15	0	0
(4)			15	0	0
(5)			15	0	0
(6)			15	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت دایان آنالیز تشخیص

تلفن : 09366021001