BIOMEDIC

mindray
healthcare within reach

BS 430

ALBUMIN

Chem	ALB	No.		Sample Type	Serum ↓
Chemistry	ALBUMIN			Print Name	ALB
Reaction Type	Endpoint ↓			Reaction Direction	Positive ↓
Pri Wave	605nm ↓			Sec Wave	700nm ↓
Unit	g/dl ↓			Decimal	0.1 ↓
Blank Time	4	5		Reaction Time	20	22

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	2 uL	uL	uL	R1	200 uL	uL
Decreased	1.5 uL	uL	uL	R2	uL	uL
Increased	4 uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

mindray
healthcare within reach

BS 430

ALKALINE PHOSPHATASE

Chem	ALP	No.		Sample Type	Serum
Chemistry	ALKALINE PHOSPHATASE			Print Name	ALP
Reaction Type	Kinetic			Reaction Direction	Positive
Pri Wave	412nm			Sec Wave	660nm
Unit	U/L			Decimal	0
Blank Time	0	0		Reaction Time	25	35

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	4 uL	uL	uL	R1	160 uL	uL
Decreased	2 uL	uL	uL	R2	40 uL	uL
Increased	8 uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear		
Factor		Replicates	2

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

mindray
healthcare within reach

BS 430

Alanine Aminotransferase

Chem	ALT	No.		Sample Type	Serum ↓
Chemistry	Alanine Aminotransferase			Print Name	ALT
Reaction Type	Kinetic ↓			Reaction Direction	negative ↓
Pri Wave	340nm ↓			Sec Wave	412nm ↓
Unit	U/L ↓			Decimal	0 ↓
Blank Time	0	0		Reaction Time	25	35

	Sample Vol	Aspirated	Diluent
Standard	15 uL	uL	uL
Decreased	5 uL	uL	uL
Increased	20 uL	uL	uL

	Reagent Vol	Diluent
R1	120 uL	uL
R2	30 uL	uL
R3	uL	uL
R4	uL	uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

mindray
healthcare within reach

BS 430

AMYLASE

Chem	AMY	No.		Sample Type	Serum
Chemistry	AMYLASE			Print Name	AMY
Reaction Type	Kinetic			Reaction Direction	Positive
Pri Wave	412nm			Sec Wave	700nm
Unit	U/L			Decimal	0
Blank Time	0	0		Reaction Time	8	18

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	4 uL	uL	uL	R1	200 uL	uL
Decreased	2 uL	uL	uL	R2	uL	uL
Increased	8 uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear		
Factor		Replicates	2

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

mindray
healthcare within reach

BS 430

Aspartate Aminotransferase

Chem	AST	No.		Sample Type	Serum
Chemistry	Aspartate Aminotransferase			Print Name	AST
Reaction Type	Kinetic			Reaction Direction	negative
Pri Wave	340nm			Sec Wave	412nm
Unit	U/L			Decimal	0
Blank Time	0	0		Reaction Time	25	35

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	15 uL	uL	uL	R1	120 uL	uL
Decreased	5 uL	uL	uL	R2	30 uL	uL
Increased	20 uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear		
Factor		Replicates	2

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

mindray
healthcare within reach

BS 430

Complement C3

Chem	C3	No.		Sample Type	Serum
Chemistry	Complement C3			Print Name	C3
Reaction Type	Endpoint			Reaction Direction	Positive
Pri Wave	340nm			Sec Wave	700nm
Unit	mg/dl			Decimal	0
Blank Time	20	22		Reaction Time	37	39

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	2 uL	uL	uL	R1	160 uL	uL
Decreased	1.5 uL	uL	uL	R2	30 uL	uL
Increased	10 uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Logit-Log 4P		
Factor		Replicates	1

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

mindray
healthcare within reach

BS 430

Complement C4

Chem	C4	No.		Sample Type	Serum ↓
Chemistry	Complement C4			Print Name	C4
Reaction Type	Endpoint ↓			Reaction Direction	Positive ↓
Pri Wave	340nm ↓			Sec Wave	700nm ↓
Unit	mg/dl ↓			Decimal	0.1 ↓
Blank Time	20	22		Reaction Time	37	39

	Sample Vol	Aspirated	Diluent
Standard	4 uL	uL	uL
Decreased	2 uL	uL	uL
Increased	10 uL	uL	uL

	Reagent Vol	Diluent
R1	160 uL	uL
R2	30 uL	uL
R3	uL	uL
R4	uL	uL

Calibration Settings

Math Model	Logit-Log 4P ↓		
Factor		Replicates	1 ↓

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

mindray
healthcare within reach

BS 430

Calcium Arsenazo III

Chem	CA	No.		Sample Type	Serum
Chemistry	Calcium Arsenazo III			Print Name	CA
Reaction Type	Endpoint			Reaction Direction	Positive
Pri Wave	660nm			Sec Wave	700nm
Unit	mg/dl			Decimal	0.1
Blank Time	4	5		Reaction Time	20	22

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	2 uL	uL	uL	R1	200 uL	uL
Decreased	1.5 uL	uL	uL	R2	uL	uL
Increased	4 uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear		
Factor		Replicates	2

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

mindray
healthcare within reach

BS 430

Cholesterol

Chem	CHOL	No.		Sample Type	Serum ↓
Chemistry	Cholesterol			Print Name	CHOL
Reaction Type	Endpoint ↓			Reaction Direction	Positive ↓
Pri Wave	505 ↓			Sec Wave	700nm ↓
Unit	mg/dl ↓			Decimal	0 ↓
Blank Time	4	5		Reaction Time	37	39

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	2 uL	uL	uL	R1	200 uL	uL
Decreased	1.5 uL	uL	uL	R2	uL	uL
Increased	10 uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

mindray
healthcare within reach

BS 430

Creatine Kinase (CKNAC)

Chem	CPK	No.		Sample Type	Serum ↓
Chemistry	Creatine Kinase (CKNAC)			Print Name	CPK
Reaction Type	Kinetic ↓			Reaction Direction	Positive ↓
Pri Wave	340nm ↓			Sec Wave	412nm ↓
Unit	U/L ↓			Decimal	0 ↓
Blank Time	0	0		Reaction Time	29	39

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	8 uL	uL	uL	R1	150 uL	uL
Decreased	2 uL	uL	uL	R2	30 uL	uL
Increased	16 uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

mindray
healthcare within reach

BS 430

CREATININE

Chem	CREAT	No.		Sample Type	Serum ↓
Chemistry	CREATININE			Print Name	CREAT
Reaction Type	Fixed-time ↓			Reaction Direction	Positive ↓
Pri Wave	505nm ↓			Sec Wave	660nm ↓
Unit	mg/dl ↓			Decimal	0.01 ↓
Blank Time	0	0		Reaction Time	24	29

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	20 uL	uL	uL	R1	100 uL	uL
Decreased	2 uL	uL	uL	R2	100 uL	uL
Increased	30 uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

mindray
healthcare within reach

BS 430

C-Reactive Protein (CRP)

Chem	CRP	No.		Sample Type	Serum ↓
Chemistry	C-Reactive Protein (CRP)			Print Name	CRP
Reaction Type	Endpoint ↓			Reaction Direction	Positive ↓
Pri Wave	340nm ↓			Sec Wave	700nm ↓
Unit	mg/L ↓			Decimal	0.1 ↓
Blank Time	20	22		Reaction Time	37	39

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	8 uL	uL	uL	R1	160 uL	uL
Decreased	4 uL	uL	uL	R2	30 uL	uL
Increased	16 uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Spline ↓		
Factor		Replicates	1 ↓

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

mindray
healthcare within reach

BS 430

GLUCOSE

Chem	GLU	No.		Sample Type	Serum ↓
Chemistry	GLUCOSE			Print Name	GLU
Reaction Type	Endpoint ↓			Reaction Direction	Positive ↓
Pri Wave	505nm ↓			Sec Wave	700nm ↓
Unit	mg/dl ↓			Decimal	0 ↓
Blank Time	4	5		Reaction Time	37	39

	Sample Vol	Aspirated	Diluent
Standard	2 uL	uL	uL
Decreased	1.5 uL	uL	uL
Increased	4 uL	uL	uL

	Reagent Vol	Diluent
R1	200 uL	uL
R2	uL	uL
R3	uL	uL
R4	uL	uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

mindray
healthcare within reach

BS 430

HDL Cholesterol

Chem	HDL	No.		Sample Type	Serum
Chemistry	HDL Cholesterol			Print Name	HDL
Reaction Type	Endpoint			Reaction Direction	Positive
Pri Wave	605nm			Sec Wave	700nm
Unit	mg/dl			Decimal	0
Blank Time	20	22		Reaction Time	37	39

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	2 uL	uL	uL	R1	150 uL	uL
Decreased	1.5 uL	uL	uL	R2	50 uL	uL
Increased	10 uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear		
Factor		Replicates	2

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

mindray
healthcare within reach

BS 430

Immunoglobulin A

Chem	IgA	No.		Sample Type	Serum
Chemistry	Immunoglobulin A			Print Name	IgA
Reaction Type	Endpoint			Reaction Direction	Positive
Pri Wave	340nm			Sec Wave	700nm
Unit	mg/dl			Decimal	0
Blank Time	20	22		Reaction Time	37	39

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	2 uL	uL	uL	R1	160 uL	uL
Decreased	1.5 uL	uL	uL	R2	30 uL	uL
Increased	10 uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Logit-Log 4P		
Factor		Replicates	1

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

mindray
healthcare within reach

BS 430

Immunoglobulin G

Chem	IgG	No.		Sample Type	Serum
Chemistry	Immunoglobulin G			Print Name	IgG
Reaction Type	Endpoint			Reaction Direction	Positive
Pri Wave	605nm			Sec Wave	0
Unit	mg/dl			Decimal	0
Blank Time	20	22		Reaction Time	37	39

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	2 uL	uL	uL	R1	160 uL	uL
Decreased	1.5 uL	uL	uL	R2	30 uL	uL
Increased	10 uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Logit-Log 4P		
Factor		Replicates	1

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

mindray
healthcare within reach

BS 430

Immunoglobulin M

Chem	IgM	No.		Sample Type	Serum
Chemistry	Immunoglobulin M			Print Name	IgM
Reaction Type	Endpoint			Reaction Direction	Positive
Pri Wave	340nm			Sec Wave	700nm
Unit	mg/dl			Decimal	0
Blank Time	20	22		Reaction Time	37	39

	Sample Vol	Aspirated	Diluent
Standard	2 uL	uL	uL
Decreased	1.5 uL	uL	uL
Increased	10 uL	uL	uL

	Reagent Vol	Diluent
R1	160 uL	uL
R2	30 uL	uL
R3	uL	uL
R4	uL	uL

Calibration Settings

Math Model	Logit-Log 4P		
Factor		Replicates	1

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

mindray
healthcare within reach

BS 430

Lactate Dehydrogenase (LDH)

Chem	LDH	No.		Sample Type	Serum ↓
Chemistry	Lactate Dehydrogenase (LDH)			Print Name	LDH
Reaction Type	Kinetic ↓			Reaction Direction	negative ↓
Pri Wave	340nm ↓			Sec Wave	412nm ↓
Unit	U/L ↓			Decimal	0 ↓
Blank Time	0	0		Reaction Time	25	35

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	3 uL	uL	uL	R1	150 uL	uL
Decreased	1.5 uL	uL	uL	R2	30 uL	uL
Increased	6 uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

mindray
healthcare within reach

BS 430

LDL Cholesterol

Chem	LDL	No.		Sample Type	Serum
Chemistry	LDL Cholesterol			Print Name	LDL
Reaction Type	Endpoint			Reaction Direction	Positive
Pri Wave	605nm			Sec Wave	700nm
Unit	mg/dl			Decimal	0
Blank Time	20	22		Reaction Time	37	39

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	2 uL	uL	uL	R1	150 uL	uL
Decreased	1.5 uL	uL	uL	R2	50 uL	uL
Increased	10 uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear		
Factor		Replicates	2

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

mindray
healthcare within reach

BS 430

Lipase

Chem	LIP	No.		Sample Type	Serum
Chemistry	Lipase			Print Name	LIP
Reaction Type	Kinetic			Reaction Direction	Positive
Pri Wave	570nm			Sec Wave	700nm
Unit	U/L			Decimal	0
Blank Time	0	0		Reaction Time	29	36

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	4 uL	uL	uL	R1	160 uL	uL
Decreased	2 uL	uL	uL	R2	40 uL	uL
Increased	8 uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear		
Factor		Replicates	2

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

mindray
healthcare within reach

BS 430

Micro Albumin

Chem	MALB	No.		Sample Type	Serum
Chemistry	Micro Albumin			Print Name	MALB
Reaction Type	Endpoint			Reaction Direction	Positive
Pri Wave	340nm			Sec Wave	700nm
Unit	mg/L			Decimal	0.1
Blank Time	20	22		Reaction Time	37	39

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	10 uL	uL	uL	R1	150 uL	uL
Decreased	2 uL	uL	uL	R2	30 uL	uL
Increased	20 uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Logit-Log 4P		
Factor		Replicates	1

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

mindray
healthcare within reach

BS 430

Creatine Kinase - MB

Chem	CKMB	No.		Sample Type	Serum
Chemistry	Creatine Kinase - MB			Print Name	CKMB
Reaction Type	Kinetic			Reaction Direction	Positive
Pri Wave	340nm			Sec Wave	412nm
Unit	U/L			Decimal	0
Blank Time	0	0		Reaction Time	29	39

	Sample Vol		Aspirated		Diluent			Reagent Vol		Diluent	
Standard	8	uL		uL		uL	R1	150	uL		uL
Decreased	2	uL		uL		uL	R2	30	uL		uL
Increased	16	uL		uL		uL	R3		uL		uL
							R4		uL		uL

Calibration Settings

Math Model	Two-point linear		
Factor		Replicates	2

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

mindray
healthcare within reach

BS 430

Magnesium Xylidyl Blue

Chem	MG	No.		Sample Type	Serum ↓
Chemistry	Magnesium Xylidyl Blue			Print Name	MG
Reaction Type	Endpoint ↓			Reaction Direction	Positive ↓
Pri Wave	505nm ↓			Sec Wave	700nm ↓
Unit	mg/dl ↓			Decimal	0.1 ↓
Blank Time	4	5		Reaction Time	20	22

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	2 uL	uL	uL	R1	200 uL	uL
Decreased	1.5 uL	uL	uL	R2	uL	uL
Increased	4 uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

mindray
healthcare within reach

BS 430

Phosphorus

Chem	PHOS	No.		Sample Type	Serum ↓
Chemistry	Phosphorus			Print Name	PHOS
Reaction Type	Endpoint ↓			Reaction Direction	Positive ↓
Pri Wave	340nm ↓			Sec Wave	700nm ↓
Unit	mg/dl ↓			Decimal	0.1 ↓
Blank Time	4	5		Reaction Time	20	22

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	2 uL	uL	uL	R1	200 uL	uL
Decreased	1.5 uL	uL	uL	R2	uL	uL
Increased	4 uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

mindray
healthcare within reach

BS 430

Triglycerides

Chem	TG	No.		Sample Type	Serum ↓
Chemistry	Triglycerides			Print Name	TG
Reaction Type	Endpoint ↓			Reaction Direction	Positive ↓
Pri Wave	505 ↓			Sec Wave	700nm ↓
Unit	mg/dl ↓			Decimal	0 ↓
Blank Time	4	5		Reaction Time	37	39

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	2 uL	uL	uL	R1	200 uL	uL
Decreased	1.5 uL	uL	uL	R2	uL	uL
Increased	10 uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

mindray
healthcare within reach

BS 430

TOTAL PROTEIN

Chem	TPRO	No.		Sample Type	Serum
Chemistry	TOTAL PROTEIN			Print Name	TPRO
Reaction Type	Endpoint			Reaction Direction	Positive
Pri Wave	546nm			Sec Wave	700nm
Unit	g/dl			Decimal	0.1
Blank Time	4	5		Reaction Time	20	22

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	5 uL	uL	uL	R1	200 uL	uL
Decreased	2 uL	uL	uL	R2	uL	uL
Increased	10 uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear		
Factor		Replicates	2

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

mindray
healthcare within reach

BS 430

URIC ACID

Chem	UA	No.		Sample Type	Serum
Chemistry	URIC ACID			Print Name	UA
Reaction Type	Endpoint			Reaction Direction	Positive
Pri Wave	546nm			Sec Wave	700nm
Unit	mg/dl			Decimal	0.01
Blank Time	0	0		Reaction Time	37	39

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	5 uL	uL	uL	R1	100 uL	uL
Decreased	1.5 uL	uL	uL	R2	100 uL	uL
Increased	10 uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear		
Factor		Replicates	2

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

mindray
healthcare within reach

BS 430

UREA

Chem	UREA	No.		Sample Type	Serum ↓
Chemistry	UREA			Print Name	UREA
Reaction Type	Fixed-time ↓			Reaction Direction	negative ↓
Pri Wave	340nm ↓			Sec Wave	412nm ↓
Unit	mg/dl ↓			Decimal	0.1 ↓
Blank Time	0	0		Reaction Time	24	29

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	2 uL	uL	uL	R1	160 uL	uL
Decreased	1.5 uL	uL	uL	R2	40 uL	uL
Increased	4 uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

شرکت دایان آنالیز تشخیص

تلفن : 09366021001

BIOMEDIC

mindray
healthcare within reach

BS 430

ZINC

Chem	ZINC	No.		Sample Type	Serum
Chemistry	ZINC			Print Name	ZINC
Reaction Type	Endpoint			Reaction Direction	Positive
Pri Wave	570nm			Sec Wave	700nm
Unit	µg/dl			Decimal	0
Blank Time	4	5		Reaction Time	19	20

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	8 uL	uL	uL	R1	160 uL	uL
Decreased	4 uL	uL	uL	R2	uL	uL
Increased	16 uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear		
Factor		Replicates	2

شرکت دایان آنالیز تشخیص

تلفن : 09366021001