BIOMEDIC

mindray
healthcare within reach

BS 480

Albumin BCG

Chem	ALB	No.		Sample Type	Serum
Chemistry	Albumin BCG			Print Name	ALB
Reaction Type	Endpoint			Reaction Direction	Positive
Pri Wave	605nm			Sec Wave	700nm
Unit	g/dl			Decimal	0.1
Blank Time	10	11		Reaction Time	24	26

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	2 uL	uL	uL	R1	200 uL	uL
Decreased	uL	uL	uL	R2	uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear		
Factor		Replicates	2

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 480

Alkaline Phosphatase

Chem	ALP	No.		Sample Type	Serum ↓
Chemistry	Alkaline Phosphatase			Print Name	ALP
Reaction Type	Kinetic ↓			Reaction Direction	Positive ↓
Pri Wave	412nm ↓			Sec Wave	412nm ↓
Unit	U/L ↓			Decimal	0 ↓
Blank Time	0	0		Reaction Time	58	78

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	3 uL	uL	uL	R1	120 uL	uL
Decreased	uL	uL	uL	R2	30 uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 480

Alanine Aminotransferase (ALT)

Chem	ALT	No.		Sample Type	Serum ↓
Chemistry	Alanine Aminotransferase (ALT)			Print Name	ALT
Reaction Type	Kinetic ↓			Reaction Direction	negative ↓
Pri Wave	340nm ↓			Sec Wave	412nm ↓
Unit	U/L ↓			Decimal	0 ↓
Blank Time	0	0		Reaction Time	58	78

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	15 uL	uL	uL	R1	120 uL	uL
Decreased	uL	uL	uL	R2	30 uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 480

Amylase CNPG3

Chem	AMY	No.		Sample Type	Serum
Chemistry	Amylase CNPG3			Print Name	AMY
Reaction Type	Kinetic			Reaction Direction	Positive
Pri Wave	412nm			Sec Wave	700nm
Unit	U/L			Decimal	0
Blank Time	0	0		Reaction Time	18	38

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	3 uL	uL	uL	R1	150 uL	uL
Decreased	uL	uL	uL	R2	uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear		
Factor		Replicates	2

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 480

Aspartate Aminotransferase (AST)

Chem	AST	No.		Sample Type	Serum ↓
Chemistry	Aspartate Aminotransferase (AST)			Print Name	AST
Reaction Type	Kinetic ↓			Reaction Direction	negative ↓
Pri Wave	340nm ↓			Sec Wave	412nm ↓
Unit	U/L ↓			Decimal	0 ↓
Blank Time	0	0		Reaction Time	58	78

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	15 uL	uL	uL	R1	120 uL	uL
Decreased	uL	uL	uL	R2	30 uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 480

Bilirubin Direct DPD

Chem	BILI D	No.		Sample Type	Serum
Chemistry	Bilirubin Direct DPD			Print Name	BILI D
Reaction Type	Endpoint			Reaction Direction	Positive
Pri Wave	546nm			Sec Wave	700nm
Unit	mg/dl			Decimal	0.01
Blank Time	47	49		Reaction Time	80	82

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	10 uL	uL	uL	R1	160 uL	uL
Decreased	uL	uL	uL	R2	40 uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear		
Factor		Replicates	2

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 480

Bilirubin Total DPD

Chem	BILI T	No.		Sample Type	Serum
Chemistry	Bilirubin Total DPD			Print Name	BILI T
Reaction Type	Endpoint			Reaction Direction	Positive
Pri Wave	546nm			Sec Wave	700nm
Unit	mg/dl			Decimal	0.01
Blank Time	47	49		Reaction Time	80	82

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	8 uL	uL	uL	R1	160 uL	uL
Decreased	uL	uL	uL	R2	40 uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear		
Factor		Replicates	2

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 480

Complement 3 (C3)

Chem	C3	No.		Sample Type	Serum ↓
Chemistry	Complement 3 (C3)			Print Name	C3
Reaction Type	Endpoint ↓			Reaction Direction	Positive ↓
Pri Wave	340nm ↓			Sec Wave	700nm ↓
Unit	mg/dl ↓			Decimal	0 ↓
Blank Time	47	49		Reaction Time	80	82

	Sample Vol		Aspirated		Diluent			Reagent Vol		Diluent	
Standard	2	uL		uL		uL	R1	160	uL		uL
Decreased		uL		uL		uL	R2	30	uL		uL
Increased		uL		uL		uL	R3		uL		uL
							R4		uL		uL

Calibration Settings

Math Model	Logit-Log 4P ↓		
Factor		Replicates	1 ↓

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 480

Complement 4 (C4)

Chem	C4	No.		Sample Type	Serum ↓
Chemistry	Complement 4 (C4)			Print Name	C4
Reaction Type	Endpoint ↓			Reaction Direction	Positive ↓
Pri Wave	340nm ↓			Sec Wave	700nm ↓
Unit	mg/dl ↓			Decimal	0.1 ↓
Blank Time	47	49		Reaction Time	80	82

	Sample Vol		Aspirated		Diluent			Reagent Vol		Diluent	
Standard	4	uL		uL		uL	R1	160	uL		uL
Decreased		uL		uL		uL	R2	30	uL		uL
Increased		uL		uL		uL	R3		uL		uL
							R4		uL		uL

Calibration Settings

Math Model	Logit-Log 4P ↓		
Factor		Replicates	1 ↓

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 480

Calcium Arsenazo III

Chem	CA	No.		Sample Type	Serum ↓
Chemistry	Calcium Arsenazo III			Print Name	CA
Reaction Type	Endpoint ↓			Reaction Direction	Positive ↓
Pri Wave	660nm ↓			Sec Wave	800nm ↓
Unit	mg/dl ↓			Decimal	0.1 ↓
Blank Time	10	11		Reaction Time	24	26

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	3 uL	uL	uL	R1	200 uL	uL
Decreased	uL	uL	uL	R2	uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 480

Cholesterol

Chem	CHOL	No.		Sample Type	Serum
Chemistry	Cholesterol			Print Name	CHOL
Reaction Type	Endpoint			Reaction Direction	Positive
Pri Wave	505nm			Sec Wave	700nm
Unit	mg/dl			Decimal	0
Blank Time	10	11		Reaction Time	80	82

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	2 uL	uL	uL	R1	200 uL	uL
Decreased	uL	uL	uL	R2	uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear		
Factor		Replicates	2

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 480

Creatine Kinase (CK-NAC)

Chem	CPK	No.		Sample Type	Serum ↓
Chemistry	Creatine Kinase (CK-NAC)			Print Name	CPK
Reaction Type	Kinetic ↓			Reaction Direction	negative ↓
Pri Wave	340nm ↓			Sec Wave	412nm ↓
Unit	U/L ↓			Decimal	0 ↓
Blank Time	0	0		Reaction Time	64	82

	Sample Vol		Aspirated		Diluent			Reagent Vol		Diluent	
Standard	9	uL		uL		uL	R1	150	uL		uL
Decreased		uL		uL		uL	R2	30	uL		uL
Increased		uL		uL		uL	R3		uL		uL
							R4		uL		uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 480

Creatinine Jaffe

Chem	CREAT	No.		Sample Type	Serum ↓
Chemistry	Creatinine Jaffe			Print Name	CREAT
Reaction Type	Kinetic ↓			Reaction Direction	Positive ↓
Pri Wave	505nm ↓			Sec Wave	700nm ↓
Unit	mg/dl ↓			Decimal	0.01 ↓
Blank Time	0	0		Reaction Time	55	65

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	24 uL	uL	uL	R1	120 uL	uL
Decreased	uL	uL	uL	R2	120 uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

www.tksmed.ir شرکت توسعه کیمیای سعادت

پشتیبانی / شکرالهی 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 480

C - reactive protein (CRP)

Chem	CRP	No.		Sample Type	Serum
Chemistry	C - reactive protein (CRP)			Print Name	CRP
Reaction Type	Endpoint			Reaction Direction	Positive
Pri Wave	340nm			Sec Wave	700nm
Unit	mg/L			Decimal	0.1
Blank Time	47	49		Reaction Time	80	82

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	8 uL	uL	uL	R1	160 uL	uL
Decreased	uL	uL	uL	R2	30 uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Logit-Log 4P		
Factor		Replicates	1

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 480

Glucose

Chem	GLU	No.		Sample Type	Serum ↓
Chemistry	Glucose			Print Name	GLU
Reaction Type	Endpoint ↓			Reaction Direction	Positive ↓
Pri Wave	505nm ↓			Sec Wave	700nm ↓
Unit	mg/dl ↓			Decimal	0 ↓
Blank Time	10	11		Reaction Time	80	82

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	2 uL	uL	uL	R1	200 uL	uL
Decreased	uL	uL	uL	R2	uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 480

HDL Cholesterol Direct

Chem	HDL	No.		Sample Type	Serum ↓
Chemistry	HDL Cholesterol Direct			Print Name	HDL
Reaction Type	Endpoint ↓			Reaction Direction	Positive ↓
Pri Wave	605nm ↓			Sec Wave	700nm ↓
Unit	mg/dl ↓			Decimal	0 ↓
Blank Time	47	49		Reaction Time	80	82

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	2 uL	uL	uL	R1	150 uL	uL
Decreased	uL	uL	uL	R2	50 uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 480

Immunoglobulin A (IgA)

Chem	IgA	No.		Sample Type	Serum
Chemistry	Immunoglobulin A (IgA)			Print Name	IgA
Reaction Type	Endpoint			Reaction Direction	Positive
Pri Wave	340nm			Sec Wave	700nm
Unit	mg/dl			Decimal	0
Blank Time	47	49		Reaction Time	80	82

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	2 uL	uL	uL	R1	160 uL	uL
Decreased	uL	uL	uL	R2	30 uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Logit-Log 4P		
Factor		Replicates	1

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 480

Immunoglobulin G (IgG)

Chem	IgG	No.		Sample Type	Serum
Chemistry	Immunoglobulin G (IgG)			Print Name	IgG
Reaction Type	Endpoint			Reaction Direction	Positive
Pri Wave	605nm			Sec Wave	800nm
Unit	mg/dl			Decimal	0
Blank Time	47	49		Reaction Time	80	82

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	2 uL	uL	uL	R1	160 uL	uL
Decreased	uL	uL	uL	R2	30 uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Logit-Log 4P		
Factor		Replicates	1

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 480

Immunoglobulin M (IgM)

Chem	IgM	No.		Sample Type	Serum
Chemistry	Immunoglobulin M (IgM)			Print Name	IgM
Reaction Type	Endpoint			Reaction Direction	Positive
Pri Wave	340nm			Sec Wave	700nm
Unit	mg/dl			Decimal	0
Blank Time	47	49		Reaction Time	80	82

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	2 uL	uL	uL	R1	160 uL	uL
Decreased	uL	uL	uL	R2	30 uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Logit-Log 4P		
Factor		Replicates	1

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 480

Lactate Dehydrogenase (LDH)

Chem	LDH	No.		Sample Type	Serum ↓
Chemistry	Lactate Dehydrogenase (LDH)			Print Name	LDH
Reaction Type	Kinetic ↓			Reaction Direction	negative ↓
Pri Wave	340nm ↓			Sec Wave	700nm ↓
Unit	U/L ↓			Decimal	0 ↓
Blank Time	0	0		Reaction Time	58	78

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	3 uL	uL	uL	R1	150 uL	uL
Decreased	uL	uL	uL	R2	30 uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

www.tksmed.ir شرکت توسعه کیمیای سعادت
09129579408 پشتیبانی / شکرالهی

BIOMEDIC

mindray
healthcare within reach

BS 480

LDL Cholesterol Direct

Chem	LDL	No.		Sample Type	Serum ↓
Chemistry	LDL Cholesterol Direct			Print Name	LDL
Reaction Type	Endpoint ↓			Reaction Direction	Positive ↓
Pri Wave	546nm ↓			Sec Wave	700nm ↓
Unit	mg/dl ↓			Decimal	0 ↓
Blank Time	47	49		Reaction Time	80	82

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	3 uL	uL	uL	R1	150 uL	uL
Decreased	uL	uL	uL	R2	50 uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 480

Lipase

Chem	LIP	No.		Sample Type	Serum ↓
Chemistry	Lipase			Print Name	LIP
Reaction Type	Kinetic ↓			Reaction Direction	Positive ↓
Pri Wave	570nm ↓			Sec Wave	700nm ↓
Unit	U/L ↓			Decimal	0 ↓
Blank Time	0	0		Reaction Time	64	76

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	3 uL	uL	uL	R1	120 uL	uL
Decreased	uL	uL	uL	R2	30 uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 480

Microalbumin

Chem	MALB	No.		Sample Type	Serum ↓
Chemistry	Microalbumin			Print Name	MALB
Reaction Type	Endpoint ↓			Reaction Direction	Positive ↓
Pri Wave	340nm ↓			Sec Wave	700nm ↓
Unit	mg/L ↓			Decimal	0.1 ↓
Blank Time	47	49		Reaction Time	80	82

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	10 uL	uL	uL	R1	160 uL	uL
Decreased	uL	uL	uL	R2	30 uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Logit-Log 4P ↓		
Factor		Replicates	1 ↓

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 480

CK - MB

Chem	MB	No.		Sample Type	Serum ↓
Chemistry	CK - MB			Print Name	MB
Reaction Type	Kinetic ↓			Reaction Direction	Positive ↓
Pri Wave	340nm ↓			Sec Wave	412nm ↓
Unit	U/L ↓			Decimal	0.1 ↓
Blank Time	0	0		Reaction Time	64	82

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	9 uL	uL	uL	R1	150 uL	uL
Decreased	uL	uL	uL	R2	30 uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 480

Magnesium Xylidyl Blue

Chem	MG	No.		Sample Type	Serum ↓
Chemistry	Magnesium Xylidyl Blue			Print Name	MG
Reaction Type	Endpoint ↓			Reaction Direction	Positive ↓
Pri Wave	546nm ↓			Sec Wave	700nm ↓
Unit	mg/dl ↓			Decimal	0.1 ↓
Blank Time	10	11		Reaction Time	24	26

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	2 uL	uL	uL	R1	200 uL	uL
Decreased	uL	uL	uL	R2	uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 480

Phosphorous UV

Chem	PHOS	No.		Sample Type	Serum
Chemistry	Phosphorous UV			Print Name	PHOS
Reaction Type	Endpoint			Reaction Direction	Positive
Pri Wave	340nm			Sec Wave	700nm
Unit	mg/dl			Decimal	0.1
Blank Time	10	11		Reaction Time	24	26

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	2 uL	uL	uL	R1	200 uL	uL
Decreased	uL	uL	uL	R2	uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear		
Factor		Replicates	2

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 480

Triglycerides

Chem	TG	No.		Sample Type	Serum ↓
Chemistry	Triglycerides			Print Name	TG
Reaction Type	Endpoint ↓			Reaction Direction	Positive ↓
Pri Wave	505nm ↓			Sec Wave	700nm ↓
Unit	mg/dl ↓			Decimal	0 ↓
Blank Time	10	11		Reaction Time	80	82

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	2 uL	uL	uL	R1	200 uL	uL
Decreased	uL	uL	uL	R2	uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 480

Total Protein

Chem	TPRO	No.		Sample Type	Serum
Chemistry	Total Protein			Print Name	TPRO
Reaction Type	Endpoint			Reaction Direction	Positive
Pri Wave	546nm			Sec Wave	700nm
Unit	g/dl			Decimal	0.1
Blank Time	10	11		Reaction Time	24	26

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	5 uL	uL	uL	R1	200 uL	uL
Decreased	uL	uL	uL	R2	uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear		
Factor		Replicates	2

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 480

Uric Acid

Chem	UA	No.		Sample Type	Serum ↓
Chemistry	Uric Acid			Print Name	UA
Reaction Type	Endpoint ↓			Reaction Direction	Positive ↓
Pri Wave	546nm ↓			Sec Wave	700nm ↓
Unit	mg/dl ↓			Decimal	0.01 ↓
Blank Time	0	0		Reaction Time	70	72

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	6 uL	uL	uL	R1	120 uL	uL
Decreased	uL	uL	uL	R2	120 uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 480

UREA

Chem	UREA	No.		Sample Type	Serum ↓
Chemistry	UREA			Print Name	UREA
Reaction Type	Kinetic ↓			Reaction Direction	negative ↓
Pri Wave	340nm ↓			Sec Wave	700nm ↓
Unit	mg/dl ↓			Decimal	0.1 ↓
Blank Time	0	0		Reaction Time	55	65

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	2 uL	uL	uL	R1	160 uL	uL
Decreased	uL	uL	uL	R2	40 uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 480

ZINC

Chem	ZINC	No.		Sample Type	Serum ↓
Chemistry	ZINC			Print Name	ZINC
Reaction Type	Endpoint ↓			Reaction Direction	Positive ↓
Pri Wave	570nm ↓			Sec Wave	700nm ↓
Unit	µg/dl ↓			Decimal	0 ↓
Blank Time	10	11		Reaction Time	24	26

	Sample Vol	Aspirated	Diluent		Reagent Vol	Diluent
Standard	8 uL	uL	uL	R1	160 uL	uL
Decreased	uL	uL	uL	R2	uL	uL
Increased	uL	uL	uL	R3	uL	uL
				R4	uL	uL

Calibration Settings

Math Model	Two-point linear ↓		
Factor		Replicates	2 ↓

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی 09129579408