BIOMEDIC

Albumin (BCG)

	BT 1000 / 2000 / 3000	
Primary Parameters	Method	End Point
	Kind Of Process	Linear
	Filter	630/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	1
	Reagent 1	300
	Reagent 2	-
	SAMPLE	
	Sample Vol	3
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	180
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	6
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	3.5 - 5
	female	3.5 - 5
	Child	3.5 - 5
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	g/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

	BT 1500 / 3500	
Primary Parameters	Method	End Point
	Kind Of Process	Linear
	Filter	630/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	1
	Reagent 1	200
	Reagent 2	-
	SAMPLE	
	Sample Vol	2
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	180
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	6
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	3.5 - 5
	female	3.5 - 5
	Child	3.5 - 5
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	g/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Alkaline phosphatase

	BT 1000 / 2000 / 3000	
Primary Parameters	Method	Kinetic
	Kind Of Process	Linear
	Filter	405/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	240
	Reagent 2	60
	SAMPLE	
	Sample Vol	6
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	180
	Incubation Time (sec)	120 - 60
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	1200
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	98 - 279
	female	98 - 279
	Child	98 - 279
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	U/L
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

	BT 1500 / 3500	
Primary Parameters	Method	Kinetic
	Kind Of Process	Linear
	Filter	405/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	160
	Reagent 2	40
	SAMPLE	
	Sample Vol	4
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	180
	Incubation Time (sec)	120 - 60
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	1200
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	98 - 279
	female	98 - 279
	Child	98 - 279
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	U/L
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

BIOMEDIC

ALT/GPT

	BT 1000 / 2000 / 3000	
Primary Parameters	Method	Kinetic
	Kind Of Process	Linear
	Filter	340/405
	Reaction Direction	Decreasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	240
	Reagent 2	60
	SAMPLE	
	Sample Vol	30
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	180
	Incubation Time (sec)	60 - 60
Check parameters	Reagent Limit (m ABS)	100
	Curve Acceptance	100
	Test limit (conc)	300
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	0 - 40
	female	0 - 40
	Child	0 - 40
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	U/L
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

	BT 1500 / 3500	
Primary Parameters	Method	Kinetic
	Kind Of Process	Linear
	Filter	340/405
	Reaction Direction	Decreasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	160
	Reagent 2	40
	SAMPLE	
	Sample Vol	20
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	180
	Incubation Time (sec)	60 - 60
Check parameters	Reagent Limit (m ABS)	100
	Curve Acceptance	100
	Test limit (conc)	300
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	0 - 40
	female	0 - 40
	Child	0 - 40
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	U/L
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Amylase - total

	BT 1000 / 2000 / 3000	
Primary Parameters	Method	Kinetic
	Kind Of Process	Linear
	Filter	405/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	1
	Reagent 1	300
	Reagent 2	-
	SAMPLE	
	Sample Vol	6
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	180
	Incubation Time (sec)	60
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	2000
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	0 - 90
	female	0 - 90
	Child	0 - 90
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	U/L
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

	BT 1500 / 3500	
Primary Parameters	Method	Kinetic
	Kind Of Process	Linear
	Filter	405/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	1
	Reagent 1	200
	Reagent 2	-
	SAMPLE	
	Sample Vol	4
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	180
	Incubation Time (sec)	60
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	2000
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	0 - 90
	female	0 - 90
	Child	0 - 90
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	U/L
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

AST/GOT

	BT 1000 / 2000 / 3000	
Primary Parameters	Method	Kinetic
	Kind Of Process	Linear
	Filter	340/405
	Reaction Direction	Decreasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	240
	Reagent 2	60
	SAMPLE	
	Sample Vol	30
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	180
	Incubation Time (sec)	60 - 60
Check parameters	Reagent Limit (m ABS)	100
	Curve Acceptance	100
	Test limit (conc)	300
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	0 - 38
	female	0 - 38
	Child	0 - 38
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	U/L
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

	BT 1500 / 3500	
Primary Parameters	Method	Kinetic
	Kind Of Process	Linear
	Filter	340/405
	Reaction Direction	Decreasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	160
	Reagent 2	40
	SAMPLE	
	Sample Vol	20
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	180
	Incubation Time (sec)	60 - 60
Check parameters	Reagent Limit (m ABS)	100
	Curve Acceptance	100
	Test limit (conc)	300
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	0 - 38
	female	0 - 38
	Child	0 - 38
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	U/L
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Complement 3

BT 1000 / 2000 / 3000		
Primary Parameters	Method	Sample Blank type (A)
	Kind Of Process	Multi-point
	Filter	340/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	280
	Reagent 2	55
	SAMPLE	
	Sample Vol	3
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	120 - 180
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	500
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	90 - 180
	female	90 - 180
	Child	90 - 180
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

BT 1500 / 3500		
Primary Parameters	Method	Sample Blank type (A)
	Kind Of Process	Multi-point
	Filter	340/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	200
	Reagent 2	40
	SAMPLE	
	Sample Vol	2
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	120 - 180
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	500
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	90 - 180
	female	90 - 180
	Child	90 - 180
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

BIOMEDIC

Complement 4

BT 1000 / 2000 / 3000		
Primary Parameters	Method	Sample Blank type (A)
	Kind Of Process	Multi-point
	Filter	340/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	280
	Reagent 2	55
	SAMPLE	
	Sample Vol	6
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	120 - 180
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	80
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	10 - 40
	female	10 - 40
	Child	10 - 40
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

BT 1500 / 3500		
Primary Parameters	Method	Sample Blank type (A)
	Kind Of Process	Multi-point
	Filter	340/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	200
	Reagent 2	40
	SAMPLE	
	Sample Vol	4
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	120 - 180
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	80
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	10 - 40
	female	10 - 40
	Child	10 - 40
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Calcium

	BT 1000 / 2000 / 3000	
Primary Parameters	Method	End Point
	Kind Of Process	Linear
	Filter	620/<No>
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	1
	Reagent 1	300
	Reagent 2	-
	SAMPLE	
	Sample Vol	5
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	180
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	20
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	8.4 - 10.4
	female	8.4 - 10.4
	Child	8.4 - 10.4
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

	BT 1500 / 3500	
Primary Parameters	Method	End Point
	Kind Of Process	Linear
	Filter	620/<No>
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	1
	Reagent 1	200
	Reagent 2	-
	SAMPLE	
	Sample Vol	3
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	180
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	20
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	8.4 - 10.4
	female	8.4 - 10.4
	Child	8.4 - 10.4
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

BIOMEDIC

Cholesterol

	BT 1000 / 2000 / 3000	
Primary Parameters	Method	End Point
	Kind Of Process	Linear
	Filter	510/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	1
	Reagent 1	300
	Reagent 2	-
	SAMPLE	
	Sample Vol	3
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	600
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	600
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	0 - 200
	female	0 - 200
	Child	0 - 200
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

	BT 1500 / 3500	
Primary Parameters	Method	End Point
	Kind Of Process	Linear
	Filter	510/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	1
	Reagent 1	200
	Reagent 2	-
	SAMPLE	
	Sample Vol	2
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	600
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	600
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	0 - 200
	female	0 - 200
	Child	0 - 200
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

BIOMEDIC

CK-NAC

	BT 1000 / 2000 / 3000	
Primary Parameters	Method	Kinetic
	Kind Of Process	Linear
	Filter	340/546
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	250
	Reagent 2	50
	SAMPLE	
	Sample Vol	15
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	180
	Incubation Time (sec)	120 - 120
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	800
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	20 - 195
	female	20 - 195
	Child	20 - 195
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	U/L
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

	BT 1500 / 3500	
Primary Parameters	Method	Kinetic
	Kind Of Process	Linear
	Filter	340/546
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	200
	Reagent 2	40
	SAMPLE	
	Sample Vol	12
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	180
	Incubation Time (sec)	120 - 120
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	800
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	20 - 195
	female	20 - 195
	Child	20 - 195
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	U/L
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Creatinine (Jaffe)

	BT 1000 / 2000 / 3000	
Primary Parameters	Method	Fixed Time
	Kind Of Process	Linear
	Filter	510/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	150
	Reagent 2	150
	SAMPLE	
	Sample Vol	30
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	60
	Incubation Time (sec)	20 - 60
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	15
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	0.7 - 1.4
	female	0.7 - 1.4
	Child	0.7 - 1.4
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

	BT 1500 / 3500	
Primary Parameters	Method	Fixed Time
	Kind Of Process	Linear
	Filter	510/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	100
	Reagent 2	100
	SAMPLE	
	Sample Vol	20
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	60
	Incubation Time (sec)	20 - 60
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	15
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	0.7 - 1.4
	female	0.7 - 1.4
	Child	0.7 - 1.4
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

C-REACTIVE PROTEIN

	BT 1000 / 2000 / 3000	
Primary Parameters	Method	Sample Blank type (A)
	Kind Of Process	Cubic spline
	Filter	340/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	280
	Reagent 2	55
	SAMPLE	
	Sample Vol	16
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	120 - 300
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	180
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	0 - 10
	female	0 - 10
	Child	0 - 10
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/l
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

	BT 1500 / 3500	
Primary Parameters	Method	Sample Blank type (A)
	Kind Of Process	Cubic spline
	Filter	340/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	200
	Reagent 2	40
	SAMPLE	
	Sample Vol	12
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	120 - 300
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	180
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	0 - 10
	female	0 - 10
	Child	0 - 10
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/l
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Glucose

BT 1000 / 2000 / 3000		
Primary Parameters	Method	End Point
	Kind Of Process	Linear
	Filter	510/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	1
	Reagent 1	300
	Reagent 2	-
	SAMPLE	
	Sample Vol	3
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	600
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	400
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	70 - 115
	female	70 - 115
	Child	70 - 115
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

BT 1500 / 3500		
Primary Parameters	Method	End Point
	Kind Of Process	Linear
	Filter	510/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	1
	Reagent 1	200
	Reagent 2	-
	SAMPLE	
	Sample Vol	2
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	600
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	400
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	70 - 115
	female	70 - 115
	Child	70 - 115
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

BIOMEDIC

HDL Cholesterol direct

BT 1000 / 2000 / 3000		
Primary Parameters	Method	Sample Blank type (A)
	Kind Of Process	Linear
	Filter	578/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	300
	Reagent 2	100
	SAMPLE	
	Sample Vol	4
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	180 - 300
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	100
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	30 - 90
	female	30 - 90
	Child	30 - 90
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

BT 1500 / 3500		
Primary Parameters	Method	Sample Blank type (A)
	Kind Of Process	Linear
	Filter	578/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	225
	Reagent 2	75
	SAMPLE	
	Sample Vol	3
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	180 - 300
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	100
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	30 - 90
	female	30 - 90
	Child	30 - 90
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Immunoglobulin A

BT 1000 / 2000 / 3000		
Primary Parameters	Method	Sample Blank type (A)
	Kind Of Process	Multi-point
	Filter	340/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	280
	Reagent 2	55
	SAMPLE	
	Sample Vol	3
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	120 - 180
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	800
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	70 - 400
	female	70 - 400
	Child	70 - 400
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

BT 1500 / 3500		
Primary Parameters	Method	Sample Blank type (A)
	Kind Of Process	Multi-point
	Filter	340/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	200
	Reagent 2	40
	SAMPLE	
	Sample Vol	2
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	120 - 180
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	800
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	70 - 400
	female	70 - 400
	Child	70 - 400
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Immunoglobulin G

BT 1000 / 2000 / 3000		
Primary Parameters	Method	Sample Blank type (A)
	Kind Of Process	Multi-point
	Filter	630/<No>
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	280
	Reagent 2	55
	SAMPLE	
	Sample Vol	3
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	120 - 180
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	3000
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	700 -1600
	female	700 -1600
	Child	700 -1600
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

BT 1500 / 3500		
Primary Parameters	Method	Sample Blank type (A)
	Kind Of Process	Multi-point
	Filter	630/<No>
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	200
	Reagent 2	40
	SAMPLE	
	Sample Vol	2
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	120 - 180
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	3000
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	700 -1600
	female	700 -1600
	Child	700 -1600
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

BIOMEDIC

Immunoglobulin M

BT 1000 / 2000 / 3000		
Primary Parameters	Method	Sample Blank type (A)
	Kind Of Process	Multi-point
	Filter	340/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	280
	Reagent 2	55
	SAMPLE	
	Sample Vol	3
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	120 - 180
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	350
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	40 -230
	female	40 -230
	Child	40 -230
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

BT 1500 / 3500		
Primary Parameters	Method	Sample Blank type (A)
	Kind Of Process	Multi-point
	Filter	340/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	200
	Reagent 2	40
	SAMPLE	
	Sample Vol	2
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	120 - 180
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	350
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	40 -230
	female	40 -230
	Child	40 -230
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Lactate Dehydrogenase

	BT 1000 / 2000 / 3000	
Primary Parameters	Method	Kinetic
	Kind Of Process	Linear
	Filter	340/700
	Reaction Direction	Decreasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	250
	Reagent 2	50
	SAMPLE	
	Sample Vol	5
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	180
	Incubation Time (sec)	120 - 60
Check parameters	Reagent Limit (m ABS)	100
	Curve Acceptance	100
	Test limit (conc)	800
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	230 - 460
	female	230 - 460
	Child	230 - 460
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	U/L
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

	BT 1500 / 3500	
Primary Parameters	Method	Kinetic
	Kind Of Process	Linear
	Filter	340/700
	Reaction Direction	Decreasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	200
	Reagent 2	40
	SAMPLE	
	Sample Vol	4
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	180
	Incubation Time (sec)	120 - 60
Check parameters	Reagent Limit (m ABS)	100
	Curve Acceptance	100
	Test limit (conc)	800
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	230 - 460
	female	230 - 460
	Child	230 - 460
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	U/L
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

LDL Cholesterol direct

	BT 1000 / 2000 / 3000	
Primary Parameters	Method	Sample Blank type (A)
	Kind Of Process	Linear
	Filter	546/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	300
	Reagent 2	100
	SAMPLE	
	Sample Vol	5
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	180 - 300
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	500
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	0 - 130
	female	0 - 130
	Child	0 - 130
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

	BT 1500 / 3500	
Primary Parameters	Method	Sample Blank type (A)
	Kind Of Process	Linear
	Filter	546/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	225
	Reagent 2	75
	SAMPLE	
	Sample Vol	4
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	180 - 300
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	500
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	0 - 130
	female	0 - 130
	Child	0 - 130
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Lipase

	BT 1000 / 2000 / 3000	
Primary Parameters	Method	Kinetic
	Kind Of Process	Linear
	Filter	578/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	240
	Reagent 2	60
	SAMPLE	
	Sample Vol	6
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	120
	Incubation Time (sec)	60 - 120
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	150
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	0 - 60
	female	0 - 60
	Child	0 - 60
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	U/L
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

	BT 1500 / 3500	
Primary Parameters	Method	Kinetic
	Kind Of Process	Linear
	Filter	578/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	160
	Reagent 2	40
	SAMPLE	
	Sample Vol	4
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	120
	Incubation Time (sec)	60 - 120
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	150
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	0 - 60
	female	0 - 60
	Child	0 - 60
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	U/L
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

BIOMEDIC

Micro Albumin

BT 1000 / 2000 / 3000		
Primary Parameters	Method	Sample Blank type (A)
	Kind Of Process	Multi-point
	Filter	340/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	280
	Reagent 2	55
	SAMPLE	
	Sample Vol	18
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	120 - 300
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	250
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	0 - 30
	female	0 - 30
	Child	0 - 30
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/l
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

BT 1500 / 3500		
Primary Parameters	Method	Sample Blank type (A)
	Kind Of Process	Multi-point
	Filter	340/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	200
	Reagent 2	40
	SAMPLE	
	Sample Vol	15
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	120 - 300
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	250
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	0 - 30
	female	0 - 30
	Child	0 - 30
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/l
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

CK-MB

	BT 1000 / 2000 / 3000	
Primary Parameters	Method	Kinetic
	Kind Of Process	Linear
	Filter	340/546
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	250
	Reagent 2	50
	SAMPLE	
	Sample Vol	15
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	180
	Incubation Time (sec)	120 -120
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	800
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	0 - 24
	female	0 - 24
	Child	0 - 24
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	U/L
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

	BT 1500 / 3500	
Primary Parameters	Method	Kinetic
	Kind Of Process	Linear
	Filter	340/546
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	200
	Reagent 2	40
	SAMPLE	
	Sample Vol	12
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	180
	Incubation Time (sec)	120 -120
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	800
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	0 - 24
	female	0 - 24
	Child	0 - 24
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	U/L
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Magnesium

	BT 1000 / 2000 / 3000	
Primary Parameters	Method	End Point
	Kind Of Process	Linear
	Filter	510/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	1
	Reagent 1	300
	Reagent 2	-
	SAMPLE	
	Sample Vol	3
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	180
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	5
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	1.6 - 2.5
	female	1.6 - 2.5
	Child	1.6 - 2.5
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

	BT 1500 / 3500	
Primary Parameters	Method	End Point
	Kind Of Process	Linear
	Filter	510/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	1
	Reagent 1	200
	Reagent 2	-
	SAMPLE	
	Sample Vol	2
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	180
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	5
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	1.6 - 2.5
	female	1.6 - 2.5
	Child	1.6 - 2.5
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

BIOMEDIC

Phosphorus

BT 1000 / 2000 / 3000		
Primary Parameters	Method	End Point
	Kind Of Process	Linear
	Filter	340/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	1
	Reagent 1	300
	Reagent 2	-
	SAMPLE	
	Sample Vol	3
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	180
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	20
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	2.5 - 5
	female	2.5 - 5
	Child	2.5 - 5
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

BT 1500 / 3500		
Primary Parameters	Method	End Point
	Kind Of Process	Linear
	Filter	340/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	1
	Reagent 1	200
	Reagent 2	-
	SAMPLE	
	Sample Vol	2
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	180
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	20
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	2.5 - 5
	female	2.5 - 5
	Child	2.5 - 5
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Triglycerides

	BT 1000 / 2000 / 3000	
Primary Parameters	Method	End Point
	Kind Of Process	Linear
	Filter	510/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	1
	Reagent 1	300
	Reagent 2	-
	SAMPLE	
	Sample Vol	3
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	600
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	600
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	0 - 160
	female	0 - 160
	Child	0 - 160
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

	BT 1500 / 3500	
Primary Parameters	Method	End Point
	Kind Of Process	Linear
	Filter	510/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	1
	Reagent 1	200
	Reagent 2	-
	SAMPLE	
	Sample Vol	2
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	600
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	600
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	0 - 160
	female	0 - 160
	Child	0 - 160
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

BIOMEDIC

Total Protein

	BT 1000 / 2000 / 3000	
Primary Parameters	Method	End Point
	Kind Of Process	Linear
	Filter	546/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	1
	Reagent 1	300
	Reagent 2	-
	SAMPLE	
	Sample Vol	8
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	300
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	12
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	6.6 - 8.3
	female	6.6 - 8.3
	Child	6.6 - 8.3
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	g/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

	BT 1500 / 3500	
Primary Parameters	Method	End Point
	Kind Of Process	Linear
	Filter	546/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	1
	Reagent 1	200
	Reagent 2	-
	SAMPLE	
	Sample Vol	5
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	300
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	12
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	6.6 - 8.3
	female	6.6 - 8.3
	Child	6.6 - 8.3
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	g/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Uric acid

BT 1000 / 2000 / 3000		
Primary Parameters	Method	End Point
	Kind Of Process	Linear
	Filter	546/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	150
	Reagent 2	150
	SAMPLE	
	Sample Vol	8
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	60 - 180
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	20
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	3.6 - 7.7
	female	3.6 - 7.7
	Child	3.6 - 7.7
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

BT 1500 / 3500		
Primary Parameters	Method	End Point
	Kind Of Process	Linear
	Filter	546/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	100
	Reagent 2	100
	SAMPLE	
	Sample Vol	5
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	60 - 180
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	20
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	3.6 - 7.7
	female	3.6 - 7.7
	Child	3.6 - 7.7
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Urea/BUN

	BT 1000 / 2000 / 3000	
Primary Parameters	Method	Fixed Time
	Kind Of Process	Linear
	Filter	340/700
	Reaction Direction	Decreasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	240
	Reagent 2	60
	SAMPLE	
	Sample Vol	4
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	60
	Incubation Time (sec)	20 - 60
Check parameters	Reagent Limit (m ABS)	100
	Curve Acceptance	100
	Test limit (conc)	200
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	15 - 45
	female	15 - 45
	Child	15 - 45
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

	BT 1500 / 3500	
Primary Parameters	Method	Fixed Time
	Kind Of Process	Linear
	Filter	340/700
	Reaction Direction	Decreasing
	REAGENT	
	Num of Reagent	2
	Reagent 1	200
	Reagent 2	50
	SAMPLE	
	Sample Vol	3
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	60
	Incubation Time (sec)	20 - 60
Check parameters	Reagent Limit (m ABS)	100
	Curve Acceptance	100
	Test limit (conc)	200
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	15 - 45
	female	15 - 45
	Child	15 - 45
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	mg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

ZINC

BT 1000 / 2000 / 3000		
Primary Parameters	Method	End Point
	Kind Of Process	Linear
	Filter	578/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	1
	Reagent 1	300
	Reagent 2	-
	SAMPLE	
	Sample Vol	15
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	180
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	400
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	60 - 120
	female	60 - 120
	Child	60 - 120
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	µg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

BT 1500 / 3500		
Primary Parameters	Method	End Point
	Kind Of Process	Linear
	Filter	578/700
	Reaction Direction	Increasing
	REAGENT	
	Num of Reagent	1
	Reagent 1	200
	Reagent 2	-
	SAMPLE	
	Sample Vol	10
	Pre Dilution	1
	Dilution	5
	Time	
	Sample Starter	Inactive
	Delay Time (sec)	0
	Reading Time (sec)	10
	Incubation Time (sec)	180
Check parameters	Reagent Limit (m ABS)	2000
	Curve Acceptance	100
	Test limit (conc)	400
	Initial ABS (m ABS)	
	Final ABS (m ABS)	
	Max ABS delta (m ABS)	
	Chec prozone	Inactive
	NORMAL RANGE	
	Male	60 - 120
	female	60 - 120
	Child	60 - 120
	Re-run hyperactive	inactive
	Re-run phatological	inactive
Secondary	Unit serum	µg/dl
	Num of needle washes	1/1
	Num of cuvette wash	2
	Additional wash	inactive
	Only needle	
	Only One wash	
	Instrument factor	1
	Shift:	0
	Reagent blank	every day
	Decimals	Auto

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸