Albumin (BCG)

Item No * ↓ | Item Name ALB | FULLNAME ALB | Optical

Data Information

Units	g/dl
Decimals	2

Analysis

Type	END Method
Main Wave Length	600
Sub Wave Length	700
Method	Bromocresol Green

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind	Vol.	Units
Sample	2	ul
Reagent 1	200	ul

Calibration

Type: Linear 1 ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability | Factor MIN | Factor MAX | CV MAX

Data Process

Read	Start	End
Main	23	25
Sub	10	11

Abs. Limit	low	High
	-3	3

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

Alkaline phosphatase

Item No	Item Name	FULLNAME	
* ↓	ALP	ALP	Optical

Data Information

Units	U/L
Decimals	0

Analysis

Type	Rate Method
Main Wave Length	405
Sub Wave Length	700
Method	DGKC

Correlation

	Slope		Intercept
Y =	1	* X +	0

Aspiration

Kind	Kind	Vol.	Units
Vol.	Sample	4	ul
	Reagent 1	160	ul
	Reagent 2	40	ul

Calibration

Type: Linear 1 ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability			
Factor MIN		Factor MAX	
CV MAX			

Data Process

Read		Start	End
	Main	39	50
	Sub	0	0

	low	High
Abs. Limit	-3	3

\# غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

ALT/GPT

Item No * ↓ | Item Name ALT | FULLNAME ALT | Optical

Data Information

Units	U/L
Decimals	0

Analysis

Type	Rate Method
Main Wave Length	340
Sub Wave Length	405
Method	IFCC

Correlation

Slope		Intercept
Y =	1	* X + 0

Aspiration

Kind	Kind	Vol.	Units
Vol.	Sample	20	ul
	Reagent 1	160	ul
	Reagent 2	40	ul

Calibration

Type: Linear 1 ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability			
Factor MIN		Factor MAX	
CV MAX			

Data Process

Read		Start	End
	Main	39	50
	Sub	0	0

	low	High
Abs. Limit	-3	3

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

Amylase - total

Item No * ↓ | Item Name AMY | FULLNAME AMY | Optical

Data Information

Units	U/L
Decimals	0

Analysis

Type	Rate Method
Main Wave Length	405
Sub Wave Length	700
Method	CNPG3

Correlation

Y = 1 (Slope) * X + 0 (Intercept)

Aspiration

Kind	Vol.	Units
Sample	4	ul
Reagent 1	200	ul

Calibration

Type: Linear 1 ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability

Factor MIN | Factor MAX

CV MAX

Data Process

Read	Start	End
Main	18	31
Sub	0	0

Abs. Limit	low	High
	-3	3

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

AST/GOT

Item No * ↓ | Item Name AST | FULLNAME AST | Optical

Data Information

Units	U/L
Decimals	0

Analysis

Type	Rate Method
Main Wave Length	340
Sub Wave Length	405
Method	IFCC

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind	Vol.	Units
Sample	20	ul
Reagent 1	160	ul
Reagent 2	40	ul

Calibration

Type: Linear 1 ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability

Factor MIN | Factor MAX

CV MAX

Data Process

Read	Start	End
Main	39	50
Sub	0	0

Abs. Limit	low	High
	-3	3

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

Complement 3

Item No * ↓ | Item Name C3 | FULLNAME C3 | Optical

Data Information

Units	mg/dl
Decimals	0

Analysis

Type	END Method
Main Wave Length	340
Sub Wave Length	700
Method	Immunoturbidimetric

Correlation

Slope		Intercept
Y =	1	* X + 0

Aspiration

Kind	Kind	Vol.	Units
Vol.	Sample	2	ul
	Reagent 1	180	ul
	Reagent 2	35	ul

Calibration

Type: Logit.log 2 ↓

Blank	0	#1	#	#2	#
#3	#	#4	#	#5	#
#6					

Calib Stability | Factor MIN | Factor MAX | CV MAX

Data Process

Read

	Start	End
Main	52	54
Sub	29	31

Abs. Limit	low	High
	-3	3

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

Complement 4

Item No * ↓ | Item Name C4 | FULLNAME C4 | Optical

Data Information

Units	mg/dl
Decimals	0

Analysis

Type	END Method
Main Wave Length	340
Sub Wave Length	700
Method	Immunoturbidimetric

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind	Vol.	Units
Sample	4	ul
Reagent 1	180	ul
Reagent 2	35	ul

Calibration

Type: Logit.log 2 ↓

Blank	0	#1	#	#2	#
#3	#	#4	#	#5	#
#6					

Calib Stability			
Factor MIN		Factor MAX	
CV MAX			

Data Process

Read

	Start	End
Main	52	54
Sub	29	31

	low	High
Abs. Limit	-3	3

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

Calcium

Item No * ↓ | Item Name CA | FULLNAME CA | Optical

Data Information

Units	mg/dl
Decimals	2

Analysis

Type	END Method
Main Wave Length	660
Sub Wave Length	800
Method	Arsenazo III

Correlation

Slope		Intercept
Y =	1	* X + 0

Aspiration

Kind	Kind	Vol.	Units
Vol.	Sample	3	ul
	Reagent 1	200	ul

Calibration

Type: Linear 1 ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability

Factor MIN | Factor MAX

CV MAX

Data Process

Read		Start	End
	Main	23	25
	Sub	10	11

	low	High
Abs. Limit	-3	3

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

Cholesterol

Item No * ↓ | Item Name CHOL | FULLNAME CHOL | Optical

Data Information

Units	mg/dl
Decimals	0

Analysis

Type	END Method
Main Wave Length	505
Sub Wave Length	700
Method	Cholesterol Oxidase

Correlation

Y = 1 (Slope) * X + 0 (Intercept)

Aspiration

Kind	Vol.	Units
Sample	2	ul
Reagent 1	200	ul

Calibration

Type: Linear 1 ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability

Factor MIN | Factor MAX

CV MAX

Data Process

Read	Start	End
Main	52	54
Sub	10	11

Abs. Limit	low	High
	-3	3

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

biolis 24i
prestige 24i

CK-NAC

Item No * ↓ | Item Name CPK | FULLNAME CPK | Optical

Data Information

Units	U/L
Decimals	0

Analysis

Type	Rate Method
Main Wave Length	340
Sub Wave Length	546
Method	IFCC

Correlation

Slope		Intercept
Y =	1	* X + 0

Aspiration

Kind	Kind	Vol.	Units
Vol.	Sample	10	ul
	Reagent 1	200	ul
	Reagent 2	40	ul

Calibration

Type: Linear 1 ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability			
Factor MIN		Factor MAX	
CV MAX			

Data Process

Read

	Start	End
Main	41	52
Sub	0	0

	low	High
Abs. Limit	-3	3

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

biolis 24i
prestige 24i

Creatinine (Jaffe)

Item No * ↓ | Item Name CREAT | FULLNAME CREAT | Optical

Data Information

Units	mg/dl
Decimals	2

Analysis

Type	Rate Method
Main Wave Length	505
Sub Wave Length	700
Method	Jaffe

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind	Vol.	Units
Sample	20	ul
Reagent 1	100	ul
Reagent 2	100	ul

Calibration

Type: Linear 1 ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability | Factor MIN | Factor MAX | CV MAX

Data Process

Read

	Start	End
Main	37	44
Sub	0	0

Abs. Limit: low -3 | High 3

\# غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

biolis 24i
prestige 24i

C-REACTIVE PROTEIN

Item No * ↓ | Item Name CRP | FULLNAME CRP | Optical

Data Information

Units	mg/L
Decimals	1

Analysis

Type	END Method
Main Wave Length	340
Sub Wave Length	700
Method	Immunoturbidimetric

Correlation

Slope		Intercept
Y =	1	* X + 0

Aspiration

Kind	Kind	Vol.	Units
Vol.	Sample	12	ul
	Reagent 1	180	ul
	Reagent 2	35	ul

Calibration

Type: Logit.log 2 ↓

Blank	0	#1	#	#2	#
#3	#	#4	#	#5	#
#6					

Calib Stability

Factor MIN | Factor MAX

CV MAX

Data Process

Read

	Start	End
Main	52	54
Sub	29	31

Abs. Limit	low	High
	-3	3

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

Glucose

Item No * ↓ | Item Name GLU | FULLNAME GLU | Optical

Data Information

Units	mg/dl
Decimals	0

Analysis

Type	END Method
Main Wave Length	505
Sub Wave Length	700
Method	Glucose oxidase

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind	Vol.	Units
Sample	2	ul
Reagent 1	200	ul

Calibration

Type: Linear 1 ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability | Factor MIN | Factor MAX | CV MAX

Data Process

Read	Start	End
Main	52	54
Sub	10	11

Abs. Limit	low	High
	-3	3

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

biolis 24i
prestige 24i

HDL Cholesterol direct

Item No * ↓ | Item Name HDL-C | FULLNAME HDL-C | Optical

Data Information

Units	mg/dl
Decimals	0

Analysis

Type	END Method
Main Wave Length	600
Sub Wave Length	700
Method	DIRECT

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind	Vol.	Units
Sample	3	ul
Reagent 1	225	ul
Reagent 2	75	ul

Calibration

Type: Linear 1 ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability | Factor MIN | Factor MAX | CV MAX

Data Process

Read

	Start	End
Main	52	54
Sub	29	31

	low	High
Abs. Limit	-3	3

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

biolis 24i
prestige 24i

Immunoglobulin A

Item No	*	Item Name	IgA	FULLNAME	IgA	Optical

Data Information

Units	mg/dl
Decimals	0

Analysis

Type	END Method
Main Wave Length	340
Sub Wave Length	700
Method	Immunoturbidimetric

Correlation

	Slope		Intercept
Y =	1	* X +	0

Aspiration

Kind	Kind	Vol.	Units
Vol.	Sample	2	ul
	Reagent 1	180	ul
	Reagent 2	35	ul

Calibration

Type: Logit.log 2

Blank	0	#1	#	#2	#
#3	#	#4	#	#5	#
#6					

Calib Stability			
Factor MIN		Factor MAX	
CV MAX			

Data Process

Read		Start	End
	Main	52	54
	Sub	29	31

	low	High
Abs. Limit	-3	3

\# غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

biolis 24i
prestige 24i

Immunoglobulin G

Item No	Item Name	FULLNAME	
* ↓	IgG	IgG	Optical

Data Information

Units	mg/dl
Decimals	0

Analysis

Type	END Method
Main Wave Length	600
Sub Wave Length	700
Method	Immunoturbidimetric

Correlation

Y = 1 * X + 0

Slope	Intercept
1	0

Aspiration

Kind	Vol.	Units
Sample	2	ul
Reagent 1	180	ul
Reagent 2	35	ul

Calibration

Type: Logit.log 2 ↓

Blank	0	#1	#	#2	#
#3	#	#4	#	#5	#
#6					

Calib Stability			
Factor MIN		Factor MAX	
CV MAX			

Data Process

Read

	Start	End
Main	52	54
Sub	29	31

Abs. Limit	low	High
	-3	3

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

biolis 24i
prestige 24i

Immunoglobulin M

Item No * ↓ | Item Name IgM | FULLNAME IgM | Optical

Data Information

Units	mg/dl
Decimals	0

Analysis

Type	END Method
Main Wave Length	340
Sub Wave Length	700
Method	Immunoturbidimetric

Correlation

Slope | Intercept

Y = 1 * X + 0

Aspiration

Kind	Kind	Vol.	Units
Vol.	Sample	2	ul
	Reagent 1	180	ul
	Reagent 2	35	ul

Calibration

Type: Logit.log 2 ↓

Blank	0	#1	#	#2	#
#3	#	#4	#	#5	#
#6					

Calib Stability

Factor MIN | Factor MAX

CV MAX

Data Process

Read

	Start	End
Main	52	54
Sub	29	31

	low	High
Abs. Limit	-3	3

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

biolis 24i
prestige 24i

Lactate Dehydrogenase

Item No	Item Name	FULLNAME	
* ↓	LDH	LDH	Optical

Data Information

Units	U/L
Decimals	0

Analysis

Type	Rate Method
Main Wave Length	340
Sub Wave Length	700
Method	(P-L) SCE

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind	Vol.	Units
Sample	4	ul
Reagent 1	200	ul
Reagent 2	40	ul

Calibration

Type: Linear 1 ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability			
Factor MIN		Factor MAX	
CV MAX			

Data Process

Read	Start	End
Main	39	50
Sub	0	0

Abs. Limit	low	High
	-3	3

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

biolis 24i
prestige 24i

LDL Cholesterol direct

Item No * ↓ | Item Name LDL-C | FULLNAME LDL-C | Optical

Data Information

Units	mg/dl
Decimals	0

Analysis

Type	END Method
Main Wave Length	546
Sub Wave Length	700
Method	DIRECT

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind	Vol.	Units
Sample	3	ul
Reagent 1	225	ul
Reagent 2	75	ul

Calibration

Type: Linear 1 ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability			
Factor MIN		Factor MAX	
CV MAX			

Data Process

Read	Start	End
Main	52	54
Sub	29	31

	low	High
Abs. Limit	-3	3

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

Lipase

Item No * ↓ | Item Name LIP | FULLNAME LIP | Optical

Data Information

Units	U/L
Decimals	1

Analysis

Type	Rate Method
Main Wave Length	570
Sub Wave Length	700
Method	Colorimetric

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind	Vol.	Units
Sample	4	ul
Reagent 1	160	ul
Reagent 2	40	ul

Calibration

Type: Linear 1 ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability			
Factor MIN		Factor MAX	
CV MAX			

Data Process

Read	Start	End
Main	43	52
Sub	0	0

Abs. Limit	low	High
	-3	3

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

Magnesium

Item No * ↓ | Item Name MG | FULLNAME MG | Optical

Data Information

Units	mg/dl
Decimals	1

Analysis

Type	END Method
Main Wave Length	505
Sub Wave Length	700
Method	Xylidyl Blue

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind	Vol.	Units
Sample	2	ul
Reagent 1	200	ul

Calibration

Type: Linear 1 ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability			
Factor MIN		Factor MAX	
CV MAX			

Data Process

Read	Start	End
Main	23	25
Sub	10	11

Abs. Limit	low	High
	-3	3

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

CK-MB

Item No * ↓ | Item Name CK-MB | FULLNAME CK-MB | Optical

Data Information

Units	U/L
Decimals	1

Analysis

Type	Rate Method
Main Wave Length	340
Sub Wave Length	546
Method	Immunoinhibition

Correlation

Slope		Intercept
Y =	1	* X + 0

Aspiration

Kind	Kind	Vol.	Units
Vol.	Sample	10	ul
	Reagent 1	200	ul
	Reagent 2	40	ul

Calibration

Type: Linear 1 ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability			
Factor MIN		Factor MAX	
CV MAX			

Data Process

Read		Start	End
	Main	41	52
	Sub	0	0

	low	High
Abs. Limit	-3	3

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

biolis 24i
prestige 24i

Micro Albumin

Item No * ↓ | Item Name MALB | FULLNAME MALB | Optical

Data Information

Units	mg/L
Decimals	1

Analysis

Type	END Method
Main Wave Length	340
Sub Wave Length	700
Method	Immunoturbidimetric

Correlation

Slope		Intercept
Y =	1	* X + 0

Aspiration

Kind	Vol.	Units
Sample	14	ul
Reagent 1	180	ul
Reagent 2	35	ul

Calibration

Type: Logit.log 2 ↓

Blank	0	#1	#	#2	#
#3	#	#4	#	#5	#
#6					

Calib Stability			
Factor MIN		Factor MAX	
CV MAX			

Data Process

Read

	Start	End
Main	52	54
Sub	29	31

	low	High
Abs. Limit	-3	3

\# غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

Phosphorus

Item No * ↓ | Item Name PHOS | FULLNAME PHOS | Optical

Data Information

Units	mg/dl
Decimals	1

Analysis

Type	END Method
Main Wave Length	340
Sub Wave Length	700
Method	Phosphomolybdate. UV

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind	Vol.	Units
Sample	2	ul
Reagent 1	200	ul

Calibration

Type: Linear 1 ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability | Factor MIN | Factor MAX | CV MAX

Data Process

Read	Start	End
Main	23	25
Sub	10	11

Abs. Limit	low	High
	-3	3

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

biolis 24i
prestige 24i

Total Protein

Item No * ↓ | Item Name T.PRO | FULLNAME T.PRO | Optical

Data Information

Units	g/dl
Decimals	2

Analysis

Type	END Method
Main Wave Length	546
Sub Wave Length	700
Method	Biuret

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind / Vol.

Kind	Vol.	Units
Sample	5	ul
Reagent 1	200	ul

Calibration

Type: Linear 1 ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability:

Factor MIN: | Factor MAX:

CV MAX:

Data Process

Read

	Start	End
Main	29	31
Sub	10	11

Abs. Limit: low -3 | High 3

\# غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

Triglycerides

Item No * ↓ | Item Name TG | FULLNAME TG | Optical

Data Information

Units	mg/dl
Decimals	0

Analysis

Type	END Method
Main Wave Length	505
Sub Wave Length	700
Method	GPO-POD

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind	Vol.	Units
Sample	2	ul
Reagent 1	200	ul

Calibration

Type: Linear 1 ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability | Factor MIN | Factor MAX | CV MAX

Data Process

Read	Start	End
Main	52	54
Sub	10	11

Abs. Limit	low	High
	-3	3

غلظت کالیبراتور

Urea/BUN

Item No * ↓ | Item Name UREA | FULLNAME UREA | Optical

Data Information

Units	mg/dl
Decimals	1

Analysis

Type	Rate Method
Main Wave Length	340
Sub Wave Length	700
Method	Urease kinetic

Correlation

Y = 1 (Slope) * X + 0 (Intercept)

Aspiration

Kind	Kind	Vol.	Units
Vol.	Sample	2	ul
	Reagent 1	160	ul
	Reagent 2	40	ul

Calibration

Type: Linear 1 ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability

Factor MIN | Factor MAX

CV MAX

Data Process

Read

	Start	End
Main	37	43
Sub	0	0

	low	High
Abs. Limit	-3	3

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

biolis 24i
prestige 24i

Uric acid

Item No	*	Item Name	UA	FULLNAME	UA	Optical

Data Information

Units	mg/dl
Decimals	1

Analysis

Type	END Method
Main Wave Length	546
Sub Wave Length	700
Method	Uricase Peroxidase

Correlation

	Slope		Intercept
Y =	1	* X +	0

Aspiration

Kind	Kind	Vol.	Units
Vol.	Sample	7	ul
	Reagent 1	140	ul
	Reagent 2	140	ul

Calibration

Type: Linear 1

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability			
Factor MIN		Factor MAX	
CV MAX			

Data Process

Read		Start	End
	Main	42	44
	Sub	0	0

	low	High
Abs. Limit	-3	3

\# غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

ZINC

Item No * ↓ | Item Name ZN | FULLNAME ZN | Optical

Data Information

Units	µg/dl
Decimals	0

Analysis

Type	END Method
Main Wave Length	570
Sub Wave Length	700
Method	Colorimetric

Correlation

Y = 1 * X + 0 (Slope: 1, Intercept: 0)

Aspiration

Kind	Vol.	Units
Sample	10	ul
Reagent 1	200	ul

Calibration

Type: Linear 1 ↓

Blank	0	#1	#	#2	
#3		#4		#5	
#6					

Calib Stability | Factor MIN | Factor MAX | CV MAX

Data Process

Read	Start	End
Main	23	25
Sub	10	11

Abs. Limit	low	High
	-3	3

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸