BIOMEDIC

DIRUI
CS 400 / CS 800

Albumin BCG

Analyze Parameters

Test Item:	ALB	Test Full Name:	ALB	Decimal Digit:	2	Unit:	g/dl		Sample Blank				
Assay:	1-point End	Test Time:	10	Point:	14	0	0	0	Control Interval:	0			
Main Wave:	600	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0					

Sample Vol.

	Serum			Urin		
Normal:	2	0	0	2	0	0
Decrement:	2	0	0	2	0	0
Increment:	4	0	0	4	0	0
Abs. Limit:	3.3			3.3	Positive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	200	*
R2:	0	*
R3:	0	*
R4:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: ALB

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	2	0	0
(2)	#	S2	2	0	0
(3)			2	0	0
(4)			2	0	0
(5)			2	0	0
(6)			2	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

DIRUI
CS 400 / CS 800

Alkaline Phosphatase

Analyze Parameters

Test Item:	ALP	Test Full Name:	ALP	Decimal Digit:	0	Unit:	U/L		Sample Blank				
Assay:	Rate A	Test Time:	10	Point:	19	29	0	0	Control Interval:	0			
Main Wave:	405	Second Wave:	660	Instrument Factor (Y=aX+b)	a=	1	b=	0					

Sample Vol.

	Serum			Urin		
Normal:	4	0	0	4	0	0
Decrement:	2	0	0	2	0	0
Increment:	8	0	0	8	0	0
Abs. Limit:	1.5			1.5	Positive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	160	*
R2:	0	*
R3:	40	*
R4:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: ALP

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	4	0	0
(2)	#	S2	4	0	0
(3)			4	0	0
(4)			4	0	0
(5)			4	0	0
(6)			4	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

DIRUI
CS 400 / CS 800

Alanine Aminotransferase (ALT)

Analyze Parameters

Test Item:	ALT	Test Full Name:	ALT	Decimal Digit:	0	Unit:	U/L		Sample Blank		
Assay:	Rate A	Test Time:	10	Point:	19	29	0	0	Control Interval:	0	
Main Wave:	340	Second Wave:	405	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin		
Normal:	20	0	0	20	0	0
Decrement:	10	0	0	10	0	0
Increment:	30	0	0	30	0	0
Abs. Limit:	0.5			0.5	Negetive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	160	*
R2:	0	*
R3:	40	*
R4:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: ALT

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	20	0	0
(2)	#	S2	20	0	0
(3)			20	0	0
(4)			20	0	0
(5)			20	0	0
(6)			20	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

DIRUI
CS 400 / CS 800

Amylase CNPG3

Analyze Parameters

Test Item:	AMY	Test Full Name:	AMY	Decimal Digit:	0	Unit:	U/L		Sample Blank				
Assay:	Rate A	Test Time:	10	Point:	4	12	0	0	Control Interval:	0			
Main Wave:	405	Second Wave:	660	Instrument Factor (Y=aX+b)	a=	1	b=	0					

Sample Vol.

	Serum			Urin		
Normal:	4	0	0	4	0	0
Decrement:	2	0	0	2	0	0
Increment:	8	0	0	8	0	0
Abs. Limit:	2			2	Positive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	200	*
R2:	0	*
R3:	0	*
R4:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: AMY

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	4	0	0
(2)	#	S2	4	0	0
(3)			4	0	0
(4)			4	0	0
(5)			4	0	0
(6)			4	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

DIRUI
CS 400 / CS 800

Aspartate Aminotransferase (AST)

Analyze Parameters

Test Item:	AST	Test Full Name:	AST	Decimal Digit:	0	Unit:	U/L		Sample Blank		
Assay:	Rate A	Test Time:	10	Point:	19	29	0	0	Control Interval:	0	
Main Wave:	340	Second Wave:	405	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin		
Normal:	20	0	0	20	0	0
Decrement:	10	0	0	10	0	0
Increment:	30	0	0	30	0	0
Abs. Limit:	0.5			0.5	Negetive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	160	*
R2:	0	*
R3:	40	*
R4:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: AST

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	20	0	0
(2)	#	S2	20	0	0
(3)			20	0	0
(4)			20	0	0
(5)			20	0	0
(6)			20	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

DIRUI
CS 400 / CS 800

Complement 3 (C3)

Analyze Parameters

Test Item:	C3	Test Full Name:	C3	Decimal Digit:	0	Unit:	mg/dl		Sample Blank				
Assay:	2-point End	Test Time:	10	Point:	14	31	0	0	Control Interval:	0			
Main Wave:	340	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0					

Sample Vol.

	Serum			Urin		
Normal:	2	0	0	2	0	0
Decrement:	2	0	0	2	0	0
Increment:	4	0	0	4	0	0
Abs. Limit:	3.3			3.3	Positive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	175	*
R2:	0	*
R3:	35	*
R4:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: C3

Caliration Methods: Logit-log 4P Point: 6 Span: 6

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	2	0	0
(2)	#	S2	2	0	0
(3)	#	S3	2	0	0
(4)	#	S4	2	0	0
(5)	#	S5	2	0	0
(6)	#	S6	2	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

DIRUI
CS 400 / CS 800

Complement 4 (C4)

Analyze Parameters

Test Item:	C4	Test Full Name:	C4	Decimal Digit:	1	Unit:	mg/dl		Sample Blank		
Assay:	2-point End	Test Time:	10	Point:	14	31	0	0	Control Interval:	0	
Main Wave:	340	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin			
Normal:	4	0	0	4	0	0	
Decrement:	2	0	0	2	0	0	
Increment:	8	0	0	8	0	0	
Abs. Limit:	3.3			3.3	Positive Reaction		
Prozone Limit:	-3.3			-3.3	Lower Limit		

Reagent

	Vol.	Pos.
R1:	175	*
R2:	0	*
R3:	35	*
R4:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: C4

Caliration Methods: Logit-log 4P | Point: 6 | Span: 6

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	4	0	0
(2)	#	S2	4	0	0
(3)	#	S3	4	0	0
(4)	#	S4	4	0	0
(5)	#	S5	4	0	0
(6)	#	S6	4	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

DIRUI
CS 400 / CS 800

Calcium Arsenazo III

Analyze Parameters

Test Item:	CA	Test Full Name:	CA	Decimal Digit:	1	Unit:	mg/dl		Sample Blank		
Assay:	1-point End	Test Time:	10	Point:	14	0	0	0	Control Interval:	0	
Main Wave:	660	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin		
Normal:	3	0	0	3	0	0
Decrement:	2	0	0	2	0	0
Increment:	6	0	0	6	0	0
Abs. Limit:	3.3			3.3	Positive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	200	*
R2:	0	*
R3:	0	*
R4:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: CA

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	3	0	0
(2)	#	S2	3	0	0
(3)			3	0	0
(4)			3	0	0
(5)			3	0	0
(6)			3	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

DIRUI
CS 400 / CS 800

Cholesterol

Analyze Parameters

Test Item:	CHOL	Test Full Name:	CHOL	Decimal Digit:	0	Unit:	mg/dl		Sample Blank		
Assay:	1-point End	Test Time:	10	Point:	31	0	0	0	Control Interval:	0	
Main Wave:	505	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin		
Normal:	2	0	0	2	0	0
Decrement:	2	0	0	2	0	0
Increment:	4	0	0	4	0	0
Abs. Limit:	3.3			3.3	Positive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	200	*
R2:	0	*
R3:	0	*
R4:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: CHOL

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	2	0	0
(2)	#	S2	2	0	0
(3)			2	0	0
(4)			2	0	0
(5)			2	0	0
(6)			2	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

DIRUI
CS 400 / CS 800

Creatine Kinase (CK-NAC)

Analyze Parameters

Test Item:	CPK	Test Full Name:	CPK	Decimal Digit:	0	Unit:	U/L		Sample Blank				
Assay:	Rate A	Test Time:	10	Point:	22	31	0	0	Control Interval:	0			
Main Wave:	340	Second Wave:	546	Instrument Factor (Y=aX+b)	a=	1	b=	0					

Sample Vol.

	Serum			Urin		
Normal:	10	0	0	10	0	0
Decrement:	5	0	0	5	0	0
Increment:	20	0	0	20	0	0
Abs. Limit:	1.5			1.5	Positive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	200	*
R2:	0	*
R3:	40	*
R4:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: CPK

Caliration Methods: Linear (2-point) | Point: 2 | Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	10	0	0
(2)	#	S2	10	0	0
(3)			10	0	0
(4)			10	0	0
(5)			10	0	0
(6)			10	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

DIRUI
CS 400 / CS 800

C - reactive protein (CRP)

Analyze Parameters

Test Item:	CRP	Test Full Name:	CRP	Decimal Digit:	1	Unit:	mg/L		Sample Blank		
Assay:	2-point End	Test Time:	10	Point:	14	31	0	0	Control Interval:	0	
Main Wave:	340	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin		
Normal:	9	0	0	9	0	0
Decrement:	4	0	0	4	0	0
Increment:	18	0	0	18	0	0
Abs. Limit:	3.3			3.3	Positive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	175	*
R2:	0	*
R3:	35	*
R4:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: CRP

Caliration Methods: Logit-log 4P | Point: 6 | Span: 6

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	9	0	0
(2)	#	S2	9	0	0
(3)	#	S3	9	0	0
(4)	#	S4	9	0	0
(5)	#	S5	9	0	0
(6)	#	S6	9	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

DIRUI
CS 400 / CS 800

Creatinine Jaffe

Analyze Parameters

Test Item:	CREAT	Test Full Name:	CREAT	Decimal Digit:	2	Unit:	mg/dl		Sample Blank		
Assay:	2-point rate	Test Time:	10	Point:	19	24	0	0	Control Interval:	0	
Main Wave:	505	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin		
Normal:	20	0	0	20	0	0
Decrement:	10	0	0	10	0	0
Increment:	30	0	0	30	0	0
Abs. Limit:	3.3			3.3	Positive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	100	*
R2:	0	*
R3:	100	*
R4:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: CREAT

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	20	0	0
(2)	#	S2	20	0	0
(3)			20	0	0
(4)			20	0	0
(5)			20	0	0
(6)			20	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

DIRUI
CS 400 / CS 800

Copper

Analyze Parameters

Test Item:	Copper	Test Full Name:	Copper	Decimal Digit:	0	Unit:	µg/dl		Sample Blank				
Assay:	1-point End	Test Time:	10	Point:	14	0	0	0	Control Interval:	0			
Main Wave:	570	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0					

Sample Vol.

	Serum			Urin		
Normal:	10	0	0	10	0	0
Decrement:	5	0	0	5	0	0
Increment:	20	0	0	20	0	0
Abs. Limit:	3.3			3.3	Positive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	200	*
R2:	0	*
R3:	0	*
R4:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: Copper

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	10	0	0
(2)	#	S2	10	0	0
(3)			10	0	0
(4)			10	0	0
(5)			10	0	0
(6)			10	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

DIRUI
CS 400 / CS 800

Glucose

Analyze Parameters

Test Item:	GLU	Test Full Name:	GLU	Decimal Digit:	0	Unit:	mg/dl		Sample Blank				
Assay:	1-point End	Test Time:	10	Point:	31	0	0	0	Control Interval:	0			
Main Wave:	505	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0					

Sample Vol.

	Serum			Urin		
Normal:	2	0	0	2	0	0
Decrement:	2	0	0	2	0	0
Increment:	4	0	0	4	0	0
Abs. Limit:	3.3			3.3	Positive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	200	*
R2:	0	*
R3:	0	*
R4:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: GLU

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	2	0	0
(2)	#	S2	2	0	0
(3)			2	0	0
(4)			2	0	0
(5)			2	0	0
(6)			2	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

DIRUI
CS 400 / CS 800

HDL Cholesterol Direct

Analyze Parameters

Test Item:	HDL	Test Full Name:	HDL	Decimal Digit:	0	Unit:	mg/dl		Sample Blank		
Assay:	2-point End	Test Time:	10	Point:	14	31	0	0	Control Interval:	0	
Main Wave:	600	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin		
Normal:	3	0	0	3	0	0
Decrement:	2	0	0	2	0	0
Increment:	6	0	0	6	0	0
Abs. Limit:	3.3			3.3	Positive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	225	*
R2:	0	*
R3:	75	*
R4:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: HDL

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	3	0	0
(2)	#	S2	3	0	0
(3)			3	0	0
(4)			3	0	0
(5)			3	0	0
(6)			3	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

DIRUI
CS 400 / CS 800

Immunoglobulin A (IgA)

Analyze Parameters

Test Item:	IgA	Test Full Name:	IgA	Decimal Digit:	0	Unit:	mg/dl		Sample Blank		
Assay:	2-point End	Test Time:	10	Point:	14	31	0	0	Control Interval:	0	
Main Wave:	340	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin		
Normal:	2	0	0	2	0	0
Decrement:	2	0	0	2	0	0
Increment:	4	0	0	4	0	0
Abs. Limit:	3.3			3.3	Positive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	175	*
R2:	0	*
R3:	35	*
R4:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: IgA

Caliration Methods: Logit-log 4P | Point: 6 | Span: 6

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	2	0	0
(2)	#	S2	2	0	0
(3)	#	S3	2	0	0
(4)	#	S4	2	0	0
(5)	#	S5	2	0	0
(6)	#	S6	2	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

DIRUI
CS 400 / CS 800

Immunoglobulin G (IgG)

Analyze Parameters

Test Item:	IgG	Test Full Name:	IgG	Decimal Digit:	0	Unit:	mg/dl		Sample Blank				
Assay:	2-point End	Test Time:	10	Point:	14	31	0	0	Control Interval:	0			
Main Wave:	600	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0					

Sample Vol.

	Serum			Urin		
Normal:	2	0	0	2	0	0
Decrement:	2	0	0	2	0	0
Increment:	4	0	0	4	0	0
Abs. Limit:	3.3			3.3	Positive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	175	*
R2:	0	*
R3:	35	*
R4:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: IgG

Caliration Methods: Logit-log 4P | Point: 6 | Span: 6

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	2	0	0
(2)	#	S2	2	0	0
(3)	#	S3	2	0	0
(4)	#	S4	2	0	0
(5)	#	S5	2	0	0
(6)	#	S6	2	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

DIRUI
CS 400 / CS 800

Immunoglobulin M (IgM)

Analyze Parameters

Test Item:	IgM	Test Full Name:	IgM	Decimal Digit:	0	Unit:	mg/dl		Sample Blank		
Assay:	2-point End	Test Time:	10	Point:	14	31	0	0	Control Interval:	0	
Main Wave:	340	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin		
Normal:	2	0	0	2	0	0
Decrement:	2	0	0	2	0	0
Increment:	4	0	0	4	0	0
Abs. Limit:	3.3			3.3	Positive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	175	*
R2:	0	*
R3:	35	*
R4:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: IgM

Caliration Methods: Logit-log 4P | Point: 6 | Span: 6

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	2	0	0
(2)	#	S2	2	0	0
(3)	#	S3	2	0	0
(4)	#	S4	2	0	0
(5)	#	S5	2	0	0
(6)	#	S6	2	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

DIRUI
CS 400 / CS 800

Lactate Dehydrogenase (LDH)

Analyze Parameters

Test Item:	LDH	Test Full Name:	LDH	Decimal Digit:	0	Unit:	U/L		Sample Blank		
Assay:	Rate A	Test Time:	10	Point:	19	29	0	0	Control Interval:	0	
Main Wave:	340	Second Wave:	405	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin		
Normal:	4	0	0	4	0	0
Decrement:	2	0	0	2	0	0
Increment:	8	0	0	8	0	0
Abs. Limit:	0.5			0.5	Negetive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	200	*
R2:	0	*
R3:	40	*
R4:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: LDH

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	4	0	0
(2)	#	S2	4	0	0
(3)			4	0	0
(4)			4	0	0
(5)			4	0	0
(6)			4	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

DIRUI
CS 400 / CS 800

LDL Cholesterol Direct

Analyze Parameters

Test Item:	LDL	Test Full Name:	LDL	Decimal Digit:	0	Unit:	mg/dl		Sample Blank		
Assay:	2-point End	Test Time:	10	Point:	14	31	0	0	Control Interval:	0	
Main Wave:	600	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin		
Normal:	3	0	0	3	0	0
Decrement:	2	0	0	2	0	0
Increment:	6	0	0	6	0	0
Abs. Limit:	3.3			3.3	Positive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	225	*
R2:	0	*
R3:	75	*
R4:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: LDL

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	3	0	0
(2)	#	S2	3	0	0
(3)			3	0	0
(4)			3	0	0
(5)			3	0	0
(6)			3	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

DIRUI
CS 400 / CS 800

Lipase

Analyze Parameters

Test Item:	LIP	Test Full Name:	LIP	Decimal Digit:	0	Unit:	U/L		Sample Blank				
Assay:	Rate A	Test Time:	10	Point:	22	28	0	0	Control Interval:	0			
Main Wave:	570	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0					

Sample Vol.

	Serum			Urin		
Normal:	4	0	0	4	0	0
Decrement:	2	0	0	2	0	0
Increment:	8	0	0	8	0	0
Abs. Limit:	2			2	Positive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	160	*
R2:	0	*
R3:	40	*
R4:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: LIP

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	4	0	0
(2)	#	S2	4	0	0
(3)			4	0	0
(4)			4	0	0
(5)			4	0	0
(6)			4	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

DIRUI
CS 400 / CS 800

Microalbumin

Analyze Parameters

Test Item:	MALB	Test Full Name:	MALB	Decimal Digit:	1	Unit:	mg/L		Sample Blank		
Assay:	2-point End	Test Time:	10	Point:	14	31	0	0	Control Interval:	0	
Main Wave:	340	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin		
Normal:	15	0	0	15	0	0
Decrement:	3	0	0	3	0	0
Increment:	25	0	0	25	0	0
Abs. Limit:	3.3			3.3	Positive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	175	*
R2:	0	*
R3:	35	*
R4:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: MALB

Caliration Methods: Logit-log 4P | Point: 6 | Span: 6

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	15	0	0
(2)	#	S2	15	0	0
(3)	#	S3	15	0	0
(4)	#	S4	15	0	0
(5)	#	S5	15	0	0
(6)	#	S6	15	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

DIRUI
CS 400 / CS 800

CK - MB

Analyze Parameters

Test Item:	CKMB	Test Full Name:	CKMB	Decimal Digit:	1	Unit:	U/L		Sample Blank		
Assay:	Rate A	Test Time:	10	Point:	22	31	0	0	Control Interval:	0	
Main Wave:	340	Second Wave:	546	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin		
Normal:	10	0	0	10	0	0
Decrement:	5	0	0	5	0	0
Increment:	20	0	0	20	0	0
Abs. Limit:	1.5			1.5	Positive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	200	*
R2:	0	*
R3:	40	*
R4:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: CKMB

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	10	0	0
(2)	#	S2	10	0	0
(3)			10	0	0
(4)			10	0	0
(5)			10	0	0
(6)			10	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

DIRUI
CS 400 / CS 800

Magnesium Xylidyl Blue

Analyze Parameters

Test Item:	MG	Test Full Name:	MG	Decimal Digit:	1	Unit:	mg/dl		Sample Blank				
Assay:	1-point End	Test Time:	10	Point:	14	0	0	0	Control Interval:	0			
Main Wave:	505	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0					

Sample Vol.

	Serum			Urin		
Normal:	2	0	0	2	0	0
Decrement:	2	0	0	2	0	0
Increment:	4	0	0	4	0	0
Abs. Limit:	3.3			3.3	Positive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	200	*
R2:	0	*
R3:	0	*
R4:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: MG

Caliration Methods: Linear (2-point) | Point: 2 | Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	2	0	0
(2)	#	S2	2	0	0
(3)			2	0	0
(4)			2	0	0
(5)			2	0	0
(6)			2	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

DIRUI
CS 400 / CS 800

Phosphorous UV

Analyze Parameters

Test Item:	PHOS	Test Full Name:	PHOS	Decimal Digit:	1	Unit:	mg/dl		Sample Blank		
Assay:	1-point End	Test Time:	10	Point:	14	0	0	0	Control Interval:	0	
Main Wave:	340	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin		
Normal:	2	0	0	2	0	0
Decrement:	2	0	0	2	0	0
Increment:	4	0	0	4	0	0
Abs. Limit:	3.3			3.3	Positive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	200	*
R2:	0	*
R3:	0	*
R4:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: PHOS

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	2	0	0
(2)	#	S2	2	0	0
(3)			2	0	0
(4)			2	0	0
(5)			2	0	0
(6)			2	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

DIRUI
CS 400 / CS 800

Triglycerides

Analyze Parameters

Test Item:	TG	Test Full Name:	TG	Decimal Digit:	0	Unit:	mg/dl		Sample Blank				
Assay:	1-point End	Test Time:	10	Point:	31	0	0	0	Control Interval:	0			
Main Wave:	505	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0					

Sample Vol.

	Serum			Urin		
Normal:	2	0	0	2	0	0
Decrement:	2	0	0	2	0	0
Increment:	4	0	0	4	0	0
Abs. Limit:	3.3			3.3	Positive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	200	*
R2:	0	*
R3:	0	*
R4:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: TG

Caliration Methods: Linear (2-point) | Point: 2 | Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	2	0	0
(2)	#	S2	2	0	0
(3)			2	0	0
(4)			2	0	0
(5)			2	0	0
(6)			2	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

DIRUI
CS 400 / CS 800

Total Protein

Analyze Parameters

Test Item:	TPRO	Test Full Name:	TPRO	Decimal Digit:	2	Unit:	g/dl		Sample Blank		
Assay:	1-point End	Test Time:	10	Point:	14	0	0	0	Control Interval:	0	
Main Wave:	546	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin		
Normal:	5	0	0	5	0	0
Decrement:	2	0	0	2	0	0
Increment:	10	0	0	10	0	0
Abs. Limit:	3.3			3.3	Positive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	200	*
R2:	0	*
R3:	0	*
R4:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: TPRO

Caliration Methods: Linear (2-point) | Point: 2 | Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	5	0	0
(2)	#	S2	5	0	0
(3)			5	0	0
(4)			5	0	0
(5)			5	0	0
(6)			5	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

DIRUI
CS 400 / CS 800

Uric Acid

Analyze Parameters

Test Item:	UA	Test Full Name:	UA	Decimal Digit:	1	Unit:	mg/dl		Sample Blank				
Assay:	1-point End	Test Time:	10	Point:	18	0	0	0	Control Interval:	0			
Main Wave:	546	Second Wave:	700	Instrument Factor (Y=aX+b)	a=	1	b=	0					

Sample Vol.

	Serum			Urin		
Normal:	5	0	0	5	0	0
Decrement:	2	0	0	2	0	0
Increment:	10	0	0	10	0	0
Abs. Limit:	3.3			3.3	Positive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	100	*
R2:	100	*
R3:	0	*
R4:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: UA

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	5	0	0
(2)	#	S2	5	0	0
(3)			5	0	0
(4)			5	0	0
(5)			5	0	0
(6)			5	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

DIRUI
CS 400 / CS 800

UREA

Analyze Parameters

Test Item:	UREA	Test Full Name:	UREA	Decimal Digit:	0	Unit:	mg/dl		Sample Blank		
Assay:	2-point rate	Test Time:	10	Point:	19	24	0	0	Control Interval:	0	
Main Wave:	340	Second Wave:	405	Instrument Factor (Y=aX+b)	a=	1	b=	0			

Sample Vol.

	Serum			Urin		
Normal:	2	0	0	2	0	0
Decrement:	2	0	0	2	0	0
Increment:	4	0	0	4	0	0
Abs. Limit:	0.1			0.1	Negetive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	160	*
R2:	0	*
R3:	40	*
R4:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: UREA

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	2	0	0
(2)	#	S2	2	0	0
(3)			2	0	0
(4)			2	0	0
(5)			2	0	0
(6)			2	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

DIRUI
CS 400 / CS 800

ZINC

Analyze Parameters

Test Item:	ZINC	Test Full Name:	ZINC	Decimal Digit:	0	Unit:	µg/dl		Sample Blank				
Assay:	1-point End	Test Time:	10	Point:	14	0	0	0	Control Interval:	0			
Main Wave:	570	Second Wave:	700	Instrument Factor (Y=aX+b)			a=	1	b=	0			

Sample Vol.

	Serum			Urin		
Normal:	10	0	0	10	0	0
Decrement:	5	0	0	5	0	0
Increment:	20	0	0	20	0	0
Abs. Limit:	3.3			3.3	Positive Reaction	
Prozone Limit:	-3.3			-3.3	Lower Limit	

Reagent

	Vol.	Pos.
R1:	200	*
R2:	0	*
R3:	0	*
R4:	0	*

Second half item of Two test analyze

Calibration Parameters

Test Item: ZINC

Caliration Methods: Linear (2-point) Point: 2 Span: 2

Drift Rate Checkup: 3.3

Discreteness Checkup: 3.3 Abs.

Sensitivity Checkup: 0

Blank Horizontal Checkuo: -3.3 - -3.3

Caliration Solution

(1)	0	S1	10	0	0
(2)	#	S2	10	0	0
(3)			10	0	0
(4)			10	0	0
(5)			10	0	0
(6)			10	0	0

Automatic Calibration

	Time Out	
Blank calibration:	0	Hour
Span calibration:	0	Hour
2 point calibration:	0	Hour
Full point calibration:	0	Hour

* توسط اپراتور وارد شود (جایگاه محلولها)

غلظت کالیبراتور

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸