HITACHI 902		
1	Test Name	ALB
2	Assay Code (Mthd)	1 point
3	Assay Code (2. Test)	0
4	Reaction Time	5
5	Assay Point 1	15
6	Assay Point 2	0
7	Assay Point 3	0
8	Assay Point 4	0
9	Wave Leng. (SUB)	700
10	Wave Leng. (MAIN)	600
11	Sample Volume	3
12	R1 VOLUME	300
13	R1 Pos.	#
14	R1 Bottle Size	Large
15	R2 VOLUME	0
16	R2 Pos. 0	0
17	R2 Bottle Size	Small
18	R3 VOLUME	0
19	R3 Pos. 0	#
20	R3 Bottle Size Small	Small
21	Calib. Type (Type)	LINEAR
22	Calib. Type (Wght)	0
23	Calib. Conc. 1	0.0
24	Calib. Pos. 1	36
25	Calib. Conc. 2	#
26	Calib. Pos. 2	##
27	Calib. Conc. 3	0
28	Calib. Pos. 3	0
29	Calib. Conc. 4	0
30	Calib. Pos. 4	0
31	Calib. Conc. 5	0
32	Calib. Pos. 5	0
33	Calib. Conc. 6	0
34	Calib. Pos. 6	0
35	S 1 ABS.	0
36	K Factor	10000
37	K 2 Factor	10000
38	K 3 Factor	10000
39	K 4 Factor	10000
40	K 5 Factor	10000
41	A Factor	0
42	B Factor	0
43	C Factor	0
44	SD Limit	0.1
45	Duplicate Limit	500
46	Sens. Limit	0
47	S 1 ABS Limit (L)	-32000
48	S 1 ABS Limit (H)	32000
49	ABS Limit	0
50	ABS Limit (D/I)	INCREASE
51	Prz. Limit	0
52	Prz. Limit (U/D)	Lower
53	Prz. (End Point)	35
54	Expect. Value (L)	3.5
55	Expect. Value (H)	5.2
56	Instr. Fact. (a)	1
57	Instr. Fact. (b)	0
58	Key Setting	#

Albumin (BCG)

BCG
Liquid, 1 Reagent
Calibrator : GENERAL CALIBRATOR (CAT NO. CA532)
Control : GENERAL Control 1 & 2

HITACHI 717	
TEST	ALB
ASSAY CODE	[1 point] [24] - [0]
SAMPLE VOL (µL)	[3] [3]
R1 VOLUME (µL)	[300] [50] - [NO]
R2 VOLUME (µL)	[0] [20] - [NO]
WAVELENGTH	[700] [600]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0.0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[g/dl]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[0] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[3.5] [5.2]
PANIC VALUE	[0] [6]
INSTRUMENT FACTOR	[1]

HITACHI 704	
TEST	ALB
ASSAY CODE	[1 point] [15] - [0]
SAMPLE VOL (µL)	[3] [3]
R1 VOLUME (µL)	[300] [50] - [NO]
R2 VOLUME (µL)	[0] [20] - [NO]
WAVELENGTH	[700] [600]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0.0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[g/dl]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[0] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[3.5] [5.2]
PANIC VALUE	[0] [6]
INSTRUMENT FACTOR	[1]

User defined
جایگاه کالیبراتور

HITACHI 902		
1	Test Name	ALP
2	Assay Code (Mthd)	Rate A
3	Assay Code (2. Test)	0
4	Reaction Time	10
5	Assay Point 1	21
6	Assay Point 2	31
7	Assay Point 3	0
8	Assay Point 4	0
9	Wave Leng. (SUB)	660
10	Wave Leng. (MAIN)	415
11	Sample Volume	5
12	R1 VOLUME	240
13	R1 Pos.	#
14	R1 Bottle Size	Large
15	R2 VOLUME	0
16	R2 Pos. 0	0
17	R2 Bottle Size	Small
18	R3 VOLUME	60
19	R3 Pos. 0	#
20	R3 Bottle Size Small	Small
21	Calib. Type (Type)	LINEAR
22	Calib. Type (Wght)	0
23	Calib. Conc. 1	0
24	Calib. Pos. 1	36
25	Calib. Conc. 2	#
26	Calib. Pos. 2	##
27	Calib. Conc. 3	0
28	Calib. Pos. 3	0
29	Calib. Conc. 4	0
30	Calib. Pos. 4	0
31	Calib. Conc. 5	0
32	Calib. Pos. 5	0
33	Calib. Conc. 6	0
34	Calib. Pos. 6	0
35	S 1 ABS.	0
36	K Factor	10000
37	K 2 Factor	10000
38	K 3 Factor	10000
39	K 4 Factor	10000
40	K 5 Factor	10000
41	A Factor	0
42	B Factor	0
43	C Factor	0
44	SD Limit	0.1
45	Duplicate Limit	500
46	Sens. Limit	0
47	S 1 ABS Limit (L)	-32000
48	S 1 ABS Limit (H)	32000
49	ABS Limit	15000
50	ABS Limit (D/I)	INCREASE
51	Prz. Limit	0
52	Prz. Limit (U/D)	Lower
53	Prz. (End Point)	35
54	Expect. Value (L)	98
55	Expect. Value (H)	279
56	Instr. Fact. (a)	1
57	Instr. Fact. (b)	0
58	Key Setting	#

Alkaline phosphatase

DGKC
Liquid, 2 Reagent
Calibrator : GENERAL CALIBRATOR (CAT NO. CA532)
Control : GENERAL Control 1 & 2

HITACHI 717	
TEST	[ALP]
ASSAY CODE	[Rate A] [30] - [45]
SAMPLE VOL (µL)	[5] [5]
R1 VOLUME (µL)	[240] [50] - [NO]
R2 VOLUME (µL)	[60] [20] - [NO]
WAVELENGTH	[660] [415]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[U/L]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[15000] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[98] [279]
PANIC VALUE	[0] [1000]
INSTRUMENT FACTOR	[1]

HITACHI 704	
TEST	[ALP]
ASSAY CODE	[Rate A] [19] - [30]
SAMPLE VOL (µL)	[5] [5]
R1 VOLUME (µL)	[240] [50] - [NO]
R2 VOLUME (µL)	[60] [20] - [NO]
WAVELENGTH	[660] [415]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[U/L]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[15000] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[98] [279]
PANIC VALUE	[0] [1000]
INSTRUMENT FACTOR	[1]

User defined
جایگاه کالیبراتور ##

HITACHI 902		
1	Test Name	ALT
2	Assay Code (Mthd)	Rate A
3	Assay Code (2. Test)	0
4	Reaction Time	10
5	Assay Point 1	21
6	Assay Point 2	31
7	Assay Point 3	0
8	Assay Point 4	0
9	Wave Leng. (SUB)	415
10	Wave Leng. (MAIN)	340
11	Sample Volume	20
12	R1 VOLUME	200
13	R1 Pos.	#
14	R1 Bottle Size	Large
15	R2 VOLUME	0
16	R2 Pos. 0	0
17	R2 Bottle Size	Small
18	R3 VOLUME	50
19	R3 Pos. 0	#
20	R3 Bottle Size Small	Small
21	Calib. Type (Type)	LINEAR
22	Calib. Type (Wght)	0
23	Calib. Conc. 1	0
24	Calib. Pos. 1	36
25	Calib. Conc. 2	#
26	Calib. Pos. 2	##
27	Calib. Conc. 3	0
28	Calib. Pos. 3	0
29	Calib. Conc. 4	0
30	Calib. Pos. 4	0
31	Calib. Conc. 5	0
32	Calib. Pos. 5	0
33	Calib. Conc. 6	0
34	Calib. Pos. 6	0
35	S 1 ABS.	0
36	K Factor	10000
37	K 2 Factor	10000
38	K 3 Factor	10000
39	K 4 Factor	10000
40	K 5 Factor	10000
41	A Factor	0
42	B Factor	0
43	C Factor	0
44	SD Limit	0.1
45	Duplicate Limit	500
46	Sens. Limit	0
47	S 1 ABS Limit (L)	-32000
48	S 1 ABS Limit (H)	32000
49	ABS Limit	5000
50	ABS Limit (D/I)	DECREASE
51	Prz. Limit	0
52	Prz. Limit (U/D)	Lower
53	Prz. (End Point)	35
54	Expect. Value (L)	0
55	Expect. Value (H)	40
56	Instr. Fact. (a)	1
57	Instr. Fact. (b)	0
58	Key Setting	#

ALT/GPT

IFCC
Liquid, 2 Reagent
Calibrator : GENERAL CALIBRATOR (CAT NO. CA532)
Control : GENERAL Control 1 & 2

HITACHI 717	
TEST	[ALT]
ASSAY CODE	[Rate A] [30] - [45]
SAMPLE VOL (µL)	[20] [20]
R1 VOLUME (µL)	[200] [50] - [NO]
R2 VOLUME (µL)	[50] [20] - [NO]
WAVELENGTH	[415] [340]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[U/L]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[5000] [DECREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[0] [40]
PANIC VALUE	[0] [350]
INSTRUMENT FACTOR	[1]

HITACHI 704	
TEST	[ALT]
ASSAY CODE	[Rate A] [20] - [30]
SAMPLE VOL (µL)	[20] [20]
R1 VOLUME (µL)	[200] [50] - [NO]
R2 VOLUME (µL)	[50] [20] - [NO]
WAVELENGTH	[415] [340]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[U/L]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[5000] [DECREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[0] [40]
PANIC VALUE	[0] [350]
INSTRUMENT FACTOR	[1]

User defined
جایگاه کالیبراتور

HITACHI 902		
1	Test Name	AMY
2	Assay Code (Mthd)	Rate A
3	Assay Code (2. Test)	0
4	Reaction Time	5
5	Assay Point 1	4
6	Assay Point 2	15
7	Assay Point 3	0
8	Assay Point 4	0
9	Wave Leng. (SUB)	660
10	Wave Leng. (MAIN)	415
11	Sample Volume	6
12	R1 VOLUME	300
13	R1 Pos.	#
14	R1 Bottle Size	Large
15	R2 VOLUME	0
16	R2 Pos. 0	0
17	R2 Bottle Size	Small
18	R3 VOLUME	0
19	R3 Pos. 0	#
20	R3 Bottle Size Small	Small
21	Calib. Type (Type)	LINEAR
22	Calib. Type (Wght)	0
23	Calib. Conc. 1	0
24	Calib. Pos. 1	36
25	Calib. Conc. 2	#
26	Calib. Pos. 2	##
27	Calib. Conc. 3	0
28	Calib. Pos. 3	0
29	Calib. Conc. 4	0
30	Calib. Pos. 4	0
31	Calib. Conc. 5	0
32	Calib. Pos. 5	0
33	Calib. Conc. 6	0
34	Calib. Pos. 6	0
35	S 1 ABS.	0
36	K Factor	10000
37	K 2 Factor	10000
38	K 3 Factor	10000
39	K 4 Factor	10000
40	K 5 Factor	10000
41	A Factor	0
42	B Factor	0
43	C Factor	0
44	SD Limit	0.1
45	Duplicate Limit	500
46	Sens. Limit	0
47	S 1 ABS Limit (L)	-32000
48	S 1 ABS Limit (H)	32000
49	ABS Limit	20000
50	ABS Limit (D/I)	INCREASE
51	Prz. Limit	0
52	Prz. Limit (U/D)	Lower
53	Prz. (End Point)	35
54	Expect. Value (L)	0
55	Expect. Value (H)	90
56	Instr. Fact. (a)	1
57	Instr. Fact. (b)	0
58	Key Setting	#

Amylase - total

CNPG3. Kinetic
Liquid, 1 Reagent
Calibrator : GENERAL CALIBRATOR (CAT NO. CA532)
Control : GENERAL Control 1 & 2

HITACHI 717	
TEST	[AMY]
ASSAY CODE	[Rate A] [5] - [20]
SAMPLE VOL (µL)	[6] [6]
R1 VOLUME (µL)	[300] [50] - [NO]
R2 VOLUME (µL)	[0] [20] - [NO]
WAVELENGTH	[660] [415]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[U/L]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[20000] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[0] [90]
PANIC VALUE	[0] [2000]
INSTRUMENT FACTOR	[1]

HITACHI 704	
TEST	[AMY]
ASSAY CODE	[Rate A] [4] - [14]
SAMPLE VOL (µL)	[6] [6]
R1 VOLUME (µL)	[300] [50] - [NO]
R2 VOLUME (µL)	[0] [20] - [NO]
WAVELENGTH	[660] [415]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[U/L]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[20000] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[0] [90]
PANIC VALUE	[0] [2000]
INSTRUMENT FACTOR	[1]

User defined
جایگاه کالیبراتور

HITACHI 902		
1	Test Name	AST
2	Assay Code (Mthd)	Rate A
3	Assay Code (2. Test)	0
4	Reaction Time	10
5	Assay Point 1	21
6	Assay Point 2	31
7	Assay Point 3	0
8	Assay Point 4	0
9	Wave Leng. (SUB)	415
10	Wave Leng. (MAIN)	340
11	Sample Volume	20
12	R1 VOLUME	200
13	R1 Pos.	#
14	R1 Bottle Size	Large
15	R2 VOLUME	0
16	R2 Pos. 0	0
17	R2 Bottle Size	Small
18	R3 VOLUME	50
19	R3 Pos. 0	#
20	R3 Bottle Size Small	Small
21	Calib. Type (Type)	LINEAR
22	Calib. Type (Wght)	0
23	Calib. Conc. 1	0
24	Calib. Pos. 1	36
25	Calib. Conc. 2	#
26	Calib. Pos. 2	##
27	Calib. Conc. 3	0
28	Calib. Pos. 3	0
29	Calib. Conc. 4	0
30	Calib. Pos. 4	0
31	Calib. Conc. 5	0
32	Calib. Pos. 5	0
33	Calib. Conc. 6	0
34	Calib. Pos. 6	0
35	S 1 ABS.	0
36	K Factor	10000
37	K 2 Factor	10000
38	K 3 Factor	10000
39	K 4 Factor	10000
40	K 5 Factor	10000
41	A Factor	0
42	B Factor	0
43	C Factor	0
44	SD Limit	0.1
45	Duplicate Limit	500
46	Sens. Limit	0
47	S 1 ABS Limit (L)	-32000
48	S 1 ABS Limit (H)	32000
49	ABS Limit	5000
50	ABS Limit (D/I)	DECREASE
51	Prz. Limit	0
52	Prz. Limit (U/D)	Lower
53	Prz. (End Point)	35
54	Expect. Value (L)	0
55	Expect. Value (H)	38
56	Instr. Fact. (a)	1
57	Instr. Fact. (b)	0
58	Key Setting	#

AST/GOT

IFCC
Liquid, 2 Reagent
Calibrator : GENERAL CALIBRATOR (CAT NO. CA532)
Control : GENERAL Control 1 & 2

HITACHI 717	
TEST	[AST]
ASSAY CODE	[Rate A] [30] - [45]
SAMPLE VOL (µL)	[20] [20]
R1 VOLUME (µL)	[200] [50] - [NO]
R2 VOLUME (µL)	[50] [20] - [NO]
WAVELENGTH	[415] [340]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[U/L]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[5000] [DECREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[0] [38]
PANIC VALUE	[0] [350]
INSTRUMENT FACTOR	[1]

HITACHI 704	
TEST	[AST]
ASSAY CODE	[Rate A] [20] - [30]
SAMPLE VOL (µL)	[20] [20]
R1 VOLUME (µL)	[200] [50] - [NO]
R2 VOLUME (µL)	[50] [20] - [NO]
WAVELENGTH	[415] [340]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[U/L]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[5000] [DECREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[0] [38]
PANIC VALUE	[0] [350]
INSTRUMENT FACTOR	[1]

User defined
جایگاه کالیبراتور

HITACHI 902		
1	Test Name	C3
2	Assay Code (Mthd)	2 Point End
3	Assay Code (2. Test)	0
4	Reaction Time	10
5	Assay Point 1	16
6	Assay Point 2	35
7	Assay Point 3	0
8	Assay Point 4	0
9	Wave Leng. (SUB)	700
10	Wave Leng. (MAIN)	340
11	Sample Volume	3
12	R1 VOLUME	250
13	R1 Pos.	#
14	R1 Bottle Size	Large
15	R2 VOLUME	0
16	R2 Pos. 0	0
17	R2 Bottle Size	Small
18	R3 VOLUME	50
19	R3 Pos. 0	#
20	R3 Bottle Size Small	Small
21	Calib. Type (Type)	LOGIT-LOG (4P)
22	Calib. Type (Wght)	0
23	Calib. Conc. 1	0
24	Calib. Pos. 1	36
25	Calib. Conc. 2	#
26	Calib. Pos. 2	##
27	Calib. Conc. 3	#
28	Calib. Pos. 3	##
29	Calib. Conc. 4	#
30	Calib. Pos. 4	##
31	Calib. Conc. 5	#
32	Calib. Pos. 5	##
33	Calib. Conc. 6	#
34	Calib. Pos. 6	##
35	S 1 ABS.	0
36	K Factor	10000
37	K 2 Factor	10000
38	K 3 Factor	10000
39	K 4 Factor	10000
40	K 5 Factor	10000
41	A Factor	0
42	B Factor	0
43	C Factor	0
44	SD Limit	999
45	Duplicate Limit	2000
46	Sens. Limit	0
47	S 1 ABS Limit (L)	-32000
48	S 1 ABS Limit (H)	32000
49	ABS Limit	0
50	ABS Limit (D/I)	INCREASE
51	Prz. Limit	0
52	Prz. Limit (U/D)	Lower
53	Prz. (End Point)	35
54	Expect. Value (L)	80
55	Expect. Value (H)	180
56	Instr. Fact. (a)	1
57	Instr. Fact. (b)	0
58	Key Setting	#

Complement 3 (C3)

IMMUNOTURB
Liquid, 2 Reagent
Calibrator : PROTEIN CALIBRATOR (CAT NO. SP131)
Control : PROTEIN Control 1 & 2

HITACHI 717	
TEST	C3
ASSAY CODE	[2 Point End] [24] - [50]
SAMPLE VOL (µL)	[3] [3]
R1 VOLUME (µL)	[0] [50] - [NO]
R2 VOLUME (µL)	[0] [20] - [NO]
WAVELENGTH	[700] [340]
CALIB. METHOD	[NONLINEAR] [1] - [6]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[#] - [##]
STD 4 CONC. POS.	[#] - [##]
STD 5 CONC. POS.	[#] - [##]
STD 6 CONC. POS.	[#] - [##]
UNIT	[mg/dl]
SD LIMIT	[999]
DUPLICATE LIMIT	[2000]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[0] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[80] [180]
PANIC VALUE	[0] [500]
INSTRUMENT FACTOR	[1]

HITACHI 704	
TEST	C3
ASSAY CODE	[2 Point End] [15] - [32]
SAMPLE VOL (µL)	[3] [3]
R1 VOLUME (µL)	[0] [50] - [NO]
R2 VOLUME (µL)	[0] [20] - [NO]
WAVELENGTH	[700] [340]
CALIB. METHOD	[NONLINEAR] [1] - [6]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[#] - [##]
STD 4 CONC. POS.	[#] - [##]
STD 5 CONC. POS.	[#] - [##]
STD 6 CONC. POS.	[#] - [##]
UNIT	[mg/dl]
SD LIMIT	[999]
DUPLICATE LIMIT	[2000]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[0] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[80] [180]
PANIC VALUE	[0] [500]
INSTRUMENT FACTOR	[1]

User defined
جایگاه کالیبراتور

HITACHI 902		
1	Test Name	C4
2	Assay Code (Mthd)	2 Point End
3	Assay Code (2. Test)	0
4	Reaction Time	10
5	Assay Point 1	16
6	Assay Point 2	35
7	Assay Point 3	0
8	Assay Point 4	0
9	Wave Leng. (SUB)	700
10	Wave Leng. (MAIN)	340
11	Sample Volume	6
12	R1 VOLUME	250
13	R1 Pos.	#
14	R1 Bottle Size	Large
15	R2 VOLUME	0
16	R2 Pos. 0	0
17	R2 Bottle Size	Small
18	R3 VOLUME	50
19	R3 Pos. 0	#
20	R3 Bottle Size Small	Small
21	Calib. Type (Type)	LOGIT-LOG (4P)
22	Calib. Type (Wght)	0
23	Calib. Conc. 1	0.0
24	Calib. Pos. 1	36
25	Calib. Conc. 2	#
26	Calib. Pos. 2	##
27	Calib. Conc. 3	#
28	Calib. Pos. 3	##
29	Calib. Conc. 4	#
30	Calib. Pos. 4	##
31	Calib. Conc. 5	#
32	Calib. Pos. 5	##
33	Calib. Conc. 6	#
34	Calib. Pos. 6	##
35	S 1 ABS.	0
36	K Factor	10000
37	K 2 Factor	10000
38	K 3 Factor	10000
39	K 4 Factor	10000
40	K 5 Factor	10000
41	A Factor	0
42	B Factor	0
43	C Factor	0
44	SD Limit	999
45	Duplicate Limit	2000
46	Sens. Limit	0
47	S 1 ABS Limit (L)	-32000
48	S 1 ABS Limit (H)	32000
49	ABS Limit	0
50	ABS Limit (D/I)	INCREASE
51	Prz. Limit	0
52	Prz. Limit (U/D)	Lower
53	Prz. (End Point)	35
54	Expect. Value (L)	10
55	Expect. Value (H)	40
56	Instr. Fact. (a)	1
57	Instr. Fact. (b)	0
58	Key Setting	#

Complement 4 (C4)

IMMUNOTURB
Liquid, 2 Reagent
Calibrator : PROTEIN CALIBRATOR (CAT NO. SP131)
Control : PROTEIN Control 1 & 2

HITACHI 717	
TEST	C4
ASSAY CODE	[2 Point End] [24] - [50]
SAMPLE VOL (µL)	[6] [6]
R1 VOLUME (µL)	[250] [50] - [NO]
R2 VOLUME (µL)	[50] [20] - [NO]
WAVELENGTH	[700] [340]
CALIB. METHOD	[NONLINEAR] [1] - [6]
STD 1 CONC. POS.	[0.0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[#] - [##]
STD 4 CONC. POS.	[#] - [##]
STD 5 CONC. POS.	[#] - [##]
STD 6 CONC. POS.	[#] - [##]
UNIT	[mg/dl]
SD LIMIT	[999]
DUPLICATE LIMIT	[2000]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[0] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[10] [40]
PANIC VALUE	[0] [80]
INSTRUMENT FACTOR	[1]

HITACHI 704	
TEST	C4
ASSAY CODE	[2 Point End] [15] - [32]
SAMPLE VOL (µL)	[6] [6]
R1 VOLUME (µL)	[250] [50] - [NO]
R2 VOLUME (µL)	[50] [20] - [NO]
WAVELENGTH	[700] [340]
CALIB. METHOD	[NONLINEAR] [1] - [6]
STD 1 CONC. POS.	[0.0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[#] - [##]
STD 4 CONC. POS.	[#] - [##]
STD 5 CONC. POS.	[#] - [##]
STD 6 CONC. POS.	[#] - [##]
UNIT	[mg/dl]
SD LIMIT	[999]
DUPLICATE LIMIT	[2000]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[0] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[10] [40]
PANIC VALUE	[0] [80]
INSTRUMENT FACTOR	[1]

User defined
جایگاه کالیبراتور

HITACHI 902		
1	Test Name	CA
2	Assay Code (Mthd)	1 point
3	Assay Code (2. Test)	0
4	Reaction Time	3
5	Assay Point 1	10
6	Assay Point 2	0
7	Assay Point 3	0
8	Assay Point 4	0
9	Wave Leng. (SUB)	700
10	Wave Leng. (MAIN)	660
11	Sample Volume	3
12	R1 VOLUME	300
13	R1 Pos.	#
14	R1 Bottle Size	Large
15	R2 VOLUME	0
16	R2 Pos. 0	0
17	R2 Bottle Size	Small
18	R3 VOLUME	0
19	R3 Pos. 0	#
20	R3 Bottle Size Small	Small
21	Calib. Type (Type)	LINEAR
22	Calib. Type (Wght)	0
23	Calib. Conc. 1	0.0
24	Calib. Pos. 1	36
25	Calib. Conc. 2	#
26	Calib. Pos. 2	##
27	Calib. Conc. 3	0
28	Calib. Pos. 3	0
29	Calib. Conc. 4	0
30	Calib. Pos. 4	0
31	Calib. Conc. 5	0
32	Calib. Pos. 5	0
33	Calib. Conc. 6	0
34	Calib. Pos. 6	0
35	S 1 ABS.	0
36	K Factor	10000
37	K 2 Factor	10000
38	K 3 Factor	10000
39	K 4 Factor	10000
40	K 5 Factor	10000
41	A Factor	0
42	B Factor	0
43	C Factor	0
44	SD Limit	0.1
45	Duplicate Limit	500
46	Sens. Limit	0
47	S 1 ABS Limit (L)	-32000
48	S 1 ABS Limit (H)	32000
49	ABS Limit	0
50	ABS Limit (D/I)	INCREASE
51	Prz. Limit	0
52	Prz. Limit (U/D)	Lower
53	Prz. (End Point)	35
54	Expect. Value (L)	8.5
55	Expect. Value (H)	10.5
56	Instr. Fact. (a)	1
57	Instr. Fact. (b)	0
58	Key Setting	#

Calcium

ARSENAZO III
Liquid, 1 Reagent
Calibrator : GENERAL CALIBRATOR (CAT NO. CA532)
Control : GENERAL Control 1 & 2

HITACHI 717	
TEST	CA
ASSAY CODE	[1 point] [20] - [0]
SAMPLE VOL (µL)	[3] [3]
R1 VOLUME (µL)	[300] [50] - [NO]
R2 VOLUME (µL)	[0] [20] - [NO]
WAVELENGTH	[700] [660]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0.0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[mg/dl]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[0] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[8.5] [10.5]
PANIC VALUE	[0] [20]
INSTRUMENT FACTOR	[1]

HITACHI 704	
TEST	CA
ASSAY CODE	[1 point] [10] - [0]
SAMPLE VOL (µL)	[3] [3]
R1 VOLUME (µL)	[300] [50] - [NO]
R2 VOLUME (µL)	[0] [20] - [NO]
WAVELENGTH	[700] [660]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0.0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[mg/dl]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[0] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[8.5] [10.5]
PANIC VALUE	[0] [20]
INSTRUMENT FACTOR	[1]

User defined
جایگاه کالیبراتور

HITACHI 902		
1	Test Name	CHOL
2	Assay Code (Mthd)	1 point
3	Assay Code (2. Test)	0
4	Reaction Time	10
5	Assay Point 1	35
6	Assay Point 2	0
7	Assay Point 3	0
8	Assay Point 4	0
9	Wave Leng. (SUB)	700
10	Wave Leng. (MAIN)	505
11	Sample Volume	3
12	R1 VOLUME	300
13	R1 Pos.	#
14	R1 Bottle Size	Large
15	R2 VOLUME	0
16	R2 Pos. 0	0
17	R2 Bottle Size	Small
18	R3 VOLUME	0
19	R3 Pos. 0	#
20	R3 Bottle Size Small	Small
21	Calib. Type (Type)	LINEAR
22	Calib. Type (Wght)	0
23	Calib. Conc. 1	0
24	Calib. Pos. 1	36
25	Calib. Conc. 2	#
26	Calib. Pos. 2	##
27	Calib. Conc. 3	0
28	Calib. Pos. 3	0
29	Calib. Conc. 4	0
30	Calib. Pos. 4	0
31	Calib. Conc. 5	0
32	Calib. Pos. 5	0
33	Calib. Conc. 6	0
34	Calib. Pos. 6	0
35	S 1 ABS.	0
36	K Factor	10000
37	K 2 Factor	10000
38	K 3 Factor	10000
39	K 4 Factor	10000
40	K 5 Factor	10000
41	A Factor	0
42	B Factor	0
43	C Factor	0
44	SD Limit	0.1
45	Duplicate Limit	500
46	Sens. Limit	0
47	S 1 ABS Limit (L)	-32000
48	S 1 ABS Limit (H)	32000
49	ABS Limit	0
50	ABS Limit (D/I)	INCREASE
51	Prz. Limit	0
52	Prz. Limit (U/D)	Lower
53	Prz. (End Point)	35
54	Expect. Value (L)	0
55	Expect. Value (H)	200
56	Instr. Fact. (a)	1
57	Instr. Fact. (b)	0
58	Key Setting	#

Cholesterol

CHOD-PO
Liquid, 1 Reagent
Calibrator : GENERAL CALIBRATOR (CAT NO. CA532)
Control : GENERAL Control 1 & 2

HITACHI 717	
TEST	[CHOL]
ASSAY CODE	[1 point] [50] - [0]
SAMPLE VOL (µL)	[3] [3]
R1 VOLUME (µL)	[300] [50] - [NO]
R2 VOLUME (µL)	[0] [20] - [NO]
WAVELENGTH	[700] [505]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[mg/dl]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[0] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[0] [200]
PANIC VALUE	[0] [600]
INSTRUMENT FACTOR	[1]

HITACHI 704	
TEST	[CHOL]
ASSAY CODE	[1 point] [32] - [0]
SAMPLE VOL (µL)	[3] [3]
R1 VOLUME (µL)	[300] [50] - [NO]
R2 VOLUME (µL)	[0] [20] - [NO]
WAVELENGTH	[700] [505]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[mg/dl]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[0] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[0] [200]
PANIC VALUE	[0] [600]
INSTRUMENT FACTOR	[1]

User defined
جایگاه کالیبراتور ##

HITACHI 902		
1	Test Name	CK-MB
2	Assay Code (Mthd)	Rate A
3	Assay Code (2. Test)	0
4	Reaction Time	10
5	Assay Point 1	24
6	Assay Point 2	35
7	Assay Point 3	0
8	Assay Point 4	0
9	Wave Leng. (SUB)	415
10	Wave Leng. (MAIN)	340
11	Sample Volume	15
12	R1 VOLUME	250
13	R1 Pos.	#
14	R1 Bottle Size	Large
15	R2 VOLUME	0
16	R2 Pos. 0	0
17	R2 Bottle Size	Small
18	R3 VOLUME	50
19	R3 Pos. 0	#
20	R3 Bottle Size Small	Small
21	Calib. Type (Type)	LINEAR
22	Calib. Type (Wght)	0
23	Calib. Conc. 1	0.0
24	Calib. Pos. 1	36
25	Calib. Conc. 2	#
26	Calib. Pos. 2	##
27	Calib. Conc. 3	0
28	Calib. Pos. 3	0
29	Calib. Conc. 4	0
30	Calib. Pos. 4	0
31	Calib. Conc. 5	0
32	Calib. Pos. 5	0
33	Calib. Conc. 6	0
34	Calib. Pos. 6	0
35	S 1 ABS.	0
36	K Factor	10000
37	K 2 Factor	10000
38	K 3 Factor	10000
39	K 4 Factor	10000
40	K 5 Factor	10000
41	A Factor	0
42	B Factor	0
43	C Factor	0
44	SD Limit	0.1
45	Duplicate Limit	500
46	Sens. Limit	0
47	S 1 ABS Limit (L)	-32000
48	S 1 ABS Limit (H)	32000
49	ABS Limit	10000
50	ABS Limit (D/I)	INCREASE
51	Prz. Limit	0
52	Prz. Limit (U/D)	Lower
53	Prz. (End Point)	35
54	Expect. Value (L)	0
55	Expect. Value (H)	24
56	Instr. Fact. (a)	1
57	Instr. Fact. (b)	0
58	Key Setting	#

CK-MB

Immunoinhibition
Liquid, 2 Reagent
Calibrator : CK-MB CALIBRATOR (CK231)
Control : CK-MB CONTROL (CK211)

HITACHI 717	
TEST	[CK-MB]
ASSAY CODE	[Rate A] [35] - [50]
SAMPLE VOL (µL)	[15] [15]
R1 VOLUME (µL)	[250] [50] - [NO]
R2 VOLUME (µL)	[50] [20] - [NO]
WAVELENGTH	[415] [340]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0.0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[U/L]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[10000] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[0] [24]
PANIC VALUE	[0] [1000]
INSTRUMENT FACTOR	[1]

HITACHI 704	
TEST	[CK-MB]
ASSAY CODE	[Rate A] [22] - [32]
SAMPLE VOL (µL)	[15] [15]
R1 VOLUME (µL)	[250] [50] - [NO]
R2 VOLUME (µL)	[50] [20] - [NO]
WAVELENGTH	[415] [340]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0.0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[U/L]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[10000] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[0] [24]
PANIC VALUE	[0] [1000]
INSTRUMENT FACTOR	[1]

User defined
جایگاه کالیبراتور

HITACHI 902		
1	Test Name	CPK
2	Assay Code (Mthd)	Rate A
3	Assay Code (2. Test)	0
4	Reaction Time	10
5	Assay Point 1	24
6	Assay Point 2	35
7	Assay Point 3	0
8	Assay Point 4	0
9	Wave Leng. (SUB)	415
10	Wave Leng. (MAIN)	340
11	Sample Volume	15
12	R1 VOLUME	250
13	R1 Pos.	#
14	R1 Bottle Size	Large
15	R2 VOLUME	0
16	R2 Pos. 0	0
17	R2 Bottle Size	Small
18	R3 VOLUME	50
19	R3 Pos. 0	#
20	R3 Bottle Size Small	Small
21	Calib. Type (Type)	LINEAR
22	Calib. Type (Wght)	0
23	Calib. Conc. 1	0
24	Calib. Pos. 1	36
25	Calib. Conc. 2	#
26	Calib. Pos. 2	##
27	Calib. Conc. 3	0
28	Calib. Pos. 3	0
29	Calib. Conc. 4	0
30	Calib. Pos. 4	0
31	Calib. Conc. 5	0
32	Calib. Pos. 5	0
33	Calib. Conc. 6	0
34	Calib. Pos. 6	0
35	S 1 ABS.	0
36	K Factor	10000
37	K 2 Factor	10000
38	K 3 Factor	10000
39	K 4 Factor	10000
40	K 5 Factor	10000
41	A Factor	0
42	B Factor	0
43	C Factor	0
44	SD Limit	0.1
45	Duplicate Limit	500
46	Sens. Limit	0
47	S 1 ABS Limit (L)	-32000
48	S 1 ABS Limit (H)	32000
49	ABS Limit	10000
50	ABS Limit (D/I)	INCREASE
51	Prz. Limit	0
52	Prz. Limit (U/D)	Lower
53	Prz. (End Point)	35
54	Expect. Value (L)	24
55	Expect. Value (H)	190
56	Instr. Fact. (a)	1
57	Instr. Fact. (b)	0
58	Key Setting	#

CK-NAC

IFCC
Liquid, 2 Reagent
Calibrator : GENERAL CALIBRATOR (CAT NO. CA532)
Control : GENERAL Control 1 & 2

HITACHI 717	
TEST	[CPK]
ASSAY CODE	[Rate A] [35] - [50]
SAMPLE VOL (µL)	[15] [15]
R1 VOLUME (µL)	[250] [50] - [NO]
R2 VOLUME (µL)	[50] [20] - [NO]
WAVELENGTH	[415] [340]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[U/L]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[10000] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[24] [190]
PANIC VALUE	[0] [1000]
INSTRUMENT FACTOR	[1]

HITACHI 704	
TEST	[CPK]
ASSAY CODE	[Rate A] [22] - [32]
SAMPLE VOL (µL)	[15] [15]
R1 VOLUME (µL)	[250] [50] - [NO]
R2 VOLUME (µL)	[50] [20] - [NO]
WAVELENGTH	[415] [340]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[U/L]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[10000] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[24] [190]
PANIC VALUE	[0] [1000]
INSTRUMENT FACTOR	[1]

User defined
جایگاه کالیبراتور ##

HITACHI 902		
1	Test Name	CREAT
2	Assay Code (Mthd)	2 point rate
3	Assay Code (2. Test)	0
4	Reaction Time	4
5	Assay Point 1	9
6	Assay Point 2	12
7	Assay Point 3	0
8	Assay Point 4	0
9	Wave Leng. (SUB)	660
10	Wave Leng. (MAIN)	505
11	Sample Volume	20
12	R1 VOLUME	120
13	R1 Pos.	#
14	R1 Bottle Size	Large
15	R2 VOLUME	120
16	R2 Pos. 0	0
17	R2 Bottle Size	Small
18	R3 VOLUME	0
19	R3 Pos. 0	#
20	R3 Bottle Size Small	Small
21	Calib. Type (Type)	LINEAR
22	Calib. Type (Wght)	0
23	Calib. Conc. 1	0.00
24	Calib. Pos. 1	36
25	Calib. Conc. 2	#
26	Calib. Pos. 2	##
27	Calib. Conc. 3	0
28	Calib. Pos. 3	0
29	Calib. Conc. 4	0
30	Calib. Pos. 4	0
31	Calib. Conc. 5	0
32	Calib. Pos. 5	0
33	Calib. Conc. 6	0
34	Calib. Pos. 6	0
35	S 1 ABS.	0
36	K Factor	10000
37	K 2 Factor	10000
38	K 3 Factor	10000
39	K 4 Factor	10000
40	K 5 Factor	10000
41	A Factor	0
42	B Factor	0
43	C Factor	0
44	SD Limit	0.1
45	Duplicate Limit	500
46	Sens. Limit	0
47	S 1 ABS Limit (L)	-32000
48	S 1 ABS Limit (H)	32000
49	ABS Limit	32000
50	ABS Limit (D/I)	INCREASE
51	Prz. Limit	0
52	Prz. Limit (U/D)	Lower
53	Prz. (End Point)	35
54	Expect. Value (L)	0.7
55	Expect. Value (H)	1.4
56	Instr. Fact. (a)	1
57	Instr. Fact. (b)	0
58	Key Setting	#

Creatinine (Jaffe)

Jaffe
Liquid, 2 Reagent
Calibrator : GENERAL CALIBRATOR (CAT NO. CA532)
Control : GENERAL Control 1 & 2

HITACHI 717	
TEST	[CREAT]
ASSAY CODE	[2 point rate] [28] - [32]
SAMPLE VOL (µL)	[20] [20]
R1 VOLUME (µL)	[150] [50] - [NO]
R2 VOLUME (µL)	[150] [20] - [NO]
WAVELENGTH	[660] [505]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0.00] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[mg/dl]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[32000] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[0.7] [1.4]
PANIC VALUE	[0.2] [15]
INSTRUMENT FACTOR	[1]

HITACHI 704	
TEST	[CREAT]
ASSAY CODE	[2 point rate] [18] - [21]
SAMPLE VOL (µL)	[20] [20]
R1 VOLUME (µL)	[150] [50] - [NO]
R2 VOLUME (µL)	[150] [20] - [NO]
WAVELENGTH	[660] [505]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0.00] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[mg/dl]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[32000] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[0.7] [1.4]
PANIC VALUE	[0.2] [15]
INSTRUMENT FACTOR	[1]

User defined
جایگاه کالیبراتور ##

HITACHI 902		
1	Test Name	CRP
2	Assay Code (Mthd)	2 Point End
3	Assay Code (2. Test)	0
4	Reaction Time	10
5	Assay Point 1	16
6	Assay Point 2	35
7	Assay Point 3	0
8	Assay Point 4	0
9	Wave Leng. (SUB)	700
10	Wave Leng. (MAIN)	340
11	Sample Volume	15
12	R1 VOLUME	260
13	R1 Pos.	#
14	R1 Bottle Size	Large
15	R2 VOLUME	0
16	R2 Pos. 0	0
17	R2 Bottle Size	Small
18	R3 VOLUME	50
19	R3 Pos. 0	#
20	R3 Bottle Size Small	Small
21	Calib. Type (Type)	Spline
22	Calib. Type (Wght)	0
23	Calib. Conc. 1	0.0
24	Calib. Pos. 1	36
25	Calib. Conc. 2	#
26	Calib. Pos. 2	##
27	Calib. Conc. 3	#
28	Calib. Pos. 3	##
29	Calib. Conc. 4	#
30	Calib. Pos. 4	##
31	Calib. Conc. 5	#
32	Calib. Pos. 5	##
33	Calib. Conc. 6	#
34	Calib. Pos. 6	##
35	S 1 ABS.	0
36	K Factor	10000
37	K 2 Factor	10000
38	K 3 Factor	10000
39	K 4 Factor	10000
40	K 5 Factor	10000
41	A Factor	0
42	B Factor	0
43	C Factor	0
44	SD Limit	999
45	Duplicate Limit	2000
46	Sens. Limit	0
47	S 1 ABS Limit (L)	-32000
48	S 1 ABS Limit (H)	32000
49	ABS Limit	0
50	ABS Limit (D/I)	INCREASE
51	Prz. Limit	0
52	Prz. Limit (U/D)	Lower
53	Prz. (End Point)	35
54	Expect. Value (L)	0
55	Expect. Value (H)	10
56	Instr. Fact. (a)	1
57	Instr. Fact. (b)	0
58	Key Setting	#

C-REACTIVE PROTEIN

IMMUNOTURB
Liquid, 2 Reagent
Calibrator : CRP CALIBRATOR (CAT NO. CP131)
Control : CRP Control 1 & 2

HITACHI 717			
TEST	[CRP]		
ASSAY CODE	[2 Point End]	[24] - [50]	
SAMPLE VOL (µL)	[15]	[15]	
R1 VOLUME (µL)	[260]	[50] - [NO]	
R2 VOLUME (µL)	[50]	[20] - [NO]	
WAVELENGTH	[700]	[340]	
CALIB. METHOD	[NONLINEAR]	[1] - [6]	
STD 1 CONC. POS.	[0.0] - [##]		
STD 2 CONC. POS.	[#] - [##]		
STD 3 CONC. POS.	[#] - [##]		
STD 4 CONC. POS.	[#] - [##]		
STD 5 CONC. POS.	[#] - [##]		
STD 6 CONC. POS.	[#] - [##]		
UNIT	[mg/l]		
SD LIMIT	[999]		
DUPLICATE LIMIT	[2000]		
SENSITIVITY LIMIT	[0]		
ABS. LIMIT (INCR./DECR.)	[0]	[INCREASE]	
PROZON LIMIT	[0]	[LOWER]	
EXPECTED VALUE	[0]	[10]	
PANIC VALUE	[0]	[180]	
INSTRUMENT FACTOR	[1]		

HITACHI 704			
TEST	[CRP]		
ASSAY CODE	[2 Point End]	[15] - [32]	
SAMPLE VOL (µL)	[15]	[15]	
R1 VOLUME (µL)	[260]	[50] - [NO]	
R2 VOLUME (µL)	[50]	[20] - [NO]	
WAVELENGTH	[700]	[340]	
CALIB. METHOD	[NONLINEAR]	[1] - [6]	
STD 1 CONC. POS.	[0.0] - [##]		
STD 2 CONC. POS.	[#] - [##]		
STD 3 CONC. POS.	[#] - [##]		
STD 4 CONC. POS.	[#] - [##]		
STD 5 CONC. POS.	[#] - [##]		
STD 6 CONC. POS.	[#] - [##]		
UNIT	[mg/l]		
SD LIMIT	[999]		
DUPLICATE LIMIT	[2000]		
SENSITIVITY LIMIT	[0]		
ABS. LIMIT (INCR./DECR.)	[0]	[INCREASE]	
PROZON LIMIT	[0]	[LOWER]	
EXPECTED VALUE	[0]	[10]	
PANIC VALUE	[0]	[180]	
INSTRUMENT FACTOR	[1]		

User defined
جایگاه کالیبراتور

HITACHI 902		
1	Test Name	GLU
2	Assay Code (Mthd)	1 point
3	Assay Code (2. Test)	0
4	Reaction Time	10
5	Assay Point 1	35
6	Assay Point 2	0
7	Assay Point 3	0
8	Assay Point 4	0
9	Wave Leng. (SUB)	700
10	Wave Leng. (MAIN)	505
11	Sample Volume	3
12	R1 VOLUME	300
13	R1 Pos.	#
14	R1 Bottle Size	Large
15	R2 VOLUME	0
16	R2 Pos. 0	0
17	R2 Bottle Size	Small
18	R3 VOLUME	0
19	R3 Pos. 0	#
20	R3 Bottle Size Small	Small
21	Calib. Type (Type)	LINEAR
22	Calib. Type (Wght)	0
23	Calib. Conc. 1	0
24	Calib. Pos. 1	36
25	Calib. Conc. 2	#
26	Calib. Pos. 2	##
27	Calib. Conc. 3	0
28	Calib. Pos. 3	0
29	Calib. Conc. 4	0
30	Calib. Pos. 4	0
31	Calib. Conc. 5	0
32	Calib. Pos. 5	0
33	Calib. Conc. 6	0
34	Calib. Pos. 6	0
35	S 1 ABS.	0
36	K Factor	10000
37	K 2 Factor	10000
38	K 3 Factor	10000
39	K 4 Factor	10000
40	K 5 Factor	10000
41	A Factor	0
42	B Factor	0
43	C Factor	0
44	SD Limit	0.1
45	Duplicate Limit	500
46	Sens. Limit	0
47	S 1 ABS Limit (L)	-32000
48	S 1 ABS Limit (H)	32000
49	ABS Limit	0
50	ABS Limit (D/I)	INCREASE
51	Prz. Limit	0
52	Prz. Limit (U/D)	Lower
53	Prz. (End Point)	35
54	Expect. Value (L)	70
55	Expect. Value (H)	110
56	Instr. Fact. (a)	1
57	Instr. Fact. (b)	0
58	Key Setting	#

Glucose

PAP
Liquid, 1 Reagent
Calibrator : GENERAL CALIBRATOR (CAT NO. CA532)
Control : GENERAL Control 1 & 2

HITACHI 717	
TEST	GLU
ASSAY CODE	[1 point] [50] - [0]
SAMPLE VOL (µL)	[3] [3]
R1 VOLUME (µL)	[300] [50] - [NO]
R2 VOLUME (µL)	[0] [20] - [NO]
WAVELENGTH	[700] [505]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[mg/dl]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[0] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[70] [110]
PANIC VALUE	[5] [400]
INSTRUMENT FACTOR	[1]

HITACHI 704	
TEST	GLU
ASSAY CODE	[1 point] [32] - [0]
SAMPLE VOL (µL)	[3] [3]
R1 VOLUME (µL)	[300] [50] - [NO]
R2 VOLUME (µL)	[0] [20] - [NO]
WAVELENGTH	[700] [505]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[mg/dl]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[0] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[70] [110]
PANIC VALUE	[5] [400]
INSTRUMENT FACTOR	[1]

User defined
جایگاه کالیبراتور

HITACHI 902		
1	Test Name	HDL-C
2	Assay Code (Mthd)	2 Point End
3	Assay Code (2. Test)	0
4	Reaction Time	10
5	Assay Point 1	16
6	Assay Point 2	35
7	Assay Point 3	0
8	Assay Point 4	0
9	Wave Leng. (SUB)	700
10	Wave Leng. (MAIN)	600
11	Sample Volume	3
12	R1 VOLUME	240
13	R1 Pos.	#
14	R1 Bottle Size	Large
15	R2 VOLUME	0
16	R2 Pos. 0	0
17	R2 Bottle Size	Small
18	R3 VOLUME	80
19	R3 Pos. 0	#
20	R3 Bottle Size Small	Small
21	Calib. Type (Type)	LINEAR
22	Calib. Type (Wght)	0
23	Calib. Conc. 1	0
24	Calib. Pos. 1	36
25	Calib. Conc. 2	#
26	Calib. Pos. 2	##
27	Calib. Conc. 3	0
28	Calib. Pos. 3	0
29	Calib. Conc. 4	0
30	Calib. Pos. 4	0
31	Calib. Conc. 5	0
32	Calib. Pos. 5	0
33	Calib. Conc. 6	0
34	Calib. Pos. 6	0
35	S 1 ABS.	0
36	K Factor	10000
37	K 2 Factor	10000
38	K 3 Factor	10000
39	K 4 Factor	10000
40	K 5 Factor	10000
41	A Factor	0
42	B Factor	0
43	C Factor	0
44	SD Limit	0.1
45	Duplicate Limit	500
46	Sens. Limit	0
47	S 1 ABS Limit (L)	-32000
48	S 1 ABS Limit (H)	32000
49	ABS Limit	0
50	ABS Limit (D/I)	INCREASE
51	Prz. Limit	0
52	Prz. Limit (U/D)	Lower
53	Prz. (End Point)	35
54	Expect. Value (L)	30
55	Expect. Value (H)	90
56	Instr. Fact. (a)	1
57	Instr. Fact. (b)	0
58	Key Setting	#

HDL Cholesterol Direct

Direct Enzymatic
Liquid, 2 Reagent
Calibrator : HDL/LDL CALIBRATOR (CAT NO. HL132)
Control : General Control 1 & 2

HITACHI 717	
TEST	HDL-C
ASSAY CODE	[2 Point End] [24] - [50]
SAMPLE VOL (µL)	[3] [3]
R1 VOLUME (µL)	[240] [50] - [NO]
R2 VOLUME (µL)	[80] [20] - [NO]
WAVELENGTH	[700] [600]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[mg/dl]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[0] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[30] [90]
PANIC VALUE	[0] [120]
INSTRUMENT FACTOR	[1]

HITACHI 704	
TEST	HDL-C
ASSAY CODE	[2 Point End] [15] - [32]
SAMPLE VOL (µL)	[3] [3]
R1 VOLUME (µL)	[240] [50] - [NO]
R2 VOLUME (µL)	[80] [20] - [NO]
WAVELENGTH	[700] [600]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[mg/dl]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[0] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[30] [90]
PANIC VALUE	[0] [120]
INSTRUMENT FACTOR	[1]

User defined
جایگاه کالیبراتور ##

HITACHI 902		
1	Test Name	IgA
2	Assay Code (Mthd)	2 Point End
3	Assay Code (2. Test)	0
4	Reaction Time	10
5	Assay Point 1	16
6	Assay Point 2	35
7	Assay Point 3	0
8	Assay Point 4	0
9	Wave Leng. (SUB)	700
10	Wave Leng. (MAIN)	340
11	Sample Volume	3
12	R1 VOLUME	250
13	R1 Pos.	#
14	R1 Bottle Size	Large
15	R2 VOLUME	0
16	R2 Pos. 0	0
17	R2 Bottle Size	Small
18	R3 VOLUME	50
19	R3 Pos. 0	#
20	R3 Bottle Size Small	Small
21	Calib. Type (Type)	LOGIT-LOG (4P)
22	Calib. Type (Wght)	0
23	Calib. Conc. 1	0
24	Calib. Pos. 1	36
25	Calib. Conc. 2	#
26	Calib. Pos. 2	##
27	Calib. Conc. 3	#
28	Calib. Pos. 3	##
29	Calib. Conc. 4	#
30	Calib. Pos. 4	##
31	Calib. Conc. 5	#
32	Calib. Pos. 5	##
33	Calib. Conc. 6	#
34	Calib. Pos. 6	##
35	S 1 ABS.	0
36	K Factor	10000
37	K 2 Factor	10000
38	K 3 Factor	10000
39	K 4 Factor	10000
40	K 5 Factor	10000
41	A Factor	0
42	B Factor	0
43	C Factor	0
44	SD Limit	999
45	Duplicate Limit	2000
46	Sens. Limit	0
47	S 1 ABS Limit (L)	-32000
48	S 1 ABS Limit (H)	32000
49	ABS Limit	0
50	ABS Limit (D/I)	INCREASE
51	Prz. Limit	0
52	Prz. Limit (U/D)	Lower
53	Prz. (End Point)	35
54	Expect. Value (L)	70
55	Expect. Value (H)	400
56	Instr. Fact. (a)	1
57	Instr. Fact. (b)	0
58	Key Setting	#

Immunoglobulin A (IgA)

IMMUNOTURB
Liquid, 2 Reagent
Calibrator : PROTEIN CALIBRATOR (CAT NO. SP131)
Control : PROTEIN Control 1 & 2

HITACHI 717	
TEST	IgA
ASSAY CODE	[2 Point End] [24] - [50]
SAMPLE VOL (µL)	[3] [3]
R1 VOLUME (µL)	[250] [50] - [NO]
R2 VOLUME (µL)	[50] [20] - [NO]
WAVELENGTH	[700] [340]
CALIB. METHOD	[NONLINEAR] [1] - [6]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[#] - [##]
STD 4 CONC. POS.	[#] - [##]
STD 5 CONC. POS.	[#] - [##]
STD 6 CONC. POS.	[#] - [##]
UNIT	[mg/dl]
SD LIMIT	[999]
DUPLICATE LIMIT	[2000]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[0] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[70] [400]
PANIC VALUE	[0] [800]
INSTRUMENT FACTOR	[1]

HITACHI 704	
TEST	IgA
ASSAY CODE	[2 Point End] [15] - [32]
SAMPLE VOL (µL)	[3] [3]
R1 VOLUME (µL)	[250] [50] - [NO]
R2 VOLUME (µL)	[50] [20] - [NO]
WAVELENGTH	[700] [340]
CALIB. METHOD	[NONLINEAR] [1] - [6]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[#] - [##]
STD 4 CONC. POS.	[#] - [##]
STD 5 CONC. POS.	[#] - [##]
STD 6 CONC. POS.	[#] - [##]
UNIT	[mg/dl]
SD LIMIT	[999]
DUPLICATE LIMIT	[2000]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[0] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[70] [400]
PANIC VALUE	[0] [800]
INSTRUMENT FACTOR	[1]

User defined
جایگاه کالیبراتور

HITACHI 902		
1	Test Name	IgG
2	Assay Code (Mthd)	2 Point End
3	Assay Code (2. Test)	0
4	Reaction Time	10
5	Assay Point 1	16
6	Assay Point 2	35
7	Assay Point 3	0
8	Assay Point 4	0
9	Wave Leng. (SUB)	Cancel
10	Wave Leng. (MAIN)	600
11	Sample Volume	3
12	R1 VOLUME	250
13	R1 Pos.	#
14	R1 Bottle Size	Large
15	R2 VOLUME	0
16	R2 Pos. 0	0
17	R2 Bottle Size	Small
18	R3 VOLUME	50
19	R3 Pos. 0	#
20	R3 Bottle Size Small	Small
21	Calib. Type (Type)	LOGIT-LOG (4P)
22	Calib. Type (Wght)	0
23	Calib. Conc. 1	0
24	Calib. Pos. 1	36
25	Calib. Conc. 2	#
26	Calib. Pos. 2	##
27	Calib. Conc. 3	#
28	Calib. Pos. 3	##
29	Calib. Conc. 4	#
30	Calib. Pos. 4	##
31	Calib. Conc. 5	#
32	Calib. Pos. 5	##
33	Calib. Conc. 6	#
34	Calib. Pos. 6	##
35	S 1 ABS.	0
36	K Factor	10000
37	K 2 Factor	10000
38	K 3 Factor	10000
39	K 4 Factor	10000
40	K 5 Factor	10000
41	A Factor	0
42	B Factor	0
43	C Factor	0
44	SD Limit	999
45	Duplicate Limit	2000
46	Sens. Limit	0
47	S 1 ABS Limit (L)	-32000
48	S 1 ABS Limit (H)	32000
49	ABS Limit	0
50	ABS Limit (D/I)	INCREASE
51	Prz. Limit	0
52	Prz. Limit (U/D)	Lower
53	Prz. (End Point)	35
54	Expect. Value (L)	700
55	Expect. Value (H)	1600
56	Instr. Fact. (a)	1
57	Instr. Fact. (b)	0
58	Key Setting	#

Immunoglobulin G (IgG)

IMMUNOTURB
Liquid, 2 Reagent
Calibrator : PROTEIN CALIBRATOR (CAT NO. SP131)
Control : PROTEIN Control 1 & 2

HITACHI 717	
TEST	IgG
ASSAY CODE	[2 Point End] [24] - [50]
SAMPLE VOL (µL)	[3] [3]
R1 VOLUME (µL)	[250] [50] - [NO]
R2 VOLUME (µL)	[50] [20] - [NO]
WAVELENGTH	[Cance] [600]
CALIB. METHOD	[NONLINEAR] [1] - [6]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[#] - [##]
STD 4 CONC. POS.	[#] - [##]
STD 5 CONC. POS.	[#] - [##]
STD 6 CONC. POS.	[#] - [##]
UNIT	[mg/dl]
SD LIMIT	[999]
DUPLICATE LIMIT	[2000]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[0] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[700] [1600]
PANIC VALUE	[100] [3800]
INSTRUMENT FACTOR	[1]

HITACHI 704	
TEST	IgG
ASSAY CODE	[2 Point End] [15] - [32]
SAMPLE VOL (µL)	[3] [3]
R1 VOLUME (µL)	[250] [50] - [NO]
R2 VOLUME (µL)	[50] [20] - [NO]
WAVELENGTH	[Cance] [600]
CALIB. METHOD	[NONLINEAR] [1] - [6]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[#] - [##]
STD 4 CONC. POS.	[#] - [##]
STD 5 CONC. POS.	[#] - [##]
STD 6 CONC. POS.	[#] - [##]
UNIT	[mg/dl]
SD LIMIT	[999]
DUPLICATE LIMIT	[2000]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[0] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[700] [1600]
PANIC VALUE	[100] [3800]
INSTRUMENT FACTOR	[1]

User defined
جایگاه کالیبراتور

HITACHI 902		
1	Test Name	IgM
2	Assay Code (Mthd)	2 Point End
3	Assay Code (2. Test)	0
4	Reaction Time	10
5	Assay Point 1	16
6	Assay Point 2	35
7	Assay Point 3	0
8	Assay Point 4	0
9	Wave Leng. (SUB)	700
10	Wave Leng. (MAIN)	340
11	Sample Volume	3
12	R1 VOLUME	250
13	R1 Pos.	#
14	R1 Bottle Size	Large
15	R2 VOLUME	0
16	R2 Pos. 0	0
17	R2 Bottle Size	Small
18	R3 VOLUME	50
19	R3 Pos. 0	#
20	R3 Bottle Size Small	Small
21	Calib. Type (Type)	LOGIT-LOG (4P)
22	Calib. Type (Wght)	0
23	Calib. Conc. 1	0
24	Calib. Pos. 1	36
25	Calib. Conc. 2	#
26	Calib. Pos. 2	##
27	Calib. Conc. 3	#
28	Calib. Pos. 3	##
29	Calib. Conc. 4	#
30	Calib. Pos. 4	##
31	Calib. Conc. 5	#
32	Calib. Pos. 5	##
33	Calib. Conc. 6	#
34	Calib. Pos. 6	##
35	S 1 ABS.	0
36	K Factor	10000
37	K 2 Factor	10000
38	K 3 Factor	10000
39	K 4 Factor	10000
40	K 5 Factor	10000
41	A Factor	0
42	B Factor	0
43	C Factor	0
44	SD Limit	999
45	Duplicate Limit	2000
46	Sens. Limit	0
47	S 1 ABS Limit (L)	-32000
48	S 1 ABS Limit (H)	32000
49	ABS Limit	0
50	ABS Limit (D/I)	INCREASE
51	Prz. Limit	0
52	Prz. Limit (U/D)	Lower
53	Prz. (End Point)	35
54	Expect. Value (L)	40
55	Expect. Value (H)	230
56	Instr. Fact. (a)	1
57	Instr. Fact. (b)	0
58	Key Setting	#

Immunoglobulin M (IgM)

IMMUNOTURB
Liquid, 2 Reagent
Calibrator : PROTEIN CALIBRATOR (CAT NO. SP131)
Control : PROTEIN Control 1 & 2

HITACHI 717	
TEST	IgM
ASSAY CODE	[2 Point End] [24] - [50]
SAMPLE VOL (µL)	[3] [3]
R1 VOLUME (µL)	[250] [50] - [NO]
R2 VOLUME (µL)	[50] [20] - [NO]
WAVELENGTH	[700] [340]
CALIB. METHOD	[NONLINEAR] [1] - [6]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[#] - [##]
STD 4 CONC. POS.	[#] - [##]
STD 5 CONC. POS.	[#] - [##]
STD 6 CONC. POS.	[#] - [##]
UNIT	[mg/dl]
SD LIMIT	[999]
DUPLICATE LIMIT	[2000]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[0] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[40] [230]
PANIC VALUE	[0] [400]
INSTRUMENT FACTOR	[1]

HITACHI 704	
TEST	IgM
ASSAY CODE	[2 Point End] [15] - [32]
SAMPLE VOL (µL)	[3] [3]
R1 VOLUME (µL)	[250] [50] - [NO]
R2 VOLUME (µL)	[50] [20] - [NO]
WAVELENGTH	[700] [340]
CALIB. METHOD	[NONLINEAR] [1] - [6]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[#] - [##]
STD 4 CONC. POS.	[#] - [##]
STD 5 CONC. POS.	[#] - [##]
STD 6 CONC. POS.	[#] - [##]
UNIT	[mg/dl]
SD LIMIT	[999]
DUPLICATE LIMIT	[2000]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[0] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[40] [230]
PANIC VALUE	[0] [400]
INSTRUMENT FACTOR	[1]

User defined
جایگاه کالیبراتور

HITACHI 902		
1	Test Name	LDH
2	Assay Code (Mthd)	Rate A
3	Assay Code (2. Test)	0
4	Reaction Time	10
5	Assay Point 1	21
6	Assay Point 2	31
7	Assay Point 3	0
8	Assay Point 4	0
9	Wave Leng. (SUB)	415
10	Wave Leng. (MAIN)	340
11	Sample Volume	5
12	R1 VOLUME	250
13	R1 Pos.	#
14	R1 Bottle Size	Large
15	R2 VOLUME	0
16	R2 Pos. 0	0
17	R2 Bottle Size	Small
18	R3 VOLUME	50
19	R3 Pos. 0	#
20	R3 Bottle Size Small	Small
21	Calib. Type (Type)	LINEAR
22	Calib. Type (Wght)	0
23	Calib. Conc. 1	0
24	Calib. Pos. 1	36
25	Calib. Conc. 2	#
26	Calib. Pos. 2	##
27	Calib. Conc. 3	0
28	Calib. Pos. 3	0
29	Calib. Conc. 4	0
30	Calib. Pos. 4	0
31	Calib. Conc. 5	0
32	Calib. Pos. 5	0
33	Calib. Conc. 6	0
34	Calib. Pos. 6	0
35	S 1 ABS.	0
36	K Factor	10000
37	K 2 Factor	10000
38	K 3 Factor	10000
39	K 4 Factor	10000
40	K 5 Factor	10000
41	A Factor	0
42	B Factor	0
43	C Factor	0
44	SD Limit	0.1
45	Duplicate Limit	500
46	Sens. Limit	0
47	S 1 ABS Limit (L)	-32000
48	S 1 ABS Limit (H)	32000
49	ABS Limit	5000
50	ABS Limit (D/I)	DECREASE
51	Prz. Limit	0
52	Prz. Limit (U/D)	Lower
53	Prz. (End Point)	35
54	Expect. Value (L)	200
55	Expect. Value (H)	400
56	Instr. Fact. (a)	1
57	Instr. Fact. (b)	0
58	Key Setting	#

Lactate Dehydrogenase

SCE
Liquid, 2 Reagent
Calibrator : GENERAL CALIBRATOR (CAT NO. CA532)
Control : GENERAL Control 1 & 2

HITACHI 717	
TEST	[LDH]
ASSAY CODE	[Rate A] [30] - [45]
SAMPLE VOL (µL)	[5] [5]
R1 VOLUME (µL)	[250] [50] - [NO]
R2 VOLUME (µL)	[50] [20] - [NO]
WAVELENGTH	[415] [340]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[U/L]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[5000] [DECREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[200] [400]
PANIC VALUE	[0] [1000]
INSTRUMENT FACTOR	[1]

HITACHI 704	
TEST	[LDH]
ASSAY CODE	[Rate A] [19] - [30]
SAMPLE VOL (µL)	[5] [5]
R1 VOLUME (µL)	[250] [50] - [NO]
R2 VOLUME (µL)	[50] [20] - [NO]
WAVELENGTH	[415] [340]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[U/L]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[5000] [DECREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[200] [400]
PANIC VALUE	[0] [1000]
INSTRUMENT FACTOR	[1]

User defined
جایگاه کالیبراتور

HITACHI 902		
1	Test Name	LDL-C
2	Assay Code (Mthd)	2 Point End
3	Assay Code (2. Test)	0
4	Reaction Time	10
5	Assay Point 1	16
6	Assay Point 2	35
7	Assay Point 3	0
8	Assay Point 4	0
9	Wave Leng. (SUB)	700
10	Wave Leng. (MAIN)	546
11	Sample Volume	4
12	R1 VOLUME	240
13	R1 Pos.	#
14	R1 Bottle Size	Large
15	R2 VOLUME	0
16	R2 Pos. 0	0
17	R2 Bottle Size	Small
18	R3 VOLUME	80
19	R3 Pos. 0	#
20	R3 Bottle Size Small	Small
21	Calib. Type (Type)	LINEAR
22	Calib. Type (Wght)	0
23	Calib. Conc. 1	0
24	Calib. Pos. 1	36
25	Calib. Conc. 2	#
26	Calib. Pos. 2	##
27	Calib. Conc. 3	0
28	Calib. Pos. 3	0
29	Calib. Conc. 4	0
30	Calib. Pos. 4	0
31	Calib. Conc. 5	0
32	Calib. Pos. 5	0
33	Calib. Conc. 6	0
34	Calib. Pos. 6	0
35	S 1 ABS.	0
36	K Factor	10000
37	K 2 Factor	10000
38	K 3 Factor	10000
39	K 4 Factor	10000
40	K 5 Factor	10000
41	A Factor	0
42	B Factor	0
43	C Factor	0
44	SD Limit	0.1
45	Duplicate Limit	500
46	Sens. Limit	0
47	S 1 ABS Limit (L)	-32000
48	S 1 ABS Limit (H)	32000
49	ABS Limit	0
50	ABS Limit (D/I)	INCREASE
51	Prz. Limit	0
52	Prz. Limit (U/D)	Lower
53	Prz. (End Point)	35
54	Expect. Value (L)	0
55	Expect. Value (H)	130
56	Instr. Fact. (a)	1
57	Instr. Fact. (b)	0
58	Key Setting	#

LDL Cholesterol Direct

Direct Enzymatic
Liquid, 2 Reagent
Calibrator : HDL/LDL CALIBRATOR (CAT NO. HL132)
Control : General Control 1 & 2

HITACHI 717	
TEST	LDL-C
ASSAY CODE	[2 Point End] [24] - [50]
SAMPLE VOL (µL)	[4] [4]
R1 VOLUME (µL)	[240] [50] - [NO]
R2 VOLUME (µL)	[80] [20] - [NO]
WAVELENGTH	[700] [546]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[mg/dl]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[0] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[0] [130]
PANIC VALUE	[0] [600]
INSTRUMENT FACTOR	[1]

HITACHI 704	
TEST	LDL-C
ASSAY CODE	[2 Point End] [15] - [32]
SAMPLE VOL (µL)	[4] [4]
R1 VOLUME (µL)	[240] [50] - [NO]
R2 VOLUME (µL)	[80] [20] - [NO]
WAVELENGTH	[700] [546]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[mg/dl]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[0] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[0] [130]
PANIC VALUE	[0] [600]
INSTRUMENT FACTOR	[1]

User defined
جایگاه کالیبراتور ##

HITACHI 902		
1	Test Name	LPS
2	Assay Code (Mthd)	Rate A
3	Assay Code (2. Test)	0
4	Reaction Time	10
5	Assay Point 1	24
6	Assay Point 2	32
7	Assay Point 3	0
8	Assay Point 4	0
9	Wave Leng. (SUB)	700
10	Wave Leng. (MAIN)	570
11	Sample Volume	6
12	R1 VOLUME	240
13	R1 Pos.	#
14	R1 Bottle Size	Large
15	R2 VOLUME	0
16	R2 Pos. 0	0
17	R2 Bottle Size	Small
18	R3 VOLUME	60
19	R3 Pos. 0	#
20	R3 Bottle Size Small	Small
21	Calib. Type (Type)	LINEAR
22	Calib. Type (Wght)	0
23	Calib. Conc. 1	0
24	Calib. Pos. 1	36
25	Calib. Conc. 2	#
26	Calib. Pos. 2	##
27	Calib. Conc. 3	0
28	Calib. Pos. 3	0
29	Calib. Conc. 4	0
30	Calib. Pos. 4	0
31	Calib. Conc. 5	0
32	Calib. Pos. 5	0
33	Calib. Conc. 6	0
34	Calib. Pos. 6	0
35	S 1 ABS.	0
36	K Factor	10000
37	K 2 Factor	10000
38	K 3 Factor	10000
39	K 4 Factor	10000
40	K 5 Factor	10000
41	A Factor	0
42	B Factor	0
43	C Factor	0
44	SD Limit	0.1
45	Duplicate Limit	500
46	Sens. Limit	0
47	S 1 ABS Limit (L)	-32000
48	S 1 ABS Limit (H)	32000
49	ABS Limit	12000
50	ABS Limit (D/I)	INCREASE
51	Prz. Limit	0
52	Prz. Limit (U/D)	Lower
53	Prz. (End Point)	35
54	Expect. Value (L)	0
55	Expect. Value (H)	60
56	Instr. Fact. (a)	1
57	Instr. Fact. (b)	0
58	Key Setting	#

Lipase

Colorimetric
Liquid, 2 Reagent
Calibrator : GENERAL CALIBRATOR (CAT NO. CA532)
Control : GENERAL Control 1 & 2

HITACHI 717	
TEST	[LPS]
ASSAY CODE	[Rate A] [35] - [45]
SAMPLE VOL (µL)	[6] [6]
R1 VOLUME (µL)	[240] [50] - [NO]
R2 VOLUME (µL)	[60] [20] - [NO]
WAVELENGTH	[700] [570]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[U/L]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[12000] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[0] [60]
PANIC VALUE	[0] [150]
INSTRUMENT FACTOR	[1]

HITACHI 704	
TEST	[LPS]
ASSAY CODE	[Rate A] [22] - [28]
SAMPLE VOL (µL)	[6] [6]
R1 VOLUME (µL)	[240] [50] - [NO]
R2 VOLUME (µL)	[60] [20] - [NO]
WAVELENGTH	[700] [570]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[U/L]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[12000] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[0] [60]
PANIC VALUE	[0] [150]
INSTRUMENT FACTOR	[1]

User defined
جایگاه کالیبراتور

HITACHI 902		
1	Test Name	MALB
2	Assay Code (Mthd)	2 Point End
3	Assay Code (2. Test)	0
4	Reaction Time	10
5	Assay Point 1	16
6	Assay Point 2	35
7	Assay Point 3	0
8	Assay Point 4	0
9	Wave Leng. (SUB)	700
10	Wave Leng. (MAIN)	340
11	Sample Volume	18
12	R1 VOLUME	250
13	R1 Pos.	#
14	R1 Bottle Size	Large
15	R2 VOLUME	0
16	R2 Pos. 0	0
17	R2 Bottle Size	Small
18	R3 VOLUME	50
19	R3 Pos. 0	#
20	R3 Bottle Size Small	Small
21	Calib. Type (Type)	LOGIT-LOG (4P)
22	Calib. Type (Wght)	0
23	Calib. Conc. 1	0.0
24	Calib. Pos. 1	36
25	Calib. Conc. 2	#
26	Calib. Pos. 2	##
27	Calib. Conc. 3	#
28	Calib. Pos. 3	##
29	Calib. Conc. 4	#
30	Calib. Pos. 4	##
31	Calib. Conc. 5	#
32	Calib. Pos. 5	##
33	Calib. Conc. 6	#
34	Calib. Pos. 6	##
35	S 1 ABS.	0
36	K Factor	10000
37	K 2 Factor	10000
38	K 3 Factor	10000
39	K 4 Factor	10000
40	K 5 Factor	10000
41	A Factor	0
42	B Factor	0
43	C Factor	0
44	SD Limit	999
45	Duplicate Limit	2000
46	Sens. Limit	0
47	S 1 ABS Limit (L)	-32000
48	S 1 ABS Limit (H)	32000
49	ABS Limit	0
50	ABS Limit (D/I)	INCREASE
51	Prz. Limit	0
52	Prz. Limit (U/D)	Lower
53	Prz. (End Point)	35
54	Expect. Value (L)	0
55	Expect. Value (H)	30
56	Instr. Fact. (a)	1
57	Instr. Fact. (b)	0
58	Key Setting	#

Micro Albumin

IMMUNOTURB
Liquid, 2 Reagent
Calibrator : MICROALBUMIN CALIBRATOR (CAT NO. MA231)
Control : MICROALBUMIN Control 1 & 2

HITACHI 717	
TEST	MALB
ASSAY CODE	[2 Point End] [24] - [50]
SAMPLE VOL (µL)	[18] [18]
R1 VOLUME (µL)	[250] [50] - [NO]
R2 VOLUME (µL)	[50] [20] - [NO]
WAVELENGTH	[700] [340]
CALIB. METHOD	[NONLINEAR] [1] - [6]
STD 1 CONC. POS.	[0.0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[#] - [##]
STD 4 CONC. POS.	[#] - [##]
STD 5 CONC. POS.	[#] - [##]
STD 6 CONC. POS.	[#] - [##]
UNIT	[mg/l]
SD LIMIT	[999]
DUPLICATE LIMIT	[2000]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[0] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[0] [30]
PANIC VALUE	[0] [300]
INSTRUMENT FACTOR	[1]

HITACHI 704	
TEST	MALB
ASSAY CODE	[2 Point End] [15] - [32]
SAMPLE VOL (µL)	[18] [18]
R1 VOLUME (µL)	[250] [50] - [NO]
R2 VOLUME (µL)	[50] [20] - [NO]
WAVELENGTH	[700] [340]
CALIB. METHOD	[NONLINEAR] [1] - [6]
STD 1 CONC. POS.	[0.0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[#] - [##]
STD 4 CONC. POS.	[#] - [##]
STD 5 CONC. POS.	[#] - [##]
STD 6 CONC. POS.	[#] - [##]
UNIT	[mg/l]
SD LIMIT	[999]
DUPLICATE LIMIT	[2000]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[0] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[0] [30]
PANIC VALUE	[0] [300]
INSTRUMENT FACTOR	[1]

User defined
جایگاه کالیبراتور

HITACHI 902		
1	Test Name	MG
2	Assay Code (Mthd)	1 point
3	Assay Code (2. Test)	0
4	Reaction Time	3
5	Assay Point 1	10
6	Assay Point 2	0
7	Assay Point 3	0
8	Assay Point 4	0
9	Wave Leng. (SUB)	700
10	Wave Leng. (MAIN)	505
11	Sample Volume	3
12	R1 VOLUME	300
13	R1 Pos.	#
14	R1 Bottle Size	Large
15	R2 VOLUME	0
16	R2 Pos. 0	0
17	R2 Bottle Size	Small
18	R3 VOLUME	0
19	R3 Pos. 0	#
20	R3 Bottle Size Small	Small
21	Calib. Type (Type)	LINEAR
22	Calib. Type (Wght)	0
23	Calib. Conc. 1	0.00
24	Calib. Pos. 1	36
25	Calib. Conc. 2	#
26	Calib. Pos. 2	##
27	Calib. Conc. 3	0
28	Calib. Pos. 3	0
29	Calib. Conc. 4	0
30	Calib. Pos. 4	0
31	Calib. Conc. 5	0
32	Calib. Pos. 5	0
33	Calib. Conc. 6	0
34	Calib. Pos. 6	0
35	S 1 ABS.	0
36	K Factor	10000
37	K 2 Factor	10000
38	K 3 Factor	10000
39	K 4 Factor	10000
40	K 5 Factor	10000
41	A Factor	0
42	B Factor	0
43	C Factor	0
44	SD Limit	0.1
45	Duplicate Limit	500
46	Sens. Limit	0
47	S 1 ABS Limit (L)	-32000
48	S 1 ABS Limit (H)	32000
49	ABS Limit	0
50	ABS Limit (D/I)	INCREASE
51	Prz. Limit	0
52	Prz. Limit (U/D)	Lower
53	Prz. (End Point)	35
54	Expect. Value (L)	1.8
55	Expect. Value (H)	2.6
56	Instr. Fact. (a)	1
57	Instr. Fact. (b)	0
58	Key Setting	#

Magnesium

XYLIDYL BLUE
Liquid, 1 Reagent
Calibrator : GENERAL CALIBRATOR (CAT NO. CA532)
Control : GENERAL Control 1 & 2

HITACHI 717			
TEST	[MG]		
ASSAY CODE	[1 point]	[20] -	[0]
SAMPLE VOL (µL)	[3]	[3]	
R1 VOLUME (µL)	[300]	[50] -	[NO]
R2 VOLUME (µL)	[0]	[20] -	[NO]
WAVELENGTH	[700]	[505]	
CALIB. METHOD	[LINEAR]	[0] -	[0]
STD 1 CONC. POS.	[0.00] -	[##]	
STD 2 CONC. POS.	[#] -	[##]	
STD 3 CONC. POS.	[0] -	[0]	
STD 4 CONC. POS.	[0] -	[0]	
STD 5 CONC. POS.	[0] -	[0]	
STD 6 CONC. POS.	[0] -	[0]	
UNIT	[mg/dl]		
SD LIMIT	[0.1]		
DUPLICATE LIMIT	[500]		
SENSITIVITY LIMIT	[0]		
ABS. LIMIT (INCR./DECR.)	[0]	[INCREASE]	
PROZON LIMIT	[0]	[LOWER]	
EXPECTED VALUE	[1.8]	[2.6]	
PANIC VALUE	[0]	[5]	
INSTRUMENT FACTOR	[1]		

HITACHI 704			
TEST	[MG]		
ASSAY CODE	[1 point]	[10] -	[0]
SAMPLE VOL (µL)	[3]	[3]	
R1 VOLUME (µL)	[300]	[50] -	[NO]
R2 VOLUME (µL)	[0]	[20] -	[NO]
WAVELENGTH	[700]	[505]	
CALIB. METHOD	[LINEAR]	[0] -	[0]
STD 1 CONC. POS.	[0.00] -	[##]	
STD 2 CONC. POS.	[#] -	[##]	
STD 3 CONC. POS.	[0] -	[0]	
STD 4 CONC. POS.	[0] -	[0]	
STD 5 CONC. POS.	[0] -	[0]	
STD 6 CONC. POS.	[0] -	[0]	
UNIT	[mg/dl]		
SD LIMIT	[0.1]		
DUPLICATE LIMIT	[500]		
SENSITIVITY LIMIT	[0]		
ABS. LIMIT (INCR./DECR.)	[0]	[INCREASE]	
PROZON LIMIT	[0]	[LOWER]	
EXPECTED VALUE	[1.8]	[2.6]	
PANIC VALUE	[0]	[5]	
INSTRUMENT FACTOR	[1]		

User defined
جایگاه کالیبراتور

HITACHI 902		
1	Test Name	PHOS
2	Assay Code (Mthd)	1 point
3	Assay Code (2. Test)	0
4	Reaction Time	4
5	Assay Point 1	12
6	Assay Point 2	0
7	Assay Point 3	0
8	Assay Point 4	0
9	Wave Leng. (SUB)	700
10	Wave Leng. (MAIN)	340
11	Sample Volume	3
12	R1 VOLUME	300
13	R1 Pos.	#
14	R1 Bottle Size	Large
15	R2 VOLUME	0
16	R2 Pos. 0	0
17	R2 Bottle Size	Small
18	R3 VOLUME	0
19	R3 Pos. 0	#
20	R3 Bottle Size Small	Small
21	Calib. Type (Type)	LINEAR
22	Calib. Type (Wght)	0
23	Calib. Conc. 1	0.0
24	Calib. Pos. 1	36
25	Calib. Conc. 2	#
26	Calib. Pos. 2	##
27	Calib. Conc. 3	0
28	Calib. Pos. 3	0
29	Calib. Conc. 4	0
30	Calib. Pos. 4	0
31	Calib. Conc. 5	0
32	Calib. Pos. 5	0
33	Calib. Conc. 6	0
34	Calib. Pos. 6	0
35	S 1 ABS.	0
36	K Factor	10000
37	K 2 Factor	10000
38	K 3 Factor	10000
39	K 4 Factor	10000
40	K 5 Factor	10000
41	A Factor	0
42	B Factor	0
43	C Factor	0
44	SD Limit	0.1
45	Duplicate Limit	500
46	Sens. Limit	0
47	S 1 ABS Limit (L)	-32000
48	S 1 ABS Limit (H)	32000
49	ABS Limit	0
50	ABS Limit (D/I)	INCREASE
51	Prz. Limit	0
52	Prz. Limit (U/D)	Lower
53	Prz. (End Point)	35
54	Expect. Value (L)	2.6
55	Expect. Value (H)	4.5
56	Instr. Fact. (a)	1
57	Instr. Fact. (b)	0
58	Key Setting	#

Phosphorus

UV
Liquid, 1 Reagent
Calibrator : GENERAL CALIBRATOR (CAT NO. CA532)
Control : GENERAL Control 1 & 2

HITACHI 717	
TEST	[PHOS]
ASSAY CODE	[1 point] [20] - [0]
SAMPLE VOL (µL)	[3] [3]
R1 VOLUME (µL)	[300] [50] - [NO]
R2 VOLUME (µL)	[0] [20] - [NO]
WAVELENGTH	[700] [340]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0.0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[mg/dl]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[0] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[2.6] [4.5]
PANIC VALUE	[0] [20]
INSTRUMENT FACTOR	[1]

HITACHI 704	
TEST	[PHOS]
ASSAY CODE	[1 point] [12] - [0]
SAMPLE VOL (µL)	[3] [3]
R1 VOLUME (µL)	[300] [50] - [NO]
R2 VOLUME (µL)	[0] [20] - [NO]
WAVELENGTH	[700] [340]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0.0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[mg/dl]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[0] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[2.6] [4.5]
PANIC VALUE	[0] [20]
INSTRUMENT FACTOR	[1]

User defined
جایگاه کالیبراتور

HITACHI 902		
1	Test Name	TG
2	Assay Code (Mthd)	1 point
3	Assay Code (2. Test)	0
4	Reaction Time	10
5	Assay Point 1	35
6	Assay Point 2	0
7	Assay Point 3	0
8	Assay Point 4	0
9	Wave Leng. (SUB)	700
10	Wave Leng. (MAIN)	505
11	Sample Volume	3
12	R1 VOLUME	300
13	R1 Pos.	#
14	R1 Bottle Size	Large
15	R2 VOLUME	0
16	R2 Pos. 0	0
17	R2 Bottle Size	Small
18	R3 VOLUME	0
19	R3 Pos. 0	#
20	R3 Bottle Size Small	Small
21	Calib. Type (Type)	LINEAR
22	Calib. Type (Wght)	0
23	Calib. Conc. 1	0
24	Calib. Pos. 1	36
25	Calib. Conc. 2	#
26	Calib. Pos. 2	##
27	Calib. Conc. 3	0
28	Calib. Pos. 3	0
29	Calib. Conc. 4	0
30	Calib. Pos. 4	0
31	Calib. Conc. 5	0
32	Calib. Pos. 5	0
33	Calib. Conc. 6	0
34	Calib. Pos. 6	0
35	S 1 ABS.	0
36	K Factor	10000
37	K 2 Factor	10000
38	K 3 Factor	10000
39	K 4 Factor	10000
40	K 5 Factor	10000
41	A Factor	0
42	B Factor	0
43	C Factor	0
44	SD Limit	0.1
45	Duplicate Limit	500
46	Sens. Limit	0
47	S 1 ABS Limit (L)	-32000
48	S 1 ABS Limit (H)	32000
49	ABS Limit	0
50	ABS Limit (D/I)	INCREASE
51	Prz. Limit	0
52	Prz. Limit (U/D)	Lower
53	Prz. (End Point)	35
54	Expect. Value (L)	0
55	Expect. Value (H)	160
56	Instr. Fact. (a)	1
57	Instr. Fact. (b)	0
58	Key Setting	#

Triglycerides

GPO
Liquid, 1 Reagent
Calibrator : GENERAL CALIBRATOR (CAT NO. CA532)
Control : GENERAL Control 1 & 2

HITACHI 717	
TEST	[TG]
ASSAY CODE	[1 point] [50] - [0]
SAMPLE VOL (µL)	[3] [3]
R1 VOLUME (µL)	[300] [50] - [NO]
R2 VOLUME (µL)	[0] [20] - [NO]
WAVELENGTH	[700] [505]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[mg/dl]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[0] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[0] [160]
PANIC VALUE	[0] [800]
INSTRUMENT FACTOR	[1]

HITACHI 704	
TEST	[TG]
ASSAY CODE	[1 point] [32] - [0]
SAMPLE VOL (µL)	[3] [3]
R1 VOLUME (µL)	[300] [50] - [NO]
R2 VOLUME (µL)	[0] [20] - [NO]
WAVELENGTH	[700] [505]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[mg/dl]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[0] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[0] [160]
PANIC VALUE	[0] [800]
INSTRUMENT FACTOR	[1]

User defined
جایگاه کالیبراتور

HITACHI 902		
1	Test Name	T.PRO
2	Assay Code (Mthd)	1 point
3	Assay Code (2. Test)	0
4	Reaction Time	5
5	Assay Point 1	16
6	Assay Point 2	0
7	Assay Point 3	0
8	Assay Point 4	0
9	Wave Leng. (SUB)	700
10	Wave Leng. (MAIN)	546
11	Sample Volume	8
12	R1 VOLUME	300
13	R1 Pos.	#
14	R1 Bottle Size	Large
15	R2 VOLUME	0
16	R2 Pos. 0	0
17	R2 Bottle Size	Small
18	R3 VOLUME	0
19	R3 Pos. 0	#
20	R3 Bottle Size Small	Small
21	Calib. Type (Type)	LINEAR
22	Calib. Type (Wght)	0
23	Calib. Conc. 1	0.0
24	Calib. Pos. 1	36
25	Calib. Conc. 2	#
26	Calib. Pos. 2	##
27	Calib. Conc. 3	0
28	Calib. Pos. 3	0
29	Calib. Conc. 4	0
30	Calib. Pos. 4	0
31	Calib. Conc. 5	0
32	Calib. Pos. 5	0
33	Calib. Conc. 6	0
34	Calib. Pos. 6	0
35	S 1 ABS.	0
36	K Factor	10000
37	K 2 Factor	10000
38	K 3 Factor	10000
39	K 4 Factor	10000
40	K 5 Factor	10000
41	A Factor	0
42	B Factor	0
43	C Factor	0
44	SD Limit	0.1
45	Duplicate Limit	500
46	Sens. Limit	0
47	S 1 ABS Limit (L)	-32000
48	S 1 ABS Limit (H)	32000
49	ABS Limit	0
50	ABS Limit (D/I)	INCREASE
51	Prz. Limit	0
52	Prz. Limit (U/D)	Lower
53	Prz. (End Point)	35
54	Expect. Value (L)	6.6
55	Expect. Value (H)	8.8
56	Instr. Fact. (a)	1
57	Instr. Fact. (b)	0
58	Key Setting	#

Total Protein

BIURET
Liquid, 1 Reagent
Calibrator : GENERAL CALIBRATOR (CAT NO. CA532)
Control : GENERAL Control 1 & 2

HITACHI 717	
TEST	T.PRO
ASSAY CODE	[1 point] [24] - [0]
SAMPLE VOL (µL)	[8] [8]
R1 VOLUME (µL)	[300] [50] - [NO]
R2 VOLUME (µL)	[0] [20] - [NO]
WAVELENGTH	[700] [546]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0.0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[g/dl]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[0] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[6.6] [8.8]
PANIC VALUE	[0] [12]
INSTRUMENT FACTOR	[1]

HITACHI 704	
TEST	T.PRO
ASSAY CODE	[1 point] [15] - [0]
SAMPLE VOL (µL)	[8] [8]
R1 VOLUME (µL)	[300] [50] - [NO]
R2 VOLUME (µL)	[0] [20] - [NO]
WAVELENGTH	[700] [546]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0.0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[g/dl]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[0] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[6.6] [8.8]
PANIC VALUE	[0] [12]
INSTRUMENT FACTOR	[1]

User defined
جایگاه کالیبراتور

HITACHI 902		
1	Test Name	UA
2	Assay Code (Mthd)	1 point
3	Assay Code (2. Test)	0
4	Reaction Time	10
5	Assay Point 1	25
6	Assay Point 2	0
7	Assay Point 3	0
8	Assay Point 4	0
9	Wave Leng. (SUB)	700
10	Wave Leng. (MAIN)	546
11	Sample Volume	8
12	R1 VOLUME	160
13	R1 Pos.	#
14	R1 Bottle Size	Large
15	R2 VOLUME	160
16	R2 Pos. 0	0
17	R2 Bottle Size	Small
18	R3 VOLUME	0
19	R3 Pos. 0	#
20	R3 Bottle Size Small	Small
21	Calib. Type (Type)	LINEAR
22	Calib. Type (Wght)	0
23	Calib. Conc. 1	0.0
24	Calib. Pos. 1	36
25	Calib. Conc. 2	#
26	Calib. Pos. 2	##
27	Calib. Conc. 3	0
28	Calib. Pos. 3	0
29	Calib. Conc. 4	0
30	Calib. Pos. 4	0
31	Calib. Conc. 5	0
32	Calib. Pos. 5	0
33	Calib. Conc. 6	0
34	Calib. Pos. 6	0
35	S 1 ABS.	0
36	K Factor	10000
37	K 2 Factor	10000
38	K 3 Factor	10000
39	K 4 Factor	10000
40	K 5 Factor	10000
41	A Factor	0
42	B Factor	0
43	C Factor	0
44	SD Limit	0.1
45	Duplicate Limit	500
46	Sens. Limit	0
47	S 1 ABS Limit (L)	-32000
48	S 1 ABS Limit (H)	32000
49	ABS Limit	0
50	ABS Limit (D/I)	INCREASE
51	Prz. Limit	0
52	Prz. Limit (U/D)	Lower
53	Prz. (End Point)	35
54	Expect. Value (L)	3.6
55	Expect. Value (H)	7.7
56	Instr. Fact. (a)	1
57	Instr. Fact. (b)	0
58	Key Setting	#

Uric acid

URICASE
Liquid, 2 Reagent
Calibrator : GENERAL CALIBRATOR (CAT NO. CA532)
Control : GENERAL Control 1 & 2

HITACHI 717				
TEST	[UA]			
ASSAY CODE	[1 point]	[35]	- [0]	
SAMPLE VOL (µL)	[8]	[8]		
R1 VOLUME (µL)	[160]	[50]	- [NO]	
R2 VOLUME (µL)	[160]	[20]	- [NO]	
WAVELENGTH	[700]	[546]		
CALIB. METHOD	[LINEAR]	[0]	- [0]	
STD 1 CONC. POS.	[0.0]	- [##]		
STD 2 CONC. POS.	[#]	- [##]		
STD 3 CONC. POS.	[0]	- [0]		
STD 4 CONC. POS.	[0]	- [0]		
STD 5 CONC. POS.	[0]	- [0]		
STD 6 CONC. POS.	[0]	- [0]		
UNIT	[mg/dl]			
SD LIMIT	[0.1]			
DUPLICATE LIMIT	[500]			
SENSITIVITY LIMIT	[0]			
ABS. LIMIT (INCR./DECR.)	[0]	[INCREASE]		
PROZON LIMIT	[0]	[LOWER]		
EXPECTED VALUE	[3.6]	[7.7]		
PANIC VALUE	[0]	[20]		
INSTRUMENT FACTOR	[1]			

HITACHI 704				
TEST	[UA]			
ASSAY CODE	[1 point]	[25]	- [0]	
SAMPLE VOL (µL)	[8]	[8]		
R1 VOLUME (µL)	[160]	[50]	- [NO]	
R2 VOLUME (µL)	[160]	[20]	- [NO]	
WAVELENGTH	[700]	[546]		
CALIB. METHOD	[LINEAR]	[0]	- [0]	
STD 1 CONC. POS.	[0.0]	- [##]		
STD 2 CONC. POS.	[#]	- [##]		
STD 3 CONC. POS.	[0]	- [0]		
STD 4 CONC. POS.	[0]	- [0]		
STD 5 CONC. POS.	[0]	- [0]		
STD 6 CONC. POS.	[0]	- [0]		
UNIT	[mg/dl]			
SD LIMIT	[0.1]			
DUPLICATE LIMIT	[500]			
SENSITIVITY LIMIT	[0]			
ABS. LIMIT (INCR./DECR.)	[0]	[INCREASE]		
PROZON LIMIT	[0]	[LOWER]		
EXPECTED VALUE	[3.6]	[7.7]		
PANIC VALUE	[0]	[20]		
INSTRUMENT FACTOR	[1]			

User defined
جایگاه کالیبراتور

HITACHI 902		
1	Test Name	UREA
2	Assay Code (Mthd)	2 point rate
3	Assay Code (2. Test)	0
4	Reaction Time	4
5	Assay Point 1	9
6	Assay Point 2	12
7	Assay Point 3	0
8	Assay Point 4	0
9	Wave Leng. (SUB)	415
10	Wave Leng. (MAIN)	340
11	Sample Volume	3
12	R1 VOLUME	240
13	R1 Pos.	#
14	R1 Bottle Size	Large
15	R2 VOLUME	60
16	R2 Pos. 0	0
17	R2 Bottle Size	Small
18	R3 VOLUME	0
19	R3 Pos. 0	#
20	R3 Bottle Size Small	Small
21	Calib. Type (Type)	LINEAR
22	Calib. Type (Wght)	0
23	Calib. Conc. 1	0.0
24	Calib. Pos. 1	36
25	Calib. Conc. 2	#
26	Calib. Pos. 2	##
27	Calib. Conc. 3	0
28	Calib. Pos. 3	0
29	Calib. Conc. 4	0
30	Calib. Pos. 4	0
31	Calib. Conc. 5	0
32	Calib. Pos. 5	0
33	Calib. Conc. 6	0
34	Calib. Pos. 6	0
35	S 1 ABS.	0
36	K Factor	10000
37	K 2 Factor	10000
38	K 3 Factor	10000
39	K 4 Factor	10000
40	K 5 Factor	10000
41	A Factor	0
42	B Factor	0
43	C Factor	0
44	SD Limit	0.1
45	Duplicate Limit	500
46	Sens. Limit	0
47	S 1 ABS Limit (L)	-32000
48	S 1 ABS Limit (H)	32000
49	ABS Limit	4000
50	ABS Limit (D/I)	DECREASE
51	Prz. Limit	0
52	Prz. Limit (U/D)	Lower
53	Prz. (End Point)	35
54	Expect. Value (L)	19
55	Expect. Value (H)	44
56	Instr. Fact. (a)	1
57	Instr. Fact. (b)	0
58	Key Setting	#

Urea/BUN

UV/KINETIC
Liquid, 2 Reagent
Calibrator : GENERAL CALIBRATOR (CAT NO. CA532)
Control : GENERAL Control 1 & 2

HITACHI 717	
TEST	[UREA]
ASSAY CODE	[2 point rate] [28] - [32]
SAMPLE VOL (µL)	[3] [3]
R1 VOLUME (µL)	[240] [50] - [NO]
R2 VOLUME (µL)	[60] [20] - [NO]
WAVELENGTH	[415] [340]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0.0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[mg/dl]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[4000] [DECREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[19] [44]
PANIC VALUE	[0] [300]
INSTRUMENT FACTOR	[1]

HITACHI 704	
TEST	[UREA]
ASSAY CODE	[2 point rate] [18] - [21]
SAMPLE VOL (µL)	[3] [3]
R1 VOLUME (µL)	[240] [50] - [NO]
R2 VOLUME (µL)	[60] [20] - [NO]
WAVELENGTH	[415] [340]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0.0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[mg/dl]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[4000] [DECREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[19] [44]
PANIC VALUE	[0] [300]
INSTRUMENT FACTOR	[1]

User defined
جایگاه کالیبراتور

HITACHI 902		
1	Test Name	ZN
2	Assay Code (Mthd)	1 point
3	Assay Code (2. Test)	0
4	Reaction Time	3
5	Assay Point 1	10
6	Assay Point 2	0
7	Assay Point 3	0
8	Assay Point 4	0
9	Wave Leng. (SUB)	700
10	Wave Leng. (MAIN)	570
11	Sample Volume	15
12	R1 VOLUME	300
13	R1 Pos.	#
14	R1 Bottle Size	Large
15	R2 VOLUME	0
16	R2 Pos. 0	0
17	R2 Bottle Size	Small
18	R3 VOLUME	0
19	R3 Pos. 0	#
20	R3 Bottle Size Small	Small
21	Calib. Type (Type)	LINEAR
22	Calib. Type (Wght)	0
23	Calib. Conc. 1	0
24	Calib. Pos. 1	36
25	Calib. Conc. 2	#
26	Calib. Pos. 2	##
27	Calib. Conc. 3	0
28	Calib. Pos. 3	0
29	Calib. Conc. 4	0
30	Calib. Pos. 4	0
31	Calib. Conc. 5	0
32	Calib. Pos. 5	0
33	Calib. Conc. 6	0
34	Calib. Pos. 6	0
35	S 1 ABS.	0
36	K Factor	10000
37	K 2 Factor	10000
38	K 3 Factor	10000
39	K 4 Factor	10000
40	K 5 Factor	10000
41	A Factor	0
42	B Factor	0
43	C Factor	0
44	SD Limit	0.1
45	Duplicate Limit	500
46	Sens. Limit	0
47	S 1 ABS Limit (L)	-32000
48	S 1 ABS Limit (H)	32000
49	ABS Limit	0
50	ABS Limit (D/I)	INCREASE
51	Prz. Limit	0
52	Prz. Limit (U/D)	Lower
53	Prz. (End Point)	35
54	Expect. Value (L)	67
55	Expect. Value (H)	127
56	Instr. Fact. (a)	1
57	Instr. Fact. (b)	0
58	Key Setting	#

ZINC

5-Br-PAPS
Liquid, 1 Reagent
Calibrator : GENERAL CALIBRATOR (CAT NO. CA532)
Control : GENERAL Control 1 & 2

HITACHI 717	
TEST	ZN
ASSAY CODE	[1 point] [20] - [0]
SAMPLE VOL (µL)	[15] [15]
R1 VOLUME (µL)	[300] [50] - [NO]
R2 VOLUME (µL)	[0] [20] - [NO]
WAVELENGTH	[700] [570]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[µg/dl]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[0] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[67] [127]
PANIC VALUE	[0] [400]
INSTRUMENT FACTOR	[1]

HITACHI 704	
TEST	ZN
ASSAY CODE	[1 point] [10] - [0]
SAMPLE VOL (µL)	[15] [15]
R1 VOLUME (µL)	[300] [50] - [NO]
R2 VOLUME (µL)	[0] [20] - [NO]
WAVELENGTH	[700] [570]
CALIB. METHOD	[LINEAR] [0] - [0]
STD 1 CONC. POS.	[0] - [##]
STD 2 CONC. POS.	[#] - [##]
STD 3 CONC. POS.	[0] - [0]
STD 4 CONC. POS.	[0] - [0]
STD 5 CONC. POS.	[0] - [0]
STD 6 CONC. POS.	[0] - [0]
UNIT	[µg/dl]
SD LIMIT	[0.1]
DUPLICATE LIMIT	[500]
SENSITIVITY LIMIT	[0]
ABS. LIMIT (INCR./DECR.)	[0] [INCREASE]
PROZON LIMIT	[0] [LOWER]
EXPECTED VALUE	[67] [127]
PANIC VALUE	[0] [400]
INSTRUMENT FACTOR	[1]

User defined
جایگاه کالیبراتور