BIOMEDIC

Albumin (BCG) — HITACHI 912

Analyze

Select Test: ALB | Analyzer Cycle Time: 10 sec | Diluent: 0 | 0

Test Name: ALB | Assay / Time / Point: 1 Point | 5 | 15 | 0

App. Code: تغییر ندهید | Wavelength (2nd / Primary): 700 | 600

Sample Volume

	Class 1			Class 2		
Normal:	3	0	0	3	0	0
Decrease:	3	0	0	3	0	0
Increase:	6	0	0	6	0	0

	Class 1	Class 2	Dec/Inc
Abs.Limit	0	0	Increase
Prozone Limit	0	0	Lower

Reagent

R1:	350	0	350	0
R2:	0	0	350	0
R3:	0	0	350	0
R4:	0	0	350	0

Twin Test:

Calibration

Calibration Type: Linear

Point: 2 | Span Point: 2

Weight: 0

SD Limit: 0.1

Duplicate Limit: 10 % 500

Sensitivity Range: -99999 - 999999

S1 Abs Range: -32000 - 32000

Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calic. Code:	#	#				
Concentration:	0.0					
Position:						
Sample Volume:	3	3	3	3	3	3
Diluted S. Vol.	0	0	0	0	0	0
Diluent Volume:	0	0	0	0	0	0

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Alkaline phosphatase

HITACHI 912

Analyze

Select Test: ALP | Analyzer Cycle Time: 10 sec | Diluent: 0 | 0

Test Name: ALP | Assay / Time / Point: Rate-A | 10 | 19 | 28 | | |

App. Code: تغییر ندهید | Wavelength (2nd / Primary): 660 | 415

Sample Volume

	Class 1			Class 2		
Normal:	5	0	0	6	0	0
Decrease:	3	0	0	3	0	0
Increase:	10	0	0	12	0	0

	Class 1	Class 2	Dec/Inc
Abs.Limit	14000	14000	Increase
Prozone Limit	0	0	Lower

Reagent

R1:	200	0	351	0
R2:	0	0	351	0
R3:	50	0	351	0
R4:	0	0	351	0

Twin Test:

Calibration

Calibration Type: Linear

Point: 2 | Span Point: 2

Weight: 0

SD Limit: 0.1

Duplicate Limit: 10 % 500

Sensitivity Range: -99999 - 999999

S1 Abs Range: -32000 - 32000

Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calic. Code:	#	#				
Concentration:	0					
Position:						
Sample Volume:	5	5	5	5	5	5
Diluted S. Vol.	0	0	0	0	0	0
Diluent Volume:	0	0	0	0	0	0

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

ALT/GPT — HITACHI 912

Analyze

Select Test: ALT | Analyzer Cycle Time: 10 sec | Diluent: 0 | 0

Test Name: ALT | Assay / Time / Point: Rate-A | 10 | 19 | 28

App. Code: تغییر ندهید | Wavelength (2nd / Primary): 415 | 340

Sample Volume

	Class 1			Class 2		
Normal:	25	0	0	25	0	0
Decrease:	5	0	0	5	0	0
Increase:	40	0	0	40	0	0

Reagent

R1:	200	0	352	0
R2:	0	0	352	0
R3:	50	0	352	0
R4:	0	0	352	0

	Class 1	Class 2	Dec/Inc
Abs.Limit	5000	5000	Decrease
Prozone Limit	0	0	Lower

Twin Test:

Calibration

Calibration Type: Linear

Point: 2 | Span Point: 2

Weight: 0

SD Limit: 0.1

Duplicate Limit: 10 % 500

Sensitivity Range: -99999 - 999999

S1 Abs Range: -32000 - 32000

Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calic. Code:	#	#				
Concentration:	0					
Position:						
Sample Volume:	25	25	25	25	25	25
Diluted S. Vol.	0	0	0	0	0	0
Diluent Volume:	0	0	0	0	0	0

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Amylase - total

HITACHI 912

Analyze

Select Test: AMY | Analyzer Cycle Time: 10 sec | Diluent: 0 | 0

Test Name: AMY | Assay / Time / Point: Rate-A | 5 | 3 | 12

App. Code: تغییر ندهید | Wavelength (2nd / Primary): 660 | 415

Sample Volume

	Class 1			Class 2		
Normal:	6	0	0	6	0	0
Decrease:	3	0	0	3	0	0
Increase:	12	0	0	12	0	0

Reagent

R1:	300	0	354	0
R2:	0	0	354	0
R3:	0	0	354	0
R4:	0	0	354	0

	Class 1	Class 2	Dec/Inc
Abs.Limit	20000	20000	Increase
Prozone Limit	0	0	Lower

Twin Test:

Calibration

Calibration Type: Linear

Point: 2 | Span Point: 2

Weight: 0

SD Limit: 0.1

Duplicate Limit: 10 % 500

Sensitivity Range: -99999 - 999999

S1 Abs Range: -32000 - 32000

Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calic. Code:	#	#				
Concentration:	0					
Position:						
Sample Volume:	6	6	6	6	6	6
Diluted S. Vol.	0	0	0	0	0	0
Diluent Volume:	0	0	0	0	0	0

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

AST/GOT

HITACHI 912

Analyze

Select Test: AST | Analyzer Cycle Time: 10 sec | Diluent: 0 | 0

Test Name: AST | Assay / Time / Point: Rate-A | 10 | 19 | 28 | | |

App. Code: تغییر ندهید | Wavelength (2nd / Primary): 415 | 340

Sample Volume

	Class 1			Class 2		
Normal:	25	0	0	25	0	0
Decrease:	5	0	0	5	0	0
Increase:	40	0	0	40	0	0

	Class 1	Class 2	Dec/Inc
Abs.Limit	5000	5000	Decrease
Prozone Limit	0	0	Lower

Reagent

R1:	200	0	358	0
R2:	0	0	358	0
R3:	50	0	358	0
R4:	0	0	358	0

Twin Test:

Calibration

Calibration Type: Linear

Point: 2 | Span Point: 2

Weight: 0

SD Limit: 0.1

Duplicate Limit: 10 % 500

Sensitivity Range: -99999 - 999999

S1 Abs Range: -32000 - 32000

Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calic. Code:	#	#				
Concentration:	0					
Position:						
Sample Volume:	25	25	25	25	25	25
Diluted S. Vol.	0	0	0	0	0	0
Diluent Volume:	0	0	0	0	0	0

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Complement 3

HITACHI 912

Analyze

Select Test: C3 | Analyzer Cycle Time: 10 sec | Diluent: 0 | 0

Test Name: C3 | Assay / Time / Point: 2 Point End | 10 | 15 | 31 | | |

App. Code: تغییر ندهید | Wavelength (2nd / Primary): 700 | 340

Sample Volume

	Class 1			Class 2		
Normal:	3	0	0	3	0	0
Decrease:	3	0	0	3	0	0
Increase:	6	0	0	6	0	0

	Class 1	Class 2	Dec/Inc
Abs.Limit	0	0	Increase
Prozone Limit	0	0	Lower

Reagent

R1:	250	0	00361	0
R2:	0	0	00361	0
R3:	50	0	00361	0
R4:	0	0	00361	0

Twin Test

Calibration

Calibration Type: Logit-Log (4P)

Point: 6 | Span Point: 6

Weight: 0

SD Limit: 999.9

Duplicate Limit: 99 % 2000

Sensitivity Range: -99999 - 999999

S1 Abs Range: -32000 - 32000

Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calic. Code:	#	#	#	#	#	#
Concentration:	0					
Position:						
Sample Volume:	3	3	3	3	3	3
Diluted S. Vol.	0	0	0	0	0	0
Diluent Volume:	0	0	0	0	0	0

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Complement 4

HITACHI 912

Analyze

Select Test: C4 | Analyzer Cycle Time: 10 sec | Diluent: 0 | 0

Test Name: C4 | Assay / Time / Point: 2 Point End | 10 | 15 | 31 | | |

App. Code: تغییر ندهید | Wavelength (2nd / Primary): 700 | 340

Sample Volume

	Class 1			Class 2		
Normal:	6	0	0	6	0	0
Decrease:	3	0	0	3	0	0
Increase:	12	0	0	12	0	0

Reagent

R1:	250	0	00362	0
R2:	0	0	00362	0
R3:	50	0	00362	0
R4:	0	0	00362	0

	Class 1	Class 2	Dec/Inc
Abs.Limit	0	0	Increase
Prozone Limit	0	0	Lower

Twin Test:

Calibration

Calibration Type: Logit-Log (4P)

Point: 6 | Span Point: 6

Weight: 0

SD Limit: 999.9

Duplicate Limit: 99 % 2000

Sensitivity Range: -99999 - 999999

S1 Abs Range: -32000 - 32000

Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calic. Code:	#	#	#	#	#	#
Concentration:	0.0					
Position:						
Sample Volume:	6	6	6	6	6	6
Diluted S. Vol.	0	0	0	0	0	0
Diluent Volume:	0	0	0	0	0	0

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Calcium

HITACHI 912

Analyze

Select Test: CA | Analyzer Cycle Time: 10 sec | Diluent: 0 | 0

Test Name: CA | Assay / Time / Point: 1 Point | 3 | 10 | 0

App. Code: تغییر ندهید | Wavelength (2nd / Primary): 800 | 660

Sample Volume

	Class 1			Class 2		
Normal:	4	0	0	4	0	0
Decrease:	3	0	0	3	0	0
Increase:	8	0	0	8	0	0

	Class 1	Class 2	Dec/Inc
Abs.Limit	0	0	Increase
Prozone Limit	0	0	Lower

Reagent

R1:	300	0	363	0
R2:	0	0	363	0
R3:	0	0	363	0
R4:	0	0	363	0

Twin Test:

Calibration

Calibration Type: Linear

Point: 2 | Span Point: 2

Weight: 0

SD Limit: 0.1

Duplicate Limit: 10 % 500

Sensitivity Range: -99999 - 999999

S1 Abs Range: -32000 - 32000

Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calic. Code:	#	#				
Concentration:	0.0					
Position:						
Sample Volume:	4	4	4	4	4	4
Diluted S. Vol.	0	0	0	0	0	0
Diluent Volume:	0	0	0	0	0	0

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Cholesterol

HITACHI 912

Analyze

Select Test: CHOL | Analyzer Cycle Time: 10 sec | Diluent: 0 | 0

Test Name: CHOL | Assay / Time / Point: 1 Point | 10 | 31 | 0

App. Code: تغییر ندهید | Wavelength (2nd / Primary): 700 | 505

Sample Volume

	Class 1			Class 2		
Normal:	3	0	0	3	0	0
Decrease:	3	0	0	3	0	0
Increase:	6	0	0	6	0	0

	Class 1	Class 2	Dec/Inc
Abs.Limit	0	0	Increase
Prozone Limit	0	0	Lower

Reagent

R1:	300	0	366	0
R2:	0	0	366	0
R3:	0	0	366	0
R4:	0	0	366	0

Twin Test:

Calibration

Calibration Type: Linear

Point: 2 | Span Point: 2

Weight: 0

SD Limit: 0.1

Duplicate Limit: 10 % 500

Sensitivity Range: -99999 - 999999

S1 Abs Range: -32000 - 32000

Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calic. Code:	#	#				
Concentration:	0					
Position:						
Sample Volume:	3	3	3	3	3	3
Diluted S. Vol.	0	0	0	0	0	0
Diluent Volume:	0	0	0	0	0	0

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

CK-NAC **HITACHI 912**

Analyze

Select Test: CPK | Analyzer Cycle Time: 10 sec | Diluent: 0 | 0

Test Name: CPK | Assay / Time / Point: Rate-A | 10 | 22 | 31

App. Code: تغییر ندهید | Wavelength (2nd / Primary): 546 | 340

Sample Volume

	Class 1			Class 2		
Normal:	10	0	0	10	0	0
Decrease:	3	0	0	3	0	0
Increase:	20	0	0	20	0	0

Reagent

R1:	200	0	368	0
R2:	0	0	368	0
R3:	40	0	368	0
R4:	0	0	368	0

	Class 1	Class 2	Dec/Inc
Abs.Limit	10000	10000	Increase
Prozone Limit	0	0	Lower

Twin Test

Calibration

Calibration Type: Linear

Point: 2 | Span Point: 2

Weight: 0

SD Limit: 0.1

Duplicate Limit: 10 % 500

Sensitivity Range: -99999 - 999999

S1 Abs Range: -32000 - 32000

Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calic. Code:	#	#				
Concentration:	0					
Position:						
Sample Volume:	10	10	10	10	10	10
Diluted S. Vol.	0	0	0	0	0	0
Diluent Volume:	0	0	0	0	0	0

www.tksmed.ir شرکت توسعه کیمیای سعادت

شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Creatinine (Jaffe)

HITACHI 912

Analyze

Select Test: CREAT | Analyzer Cycle Time: 10 sec | Diluent: 0 | 0

Test Name: CREAT | Assay / Time / Point: 2 Point Rate | 4 | 8 | 12 | | |

App. Code: تغییر ندهید | Wavelength (2nd / Primary): 700 | 505

Sample Volume

	Class 1			Class 2		
Normal:	24	0	0	24	0	0
Decrease:	3	0	0	3	0	0
Increase:	40	0	0	40	0	0

	Class 1	Class 2	Dec/Inc
Abs.Limit	32000	32000	Increase
Prozone Limit	0	0	Lower

Reagent

R1:	120	0	370	0
R2:	120	0	370	0
R3:	0	0	370	0
R4:	0	0	370	0

Twin Test

Calibration

Calibration Type: Linear

Point: 2 | Span Point: 2

Weight: 0

SD Limit: 0.1

Duplicate Limit: 10 % 500

Sensitivity Range: -99999 - 999999

S1 Abs Range: -32000 - 32000

Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calic. Code:	#	#				
Concentration:	0.00					
Position:						
Sample Volume:	24	24	24	24	24	24
Diluted S. Vol.	0	0	0	0	0	0
Diluent Volume:	0	0	0	0	0	0

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

C-REACTIVE PROTEIN — HITACHI 912

Analyze

Select Test: CRP | Analyzer Cycle Time: 10 sec | Diluent: 0 | 0

Test Name: CRP | Assay / Time / Point: 2 Point End | 10 | 15 | 31

App. Code: تغییر ندهید | Wavelength (2nd / Primary): 700 | 340

Sample Volume

	Class 1			Class 2		
Normal:	15	0	0	10	0	0
Decrease:	3	0	0	3	0	0
Increase:	30	0	0	30	0	0

Reagent

R1:	250	0	371	0
R2:	0	0	371	0
R3:	50	0	371	0
R4:	0	0	371	0

	Class 1	Class 2	Dec/Inc
Abs.Limit	0	0	Increase
Prozone Limit	0	0	Lower

Twin Test:

Calibration

Calibration Type: Logit-Log (4P)

Point: 6 | Span Point: 6

Weight: 0

SD Limit: 999.9

Duplicate Limit: 99 % 2000

Sensitivity Range: -99999 - 999999

S1 Abs Range: -32000 - 32000

Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calic. Code:	#	#	#	#	#	#
Concentration:	0.0					
Position:						
Sample Volume:	15	15	15	15	15	15
Diluted S. Vol.	0	0	0	0	0	0
Diluent Volume:	0	0	0	0	0	0

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Glucose

HITACHI 912

Analyze

Select Test: GLU | Analyzer Cycle Time: 10 sec | Diluent: 0 | 0

Test Name: GLU | Assay / Time / Point: 1 Point | 10 | 31 | 0

App. Code: تغییر ندهید | Wavelength (2nd / Primary): 700 | 505

Sample Volume

	Class 1			Class 2		
Normal:	3	0	0	3	0	0
Decrease:	3	0	0	3	0	0
Increase:	6	0	0	6	0	0

	Class 1	Class 2	Dec/Inc
Abs.Limit	0	0	Increase
Prozone Limit	0	0	Lower

Reagent

R1:	300	0	379	0
R2:	0	0	379	0
R3:	0	0	379	0
R4:	0	0	379	0

Twin Test

Calibration

Calibration Type: Linear

Point: 2 | Span Point: 2

Weight: 0

SD Limit: 0.1

Duplicate Limit: 10 % 500

Sensitivity Range: -99999 - 999999

S1 Abs Range: -32000 - 32000

Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calic. Code:	#	#				
Concentration:	0					
Position:						
Sample Volume:	3	3	3	3	3	3
Diluted S. Vol.	0	0	0	0	0	0
Diluent Volume:	0	0	0	0	0	0

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

HDL Cholesterol Direct

HITACHI 912

Analyze

Select Test: HDL-C | Analyzer Cycle Time: 10 sec | Diluent: 0 | 0

Test Name: HDL-C | Assay / Time / Point: 2 Point End | 10 | 15 | 31 | | |

App. Code: تغییر ندهید | Wavelength (2nd / Primary): 700 | 600

Sample Volume

	Class 1			Class 2		
Normal:	3	0	0	3	0	0
Decrease:	3	0	0	3	0	0
Increase:	6	0	0	6	0	0

	Class 1	Class 2	Dec/Inc
Abs.Limit	0	0	Increase
Prozone Limit	0	0	Lower

Reagent

R1:	240	0	00381	0
R2:	0	0	00381	0
R3:	80	0	00381	0
R4:	0	0	00381	0

Twin Test:

Calibration

Calibration Type: Linear

Point: 2 | Span Point: 2

Weight: 0

SD Limit: 0.1

Duplicate Limit: 10 % 500

Sensitivity Range: -99999 - 999999

S1 Abs Range: -32000 - 32000

Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calic. Code:	#	#				
Concentration:	0					
Position:						
Sample Volume:	3	3	3	3	3	3
Diluted S. Vol.	0	0	0	0	0	0
Diluent Volume:	0	0	0	0	0	0

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Immunoglobulin A — HITACHI 912

Analyze

Select Test: IgA | Analyzer Cycle Time: 10 sec | Diluent: 0 | 0

Test Name: IgA | Assay / Time / Point: 2 Point End | 10 | 15 | 31 | | |

App. Code: تغییر ندهید | Wavelength (2nd / Primary): 700 | 340

Sample Volume

	Class 1			Class 2		
Normal:	3	0	0	3	0	0
Decrease:	3	0	0	3	0	0
Increase:	6	0	0	6	0	0

	Class 1	Class 2	Dec/Inc
Abs.Limit	0	0	Increase
Prozone Limit	0	0	Lower

Reagent

R1:	250	0	00383	0
R2:	0	0	00383	0
R3:	50	0	00383	0
R4:	0	0	00383	0

Twin Test:

Calibration

Calibration Type: Logit-Log (4P)

Point: 6 | Span Point: 6

Weight: 0

SD Limit: 999.9

Duplicate Limit: 99 % 2000

Sensitivity Range: -99999 - 999999

S1 Abs Range: -32000 - 32000

Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calic. Code:	#	#	#	#	#	#
Concentration:	0					
Position:						
Sample Volume:	3	3	3	3	3	3
Diluted S. Vol.	0	0	0	0	0	0
Diluent Volume:	0	0	0	0	0	0

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Immunoglobulin G

HITACHI 912

Analyze

Select Test: IgG | Analyzer Cycle Time: 10 sec | Diluent: 0 | 0

Test Name: IgG | Assay / Time / Point: 2 Point End | 10 | 15 | 31 | | |

App. Code: تغییر ندهید | Wavelength (2nd / Primary): Cancel | 600

Sample Volume

	Class 1			Class 2		
Normal:	3	0	0	3	0	0
Decrease:	3	0	0	3	0	0
Increase:	6	0	0	6	0	0

	Class 1	Class 2	Dec/Inc
Abs.Limit	0	0	Increase
Prozone Limit	0	0	Lower

Reagent

R1:	250	0	00384	0
R2:	0	0	00384	0
R3:	50	0	00384	0
R4:	0	0	00384	0

Twin Test

Calibration

Calibration Type: Logit-Log (4P)

Point: 6 | Span Point: 6

Weight: 0

SD Limit: 999.9

Duplicate Limit: 99 % 2000

Sensitivity Range: -99999 - 999999

S1 Abs Range: -32000 - 32000

Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calic. Code:	#	#	#	#	#	#
Concentration:	0					
Position:						
Sample Volume:	3	3	3	3	3	3
Diluted S. Vol.	0	0	0	0	0	0
Diluent Volume:	0	0	0	0	0	0

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Immunoglobulin M

HITACHI 912

Analyze

Select Test: IgM | Analyzer Cycle Time: 10 sec | Diluent: 0 | 0

Test Name: IgM | Assay / Time / Point: 2 Point End | 10 | 15 | 31 | | |

App. Code: تغییر ندهید | Wavelength (2nd / Primary): 700 | 340

Sample Volume

	Class 1			Class 2		
Normal:	3	0	0	3	0	0
Decrease:	3	0	0	3	0	0
Increase:	6	0	0	6	0	0

	Class 1	Class 2	Dec/Inc
Abs.Limit	0	0	Increase
Prozone Limit	0	0	Lower

Reagent

R1:	250	0	00385	0
R2:	0	0	00385	0
R3:	50	0	00385	0
R4:	0	0	00385	0

Twin Test:

Calibration

Calibration Type: Logit-Log (4P)

Point: 6 | Span Point: 6

Weight: 0

SD Limit: 999.9

Duplicate Limit: 99 % 2000

Sensitivity Range: -99999 - 999999

S1 Abs Range: -32000 - 32000

Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calic. Code:	#	#	#	#	#	#
Concentration:	0					
Position:						
Sample Volume:	3	3	3	3	3	3
Diluted S. Vol.	0	0	0	0	0	0
Diluent Volume:	0	0	0	0	0	0

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Lactate Dehydrogenase — HITACHI 912

Analyze

Select Test: LDH | Analyzer Cycle Time: 10 sec | Diluent: 0 | 0

Test Name: LDH | Assay / Time / Point: Rate-A | 10 | 19 | 28 | | |

App. Code: تغییر ندهید | Wavelength (2nd / Primary): 415 | 340

Sample Volume

	Class 1			Class 2		
Normal:	4	0	0	4	0	0
Decrease:	3	0	0	3	0	0
Increase:	8	0	0	8	0	0

Reagent

R1:	200	0	388	0
R2:	0	0	388	0
R3:	40	0	388	0
R4:	0	0	388	0

	Class 1	Class 2	Dec/Inc
Abs.Limit	3000	3000	Decrease
Prozone Limit	0	0	Lower

Twin Test:

Calibration

Calibration Type: Linear

Point: 2 | Span Point: 2

Weight: 0

SD Limit: 0.1

Duplicate Limit: 10 % 500

Sensitivity Range: -99999 - 999999

S1 Abs Range: -32000 - 32000

Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calic. Code:	#	#				
Concentration:	0					
Position:						
Sample Volume:	4	4	4	4	4	4
Diluted S. Vol.	0	0	0	0	0	0
Diluent Volume:	0	0	0	0	0	0

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

LDL Cholesterol Direct

HITACHI 912

Analyze

Select Test: LDL-C | Analyzer Cycle Time: 10 sec | Diluent: 0 | 0

Test Name: LDL-C | Assay / Time / Point: 2 Point End | 10 | 15 | 31

App. Code: تغییر ندهید | Wavelength (2nd / Primary): 700 | 546

Sample Volume

	Class 1			Class 2		
Normal:	4	0	0	3	0	0
Decrease:	3	0	0	3	0	0
Increase:	8	0	0	6	0	0

	Class 1	Class 2	Dec/Inc
Abs.Limit	0	0	Increase
Prozone Limit	0	0	Lower

Reagent

R1:	240	0	00389	0
R2:	0	0	00389	0
R3:	80	0	00389	0
R4:	0	0	00389	0

Twin Test

Calibration

Calibration Type: Linear

Point: 2 | Span Point: 2

Weight: 0

SD Limit: 0.1

Duplicate Limit: 10 % 500

Sensitivity Range: -99999 - 999999

S1 Abs Range: -32000 - 32000

Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calic. Code:	#	#				
Concentration:	0					
Position:						
Sample Volume:	4	4	4	4	4	4
Diluted S. Vol.	0	0	0	0	0	0
Diluent Volume:	0	0	0	0	0	0

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Lipase

HITACHI 912

Analyze

Select Test: LIP | Analyzer Cycle Time: 10 sec | Diluent: 0 | 0

Test Name: LIP | Assay / Time / Point: Rate-A | 10 | 21 | 27

App. Code: تغییر ندهید | Wavelength (2nd / Primary): 700 | 570

Sample Volume

	Class 1			Class 2		
Normal:	6	0	0	6	0	0
Decrease:	3	0	0	3	0	0
Increase:	12	0	0	12	0	0

	Class 1	Class 2	Dec/Inc
Abs.Limit	10000	10000	Increase
Prozone Limit	0	0	Lower

Reagent

R1:	240	0	390	0
R2:	0	0	390	0
R3:	60	0	390	0
R4:	0	0	390	0

Twin Test

Calibration

Calibration Type: Linear

Point: 2 | Span Point: 2

Weight: 0

SD Limit: 0.1

Duplicate Limit: 10 % 500

Sensitivity Range: -99999 - 999999

S1 Abs Range: -32000 - 32000

Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calic. Code:	#	#				
Concentration:	0					
Position:						
Sample Volume:	6	6	6	6	6	6
Diluted S. Vol.	0	0	0	0	0	0
Diluent Volume:	0	0	0	0	0	0

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Micro Albumin — HITACHI 912

Analyze

Select Test: MALB | Analyzer Cycle Time: 10 sec | Diluent: 0 | 0

Test Name: MALB | Assay / Time / Point: 2 Point End | 10 | 15 | 31 | | |

App. Code: تغییر ندهید | Wavelength (2nd / Primary): 700 | 340

Sample Volume

	Class 1			Class 2		
Normal:	18	0	0	18	0	0
Decrease:	3	0	0	3	0	0
Increase:	35	0	0	35	0	0

	Class 1	Class 2	Dec/Inc
Abs.Limit	0	0	Increase
Prozone Limit	0	0	Lower

Reagent

R1:	250	0	392	0
R2:	0	0	392	0
R3:	50	0	392	0
R4:	0	0	392	0

Twin Test:

Calibration

Calibration Type: Logit-Log (4P)

Point: 6 | Span Point: 6

Weight: 0

SD Limit: 999.9

Duplicate Limit: 99 % 2000

Sensitivity Range: -99999 - 999999

S1 Abs Range: -32000 - 32000

Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calic. Code:	#	#	#	#	#	#
Concentration:	0.0					
Position:						
Sample Volume:	18	18	18	18	18	18
Diluted S. Vol.	0	0	0	0	0	0
Diluent Volume:	0	0	0	0	0	0

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

CK-MB **HITACHI 912**

Analyze

Select Test: CK-MB | Analyzer Cycle Time: 10 sec | Diluent: 0 | 0

Test Name: CK-MB | Assay / Time / Point: Rate-A | 10 | 22 | 31 | | |

App. Code: تغییر ندهید | Wavelength (2nd / Primary): 546 | 340

Sample Volume

	Class 1			Class 2		
Normal:	10	0	0	10	0	0
Decrease:	3	0	0	3	0	0
Increase:	20	0	0	20	0	0

	Class 1	Class 2	Dec/Inc
Abs.Limit	10000	10000	Increase
Prozone Limit	0	0	Lower

Reagent

R1:	200	0	367	0
R2:	0	0	367	0
R3:	40	0	367	0
R4:	0	0	367	0

Twin Test

Calibration

Calibration Type: Linear

Point: 2 | Span Point: 2

Weight: 0

SD Limit: 0.1

Duplicate Limit: 10 % 500

Sensitivity Range: -99999 - 999999

S1 Abs Range: -32000 - 32000

Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calic. Code:	#	#				
Concentration:	0.0					
Position:						
Sample Volume:	10	10	10	10	10	10
Diluted S. Vol.	0	0	0	0	0	0
Diluent Volume:	0	0	0	0	0	0

شرکت توسعه کیمیای سعادت www.tksmed.ir

شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Magnesium

HITACHI 912

Analyze

Select Test: MG | Analyzer Cycle Time: 10 sec | Diluent: 0 | 0

Test Name: MG | Assay / Time / Point: 1 Point | 3 | 10 | 0

App. Code: تغییر ندهید | Wavelength (2nd / Primary): 660 | 505

Sample Volume

	Class 1			Class 2		
Normal:	3	0	0	3	0	0
Decrease:	3	0	0	3	0	0
Increase:	6	0	0	6	0	0

	Class 1	Class 2	Dec/Inc
Abs.Limit	0	0	Increase
Prozone Limit	0	0	Lower

Reagent

R1:	300	0	393	0
R2:	0	0	393	0
R3:	0	0	393	0
R4:	0	0	393	0

Twin Test:

Calibration

Calibration Type: Linear

Point: 2 | Span Point: 2

Weight: 0

SD Limit: 0.1

Duplicate Limit: 10 % 500

Sensitivity Range: -99999 - 999999

S1 Abs Range: -32000 - 32000

Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calic. Code:	#	#				
Concentration:	0.0					
Position:						
Sample Volume:	3	3	3	3	3	3
Diluted S. Vol.	0	0	0	0	0	0
Diluent Volume:	0	0	0	0	0	0

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Phosphorus

HITACHI 912

Analyze

Select Test: PHOS | Analyzer Cycle Time: 10 sec | Diluent: 0 | 0

Test Name: PHOS | Assay / Time / Point: 1 Point | 5 | 15 | 0

App. Code: تغییر ندهید | Wavelength (2nd / Primary): 700 | 340

Sample Volume

	Class 1			Class 2		
Normal:	3	0	0	3	0	0
Decrease:	3	0	0	3	0	0
Increase:	6	0	0	6	0	0

	Class 1	Class 2	Dec/Inc
Abs.Limit	0	0	Increase
Prozone Limit	0	0	Lower

Reagent

R1:	300	0	395	0
R2:	0	0	395	0
R3:	0	0	395	0
R4:	0	0	395	0

Twin Test:

Calibration

Calibration Type: Linear

Point: 2 | Span Point: 2

Weight: 0

SD Limit: 0.1

Duplicate Limit: 10 % 500

Sensitivity Range: -99999 - 999999

S1 Abs Range: -32000 - 32000

Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calic. Code:	#	#				
Concentration:	0.0					
Position:						
Sample Volume:	3	3	3	3	3	3
Diluted S. Vol.	0	0	0	0	0	0
Diluent Volume:	0	0	0	0	0	0

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Triglycerides — HITACHI 912

Analyze

Select Test: TG | Analyzer Cycle Time: 10 sec | Diluent: 0 | 0

Test Name: TG | Assay / Time / Point: 1 Point | 10 | 31 | 0

App. Code: تغییر ندهید | Wavelength (2nd / Primary): 700 | 505

Sample Volume

	Class 1			Class 2		
Normal:	3	0	0	3	0	0
Decrease:	3	0	0	3	0	0
Increase:	6	0	0	6	0	0

	Class 1	Class 2	Dec/Inc
Abs.Limit	0	0	Increase
Prozone Limit	0	0	Lower

Reagent

R1:	300	0	398	0
R2:	0	0	398	0
R3:	0	0	398	0
R4:	0	0	398	0

Twin Test:

Calibration

Calibration Type: Linear

Point: 2 | Span Point: 2

Weight: 0

SD Limit: 0.1

Duplicate Limit: 10 % 500

Sensitivity Range: -99999 - 999999

S1 Abs Range: -32000 - 32000

Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calic. Code:	#	#				
Concentration:	0					
Position:						
Sample Volume:	3	3	3	3	3	3
Diluted S. Vol.	0	0	0	0	0	0
Diluent Volume:	0	0	0	0	0	0

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Total Protein

HITACHI 912

Analyze

Select Test: T.PRO | Analyzer Cycle Time: 10 sec | Diluent: 0 | 0

Test Name: T.PRO | Assay / Time / Point: 1 Point | 5 | 15 | 0

App. Code: تغییر ندهید | Wavelength (2nd / Primary): 700 | 546

Sample Volume

	Class 1			Class 2		
Normal:	8	0	0	8	0	0
Decrease:	3	0	0	3	0	0
Increase:	16	0	0	16	0	0

Reagent

R1:	300	0	397	0
R2:	0	0	397	0
R3:	0	0	397	0
R4:	0	0	397	0

	Class 1	Class 2	Dec/Inc
Abs.Limit	0	0	Increase
Prozone Limit	0	0	Lower

Twin Test:

Calibration

Calibration Type: Linear

Point: 2 | Span Point: 2

Weight: 0

SD Limit: 0.1

Duplicate Limit: 10 % 500

Sensitivity Range: -99999 - 999999

S1 Abs Range: -32000 - 32000

Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calic. Code:	#	#				
Concentration:	0.0					
Position:						
Sample Volume:	8	8	8	8	8	8
Diluted S. Vol.	0	0	0	0	0	0
Diluent Volume:	0	0	0	0	0	0

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Uric acid — HITACHI 912

Analyze

Select Test: UA | Analyzer Cycle Time: 10 sec | Diluent: 0 | 0

Test Name: UA | Assay / Time / Point: 1 Point | 5 | 15 | 0

App. Code: تغییر ندهید | Wavelength (2nd / Primary): 700 | 546

Sample Volume

	Class 1			Class 2		
Normal:	6	0	0	6	0	0
Decrease:	3	0	0	3	0	0
Increase:	12	0	0	12	0	0

	Class 1	Class 2	Dec/Inc
Abs.Limit	0	0	Increase
Prozone Limit	0	0	Lower

Reagent

R1:	120	0	402	0
R2:	120	0	402	0
R3:	0	0	402	0
R4:	0	0	402	0

Twin Test:

Calibration

Calibration Type: Linear

Point: 2 | Span Point: 2

Weight: 0

SD Limit: 0.1

Duplicate Limit: 10 % 500

Sensitivity Range: -99999 - 999999

S1 Abs Range: -32000 - 32000

Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calic. Code:	#	#				
Concentration:	0.0					
Position:						
Sample Volume:	6	6	6	6	6	6
Diluted S. Vol.	0	0	0	0	0	0
Diluent Volume:	0	0	0	0	0	0

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Urea/BUN

HITACHI 912

Analyze

Select Test: UREA | Analyzer Cycle Time: 10 sec | Diluent: 0 | 0

Test Name: UREA | Assay / Time / Point: 2 Point Rate | 4 | 8 | 12 | | |

App. Code: تغییر ندهید | Wavelength (2nd / Primary): 415 | 340

Sample Volume

	Class 1			Class 2		
Normal:	3	0	0	3	0	0
Decrease:	3	0	0	3	0	0
Increase:	6	0	0	6	0	0

	Class 1	Class 2	Dec/Inc
Abs.Limit	3000	3000	Decrease
Prozone Limit	0	0	Lower

Reagent

R1:	240	0	403	0
R2:	60	0	403	0
R3:	0	0	403	0
R4:	0	0	403	0

Twin Test:

Calibration

Calibration Type: Linear

Point: 2 | Span Point: 2

Weight: 0

SD Limit: 0.1

Duplicate Limit: 10 % 500

Sensitivity Range: -99999 - 999999

S1 Abs Range: -32000 - 32000

Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calic. Code:	#	#				
Concentration:	0					
Position:						
Sample Volume:	3	3	3	3	3	3
Diluted S. Vol.	0	0	0	0	0	0
Diluent Volume:	0	0	0	0	0	0

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

ZINC **HITACHI 912**

Analyze

Select Test: ZN | Analyzer Cycle Time: 10 sec | Diluent: 0 | 0

Test Name: ZN | Assay / Time / Point: 1 Point | 3 | 10 | 0

App. Code: تغییر ندهید | Wavelength (2nd / Primary): 700 | 570

Sample Volume

	Class 1			Class 2		
Normal:	15	0	0	15	0	0
Decrease:	5	0	0	5	0	0
Increase:	30	0	0	30	0	0

	Class 1	Class 2	Dec/Inc
Abs.Limit	0	0	Increase
Prozone Limit	0	0	Lower

Reagent

R1:	300	0	404	0
R2:	0	0	404	0
R3:	0	0	404	0
R4:	0	0	404	0

Twin Test

Calibration

Calibration Type: Linear

Point: 2 | Span Point: 2

Weight: 0

SD Limit: 0.1

Duplicate Limit: 10 % 500

Sensitivity Range: -99999 - 999999

S1 Abs Range: -32000 - 32000

Others

<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Calic. Code:	#	#				
Concentration:	0					
Position:						
Sample Volume:	15	15	15	15	15	15
Diluted S. Vol.	0	0	0	0	0	0
Diluent Volume:	0	0	0	0	0	0

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸