BIOMEDIC

BS 200

mindray
healthcare within reach

Parameters	Glucose	UREA	Creatinine Jaffe	Uric Acid	Cholesterol	Triglycerides	HDL Cholesterol Direct	LDL Cholesterol Direct	Alanine Aminotransferase (ALT)	Aspartate Aminotransferase (AST)	Alkaline Phosphatase	Creatine Kinase (CK-NAC)	CK - MB	Lactate Dehydrogenase (LDH)	Albumin BCG	Total Protein	Calcium Arsenazo III	Amylase CNPG3	Lipase	Phosphorous UV	Magnesium Xylidyl Blue	ZINC	Copper	C - reactive protein (CRP)	Microalbumin	Complement 3 (C3)	Complement 4 (C4)	Immunoglobulin A (IgA)	Immunoglobulin G (IgG)	Immunoglobulin M (IgM)
Test	GLU	UREA	CRT	UA	CHOL	TG	HDL	LDL	ALT	AST	ALP	CPK	MB	LDH	ALB	TPRO	CA	AMY	LPS	PHOS	MG	ZINC	CU	CRP	MALB	C3	C4	IgA	IgG	IgM
No.	#	#	#	#	#	#	#	#	#	#	#	#	#	#	#	#	#	#	#	#	#	#	#	#	#	#	#	#	#	#
Full Name	Glucose	UREA	Creatinine	Uric Acid	Cholesterol	Triglycerides	HDL	LDL	ALT	AST	ALP	CPK	CK - MB	LDH	Albumin BCG	Total Protein	Calcium	Amylase	Lipase	Phosphorous	Magnesium	ZINC	Copper	CRP	MALB	C3	C4	IgA	IgG	IgM
Standard No.																														
Reac. Type	Endpoint	Fixed-time	Fixed-time	Endpoint	Endpoint	Endpoint	Endpoint	Endpoint	Kinetic	Kinetic	Kinetic	Kinetic	Kinetic	Kinetic	Endpoint	Endpoint	Endpoint	Kinetic	Kinetic	Endpoint	Endpoint	Endpoint	Endpoint	Endpoint	Endpoint	Endpoint	Endpoint	Endpoint	Endpoint	Endpoint
Pri. Wave	510 nm	340 nm	510 nm	546 nm	510 nm	510 nm	578 nm	546 nm	340 nm	340 nm	405 nm	340 nm	340 nm	340 nm	630 nm	546 nm	670 nm	405 nm	578 nm	340 nm	546 nm	578 nm	578 nm	340 nm	340 nm	340 nm	340 nm	340 nm	630 nm	340 nm
Sec. Wave																														
Direction	Increase	Decrease	Increase	Increase	Increase	Increase	Increase	Increase	Decrease	Decrease	Increase	Increase	Increase	Decrease	Increase	Increase	Increase	Increase	Increase	Increase	Increase	Increase	Increase	Increase	Increase	Increase	Increase	Increase	Increase	Increase
Reac. Time	0 - 25	3 - 9	3 - 9	0 - 10	0 - 25	0 - 25	-1 - 25	-1 - 25	4 - 15	4 - 15	4 - 15	10 - 20	10 - 20	4 - 15	0 - 10	0 - 10	0 - 10	3 - 14	10 - 20	0 - 10	0 - 10	0 - 10	0 - 10	-1 - 18	-1 - 18	-1 - 18	-1 - 18	-1 - 18	-1 - 18	-1 - 18	-1 - 18
Incuba. Time		5	5	5			15	15	5	5	5	10	10	10					5					15	15	10	10	10	10	10
Unit	mg/dl	mg/dl	mg/dl	mg/dl	mg/dl	mg/dl	mg/dl	mg/dl	mg/dl	mg/dl	mg/dl	U/L	U/L	U/L	g/dl	g/dl	mg/dl	U/L	U/L	mg/dl	mg/dl	µg/dl	µg/dl	mg/l	mg/l	mg/dl	mg/dl	mg/dl	mg/dl	mg/dl
Precision	Integer	0.1	0.01	0.01	Integer	Integer	Integer	Integer	Integer	Integer	Integer	Integer	0.1	Integer	0.1	0.1	0.1	Integer	Integer	0.1	0.01	Integer	Integer	0.1	0.1	Integer	0.1	Integer	Integer	Integer
R1	300	240	180	180	300	300	240	240	240	240	240	250	250	250	300	300	300	300	240	300	300	300	300	250	250	250	250	250	250	250
R2	0	60	180	180	0	0	80	80	60	60	60	50	50	50	0	0	0	0	60	0	0	0	0	50	50	50	50	50	50	50
Sample Volume	3	3	36	9	3	3	3	4	30	30	6	15	15	5	3	6	3	6	6	3	3	15	15	15	18	3	6	3	3	3
R1 Blank																														
Mixed Rgt. Blank																														
Linearity Range	0 - 400	0 - 200	0 - 20	0 - 20	0 - 600	0 - 600	0 - 120	0 - 600	0 - 1000	0 - 300	0 - 300	0 - 800	0 - 800	0 - 1000	0 - 5	0 - 12	0 - 20	0 - 1500	0 - 150	0 - 20	0 - 7	0 - 400	0 - 400	0 - 180	0 - 280	0 - 500	0 - 80	0 - 800	0 - 4000	0 - 350
Linearity Limit									0.2	0.2	0.2	0.2	0.2	0.2				0.2	0.2											
Substrate Limit																														
Factor																														
Compensate																														
Slope	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
Intercept	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0
Calibration																														
Rule	Two-Point Linear	Two-Point Linear	Two-Point Linear	Two-Point Linear	Two-Point Linear	Two-Point Linear	Two-Point Linear	Two-Point Linear	Two-Point Linear	Two-Point Linear	Two-Point Linear	Two-Point Linear	Two-Point Linear	Two-Point Linear	Two-Point Linear	Two-Point Linear	Two-Point Linear	Two-Point Linear	Two-Point Linear	Two-Point Linear	Two-Point Linear	Two-Point Linear	Two-Point Linear	Logit-Log 4P	Logit-Log 4P	Logit-Log 4P	Logit-Log 4P	Logit-Log 4P	Logit-Log 4P	Logit-Log 4P

شرکت توسعه کیمیای سعادت www.tksmed.ir
شکرالهی ۰۹۱۲۹۵۷۹۴۰۸