BIOMEDIC

mindray
healthcare within reach

BS 430

# ALBUMIN

| | | | | | |
|---|---|---|---|---|---|
| Chem | ALB | No. | | Sample Type | Serum |
| Chemistry | ALBUMIN | | | Print Name | ALB |
| Reaction Type | Endpoint | | | Reaction Direction | Positive |
| Pri Wave | 605nm | | | Sec Wave | 700nm |
| Unit | g/dl | | | Decimal | 0.1 |
| Blank Time | 4 | 5 | | Reaction Time | 20 | 22 |

| | Sample Vol | Aspirated | Diluent | | Reagent Vol | Diluent |
|---|---|---|---|---|---|---|
| Standard | 2 uL | uL | uL | R1 | 200 uL | uL |
| Decreased | 1.5 uL | uL | uL | R2 | uL | uL |
| Increased | 4 uL | uL | uL | R3 | uL | uL |
| | | | | R4 | uL | uL |

**Calibration Settings**

| | | | |
|---|---|---|---|
| Math Model | Two-point linear | | |
| Factor | | Replicates | 2 |

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی 09129579409 - 09129579408

BIOMEDIC

mindray
healthcare within reach

**BS 430**

# ALKALINE PHOSPHATASE

| | | | | | |
|---|---|---|---|---|---|
| Chem | ALP | No. | | Sample Type | Serum ↓ |
| Chemistry | ALKALINE PHOSPHATASE | | | Print Name | ALP |
| Reaction Type | Kinetic ↓ | | | Reaction Direction | Positive ↓ |
| Pri Wave | 412nm ↓ | | | Sec Wave | 660nm ↓ |
| Unit | U/L ↓ | | | Decimal | 0 ↓ |
| Blank Time | 0 | 0 | | Reaction Time | 25 | 35 |

| | Sample Vol | Aspirated | Diluent | | Reagent Vol | Diluent |
|---|---|---|---|---|---|---|
| Standard | 4 uL | uL | uL | R1 | 160 uL | uL |
| Decreased | 2 uL | uL | uL | R2 | 40 uL | uL |
| Increased | 8 uL | uL | uL | R3 | uL | uL |
| | | | | R4 | uL | uL |

**Calibration Settings**

| | | | |
|---|---|---|---|
| Math Model | Two-point linear ↓ | | |
| Factor | | Replicates | 2 ↓ |

**www.tksmed.ir شرکت توسعه کیمیای سعادت**

**09129579408 - 09129579409 پشتیبانی**

BIOMEDIC

mindray
healthcare within reach

BS 430

# Alanine Aminotransferase

| | | | | | |
|---|---|---|---|---|---|
| Chem | ALT | No. | | Sample Type | Serum |
| Chemistry | Alanine Aminotransferase | | | Print Name | ALT |
| Reaction Type | Kinetic | | | Reaction Direction | negative |
| Pri Wave | 340nm | | | Sec Wave | 412nm |
| Unit | U/L | | | Decimal | 0 |
| Blank Time | 0 | 0 | | Reaction Time | 25 | 35 |

| | Sample Vol | Aspirated | Diluent |
|---|---|---|---|
| Standard | 15 uL | uL | uL |
| Decreased | 5 uL | uL | uL |
| Increased | 20 uL | uL | uL |

| | Reagent Vol | Diluent |
|---|---|---|
| R1 | 120 uL | uL |
| R2 | 30 uL | uL |
| R3 | uL | uL |
| R4 | uL | uL |

**Calibration Settings**

| | | | |
|---|---|---|---|
| Math Model | Two-point linear | | |
| Factor | | Replicates | 2 |

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی 09129579409 - 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 430

# AMYLASE

| | | | | | |
|---|---|---|---|---|---|
| Chem | AMY | No. | | Sample Type | Serum |
| Chemistry | AMYLASE | | | Print Name | AMY |
| Reaction Type | Kinetic | | | Reaction Direction | Positive |
| Pri Wave | 412nm | | | Sec Wave | 700nm |
| Unit | U/L | | | Decimal | 0 |
| Blank Time | 0 | 0 | | Reaction Time | 8 | 18 |

| | Sample Vol | Aspirated | Diluent | | Reagent Vol | Diluent |
|---|---|---|---|---|---|---|
| Standard | 4 uL | uL | uL | R1 | 200 uL | uL |
| Decreased | 2 uL | uL | uL | R2 | uL | uL |
| Increased | 8 uL | uL | uL | R3 | uL | uL |
| | | | | R4 | uL | uL |

**Calibration Settings**

| | | | |
|---|---|---|---|
| Math Model | Two-point linear | | |
| Factor | | Replicates | 2 |

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی 09129579409 - 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 430

# Aspartate Aminotransferase

| | | | | | |
|---|---|---|---|---|---|
| Chem | AST | No. | | Sample Type | Serum |
| Chemistry | Aspartate Aminotransferase | | | Print Name | AST |
| Reaction Type | Kinetic | | | Reaction Direction | negative |
| Pri Wave | 340nm | | | Sec Wave | 412nm |
| Unit | U/L | | | Decimal | 0 |
| Blank Time | 0 | 0 | | Reaction Time | 25 | 35 |

| | Sample Vol | Aspirated | Diluent |
|---|---|---|---|
| Standard | 15 uL | uL | uL |
| Decreased | 5 uL | uL | uL |
| Increased | 20 uL | uL | uL |

| | Reagent Vol | Diluent |
|---|---|---|
| R1 | 120 uL | uL |
| R2 | 30 uL | uL |
| R3 | uL | uL |
| R4 | uL | uL |

**Calibration Settings**

| | | | |
|---|---|---|---|
| Math Model | Two-point linear | | |
| Factor | | Replicates | 2 |

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی 09129579409 - 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 430

# Complement C3

| | | | | | |
|---|---|---|---|---|---|
| Chem | C3 | No. | | Sample Type | Serum |
| Chemistry | Complement C3 | | | Print Name | C3 |
| Reaction Type | Endpoint | | | Reaction Direction | Positive |
| Pri Wave | 340nm | | | Sec Wave | 700nm |
| Unit | mg/dl | | | Decimal | 0 |
| Blank Time | 20 | 22 | | Reaction Time | 37 | 39 |

| | Sample Vol | Aspirated | Diluent | | Reagent Vol | Diluent |
|---|---|---|---|---|---|---|
| Standard | 2 uL | uL | uL | R1 | 160 uL | uL |
| Decreased | 1.5 uL | uL | uL | R2 | 30 uL | uL |
| Increased | 10 uL | uL | uL | R3 | uL | uL |
| | | | | R4 | uL | uL |

**Calibration Settings**

| | | | |
|---|---|---|---|
| Math Model | Logit-Log 4P | | |
| Factor | | Replicates | 1 |

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی 09129579409 - 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 430

# Complement C4

| | | | | | |
|---|---|---|---|---|---|
| Chem | C4 | No. | | Sample Type | Serum |
| Chemistry | Complement C4 | | | Print Name | C4 |
| Reaction Type | Endpoint | | | Reaction Direction | Positive |
| Pri Wave | 340nm | | | Sec Wave | 700nm |
| Unit | mg/dl | | | Decimal | 0.1 |
| Blank Time | 20 | 22 | | Reaction Time | 37 | 39 |

| | Sample Vol | Aspirated | Diluent | | Reagent Vol | Diluent |
|---|---|---|---|---|---|---|
| Standard | 4 uL | uL | uL | R1 | 160 uL | uL |
| Decreased | 2 uL | uL | uL | R2 | 30 uL | uL |
| Increased | 10 uL | uL | uL | R3 | uL | uL |
| | | | | R4 | uL | uL |

**Calibration Settings**

| | | | |
|---|---|---|---|
| Math Model | Logit-Log 4P | | |
| Factor | | Replicates | 1 |

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی 09129579409 - 09129579408

BIOMEDIC

mindray
healthcare within reach

**BS 430**

# Calcium Arsenazo III

| | | | | | |
|---|---|---|---|---|---|
| Chem | CA | No. | | Sample Type | Serum |
| Chemistry | Calcium Arsenazo III | | | Print Name | CA |
| Reaction Type | Endpoint | | | Reaction Direction | Positive |
| Pri Wave | 660nm | | | Sec Wave | 700nm |
| Unit | mg/dl | | | Decimal | 0.1 |
| Blank Time | 4 | 5 | | Reaction Time | 20 | 22 |

| | Sample Vol | Aspirated | Diluent | | Reagent Vol | Diluent |
|---|---|---|---|---|---|---|
| Standard | 2 uL | uL | uL | R1 | 200 uL | uL |
| Decreased | 1.5 uL | uL | uL | R2 | uL | uL |
| Increased | 4 uL | uL | uL | R3 | uL | uL |
| | | | | R4 | uL | uL |

**Calibration Settings**

| | | | |
|---|---|---|---|
| Math Model | Two-point linear | | |
| Factor | | Replicates | 2 |

**www.tksmed.ir شرکت توسعه کیمیای سعادت**

**09129579408 - 09129579409 پشتیبانی**

BIOMEDIC

mindray
healthcare within reach

BS 430

# Cholesterol

| | | | | | |
|---|---|---|---|---|---|
| Chem | CHOL | No. | | Sample Type | Serum |
| Chemistry | Cholesterol | | | Print Name | CHOL |
| Reaction Type | Endpoint | | | Reaction Direction | Positive |
| Pri Wave | 505 | | | Sec Wave | 700nm |
| Unit | mg/dl | | | Decimal | 0 |
| Blank Time | 4 | 5 | | Reaction Time | 37 | 39 |

| | Sample Vol | Aspirated | Diluent | | Reagent Vol | Diluent |
|---|---|---|---|---|---|---|
| Standard | 2 uL | uL | uL | R1 | 200 uL | uL |
| Decreased | 1.5 uL | uL | uL | R2 | uL | uL |
| Increased | 10 uL | uL | uL | R3 | uL | uL |
| | | | | R4 | uL | uL |

**Calibration Settings**

| | | | |
|---|---|---|---|
| Math Model | Two-point linear | | |
| Factor | | Replicates | 2 |

**www.tksmed.ir شرکت توسعه کیمیای سعادت**

**09129579408 - 09129579409 پشتیبانی**

BIOMEDIC

mindray
healthcare within reach

BS 430

# Creatine Kinase (CKNAC)

| | | | | | |
|---|---|---|---|---|---|
| Chem | CPK | No. | | Sample Type | Serum |
| Chemistry | Creatine Kinase (CKNAC) | | | Print Name | CPK |
| Reaction Type | Kinetic | | | Reaction Direction | Positive |
| Pri Wave | 340nm | | | Sec Wave | 412nm |
| Unit | U/L | | | Decimal | 0 |
| Blank Time | 0 | 0 | | Reaction Time | 29 | 39 |

| | Sample Vol | Aspirated | Diluent | | Reagent Vol | Diluent |
|---|---|---|---|---|---|---|
| Standard | 8 uL | uL | uL | R1 | 150 uL | uL |
| Decreased | 2 uL | uL | uL | R2 | 30 uL | uL |
| Increased | 16 uL | uL | uL | R3 | uL | uL |
| | | | | R4 | uL | uL |

**Calibration Settings**

| | | | |
|---|---|---|---|
| Math Model | Two-point linear | | |
| Factor | | Replicates | 2 |


**www.tksmed.ir شرکت توسعه کیمیای سعادت**

**09129579408 - 09129579409 پشتیبانی**

BIOMEDIC

mindray
healthcare within reach

BS 430

# CREATININE

| | | | | | |
|---|---|---|---|---|---|
| Chem | CREAT | No. | | Sample Type | Serum |
| Chemistry | CREATININE | | | Print Name | CREAT |
| Reaction Type | Fixed-time | | | Reaction Direction | Positive |
| Pri Wave | 505nm | | | Sec Wave | 660nm |
| Unit | mg/dl | | | Decimal | 0.01 |
| Blank Time | 0 | 0 | | Reaction Time | 24 | 29 |

| | Sample Vol | | Aspirated | | Diluent | | Reagent Vol | | Diluent | |
|---|---|---|---|---|---|---|---|---|---|---|
| Standard | 20 | uL | | uL | | uL | R1 | 100 | uL | | uL |
| Decreased | 2 | uL | | uL | | uL | R2 | 100 | uL | | uL |
| Increased | 30 | uL | | uL | | uL | R3 | | uL | | uL |
| | | | | | | | R4 | | uL | | uL |

**Calibration Settings**

| | | | |
|---|---|---|---|
| Math Model | Two-point linear | | |
| Factor | | Replicates | 2 |


شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی 09129579409 - 09129579408

BIOMEDIC

mindray
healthcare within reach

**BS 430**

# C-Reactive Protein (CRP)

| | | | | | |
|---|---|---|---|---|---|
| Chem | CRP | No. | | Sample Type | Serum |
| Chemistry | C-Reactive Protein (CRP) | | | Print Name | CRP |
| Reaction Type | Endpoint | | | Reaction Direction | Positive |
| Pri Wave | 340nm | | | Sec Wave | 700nm |
| Unit | mg/L | | | Decimal | 0.1 |
| Blank Time | 20 | 22 | | Reaction Time | 37 | 39 |

| | Sample Vol | Aspirated | Diluent | | Reagent Vol | Diluent |
|---|---|---|---|---|---|---|
| Standard | 8 uL | uL | uL | R1 | 160 uL | uL |
| Decreased | 4 uL | uL | uL | R2 | 30 uL | uL |
| Increased | 16 uL | uL | uL | R3 | uL | uL |
| | | | | R4 | uL | uL |

**Calibration Settings**

| | | | |
|---|---|---|---|
| Math Model | Spline | | |
| Factor | | Replicates | 1 |

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی 09129579408 - 09129579409

BIOMEDIC

mindray
healthcare within reach

BS 430

# GLUCOSE

| | | | | | |
|---|---|---|---|---|---|
| Chem | GLU | No. | | Sample Type | Serum |
| Chemistry | GLUCOSE | | | Print Name | GLU |
| Reaction Type | Endpoint | | | Reaction Direction | Positive |
| Pri Wave | 505nm | | | Sec Wave | 700nm |
| Unit | mg/dl | | | Decimal | 0 |
| Blank Time | 4 | 5 | | Reaction Time | 37 | 39 |

| | Sample Vol | Aspirated | Diluent | | Reagent Vol | Diluent |
|---|---|---|---|---|---|---|
| Standard | 2 uL | uL | uL | R1 | 200 uL | uL |
| Decreased | 1.5 uL | uL | uL | R2 | uL | uL |
| Increased | 4 uL | uL | uL | R3 | uL | uL |
| | | | | R4 | uL | uL |

**Calibration Settings**

| | | | |
|---|---|---|---|
| Math Model | Two-point linear | | |
| Factor | | Replicates | 2 |

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی 09129579409 - 09129579408

BIOMEDIC

mindray
healthcare within reach

**BS 430**

# HDL Cholesterol

| | | | | | |
|---|---|---|---|---|---|
| Chem | HDL | No. | | Sample Type | Serum |
| Chemistry | HDL Cholesterol | | | Print Name | HDL |
| Reaction Type | Endpoint | | | Reaction Direction | Positive |
| Pri Wave | 605nm | | | Sec Wave | 700nm |
| Unit | mg/dl | | | Decimal | 0 |
| Blank Time | 20 | 22 | | Reaction Time | 37 | 39 |

| | Sample Vol | Aspirated | Diluent | | Reagent Vol | Diluent |
|---|---|---|---|---|---|---|
| Standard | 2 uL | uL | uL | R1 | 150 uL | uL |
| Decreased | 1.5 uL | uL | uL | R2 | 50 uL | uL |
| Increased | 10 uL | uL | uL | R3 | uL | uL |
| | | | | R4 | uL | uL |

**Calibration Settings**

| | | | |
|---|---|---|---|
| Math Model | Two-point linear | | |
| Factor | | Replicates | 2 |

**www.tksmed.ir شرکت توسعه کیمیای سعادت**

**09129579408 - 09129579409 پشتیبانی**

BIOMEDIC

mindray
healthcare within reach

**BS 430**

# Immunoglobulin A

| | | | | | |
|---|---|---|---|---|---|
| Chem | IgA | No. | | Sample Type | Serum |
| Chemistry | Immunoglobulin A | | | Print Name | IgA |
| Reaction Type | Endpoint | | | Reaction Direction | Positive |
| Pri Wave | 340nm | | | Sec Wave | 700nm |
| Unit | mg/dl | | | Decimal | 0 |
| Blank Time | 20 | 22 | | Reaction Time | 37 | 39 |

| | Sample Vol | Aspirated | Diluent | | Reagent Vol | Diluent |
|---|---|---|---|---|---|---|
| Standard | 2 uL | uL | uL | R1 | 160 uL | uL |
| Decreased | 1.5 uL | uL | uL | R2 | 30 uL | uL |
| Increased | 10 uL | uL | uL | R3 | uL | uL |
| | | | | R4 | uL | uL |

**Calibration Settings**

| | | | |
|---|---|---|---|
| Math Model | Logit-Log 4P | | |
| Factor | | Replicates | 1 |

**www.tksmed.ir شرکت توسعه کیمیای سعادت**

**09129579408 - 09129579409 پشتیبانی**

BIOMEDIC

mindray
healthcare within reach

BS 430

# Immunoglobulin G

| | | | | | |
|---|---|---|---|---|---|
| Chem | IgG | No. | | Sample Type | Serum |
| Chemistry | Immunoglobulin G | | | Print Name | IgG |
| Reaction Type | Endpoint | | | Reaction Direction | Positive |
| Pri Wave | 605nm | | | Sec Wave | 0 |
| Unit | mg/dl | | | Decimal | 0 |
| Blank Time | 20 | 22 | | Reaction Time | 37 | 39 |

| | Sample Vol | Aspirated | Diluent | | Reagent Vol | Diluent |
|---|---|---|---|---|---|---|
| Standard | 2 uL | uL | uL | R1 | 160 uL | uL |
| Decreased | 1.5 uL | uL | uL | R2 | 30 uL | uL |
| Increased | 10 uL | uL | uL | R3 | uL | uL |
| | | | | R4 | uL | uL |

**Calibration Settings**

| | | | |
|---|---|---|---|
| Math Model | Logit-Log 4P | | |
| Factor | | Replicates | 1 |

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی 09129579409 - 09129579408

BIOMEDIC

mindray
healthcare within reach

BS 430

# Immunoglobulin M

| | | | | | |
|---|---|---|---|---|---|
| Chem | IgM | No. | | Sample Type | Serum |
| Chemistry | Immunoglobulin M | | | Print Name | IgM |
| Reaction Type | Endpoint | | | Reaction Direction | Positive |
| Pri Wave | 340nm | | | Sec Wave | 700nm |
| Unit | mg/dl | | | Decimal | 0 |
| Blank Time | 20 | 22 | | Reaction Time | 37 | 39 |

| | Sample Vol | Aspirated | Diluent |
|---|---|---|---|
| Standard | 2 uL | uL | uL |
| Decreased | 1.5 uL | uL | uL |
| Increased | 10 uL | uL | uL |

| | Reagent Vol | Diluent |
|---|---|---|
| R1 | 160 uL | uL |
| R2 | 30 uL | uL |
| R3 | uL | uL |
| R4 | uL | uL |

**Calibration Settings**

| | | | |
|---|---|---|---|
| Math Model | Logit-Log 4P | | |
| Factor | | Replicates | 1 |

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی 09129579409 - 09129579408

BIOMEDIC

mindray
healthcare within reach

**BS 430**

# Lactate Dehydrogenase (LDH)

| | | | | | |
|---|---|---|---|---|---|
| Chem | LDH | No. | | Sample Type | Serum ↓ |
| Chemistry | Lactate Dehydrogenase (LDH) | | | Print Name | LDH |
| Reaction Type | Kinetic ↓ | | | Reaction Direction | negative ↓ |
| Pri Wave | 340nm ↓ | | | Sec Wave | 412nm ↓ |
| Unit | U/L ↓ | | | Decimal | 0 ↓ |
| Blank Time | 0 | 0 | | Reaction Time | 25 | 35 |

| | Sample Vol | Aspirated | Diluent | | Reagent Vol | Diluent |
|---|---|---|---|---|---|---|
| Standard | 3 uL | uL | uL | R1 | 150 uL | uL |
| Decreased | 1.5 uL | uL | uL | R2 | 30 uL | uL |
| Increased | 6 uL | uL | uL | R3 | uL | uL |
| | | | | R4 | uL | uL |

**Calibration Settings**

| | | | |
|---|---|---|---|
| Math Model | Two-point linear ↓ | | |
| Factor | | Replicates | 2 ↓ |

**www.tksmed.ir شرکت توسعه کیمیای سعادت**

**09129579408 - 09129579409 پشتیبانی**

BIOMEDIC

mindray
healthcare within reach

**BS 430**

# LDL Cholesterol

| | | | | | |
|---|---|---|---|---|---|
| Chem | LDL | No. | | Sample Type | Serum |
| Chemistry | LDL Cholesterol | | | Print Name | LDL |
| Reaction Type | Endpoint | | | Reaction Direction | Positive |
| Pri Wave | 605nm | | | Sec Wave | 700nm |
| Unit | mg/dl | | | Decimal | 0 |
| Blank Time | 20 | 22 | | Reaction Time | 37 | 39 |

| | Sample Vol | Aspirated | Diluent | | Reagent Vol | Diluent |
|---|---|---|---|---|---|---|
| Standard | 2 uL | uL | uL | R1 | 150 uL | uL |
| Decreased | 1.5 uL | uL | uL | R2 | 50 uL | uL |
| Increased | 10 uL | uL | uL | R3 | uL | uL |
| | | | | R4 | uL | uL |

**Calibration Settings**

| | | | |
|---|---|---|---|
| Math Model | Two-point linear | | |
| Factor | | Replicates | 2 |

**www.tksmed.ir شرکت توسعه کیمیای سعادت**

**پشتیبانی 09129579408 - 09129579409**

BIOMEDIC

mindray
healthcare within reach

**BS 430**

# Lipase

| | | | | | |
|---|---|---|---|---|---|
| Chem | LIP | No. | | Sample Type | Serum |
| Chemistry | Lipase | | | Print Name | LIP |
| Reaction Type | Kinetic | | | Reaction Direction | Positive |
| Pri Wave | 570nm | | | Sec Wave | 700nm |
| Unit | U/L | | | Decimal | 0 |
| Blank Time | 0 | 0 | | Reaction Time | 29 | 36 |

| | Sample Vol | | Aspirated | | Diluent | | | Reagent Vol | | Diluent | |
|---|---|---|---|---|---|---|---|---|---|---|---|
| Standard | 4 | uL | | uL | | uL | R1 | 160 | uL | | uL |
| Decreased | 2 | uL | | uL | | uL | R2 | 40 | uL | | uL |
| Increased | 8 | uL | | uL | | uL | R3 | | uL | | uL |
| | | | | | | | R4 | | uL | | uL |

**Calibration Settings**

| | | | |
|---|---|---|---|
| Math Model | Two-point linear | | |
| Factor | | Replicates | 2 |

**www.tksmed.ir شرکت توسعه کیمیای سعادت**

**09129579408 - 09129579409 پشتیبانی**

BIOMEDIC

mindray
healthcare within reach

BS 430

# Micro Albumin

| | | | | | |
|---|---|---|---|---|---|
| Chem | MALB | No. | | Sample Type | Serum |
| Chemistry | Micro Albumin | | | Print Name | MALB |
| Reaction Type | Endpoint | | | Reaction Direction | Positive |
| Pri Wave | 340nm | | | Sec Wave | 700nm |
| Unit | mg/L | | | Decimal | 0.1 |
| Blank Time | 20 | 22 | | Reaction Time | 37 | 39 |

| | Sample Vol | Aspirated | Diluent | | Reagent Vol | Diluent |
|---|---|---|---|---|---|---|
| Standard | 10 uL | uL | uL | R1 | 150 uL | uL |
| Decreased | 2 uL | uL | uL | R2 | 30 uL | uL |
| Increased | 20 uL | uL | uL | R3 | uL | uL |
| | | | | R4 | uL | uL |

**Calibration Settings**

| | | | |
|---|---|---|---|
| Math Model | Logit-Log 4P | | |
| Factor | | Replicates | 1 |

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی 09129579409 - 09129579408

BIOMEDIC

mindray
healthcare within reach

**BS 430**

# Creatine Kinase - MB

| | | | | | |
|---|---|---|---|---|---|
| Chem | CKMB | No. | | Sample Type | Serum |
| Chemistry | Creatine Kinase - MB | | | Print Name | CKMB |
| Reaction Type | Kinetic | | | Reaction Direction | Positive |
| Pri Wave | 340nm | | | Sec Wave | 412nm |
| Unit | U/L | | | Decimal | 0 |
| Blank Time | 0 | 0 | | Reaction Time | 29 | 39 |

| | Sample Vol | Aspirated | Diluent | | Reagent Vol | Diluent |
|---|---|---|---|---|---|---|
| Standard | 8 uL | uL | uL | R1 | 150 uL | uL |
| Decreased | 2 uL | uL | uL | R2 | 30 uL | uL |
| Increased | 16 uL | uL | uL | R3 | uL | uL |
| | | | | R4 | uL | uL |

**Calibration Settings**

| | | | |
|---|---|---|---|
| Math Model | Two-point linear | | |
| Factor | | Replicates | 2 |

**www.tksmed.ir شرکت توسعه کیمیای سعادت**

**09129579408 - 09129579409 پشتیبانی**

BIOMEDIC

mindray
healthcare within reach

**BS 430**

# Magnesium Xylidyl Blue

| | | | | | |
|---|---|---|---|---|---|
| Chem | MG | No. | | Sample Type | Serum |
| Chemistry | Magnesium Xylidyl Blue | | | Print Name | MG |
| Reaction Type | Endpoint | | | Reaction Direction | Positive |
| Pri Wave | 505nm | | | Sec Wave | 700nm |
| Unit | mg/dl | | | Decimal | 0.1 |
| Blank Time | 4 | 5 | | Reaction Time | 20 | 22 |

| | Sample Vol | Aspirated | Diluent | | Reagent Vol | Diluent |
|---|---|---|---|---|---|---|
| Standard | 2 uL | uL | uL | R1 | 200 uL | uL |
| Decreased | 1.5 uL | uL | uL | R2 | uL | uL |
| Increased | 4 uL | uL | uL | R3 | uL | uL |
| | | | | R4 | uL | uL |

**Calibration Settings**

| | | | |
|---|---|---|---|
| Math Model | Two-point linear | | |
| Factor | | Replicates | 2 |

**www.tksmed.ir شرکت توسعه کیمیای سعادت**

**09129579408 - 09129579409 پشتیبانی**

BIOMEDIC

mindray
healthcare within reach

BS 430

# Phosphorus

| | | | | | |
|---|---|---|---|---|---|
| Chem | PHOS | No. | | Sample Type | Serum |
| Chemistry | Phosphorus | | | Print Name | PHOS |
| Reaction Type | Endpoint | | | Reaction Direction | Positive |
| Pri Wave | 340nm | | | Sec Wave | 700nm |
| Unit | mg/dl | | | Decimal | 0.1 |
| Blank Time | 4 | 5 | | Reaction Time | 20 | 22 |

| | Sample Vol | Aspirated | Diluent | | Reagent Vol | Diluent |
|---|---|---|---|---|---|---|
| Standard | 2 uL | uL | uL | R1 | 200 uL | uL |
| Decreased | 1.5 uL | uL | uL | R2 | uL | uL |
| Increased | 4 uL | uL | uL | R3 | uL | uL |
| | | | | R4 | uL | uL |

**Calibration Settings**

| | | | |
|---|---|---|---|
| Math Model | Two-point linear | | |
| Factor | | Replicates | 2 |

**www.tksmed.ir شرکت توسعه کیمیای سعادت**

**09129579408 - 09129579409 پشتیبانی**

BIOMEDIC

mindray
healthcare within reach

BS 430

# Triglycerides

| | | | | | |
|---|---|---|---|---|---|
| Chem | TG | No. | | Sample Type | Serum |
| Chemistry | Triglycerides | | | Print Name | TG |
| Reaction Type | Endpoint | | | Reaction Direction | Positive |
| Pri Wave | 505 | | | Sec Wave | 700nm |
| Unit | mg/dl | | | Decimal | 0 |
| Blank Time | 4 | 5 | | Reaction Time | 37 | 39 |

| | Sample Vol | Aspirated | Diluent | | Reagent Vol | Diluent |
|---|---|---|---|---|---|---|
| Standard | 2 uL | uL | uL | R1 | 200 uL | uL |
| Decreased | 1.5 uL | uL | uL | R2 | uL | uL |
| Increased | 10 uL | uL | uL | R3 | uL | uL |
| | | | | R4 | uL | uL |

**Calibration Settings**

| | | | |
|---|---|---|---|
| Math Model | Two-point linear | | |
| Factor | | Replicates | 2 |

**www.tksmed.ir شرکت توسعه کیمیای سعادت**

**09129579408 - 09129579409 پشتیبانی**

BIOMEDIC

mindray
healthcare within reach

BS 430

# TOTAL PROTEIN

| | | | | | |
|---|---|---|---|---|---|
| Chem | TPRO | No. | | Sample Type | Serum |
| Chemistry | TOTAL PROTEIN | | | Print Name | TPRO |
| Reaction Type | Endpoint | | | Reaction Direction | Positive |
| Pri Wave | 546nm | | | Sec Wave | 700nm |
| Unit | g/dl | | | Decimal | 0.1 |
| Blank Time | 4 | 5 | | Reaction Time | 20 | 22 |

| | Sample Vol | Aspirated | Diluent | | Reagent Vol | Diluent |
|---|---|---|---|---|---|---|
| Standard | 5 uL | uL | uL | R1 | 200 uL | uL |
| Decreased | 2 uL | uL | uL | R2 | uL | uL |
| Increased | 10 uL | uL | uL | R3 | uL | uL |
| | | | | R4 | uL | uL |

**Calibration Settings**

| | | | |
|---|---|---|---|
| Math Model | Two-point linear | | |
| Factor | | Replicates | 2 |

**www.tksmed.ir شرکت توسعه کیمیای سعادت**

**09129579408 - 09129579409 پشتیبانی**

BIOMEDIC

mindray
healthcare within reach

BS 430

# URIC ACID

| | | | | | |
|---|---|---|---|---|---|
| Chem | UA | No. | | Sample Type | Serum |
| Chemistry | URIC ACID | | | Print Name | UA |
| Reaction Type | Endpoint | | | Reaction Direction | Positive |
| Pri Wave | 546nm | | | Sec Wave | 700nm |
| Unit | mg/dl | | | Decimal | 0.01 |
| Blank Time | 0 | 0 | | Reaction Time | 37 | 39 |

| | Sample Vol | Aspirated | Diluent | | Reagent Vol | Diluent |
|---|---|---|---|---|---|---|
| Standard | 5 uL | uL | uL | R1 | 100 uL | uL |
| Decreased | 1.5 uL | uL | uL | R2 | 100 uL | uL |
| Increased | 10 uL | uL | uL | R3 | uL | uL |
| | | | | R4 | uL | uL |

**Calibration Settings**

| | | | |
|---|---|---|---|
| Math Model | Two-point linear | | |
| Factor | | Replicates | 2 |


**www.tksmed.ir شرکت توسعه کیمیای سعادت**

**09129579408 - 09129579409 پشتیبانی**

BIOMEDIC

mindray
healthcare within reach

**BS 430**

# UREA

| | | | | | |
|---|---|---|---|---|---|
| Chem | UREA | No. | | Sample Type | Serum |
| Chemistry | UREA | | | Print Name | UREA |
| Reaction Type | Fixed-time | | | Reaction Direction | negative |
| Pri Wave | 340nm | | | Sec Wave | 412nm |
| Unit | mg/dl | | | Decimal | 0.1 |
| Blank Time | 0 | 0 | | Reaction Time | 24 | 29 |

| | Sample Vol | Aspirated | Diluent | | Reagent Vol | Diluent |
|---|---|---|---|---|---|---|
| Standard | 2 uL | uL | uL | R1 | 160 uL | uL |
| Decreased | 1.5 uL | uL | uL | R2 | 40 uL | uL |
| Increased | 4 uL | uL | uL | R3 | uL | uL |
| | | | | R4 | uL | uL |

**Calibration Settings**

| | | | |
|---|---|---|---|
| Math Model | Two-point linear | | |
| Factor | | Replicates | 2 |

**www.tksmed.ir شرکت توسعه کیمیای سعادت**

**09129579408 - 09129579409 پشتیبانی**

BIOMEDIC

mindray
healthcare within reach

BS 430

# ZINC

| | | | | | |
|---|---|---|---|---|---|
| Chem | ZINC | No. | | Sample Type | Serum |
| Chemistry | ZINC | | | Print Name | ZINC |
| Reaction Type | Endpoint | | | Reaction Direction | Positive |
| Pri Wave | 570nm | | | Sec Wave | 700nm |
| Unit | µg/dl | | | Decimal | 0 |
| Blank Time | 4 | 5 | | Reaction Time | 19 | 20 |

| | Sample Vol | Aspirated | Diluent | | Reagent Vol | Diluent |
|---|---|---|---|---|---|---|
| Standard | 8 uL | uL | uL | R1 | 160 uL | uL |
| Decreased | 4 uL | uL | uL | R2 | uL | uL |
| Increased | 16 uL | uL | uL | R3 | uL | uL |
| | | | | R4 | uL | uL |

**Calibration Settings**

| | | | |
|---|---|---|---|
| Math Model | Two-point linear | | |
| Factor | | Replicates | 2 |

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی 09129579409 - 09129579408