BIOMEDIC

OLYMPUS

Albumin BCG

Specific Test Parameters

General | LIH | ISE | Range

Test Name: ALB | < | > | Type: Serum | Operation: Yes

Sample:	Volume	2 µL	Dilution	0 µL	Pre-Dilution Rate:	1		
Reagents:	R1 Volume	200 µL	Dilution	0 µL		Min OD		Max OD
	R2 Volume	µL	Dilution	0 µL		L -2		H 2.5
					Reagent OD limit:			
Wavelength:	Pri.	600	Sec.	700		First L -2		First H 2.5
Method:		END				Last L -2		Last H 2.5
Reaction slope:		+			Dynamic Range:			
Measuring Point 1:	First	0	Last	20		L 0		H 5
Measuring Point 2:	First		Last		Correlation Factor:			
Linearity :		%				A 1		B 0
No Lag Time:					On-board stability period:			99

Specific Test Parameters

General | ISE

Test Name: ALB | < | > | Type: Serum

Calibration Type: AB | Formula: Y=AX+B | Counts: 2 | Process: CONC

	Cal. No.	OD	CONC	Factor/OD-L	Factor/OD-H
Point 1:	#		*	-9999999	9999999
Point 2:				-9999999	9999999
Point 3:				-9999999	9999999
Point 4:				-9999999	9999999
Point 5:				-9999999	9999999
Point 6:				-9999999	9999999
Point 7:					

1-Point Cal. Point:

MB Type Factor:

Calibration Stability Period: 99

User defined

* Calibrator concentration

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

OLYMPUS

ALKALINE PHOSPHATASE

Specific Test Parameters

General | LIH | ISE | Range

Test Name: ALK | < | > | Type: Serum | Operation: Yes

Sample:	Volume	4	µL	Dilution	0	µL	Pre-Dilution Rate:	1
Reagents:	R1 Volume	160	µL	Dilution	0	µL	Min OD	Max OD
	R2 Volume	40	µL	Dilution	0	µL	L -2	H 2.5
							Reagent OD limit:	
Wavelength:	Pri.	410		Sec.	660		First L -2	First H 2.5
Method:		RATE					Last L -2	Last H 2.5
Reaction slope:		+					Dynamic Range:	
Measuring Point 1:	First	14		Last	24		L 2	H 1000
Measuring Point 2:	First			Last			Correlation Factor:	
Linearity :		20	%				A 1	B 0
No Lag Time:							On-board stability period:	99

Specific Test Parameters

General | ISE

Test Name: ALK | < | > | Type: Serum

Calibration Type: AB | Formula: Y=AX+B | Counts: 2 | Process: CONC

	Cal. No.	OD	CONC	Factor/OD-L	Factor/OD-H
Point 1:	#		*	-9999999	9999999
Point 2:				-9999999	9999999
Point 3:				-9999999	9999999
Point 4:				-9999999	9999999
Point 5:				-9999999	9999999
Point 6:				-9999999	9999999
Point 7:					

1-Point Cal. Point:

MB Type Factor:

Calibration Stability Period: 99

User defined

* Calibrator concentration

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

OLYMPUS

Alanine Aminotransferase (ALT)

Specific Test Parameters

General | LIH | ISE | Range

Test Name: ALT | < | > | Type: Serum | Operation: Yes

Sample:	Volume	20 µL	Dilution	0 µL	Pre-Dilution Rate:	1		
Reagents:	R1 Volume	160 µL	Dilution	0 µL		Min OD		Max OD
	R2 Volume	40 µL	Dilution	0 µL		L -2		H 2.5
					Reagent OD limit:			
Wavelength:	Pri.	340	Sec.	700	First L	-2	First H	2.5
Method:		RATE			Last L	-2	Last H	2.5
Reaction slope:		-			Dynamic Range:			
Measuring Point 1:	First	14	Last	24	L	1	H	400
Measuring Point 2:	First		Last		Correlation Factor:			
Linearity :		20 %			A	1	B	0
No Lag Time:					On-board stability period:			99

Specific Test Parameters

General | ISE

Test Name: ALT | < | > | Type: Serum

Calibration Type: AB | Formula: Y=AX+B | Counts: 2 | Process: CONC

	Cal. No.	OD	CONC	Factor/OD-L	Factor/OD-H
Point 1:	#		*	-9999999	9999999
Point 2:				-9999999	9999999
Point 3:				-9999999	9999999
Point 4:				-9999999	9999999
Point 5:				-9999999	9999999
Point 6:				-9999999	9999999
Point 7:					

1-Point Cal. Point:

MB Type Factor:

Calibration Stability Period: 99

User defined

* Calibrator concentration

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

OLYMPUS

α- AMYLASE

Specific Test Parameters

General | LIH | ISE | Range

Test Name: AMY | < | > | Type: Serum | Operation: Yes

Sample:	Volume	4 µL	Dilution	0 µL	Pre-Dilution Rate:	1		
Reagents:	R1 Volume	200 µL	Dilution	0 µL		Min OD		Max OD
	R2 Volume	µL	Dilution	0 µL		L -2		H 2.5
					Reagent OD limit:			
Wavelength:	Pri.	410	Sec.	660		First L -2		First H 2.5
Method:		RATE				Last L -2		Last H 2.5
Reaction slope:		+			Dynamic Range:			
Measuring Point 1:	First	3	Last	12		L 2		H 2000
Measuring Point 2:	First		Last		Correlation Factor:			
Linearity :		20 %				A 1		B 0
No Lag Time:					On-board stability period:			99

Specific Test Parameters

General | ISE

Test Name: AMY | < | > | Type: Serum

Calibration Type: AB | Formula: Y=AX+B | Counts: 2 | Process: CONC

	Cal. No.	OD	CONC	Factor/OD-L	Factor/OD-H
Point 1:	#		*	-9999999	9999999
Point 2:				-9999999	9999999
Point 3:				-9999999	9999999
Point 4:				-9999999	9999999
Point 5:				-9999999	9999999
Point 6:				-9999999	9999999
Point 7:					

1-Point Cal. Point:

MB Type Factor:

Calibration Stability Period: 99

User defined

* Calibrator concentration

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

OLYMPUS

Aspartate Aminotransferase (AST)

Specific Test Parameters

General | LIH | ISE | Range

Test Name: AST | < | > | Type: Serum | Operation: Yes

Sample:	Volume	20	µL	Dilution	0	µL	Pre-Dilution Rate:	1
Reagents:	R1 Volume	160	µL	Dilution	0	µL	Min OD	Max OD
	R2 Volume	40	µL	Dilution	0	µL	L -2	H 2.5
							Reagent OD limit:	
Wavelength:	Pri.	340		Sec.	700		First L -2	First H 2.5
Method:		RATE					Last L -2	Last H 2.5
Reaction slope:		-					Dynamic Range:	
Measuring Point 1:	First	14		Last	24		L 1	H 400
Measuring Point 2:	First			Last			Correlation Factor:	
Linearity :		20	%				A 1	B 0
No Lag Time:							On-board stability period:	99

Specific Test Parameters

General | ISE

Test Name: AST | < | > | Type: Serum

Calibration Type: AB | Formula: Y=AX+B | Counts: 2 | Process: CONC

	Cal. No.	OD	CONC	Factor/OD-L	Factor/OD-H
Point 1:	#		*	-9999999	9999999
Point 2:				-9999999	9999999
Point 3:				-9999999	9999999
Point 4:				-9999999	9999999
Point 5:				-9999999	9999999
Point 6:				-9999999	9999999
Point 7:					

1-Point Cal. Point:

MB Type Factor:

Calibration Stability Period: 99

User defined

* Calibrator concentration

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

OLYMPUS

Complement 3 (C3)

Specific Test Parameters

General | LIH | ISE | Range

Test Name: C3 | < | > | Type: Serum | Operation: Yes

Sample:	Volume	2 µL	Dilution	0 µL	Pre-Dilution Rate:	1		
Reagents:	R1 Volume	170 µL	Dilution	0 µL		Min OD		Max OD
	R2 Volume	34 µL	Dilution	0 µL		L -2		H 2.5
					Reagent OD limit:			
Wavelength:	Pri.	340	Sec.	700	First L	-2	First H	2.5
Method:		END			Last L	-2	Last H	2.5
Reaction slope:		+			Dynamic Range:			
Measuring Point 1:	First	0	Last	27	L	3	H	500
Measuring Point 2:	First		Last	10	Correlation Factor:			
Linearity :		%			A	1	B	0
No Lag Time:					On-board stability period:			99

Specific Test Parameters

General | ISE

Test Name: C3 | < | > | Type: Serum

Calibration Type: 6AB | Formula: POLYGONAL | Counts: 1 | Process: CONC

	Cal. No.	OD	CONC	Factor/OD-L	Factor/OD-H
Point 1:	#		0	-2.0000	2.5000
Point 2:	#		1/16*	-2.0000	2.5000
Point 3:	#		1/8*	-2.0000	2.5000
Point 4:	#		1/4*	-2.0000	2.5000
Point 5:	#		1/2*	-2.0000	2.5000
Point 6:	#		*	-2.0000	2.5000
Point 7:					

1-Point Cal. Point:

MB Type Factor:

Calibration Stability Period: 99

User defined

* Calibrator concentration

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

OLYMPUS

Complement 4 (C4)

Specific Test Parameters

General | LIH | ISE | Range

Test Name: C4 | < | > | Type: Serum | Operation: Yes

Sample:	Volume	4 µL	Dilution	0 µL	Pre-Dilution Rate:	1		
Reagents:	R1 Volume	170 µL	Dilution	0 µL		Min OD		Max OD
	R2 Volume	34 µL	Dilution	0 µL		L -2		H 2.5
					Reagent OD limit:			
Wavelength:	Pri.	340	Sec.	700	First L	-2	First H	2.5
Method:		END			Last L	-2	Last H	2.5
Reaction slope:		+			Dynamic Range:			
Measuring Point 1:	First	0	Last	27	L	1	H	80
Measuring Point 2:	First		Last	10	Correlation Factor:			
Linearity :		%			A	1	B	0
No Lag Time:					On-board stability period:			99

Specific Test Parameters

General | ISE

Test Name: C4 | < | > | Type: Serum

Calibration Type: 6AB | Formula: POLYGONAL | Counts: 1 | Process: CONC

	Cal. No.	OD	CONC	Factor/OD-L	Factor/OD-H
Point 1:	#		0	-2.0000	2.5000
Point 2:	#		1/16*	-2.0000	2.5000
Point 3:	#		1/8*	-2.0000	2.5000
Point 4:	#		1/4*	-2.0000	2.5000
Point 5:	#		1/2*	-2.0000	2.5000
Point 6:	#		*	-2.0000	2.5000
Point 7:					

1-Point Cal. Point:

MB Type Factor:

Calibration Stability Period: 99

User defined

* Calibrator concentration

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

OLYMPUS

Calcium Arsenazo III

Specific Test Parameters

General | LIH | ISE | Range

Test Name: CA | < | > | Type: Serum | Operation: Yes

Sample:	Volume	3 µL	Dilution	0 µL	Pre-Dilution Rate:	1		
Reagents:	R1 Volume	200 µL	Dilution	0 µL		Min OD		Max OD
	R2 Volume	µL	Dilution	0 µL		L -2		H 2.5
					Reagent OD limit:			
Wavelength:	Pri.	660	Sec.	800		First L -2		First H 2.5
Method:		END				Last L -2		Last H 2.5
Reaction slope:		+			Dynamic Range:			
Measuring Point 1:	First	0	Last	10		L 2		H 20
Measuring Point 2:	First		Last		Correlation Factor:			
Linearity :		%				A 1		B 0
No Lag Time:					On-board stability period:			99

Specific Test Parameters

General | ISE

Test Name: CA | < | > | Type: Serum

Calibration Type: AB | Formula: Y=AX+B | Counts: 2 | Process: CONC

	Cal. No.	OD	CONC	Factor/OD-L	Factor/OD-H
Point 1:	#		*	-9999999	9999999
Point 2:				-9999999	9999999
Point 3:				-9999999	9999999
Point 4:				-9999999	9999999
Point 5:				-9999999	9999999
Point 6:				-9999999	9999999
Point 7:					

1-Point Cal. Point:

MB Type Factor:

Calibration Stability Period: 99

User defined

* Calibrator concentration

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

OLYMPUS

Cholesterol

Specific Test Parameters

General | LIH | ISE | Range

Test Name: CHOL | < | > | Type: Serum | Operation: Yes

Sample:	Volume	2 µL	Dilution	0 µL	Pre-Dilution Rate:	1		
Reagents:	R1 Volume	200 µL	Dilution	0 µL		Min OD		Max OD
	R2 Volume	µL	Dilution	0 µL		L -2		H 2.5
					Reagent OD limit:			
Wavelength:	Pri.	520	Sec.	700	First L	-2	First H	2.5
Method:		END			Last L	-2	Last H	2.5
Reaction slope:		+			Dynamic Range:			
Measuring Point 1:	First	0	Last	27	L	2	H	600
Measuring Point 2:	First		Last		Correlation Factor:			
Linearity :		%			A	1	B	0
No Lag Time:					On-board stability period:			99

Specific Test Parameters

General | ISE

Test Name: CHOL | < | > | Type: Serum

Calibration Type: AB | Formula: Y=AX+B | Counts: 2 | Process: CONC

	Cal. No.	OD	CONC	Factor/OD-L	Factor/OD-H
Point 1:	#		*	-9999999	9999999
Point 2:				-9999999	9999999
Point 3:				-9999999	9999999
Point 4:				-9999999	9999999
Point 5:				-9999999	9999999
Point 6:				-9999999	9999999
Point 7:					

1-Point Cal. Point:

MB Type Factor:

Calibration Stability Period: 99

User defined

* Calibrator concentration

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

OLYMPUS

Creatine Kinase (CK-NAC)

Specific Test Parameters

General | LIH | ISE | Range

Test Name: CPK | < | > | Type: Serum | Operation: Yes

Sample:	Volume	8 µL	Dilution	0 µL	Pre-Dilution Rate:	1		
Reagents:	R1 Volume	150 µL	Dilution	0 µL		Min OD		Max OD
	R2 Volume	30 µL	Dilution	0 µL		L -2		H 2.5
					Reagent OD limit:			
Wavelength:	Pri.	340	Sec.	540		First L -2		First H 2.5
Method:		RATE				Last L -2		Last H 2.5
Reaction slope:		+			Dynamic Range:			
Measuring Point 1:	First	17	Last	27		L 2		H 1200
Measuring Point 2:	First		Last		Correlation Factor:			
Linearity :		20 %				A 1		B 0
No Lag Time:					On-board stability period:			99

Specific Test Parameters

General | ISE

Test Name: CPK | < | > | Type: Serum

Calibration Type: AB | Formula: Y=AX+B | Counts: 2 | Process: CONC

	Cal. No.	OD	CONC	Factor/OD-L	Factor/OD-H
Point 1:	#		*	-9999999	9999999
Point 2:				-9999999	9999999
Point 3:				-9999999	9999999
Point 4:				-9999999	9999999
Point 5:				-9999999	9999999
Point 6:				-9999999	9999999
Point 7:					

1-Point Cal. Point:

MB Type Factor:

Calibration Stability Period: 99

User defined

* Calibrator concentration

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

OLYMPUS

Creatinine Jaffe

Specific Test Parameters

General | LIH | ISE | Range

Test Name: CREAT | < | > | Type: Serum | Operation: Yes

Sample:	Volume	20 µL	Dilution	0 µL	Pre-Dilution Rate:	1		
Reagents:	R1 Volume	100 µL	Dilution	0 µL		Min OD		Max OD
	R2 Volume	100 µL	Dilution	0 µL		L -2		H 2.5
					Reagent OD limit:			
Wavelength:	Pri.	520	Sec.	700	First L	-2	First H	2.5
Method:		FIXED			Last L	-2	Last H	2.5
Reaction slope:		+			Dynamic Range:			
Measuring Point 1:	First	13	Last	19	L	0.1	H	15
Measuring Point 2:	First		Last		Correlation Factor:			
Linearity :		%			A	1	B	0
No Lag Time:					On-board stability period:			99

Specific Test Parameters

General | ISE

Test Name: CREAT | < | > | Type: Serum

Calibration Type: AB | Formula: Y=AX+B | Counts: 2 | Process: CONC

	Cal. No.	OD	CONC	Factor/OD-L	Factor/OD-H
Point 1:	#		*	-9999999	9999999
Point 2:				-9999999	9999999
Point 3:				-9999999	9999999
Point 4:				-9999999	9999999
Point 5:				-9999999	9999999
Point 6:				-9999999	9999999
Point 7:					

1-Point Cal. Point:

MB Type Factor:

Calibration Stability Period: 99

User defined

* Calibrator concentration

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

OLYMPUS

C-Reactive Protein (CRP)

Specific Test Parameters

General | LIH | ISE | Range

Test Name: CRP | < | > | Type: Serum | Operation: Yes

Sample:	Volume	8 µL	Dilution	0 µL	Pre-Dilution Rate:	1		
Reagents:	R1 Volume	160 µL	Dilution	0 µL		Min OD		Max OD
	R2 Volume	30 µL	Dilution	0 µL		L -2		H 2.5
					Reagent OD limit:			
Wavelength:	Pri.	340	Sec.	700	First L	-2	First H	2.5
Method:		END			Last L	-2	Last H	2.5
Reaction slope:		+			Dynamic Range:			
Measuring Point 1:	First	0	Last	27	L	1	H	180
Measuring Point 2:	First		Last	10	Correlation Factor:			
Linearity :		%			A	1	B	0
No Lag Time:					On-board stability period:			99

Specific Test Parameters

General | ISE

Test Name: CRP | < | > | Type: Serum

Calibration Type: 6AB | Formula: POLYGONAL | Counts: 1 | Process: CONC

	Cal. No.	OD	CONC	Factor/OD-L	Factor/OD-H
Point 1:	#		0	-2.0000	2.5000
Point 2:	#		1/16*	-2.0000	2.5000
Point 3:	#		1/8*	-2.0000	2.5000
Point 4:	#		1/4*	-2.0000	2.5000
Point 5:	#		1/2*	-2.0000	2.5000
Point 6:	#		*	-2.0000	2.5000
Point 7:					

1-Point Cal. Point:

MB Type Factor:

Calibration Stability Period: 99

User defined

* Calibrator concentration

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

OLYMPUS

GLUCOSE

Specific Test Parameters

General | LIH | ISE | Range

Test Name: GLU | < | > | Type: Serum | Operation: Yes

Sample:	Volume	2 µL	Dilution	0 µL	Pre-Dilution Rate:	1		
Reagents:	R1 Volume	200 µL	Dilution	0 µL		Min OD		Max OD
	R2 Volume	µL	Dilution	0 µL		L -2		H 2.5
					Reagent OD limit:			
Wavelength:	Pri.	520	Sec.	700	First L	-2	First H	2.5
Method:		END			Last L	-2	Last H	2.5
Reaction slope:		+			Dynamic Range:			
Measuring Point 1:	First	0	Last	27	L	5	H	400
Measuring Point 2:	First		Last		Correlation Factor:			
Linearity :		%			A	1	B	0
No Lag Time:					On-board stability period:			99

Specific Test Parameters

General | ISE

Test Name: GLU | < | > | Type: Serum

Calibration Type: AB | Formula: Y=AX+B | Counts: 2 | Process: CONC

	Cal. No.	OD	CONC	Factor/OD-L	Factor/OD-H
Point 1:	#		*	-9999999	9999999
Point 2:				-9999999	9999999
Point 3:				-9999999	9999999
Point 4:				-9999999	9999999
Point 5:				-9999999	9999999
Point 6:				-9999999	9999999
Point 7:					

1-Point Cal. Point:

MB Type Factor:

Calibration Stability Period: 99

User defined

* Calibrator concentration

www.tksmed.ir شرکت توسعه کیمیای سعادت

پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

OLYMPUS

HDL Cholesterol

Specific Test Parameters

General | LIH | ISE | Range

Test Name: HDL | < | > | Type: Serum | Operation: Yes

Sample:	Volume	2 µL	Dilution	0 µL	Pre-Dilution Rate:	1		
Reagents:	R1 Volume	150 µL	Dilution	0 µL		Min OD		Max OD
	R2 Volume	50 µL	Dilution	0 µL		L -2		H 2.5
					Reagent OD limit:			
Wavelength:	Pri.	600	Sec.	700	First L	-2	First H	2.5
Method:		END			Last L	-2	Last H	2.5
Reaction slope:		+			Dynamic Range:			
Measuring Point 1:	First	0	Last	27	L	2	H	100
Measuring Point 2:	First		Last	10	Correlation Factor:			
Linearity :		%			A	1	B	0
No Lag Time:					On-board stability period:			99

Specific Test Parameters

General | ISE

Test Name: HDL | < | > | Type: Serum

Calibration Type: AB | Formula: Y=AX+B | Counts: 2 | Process: CONC

	Cal. No.	OD	CONC	Factor/OD-L	Factor/OD-H
Point 1:	#		*	-9999999	9999999
Point 2:				-9999999	9999999
Point 3:				-9999999	9999999
Point 4:				-9999999	9999999
Point 5:				-9999999	9999999
Point 6:				-9999999	9999999
Point 7:					

1-Point Cal. Point:

MB Type Factor:

Calibration Stability Period: 99

User defined

* Calibrator concentration

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

OLYMPUS

Immunoglobulin A (IgA)

Specific Test Parameters

General | LIH | ISE | Range

Test Name: IgA | < | > | Type: Serum | Operation: Yes

Sample:	Volume	2 µL	Dilution	0 µL	Pre-Dilution Rate:	1		
Reagents:	R1 Volume	170 µL	Dilution	0 µL		Min OD		Max OD
	R2 Volume	34 µL	Dilution	0 µL		L -2		H 2.5
					Reagent OD limit:			
Wavelength:	Pri.	340	Sec.	700	First L	-2	First H	2.5
Method:		END			Last L	-2	Last H	2.5
Reaction slope:		+			Dynamic Range:			
Measuring Point 1:	First	0	Last	27	L	3	H	800
Measuring Point 2:	First		Last	10	Correlation Factor:			
Linearity :		%			A	1	B	0
No Lag Time:					On-board stability period:			99

Specific Test Parameters

General | ISE

Test Name: IgA | < | > | Type: Serum

Calibration Type: 6AB | Formula: POLYGONAL | Counts: 1 | Process: CONC

	Cal. No.	OD	CONC	Factor/OD-L	Factor/OD-H
Point 1:	#		0	-2.0000	2.5000
Point 2:	#		1/16*	-2.0000	2.5000
Point 3:	#		1/8*	-2.0000	2.5000
Point 4:	#		1/4*	-2.0000	2.5000
Point 5:	#		1/2*	-2.0000	2.5000
Point 6:	#		*	-2.0000	2.5000
Point 7:					

1-Point Cal. Point:

MB Type Factor:

Calibration Stability Period: 99

User defined

* Calibrator concentration

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

OLYMPUS

Immunoglobulin G (IgG)

Specific Test Parameters

General | LIH | ISE | Range

Test Name: IgG | < | > | Type: Serum | Operation: Yes

Sample:	Volume	2 µL	Dilution	0 µL	Pre-Dilution Rate:	1		
Reagents:	R1 Volume	170 µL	Dilution	0 µL		Min OD		Max OD
	R2 Volume	34 µL	Dilution	0 µL		L -2		H 2.5
					Reagent OD limit:			
Wavelength:	Pri.	600	Sec.			First L -2		First H 2.5
Method:		END				Last L -2		Last H 2.5
Reaction slope:		+			Dynamic Range:			
Measuring Point 1:	First	0	Last	27		L 20		H 3000
Measuring Point 2:	First		Last	10	Correlation Factor:			
Linearity :		%				A 1		B 0
No Lag Time:					On-board stability period:			99

Specific Test Parameters

General | ISE

Test Name: IgG | < | > | Type: Serum

Calibration Type: 6AB | Formula: POLYGONAL | Counts: 1 | Process: CONC

	Cal. No.	OD	CONC	Factor/OD-L	Factor/OD-H
Point 1:	#		0	-2.0000	2.5000
Point 2:	#		1/16*	-2.0000	2.5000
Point 3:	#		1/8*	-2.0000	2.5000
Point 4:	#		1/4*	-2.0000	2.5000
Point 5:	#		1/2*	-2.0000	2.5000
Point 6:	#		*	-2.0000	2.5000
Point 7:					

1-Point Cal. Point:

MB Type Factor:

Calibration Stability Period: 99

User defined

* Calibrator concentration

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

OLYMPUS

Immunoglobulin M (IgM)

Specific Test Parameters

General | LIH | ISE | Range

Test Name: IgM | < | > | Type: Serum | Operation: Yes

Sample:	Volume	2 µL	Dilution	0 µL	Pre-Dilution Rate:	1		
Reagents:	R1 Volume	170 µL	Dilution	0 µL		Min OD		Max OD
	R2 Volume	34 µL	Dilution	0 µL		L -2		H 2.5
					Reagent OD limit:			
Wavelength:	Pri.	340	Sec.	700	First L	-2	First H	2.5
Method:		END			Last L	-2	Last H	2.5
Reaction slope:		+			Dynamic Range:			
Measuring Point 1:	First	0	Last	27	L	3	H	350
Measuring Point 2:	First		Last	10	Correlation Factor:			
Linearity :		%			A	1	B	0
No Lag Time:					On-board stability period:			99

Specific Test Parameters

General | ISE

Test Name: IgM | < | > | Type: Serum

Calibration Type: 6AB | Formula: POLYGONAL | Counts: 1 | Process: CONC

	Cal. No.	OD	CONC	Factor/OD-L	Factor/OD-H
Point 1:	#		0	-2.0000	2.5000
Point 2:	#		1/16*	-2.0000	2.5000
Point 3:	#		1/8*	-2.0000	2.5000
Point 4:	#		1/4*	-2.0000	2.5000
Point 5:	#		1/2*	-2.0000	2.5000
Point 6:	#		*	-2.0000	2.5000
Point 7:					

1-Point Cal. Point:

MB Type Factor:

Calibration Stability Period: 99

User defined

* Calibrator concentration

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

OLYMPUS

Lactate Dehydrogenase (LDH)

Specific Test Parameters

General | LIH | ISE | Range

Test Name: LDH | < | > | Type: Serum | Operation: Yes

Sample:	Volume	3	µL	Dilution	0	µL	Pre-Dilution Rate:	1
Reagents:	R1 Volume	150	µL	Dilution	0	µL	Min OD	Max OD
	R2 Volume	30	µL	Dilution	0	µL	L -2	H 2.5
							Reagent OD limit:	
Wavelength:	Pri.	340		Sec.	410		First L -2	First H 2.5
Method:		RATE					Last L -2	Last H 2.5
Reaction slope:		-					Dynamic Range:	
Measuring Point 1:	First	14		Last	24		L 2	H 1000
Measuring Point 2:	First			Last			Correlation Factor:	
Linearity :		20	%				A 1	B 0
No Lag Time:							On-board stability period:	99

Specific Test Parameters

General | ISE

Test Name: LDH | < | > | Type: Serum

Calibration Type: AB | Formula: Y=AX+B | Counts: 2 | Process: CONC

	Cal. No.	OD	CONC	Factor/OD-L	Factor/OD-H
Point 1:	#		*	-9999999	9999999
Point 2:				-9999999	9999999
Point 3:				-9999999	9999999
Point 4:				-9999999	9999999
Point 5:				-9999999	9999999
Point 6:				-9999999	9999999
Point 7:					

1-Point Cal. Point:

MB Type Factor:

Calibration Stability Period: 99

User defined

* Calibrator concentration

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

OLYMPUS

LDL Cholesterol

Specific Test Parameters

General | LIH | ISE | Range

Test Name: LDL | < | > | Type: Serum | Operation: Yes

Sample:	Volume	3 µL	Dilution	0 µL	Pre-Dilution Rate:	1		
Reagents:	R1 Volume	150 µL	Dilution	0 µL		Min OD		Max OD
	R2 Volume	50 µL	Dilution	0 µL		L -2		H 2.5
					Reagent OD limit:			
Wavelength:	Pri.	540	Sec.	700		First L -2		First H 2.5
Method:		END				Last L -2		Last H 2.5
Reaction slope:		+			Dynamic Range:			
Measuring Point 1:	First	0	Last	27		L 2		H 700
Measuring Point 2:	First		Last	10	Correlation Factor:			
Linearity :		%				A 1		B 0
No Lag Time:					On-board stability period:			99

Specific Test Parameters

General | ISE

Test Name: LDL | < | > | Type: Serum

Calibration Type: AB | Formula: Y=AX+B | Counts: 2 | Process: CONC

	Cal. No.	OD	CONC	Factor/OD-L	Factor/OD-H
Point 1:	#		*	-9999999	9999999
Point 2:				-9999999	9999999
Point 3:				-9999999	9999999
Point 4:				-9999999	9999999
Point 5:				-9999999	9999999
Point 6:				-9999999	9999999
Point 7:					

1-Point Cal. Point:

MB Type Factor:

Calibration Stability Period: 99

User defined

* Calibrator concentration

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

OLYMPUS

Lipase

Specific Test Parameters

General | LIH | ISE | Range

Test Name: LIP | < | > | Type: Serum | Operation: Yes

Sample:	Volume	4 µL	Dilution	0 µL	Pre-Dilution Rate:	1		
Reagents:	R1 Volume	160 µL	Dilution	0 µL		Min OD		Max OD
	R2 Volume	40 µL	Dilution	0 µL		L -2		H 2.5
					Reagent OD limit:			
Wavelength:	Pri.	570	Sec.	700		First L -2		First H 2.5
Method:		RATE				Last L -2		Last H 2.5
Reaction slope:		+			Dynamic Range:			
Measuring Point 1:	First	17	Last	25		L 1		H 150
Measuring Point 2:	First		Last		Correlation Factor:			
Linearity :		20 %				A 1		B 0
No Lag Time:					On-board stability period:			99

Specific Test Parameters

General | ISE

Test Name: LIP | < | > | Type: Serum

Calibration Type: AB | Formula: Y=AX+B | Counts: 2 | Process: CONC

	Cal. No.	OD	CONC	Factor/OD-L	Factor/OD-H
Point 1:	#		*	-9999999	9999999
Point 2:				-9999999	9999999
Point 3:				-9999999	9999999
Point 4:				-9999999	9999999
Point 5:				-9999999	9999999
Point 6:				-9999999	9999999
Point 7:					

1-Point Cal. Point:

MB Type Factor:

Calibration Stability Period: 99

User defined

* Calibrator concentration

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

OLYMPUS

Micro Albumin

Specific Test Parameters

General | LIH | ISE | Range

Test Name: MALB | < | > | Type: Serum | Operation: Yes

Sample:	Volume	12 µL	Dilution	0 µL	Pre-Dilution Rate:	1		
Reagents:	R1 Volume	150 µL	Dilution	0 µL		Min OD		Max OD
	R2 Volume	30 µL	Dilution	0 µL		L -2		H 2.5
					Reagent OD limit:			
Wavelength:	Pri.	340	Sec.	700	First L	-2	First H	2.5
Method:		END			Last L	-2	Last H	2.5
Reaction slope:		+			Dynamic Range:			
Measuring Point 1:	First	0	Last	27	L	2	H	250
Measuring Point 2:	First		Last	10	Correlation Factor:			
Linearity :		%			A	1	B	0
No Lag Time:					On-board stability period:			99

Specific Test Parameters

General | ISE

Test Name: MALB | < | > | Type: Serum

Calibration Type: 6AB | Formula: POLYGONAL | Counts: 1 | Process: CONC

	Cal. No.	OD	CONC	Factor/OD-L	Factor/OD-H
Point 1:	#		0	-2.0000	2.5000
Point 2:	#		1/16*	-2.0000	2.5000
Point 3:	#		1/8*	-2.0000	2.5000
Point 4:	#		1/4*	-2.0000	2.5000
Point 5:	#		1/2*	-2.0000	2.5000
Point 6:	#		*	-2.0000	2.5000
Point 7:					

1-Point Cal. Point:

MB Type Factor:

Calibration Stability Period: 99

User defined

* Calibrator concentration

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

OLYMPUS

CK - MB

Specific Test Parameters

General | LIH | ISE | Range

Test Name: CKMB | < | > | Type: Serum | Operation: Yes

Sample:	Volume	8	µL	Dilution	0	µL	Pre-Dilution Rate:	1
Reagents:	R1 Volume	150	µL	Dilution	0	µL	Min OD	Max OD
	R2 Volume	30	µL	Dilution	0	µL	L -2	H 2.5
							Reagent OD limit:	
Wavelength:	Pri.	340		Sec.	540		First L -2	First H 2.5
Method:		RATE					Last L -2	Last H 2.5
Reaction slope:		+					Dynamic Range:	
Measuring Point 1:	First	17		Last	27		L 2	H 800
Measuring Point 2:	First			Last			Correlation Factor:	
Linearity :		20	%				A 1	B 0
No Lag Time:							On-board stability period:	99

Specific Test Parameters

General | ISE

Test Name: CKMB | < | > | Type: Serum

Calibration Type: AB | Formula: Y=AX+B | Counts: 2 | Process: CONC

	Cal. No.	OD	CONC	Factor/OD-L	Factor/OD-H
Point 1:	#		*	-9999999	9999999
Point 2:				-9999999	9999999
Point 3:				-9999999	9999999
Point 4:				-9999999	9999999
Point 5:				-9999999	9999999
Point 6:				-9999999	9999999
Point 7:					

1-Point Cal. Point:

MB Type Factor:

Calibration Stability Period: 99

User defined

* Calibrator concentration

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

OLYMPUS

PHOSPHORUS

Specific Test Parameters

General | LIH | ISE | Range

Test Name: PHOS | < | > | Type: Serum | Operation: Yes

Sample:	Volume	2 µL	Dilution	0 µL	Pre-Dilution Rate:	1		
Reagents:	R1 Volume	200 µL	Dilution	0 µL		Min OD		Max OD
	R2 Volume	µL	Dilution	0 µL		L -2		H 2.5
					Reagent OD limit:			
Wavelength:	Pri.	340	Sec.	700		First L -2		First H 2.5
Method:		END				Last L -2		Last H 2.5
Reaction slope:		+			Dynamic Range:			
Measuring Point 1:	First	0	Last	10		L 1		H 20
Measuring Point 2:	First		Last		Correlation Factor:			
Linearity :		%				A 1		B 0
No Lag Time:					On-board stability period:			99

Specific Test Parameters

General | ISE

Test Name: PHOS | < | > | Type: Serum

Calibration Type: AB | Formula: Y=AX+B | Counts: 2 | Process: CONC

	Cal. No.	OD	CONC	Factor/OD-L	Factor/OD-H
Point 1:	#		*	-9999999	9999999
Point 2:				-9999999	9999999
Point 3:				-9999999	9999999
Point 4:				-9999999	9999999
Point 5:				-9999999	9999999
Point 6:				-9999999	9999999
Point 7:					

1-Point Cal. Point:

MB Type Factor:

Calibration Stability Period: 99

User defined

* Calibrator concentration

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

OLYMPUS

Triglycerides

Specific Test Parameters

General | LIH | ISE | Range

Test Name: TG | < | > | Type: Serum | Operation: Yes

Sample:	Volume	2 µL	Dilution	0 µL	Pre-Dilution Rate:	1		
Reagents:	R1 Volume	200 µL	Dilution	0 µL		Min OD		Max OD
	R2 Volume	µL	Dilution	0 µL		L -2		H 2.5
					Reagent OD limit:			
Wavelength:	Pri.	520	Sec.	700		First L -2		First H 2.5
Method:		END				Last L -2		Last H 2.5
Reaction slope:		+				Dynamic Range:		
Measuring Point 1:	First	0	Last	27		L 2		H 600
Measuring Point 2:	First		Last			Correlation Factor:		
Linearity :		%				A 1		B 0
No Lag Time:						On-board stability period:		99

Specific Test Parameters

General | ISE

Test Name: TG | < | > | Type: Serum

Calibration Type: AB | Formula: Y=AX+B | Counts: 2 | Process: CONC

	Cal. No.	OD	CONC	Factor/OD-L	Factor/OD-H
Point 1:	#		*	-9999999	9999999
Point 2:				-9999999	9999999
Point 3:				-9999999	9999999
Point 4:				-9999999	9999999
Point 5:				-9999999	9999999
Point 6:				-9999999	9999999
Point 7:					

1-Point Cal. Point:

MB Type Factor:

Calibration Stability Period: 99

User defined

* Calibrator concentration

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

OLYMPUS

TOTAL PROTEIN

Specific Test Parameters

General | LIH | ISE | Range

Test Name: TPRO | < | > | Type: Serum | Operation: Yes

Sample:	Volume	5 µL	Dilution	0 µL	Pre-Dilution Rate:	1		
Reagents:	R1 Volume	200 µL	Dilution	0 µL		Min OD		Max OD
	R2 Volume	µL	Dilution	0 µL		L -2		H 2.5
					Reagent OD limit:			
Wavelength:	Pri.	540	Sec.	700		First L -2		First H 2.5
Method:		END				Last L -2		Last H 2.5
Reaction slope:		+			Dynamic Range:			
Measuring Point 1:	First	0	Last	20		L 0		H 12
Measuring Point 2:	First		Last		Correlation Factor:			
Linearity :		%				A 1		B 0
No Lag Time:					On-board stability period:			99

Specific Test Parameters

General | ISE

Test Name: TPRO | < | > | Type: Serum

Calibration Type: AB | Formula: Y=AX+B | Counts: 2 | Process: CONC

	Cal. No.	OD	CONC	Factor/OD-L	Factor/OD-H
Point 1:	#		*	-9999999	9999999
Point 2:				-9999999	9999999
Point 3:				-9999999	9999999
Point 4:				-9999999	9999999
Point 5:				-9999999	9999999
Point 6:				-9999999	9999999
Point 7:					

1-Point Cal. Point:

MB Type Factor:

Calibration Stability Period: 99

User defined

* Calibrator concentration

www.tksmed.ir شرکت توسعه کیمیای سعادت

پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

OLYMPUS

URIC ACID

Specific Test Parameters

General | LIH | ISE | Range

Test Name: UA | < | > | Type: Serum | Operation: Yes

Sample:	Volume	6 µL	Dilution	0 µL	Pre-Dilution Rate:	1		
Reagents:	R1 Volume	120 µL	Dilution	0 µL		Min OD		Max OD
	R2 Volume	120 µL	Dilution	0 µL		L -2		H 2.5
					Reagent OD limit:			
Wavelength:	Pri.	540	Sec.	700	First L	-2	First H	2.5
Method:		END			Last L	-2	Last H	2.5
Reaction slope:		+			Dynamic Range:			
Measuring Point 1:	First	0	Last	20	L	1	H	20
Measuring Point 2:	First		Last		Correlation Factor:			
Linearity :		%			A	1	B	0
No Lag Time:					On-board stability period:			99

Specific Test Parameters

General | ISE

Test Name: UA | < | > | Type: Serum

Calibration Type: AB | Formula: Y=AX+B | Counts: 2 | Process: CONC

	Cal. No.	OD	CONC	Factor/OD-L	Factor/OD-H
Point 1:	#		*	-9999999	9999999
Point 2:				-9999999	9999999
Point 3:				-9999999	9999999
Point 4:				-9999999	9999999
Point 5:				-9999999	9999999
Point 6:				-9999999	9999999
Point 7:					

1-Point Cal. Point:

MB Type Factor:

Calibration Stability Period: 99

User defined

* Calibrator concentration

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

OLYMPUS

UREA

Specific Test Parameters

General | LIH | ISE | Range

Test Name: UREA | < | > | Type: Serum | Operation: Yes

Sample:	Volume	2	µL	Dilution	0	µL	Pre-Dilution Rate:	1
Reagents:	R1 Volume	160	µL	Dilution	0	µL	Min OD	Max OD
	R2 Volume	40	µL	Dilution	0	µL	L -2	H 2.5
							Reagent OD limit:	
Wavelength:	Pri.	340		Sec.	410		First L -2	First H 2.5
Method:		FIXED					Last L -2	Last H 2.5
Reaction slope:		-					Dynamic Range:	
Measuring Point 1:	First	13		Last	19		L 1	H 300
Measuring Point 2:	First			Last			Correlation Factor:	
Linearity :			%				A 1	B 0
No Lag Time:							On-board stability period:	99

Specific Test Parameters

General | ISE

Test Name: UREA | < | > | Type: Serum

Calibration Type: AB | Formula: Y=AX+B | Counts: 2 | Process: CONC

	Cal. No.	OD	CONC	Factor/OD-L	Factor/OD-H
Point 1:	#		*	-9999999	9999999
Point 2:				-9999999	9999999
Point 3:				-9999999	9999999
Point 4:				-9999999	9999999
Point 5:				-9999999	9999999
Point 6:				-9999999	9999999
Point 7:					

1-Point Cal. Point:

MB Type Factor:

Calibration Stability Period: 99

User defined

* Calibrator concentration

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

OLYMPUS

ZINC

Specific Test Parameters

General | LIH | ISE | Range

Test Name: ZINC | < | > | Type: Serum | Operation: Yes

Sample:	Volume	8 µL	Dilution	0 µL	Pre-Dilution Rate:	1		
Reagents:	R1 Volume	160 µL	Dilution	0 µL		Min OD		Max OD
	R2 Volume	µL	Dilution	0 µL		L -2		H 2.5
					Reagent OD limit:			
Wavelength:	Pri.	570	Sec.	700		First L -2		First H 2.5
Method:		END				Last L -2		Last H 2.5
Reaction slope:		+			Dynamic Range:			
Measuring Point 1:	First	0	Last	9		L 20		H 400
Measuring Point 2:	First		Last		Correlation Factor:			
Linearity :		%				A 1		B 0
No Lag Time:					On-board stability period:			99

Specific Test Parameters

General | ISE

Test Name: ZINC | < | > | Type: Serum

Calibration Type: AB | Formula: Y=AX+B | Counts: 2 | Process: CONC

	Cal. No.	OD	CONC	Factor/OD-L	Factor/OD-H
Point 1:	#		*	-9999999	9999999
Point 2:				-9999999	9999999
Point 3:				-9999999	9999999
Point 4:				-9999999	9999999
Point 5:				-9999999	9999999
Point 6:				-9999999	9999999
Point 7:					

1-Point Cal. Point:

MB Type Factor:

Calibration Stability Period: 99

User defined

* Calibrator concentration

www.tksmed.ir شرکت توسعه کیمیای سعادت
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸