BIOMEDIC

Kayto 420

Albumin BCG

Test No	#			Full Name	ALB

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	2	0	1
Increment	4	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	200
Reagent 2	0
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point			
Wavelength 1	600	700		
Blank type	Pre-Blank	130	150	
Test time	280	300	Sec	
Calculation	Multi-pt.Linear Calibration			

Limited Range

ABS Range	0	0		
Linear Range	0	5	g/dl	
Reagent Blank	0	0		
Linear Limit	0	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	g/dl	B	0
Precision	X.XX	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 420

Alkaline Phosphatase

Test No	#			Full Name	ALP

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	4	0	1
Increment	8	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	160
Reagent 2	40
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	Kinetiec			
Wavelength 1	405	670		
Blank type	NO	0	0	
Test time	60	240	Sec	
Calculation	Multi-pt.Linear Calibration			

Limited Range

ABS Range	0	0		
Linear Range	3	1000	U/L	
Reagent Blank	0	0		
Linear Limit	20	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	U/L	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 420

Alanine Aminotransferase (ALT)

Test No	#			Full Name	ALT

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	20	0	1
Increment	30	0	1
Decrement	10	0	1

Reagent Volume (µl)

Reagent 1	160
Reagent 2	40
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	Kinetiec			
Wavelength 1	340	405		
Blank type	NO	0	0	
Test time	60	240	Sec	
Calculation	Multi-pt.Linear Calibration			

Limited Range

ABS Range	0	0		
Linear Range	1	300	U/L	
Reagent Blank	0	0		
Linear Limit	20	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	U/L	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 420

Amylase CNPG3

Test No	#			Full Name	AMY

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	4	0	1
Increment	8	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	200
Reagent 2	0
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	Kinetiec			
Wavelength 1	405	670		
Blank type	NO	0	0	
Test time	60	240	Sec	
Calculation	Multi-pt.Linear Calibration			

Limited Range

ABS Range	0	0		
Linear Range	1	2000	U/L	
Reagent Blank	0	0		
Linear Limit	20	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	U/L	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 420

Aspartate Aminotransferase (AST)

Test No	#			Full Name	AST

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	20	0	1
Increment	30	0	1
Decrement	10	0	1

Reagent Volume (µl)

Reagent 1	160
Reagent 2	40
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	Kinetiec			
Wavelength 1	340	405		
Blank type	NO	0	0	
Test time	60	240	Sec	
Calculation	Multi-pt.Linear Calibration			

Limited Range

ABS Range	0	0		
Linear Range	1	300	U/L	
Reagent Blank	0	0		
Linear Limit	20	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	U/L	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 420

Complement 3 (C3)

Test No	#			Full Name	C3

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	2	0	1
Increment	4	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	180
Reagent 2	35
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point			
Wavelength 1	340	700		
Blank type	Pre-Blank	420	450	
Test time	280	300	Sec	
Calculation	Logistic-Log 4P			

Limited Range

ABS Range	0	0		
Linear Range	1	500	mg/dl	
Reagent Blank	0	0		
Linear Limit	0	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	mg/dl	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 420

Complement 4 (C4)

Test No	#			Full Name	C4

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	4	0	1
Increment	8	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	180
Reagent 2	35
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point			
Wavelength 1	340	700		
Blank type	Pre-Blank	420	450	
Test time	280	300	Sec	
Calculation	Logistic-Log 4P			

Limited Range

ABS Range	0	0		
Linear Range	1	80	mg/dl	
Reagent Blank	0	0		
Linear Limit	0	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	mg/dl	B	0
Precision	X.X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 420

Calcium Arsenazo III

Test No	#			Full Name	CA

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	3	0	1
Increment	6	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	200
Reagent 2	0
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point				
Wavelength 1	670	795			
Blank type	Pre-Blank	130	150		
Test time	280	300	Sec		
Calculation	Multi-pt.Linear Calibration				

Limited Range

ABS Range	0	0		
Linear Range	1	20	mg/dl	
Reagent Blank	0	0		
Linear Limit	0	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	mg/dl	B	0
Precision	X.X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 420

Cholesterol

Test No	#			Full Name	CHOL

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	2	0	1
Increment	4	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	200
Reagent 2	0
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point			
Wavelength 1	505	700		
Blank type	Pre-Blank	130	- 150	
Test time	580	600	Sec	
Calculation	Multi-pt.Linear Calibration			

Limited Range

ABS Range	0	- 0		
Linear Range	5	- 700		mg/dl
Reagent Blank	0	- 0		
Linear Limit	0	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	mg/dl	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 420

Creatine Kinase (CK-NAC)

Test No	#			Full Name	CPK

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	10	0	1
Increment	20	0	1
Decrement	5	0	1

Reagent Volume (µl)

Reagent 1	200
Reagent 2	40
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	Kinetiec			
Wavelength 1	340	540		
Blank type	NO	0	0	
Test time	120	300	Sec	
Calculation	Multi-pt.Linear Calibration			

Limited Range

ABS Range	0	0		
Linear Range	1	800	U/L	
Reagent Blank	0	0		
Linear Limit	20	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	U/L	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 420

Creatinine Jaffe

Test No	#			Full Name	CREAT

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	20	0	1
Increment	30	0	1
Decrement	10	0	1

Reagent Volume (µl)

Reagent 1	100
Reagent 2	100
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	Kinetiec			
Wavelength 1	505	700		
Blank type	NO	0	0	
Test time	40	120	Sec	
Calculation	Multi-pt.Linear Calibration			

Limited Range

ABS Range	0	0		
Linear Range	0	15	mg/dl	
Reagent Blank	0	0		
Linear Limit	20	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	mg/dl	B	0
Precision	X.XX	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 420

C - reactive protein (CRP)

Test No	#			Full Name	CRP

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	9	0	1
Increment	18	0	1
Decrement	4	0	1

Reagent Volume (µl)

Reagent 1	180
Reagent 2	35
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point			
Wavelength 1	340	700		
Blank type	Pre-Blank	420	450	
Test time	280	300	Sec	
Calculation	Spline			

Limited Range

ABS Range	0	0		
Linear Range	0	160	mg/L	
Reagent Blank	0	0		
Linear Limit	0	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	mg/L	B	0
Precision	X.X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 420

Glucose

Test No	#			Full Name	GLU

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	2	0	1
Increment	4	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	200
Reagent 2	0
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point			
Wavelength 1	505	700		
Blank type	Pre-Blank	130	150	
Test time	580	600	Sec	
Calculation	Multi-pt.Linear Calibration			

Limited Range

ABS Range	0	0		
Linear Range	5	400	mg/dl	
Reagent Blank	0	0		
Linear Limit	0	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	mg/dl	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 420

HDL Cholesterol Direct

Test No	#			Full Name	HDL

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	3	0	1
Increment	6	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	225
Reagent 2	75
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point				
Wavelength 1	600	700			
Blank type	Pre-Blank	420	450		
Test time	280	300	Sec		
Calculation	Multi-pt.Linear Calibration				

Limited Range

ABS Range	0	0		
Linear Range	1	100	mg/dl	
Reagent Blank	0	0		
Linear Limit	0	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	mg/dl	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 420

Immunoglobulin A (IgA)

Test No	#			Full Name	IgA

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	2	0	1
Increment	4	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	180
Reagent 2	35
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point			
Wavelength 1	340	700		
Blank type	Pre-Blank	420	450	
Test time	280	300	Sec	
Calculation	Logistic-Log 4P			

Limited Range

ABS Range	0	0		
Linear Range	1	800	mg/dl	
Reagent Blank	0	0		
Linear Limit	0	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	mg/dl	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 420

Immunoglobulin G (IgG)

Test No	#			Full Name	IgG

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	2	0	1
Increment	4	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	180
Reagent 2	35
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point			
Wavelength 1	600	795		
Blank type	Pre-Blank	420	450	
Test time	280	300	Sec	
Calculation	Logistic-Log 4P			

Limited Range

ABS Range	0	0		
Linear Range	5	3000	mg/dl	
Reagent Blank	0	0		
Linear Limit	0	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	mg/dl	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 420

Immunoglobulin M (IgM)

Test No	#			Full Name	IgM

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	2	0	1
Increment	4	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	180
Reagent 2	35
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point			
Wavelength 1	340	700		
Blank type	Pre-Blank	420	450	
Test time	280	300	Sec	
Calculation	Logistic-Log 4P			

Limited Range

ABS Range	0	0		
Linear Range	1	300	mg/dl	
Reagent Blank	0	0		
Linear Limit	0	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	mg/dl	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 420

Lactate Dehydrogenase (LDH)

Test No	#			Full Name	LDH

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	4	0	1
Increment	8	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	200
Reagent 2	40
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	Kinetiec			
Wavelength 1	340	700		
Blank type	NO	0	0	
Test time	60	240	Sec	
Calculation	Multi-pt.Linear Calibration			

Limited Range

ABS Range	0	0		
Linear Range	1	1000	U/L	
Reagent Blank	0	0		
Linear Limit	20	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	U/L	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 420

LDL Cholesterol Direct

Test No	#			Full Name	LDL

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	3	0	1
Increment	6	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	225
Reagent 2	75
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point				
Wavelength 1	540	700			
Blank type	Pre-Blank	420	450		
Test time	280	300	Sec		
Calculation	Multi-pt.Linear Calibration				

Limited Range

ABS Range	0	0		
Linear Range	1	500	mg/dl	
Reagent Blank	0	0		
Linear Limit	0	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	mg/dl	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 420

Lipase

Test No	#			Full Name	LIP

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	4	0	1
Increment	8	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	160
Reagent 2	40
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	Kinetiec			
Wavelength 1	570	700		
Blank type	NO	0	0	
Test time	120	240	Sec	
Calculation	Multi-pt.Linear Calibration			

Limited Range

ABS Range	0	0		
Linear Range	1	150	U/L	
Reagent Blank	0	0		
Linear Limit	20	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	U/L	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 420

Microalbumin

Test No	#			Full Name	MALB		

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	15	0	1
Increment	25	0	1
Decrement	3	0	1

Reagent Volume (µl)

Reagent 1	180
Reagent 2	35
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point			
Wavelength 1	340	700		
Blank type	Pre-Blank	420	450	
Test time	280	300	Sec	
Calculation	Logistic-Log 4P			

Limited Range

ABS Range	0	0		
Linear Range	0	250	mg/L	
Reagent Blank	0	0		
Linear Limit	0	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	mg/L	B	0
Precision	X.X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 420

CK - MB

Test No	#			Full Name	CKMB

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	10	0	1
Increment	20	0	1
Decrement	5	0	1

Reagent Volume (µl)

Reagent 1	200
Reagent 2	40
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	Kinetiec			
Wavelength 1	340	540		
Blank type	NO	0	0	
Test time	120	300	Sec	
Calculation	Multi-pt.Linear Calibration			

Limited Range

ABS Range	0	0		
Linear Range	1	800	U/L	
Reagent Blank	0	0		
Linear Limit	20	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	U/L	B	0
Precision	X.X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 420

Magnesium Xylidyl Blue

Test No	#			Full Name	MG		
Sample Volume (µl)				**Reagent Volume (µl)**			
	Test Volume	Dilution Volume	Dilution Ratio	Reagent 1	200		
Standard	2	0	1	Reagent 2	0		
Increment	4	0	1	Reagent 3	0		
Decrement	2	0	1	Reagent 4	0		
Test Parameter				**Limited Range**			
Method type	End-point			ABS Range	0 -	0	
Wavelength 1	505	670		Linear Range	0 -	5	mg/dl
Blank type	Pre-Blank	130 -	150	Reagent Blank	0 -	0	
Test time	280	300	Sec	Linear Limit	0	Substrat Exhaust	0
Calculation	Multi-pt.Linear Calibration			Inc. Judgment	0	Dec. Judgment	0
Result Parameter				**Prozone**			
Unit	mg/dl	B	0	Q1	0	Q2	0
Precision	X.X	K	1	Q3	0	Q4	0
				PC	0	ABS	0

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 420

Phosphorous UV

Test No	#			Full Name	PHOS

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	2	0	1
Increment	4	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	200
Reagent 2	0
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point			
Wavelength 1	340	700		
Blank type	Pre-Blank	130	150	
Test time	280	300	Sec	
Calculation	Multi-pt.Linear Calibration			

Limited Range

ABS Range	0	0		
Linear Range	0	20	mg/dl	
Reagent Blank	0	0		
Linear Limit	0	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	mg/dl	B	0
Precision	X.X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 420

Triglycerides

Test No	#			Full Name	TG

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	2	0	1
Increment	4	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	200
Reagent 2	0
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point			
Wavelength 1	505	700		
Blank type	Pre-Blank	130	150	
Test time	580	600	Sec	
Calculation	Multi-pt.Linear Calibration			

Limited Range

ABS Range	0	0		
Linear Range	5	700	mg/dl	
Reagent Blank	0	0		
Linear Limit	0	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	mg/dl	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 420

Total Protein

Test No	#			Full Name	TPRO

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	5	0	1
Increment	10	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	200
Reagent 2	0
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point			
Wavelength 1	540	700		
Blank type	Pre-Blank	130	150	
Test time	280	300	Sec	
Calculation	Multi-pt.Linear Calibration			

Limited Range

ABS Range	0	0		
Linear Range	0	12	g/dl	
Reagent Blank	0	0		
Linear Limit	0	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	g/dl	B	0
Precision	X.XX	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 420

Uric Acid

Test No	#			Full Name	UA		

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	5	0	1
Increment	10	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	100
Reagent 2	100
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point			
Wavelength 1	540	700		
Blank type	NO	0	0	
Test time	180	200	Sec	
Calculation	Multi-pt.Linear Calibration			

Limited Range

ABS Range	0	0		
Linear Range	0	20	mg/dl	
Reagent Blank	0	0		
Linear Limit	0	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	mg/dl	B	0
Precision	X.X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 420

UREA

Test No	#			Full Name	UREA		

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	2	0	1
Increment	4	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	160
Reagent 2	40
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	Kinetiec			
Wavelength 1	340	405		
Blank type	NO	0	0	
Test time	40	120	Sec	
Calculation	Multi-pt.Linear Calibration			

Limited Range

ABS Range	0	0		
Linear Range	1	300	mg/dl	
Reagent Blank	0	0		
Linear Limit	20	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	mg/dl	B	0
Precision	X.X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 420

ZINC

Test No	#			Full Name	ZINC

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	10	0	1
Increment	20	0	1
Decrement	5	0	1

Reagent Volume (µl)

Reagent 1	200
Reagent 2	0
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point			
Wavelength 1	570	700		
Blank type	Pre-Blank	130	150	
Test time	160	180	Sec	
Calculation	Multi-pt.Linear Calibration			

Limited Range

ABS Range	0	0		
Linear Range	20	400	µg/dl	
Reagent Blank	0	0		
Linear Limit	0	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	µg/dl	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir

پشتیبانی / شکرالهی 09129579408