BIOMEDIC

Kayto 800

Test No	#			Full Name	ALBUMIN

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	2	0	1
Increment	4	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	200
Reagent 2	0
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point			
Wavelength 1	635	700		
Blank type	Pre-Blank	32	-	50
Test time	280	300	Sec	
Calculation	Multi-pt.Linear Calibration			

Limited Range

ABS Range	0	-	0	
Linear Range	0	-	6	g/dl
Reagent Blank	0	-	0	
Linear Limit	0	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	g/dl	B	0
Precision	X.XX	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 800

Test No	#			Full Name	ALP

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	4	0	1
Increment	8	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	160
Reagent 2	40
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	Kinetiec			
Wavelength 1	405	700		
Blank type	NO	-		
Test time	60	240	Sec	
Calculation	Multi-pt.Linear Calibration			

Limited Range

ABS Range	0	- 0		
Linear Range	1	- 1000	U/L	
Reagent Blank	0	- 0		
Linear Limit	20	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	U/L	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Rayto 800

Test No	#			Full Name	ALT

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	20	0	1
Increment	25	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	160
Reagent 2	40
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	Kinetiec		
Wavelength 1	340	405	
Blank type	NO	-	
Test time	60	240	Sec
Calculation	Multi-pt.Linear Calibration		

Limited Range

ABS Range	0	- 0		
Linear Range	0	- 300	U/L	
Reagent Blank	0	- 0		
Linear Limit	20	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	U/L	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Rayto 800

Test No	#			Full Name	Amylase

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	4	0	1
Increment	8	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	200
Reagent 2	0
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	Kinetiec			
Wavelength 1	405	670		
Blank type	NO	-		
Test time	60	240	Sec	
Calculation	Multi-pt.Linear Calibration			

Limited Range

ABS Range	0	-	0		
Linear Range	1	-	1500	U/L	
Reagent Blank	0	-	0		
Linear Limit	20	Substrat Exhaust	0		
Inc. Judgment	0	Dec. Judgment	0		

Result Parameter

Unit	U/L	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 800

Test No	#			Full Name	AST

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	20	0	1
Increment	25	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	160
Reagent 2	40
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	Kinetiec		
Wavelength 1	340	405	
Blank type	NO	-	
Test time	60	240	Sec
Calculation	Multi-pt.Linear Calibration		

Limited Range

ABS Range	0	- 0		
Linear Range	0	- 300	U/L	
Reagent Blank	0	- 0		
Linear Limit	20	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	U/L	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Rayto 800

Test No	#			Full Name	C3

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	2	0	1
Increment	4	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	180
Reagent 2	35
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point			
Wavelength 1	340	700		
Blank type	Pre-Blank	250	-	270
Test time	280	300	Sec	
Calculation	Logistic-Log 4P			

Limited Range

ABS Range	0	-	0	
Linear Range	1	-	500	mg/dl
Reagent Blank	0	-	0	
Linear Limit	0	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	mg/dl	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 800

Test No	#	Full Name	C4

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	4	0	1
Increment	8	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	180
Reagent 2	35
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point			
Wavelength 1	340	700		
Blank type	Pre-Blank	250	-	270
Test time	280	300	Sec	
Calculation	Logistic-Log 4P			

Limited Range

ABS Range	0	-	0	
Linear Range	1	-	80	mg/dl
Reagent Blank	0	-	0	
Linear Limit	0	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	mg/dl	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 800

Test No	#			Full Name	Calcium

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	2	0	1
Increment	4	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	200
Reagent 2	0
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point			
Wavelength 1	670	795		
Blank type	NO	32	- 50	
Test time	180	200	Sec	
Calculation	Multi-pt.Linear Calibration			

Limited Range

ABS Range	0	- 0		
Linear Range	0	- 20	mg/dl	
Reagent Blank	0	- 0		
Linear Limit	0	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	mg/dl	B	0
Precision	X.X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Rayto 800

Test No	#			Full Name	Cholesterol

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	2	0	1
Increment	4	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	200
Reagent 2	0
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point			
Wavelength 1	505	700		
Blank type	Pre-Blank	32	-	50
Test time	510	530	Sec	
Calculation	Multi-pt.Linear Calibration			

Limited Range

ABS Range	0	-	0	
Linear Range	1	-	600	mg/dl
Reagent Blank	0	-	0	
Linear Limit	0	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	mg/dl	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Rayto 800

Test No	#			Full Name	CPK

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	12	0	1
Increment	20	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	200
Reagent 2	40
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point		
Wavelength 1	340	540	
Blank type	NO	-	
Test time	120	300	Sec
Calculation	Multi-pt.Linear Calibration		

Limited Range

ABS Range	0	- 0		
Linear Range	0	- 1000		U/L
Reagent Blank	0	- 0		
Linear Limit	0		Substrat Exhaust	0
Inc. Judgment	0		Dec. Judgment	0

Result Parameter

Unit	U/L	B	0
Precision	X.XX	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 800

Test No	#			Full Name	Creatinine

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	24	0	1
Increment	25	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	120
Reagent 2	120
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	Kinetiec		
Wavelength 1	505	670	
Blank type	NO	-	
Test time	40	120	Sec
Calculation	Multi-pt.Linear Calibration		

Limited Range

ABS Range	0	- 0		
Linear Range	0	- 15		mg/dl
Reagent Blank	0	- 0		
Linear Limit	20		Substrat Exhaust	0
Inc. Judgment	0		Dec. Judgment	0

Result Parameter

Unit	mg/dl	B	0
Precision	X.XX	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Rayto 800

Test No	#			Full Name	CRP

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	12	0	1
Increment	25	0	1
Decrement	5	0	1

Reagent Volume (µl)

Reagent 1	200
Reagent 2	40
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point			
Wavelength 1	340	700		
Blank type	Pre-Blank	250	- 270	
Test time	280	300	Sec	
Calculation	Logistic-Log 4P			

Limited Range

ABS Range	0	- 0		
Linear Range	1	- 180		mg/L
Reagent Blank	0	- 0		
Linear Limit	0	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	mg/L	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Rayto 800

Test No	#			Full Name	Glucose

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	2	0	1
Increment	4	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	200
Reagent 2	0
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point			
Wavelength 1	505	700		
Blank type	Pre-Blank	32	-	50
Test time	510	530	Sec	
Calculation	Multi-pt.Linear Calibration			

Limited Range

ABS Range	0	-	0	
Linear Range	1	-	400	mg/dl
Reagent Blank	0	-	0	
Linear Limit	0	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	mg/dl	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 800

Test No	#			Full Name	HDL

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	2	0	1
Increment	4	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	150
Reagent 2	50
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point			
Wavelength 1	600	700		
Blank type	Pre-Blank	250	- 270	
Test time	280	300	Sec	
Calculation	Multi-pt.Linear Calibration			

Limited Range

ABS Range	0	- 0		
Linear Range	1	- 120		mg/dl
Reagent Blank	0	- 0		
Linear Limit	0	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	mg/dl	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Rayto 800

Test No	#	Full Name	IGA

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	2	0	1
Increment	4	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	180
Reagent 2	35
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point			
Wavelength 1	340	700		
Blank type	Pre-Blank	250	-	270
Test time	280	300	Sec	
Calculation	Logistic-Log 4P			

Limited Range

ABS Range	0	-	0	
Linear Range	1	-	800	mg/dl
Reagent Blank	0	-	0	
Linear Limit	0	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	mg/dl	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 800

Test No	#			Full Name	IGG

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	2	0	1
Increment	4	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	180
Reagent 2	35
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point			
Wavelength 1	600	795		
Blank type	Pre-Blank	250	-	270
Test time	280	300	Sec	
Calculation	Logistic-Log 4P			

Limited Range

ABS Range	0	-	0	
Linear Range	1	-	3500	mg/dl
Reagent Blank	0	-	0	
Linear Limit	0	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	mg/dl	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 800

Test No	#	Full Name	IGM

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	2	0	1
Increment	4	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	180
Reagent 2	35
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point			
Wavelength 1	340	700		
Blank type	Pre-Blank	250	-	270
Test time	280	300	Sec	
Calculation	Logistic-Log 4P			

Limited Range

ABS Range	0	-	0	
Linear Range	1	-	400	mg/dl
Reagent Blank	0	-	0	
Linear Limit	0	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	mg/dl	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Rayto 800

Test No	#			Full Name	LDH

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	3	0	1
Increment	6	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	150
Reagent 2	30
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	Kinetiec		
Wavelength 1	340	700	
Blank type	NO	-	
Test time	60	240	Sec
Calculation	Multi-pt.Linear Calibration		

Limited Range

ABS Range	0	- 0		
Linear Range	1	- 1000		U/L
Reagent Blank	0	- 0		
Linear Limit	20		Substrat Exhaust	0
Inc. Judgment	0		Dec. Judgment	0

Result Parameter

Unit	U/L	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 800

Test No	#			Full Name	LDL

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	3	0	1
Increment	6	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	150
Reagent 2	50
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point			
Wavelength 1	540	700		
Blank type	Pre-Blank	250	- 270	
Test time	280	300	Sec	
Calculation	Multi-pt.Linear Calibration			

Limited Range

ABS Range	0	- 0		
Linear Range	1	- 600		mg/dl
Reagent Blank	0	- 0		
Linear Limit	0	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	mg/dl	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Kayto 800

Test No	#			Full Name	Lipase

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	4	0	1
Increment	8	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	160
Reagent 2	40
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	Kinetiec		
Wavelength 1	570	700	
Blank type	NO	-	
Test time	120	240	Sec
Calculation	Multi-pt.Linear Calibration		

Limited Range

ABS Range	0	- 0		
Linear Range	1	- 150		U/L
Reagent Blank	0	- 0		
Linear Limit	0		Substrat Exhaust	0
Inc. Judgment	0		Dec. Judgment	0

Result Parameter

Unit	U/L	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Rayto 800

Test No	#		Full Name	CK-MB

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	12	0	1
Increment	20	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	200
Reagent 2	40
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point		
Wavelength 1	340	540	
Blank type	NO	-	
Test time	120	300	Sec
Calculation	Multi-pt.Linear Calibration		

Limited Range

ABS Range	0	- 0		
Linear Range	0	- 1000		U/L
Reagent Blank	0	- 0		
Linear Limit	0		Substrat Exhaust	0
Inc. Judgment	0		Dec. Judgment	0

Result Parameter

Unit	U/L	B	0
Precision	X.XX	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Rayto 800

Test No	#	Full Name	Magnesium

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	2	0	1
Increment	4	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	200
Reagent 2	0
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point			
Wavelength 1	505	700		
Blank type	Pre-Blank	32	-	50
Test time	120	150	Sec	
Calculation	Multi-pt.Linear Calibration			

Limited Range

ABS Range	0	-	0		
Linear Range	0.1	-	5	mg/dl	
Reagent Blank	0	-	0		
Linear Limit	0	Substrat Exhaust	0		
Inc. Judgment	0	Dec. Judgment	0		

Result Parameter

Unit	mg/dl	B	0
Precision	X.XX	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Rayto 800

Test No	#			Full Name	Micro Albumin

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	14	0	1
Increment	24	0	1
Decrement	7	0	1

Reagent Volume (µl)

Reagent 1	200
Reagent 2	40
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point			
Wavelength 1	340	700		
Blank type	Pre-Blank	250	- 270	
Test time	280	300	Sec	
Calculation	Logistic-Log 4P			

Limited Range

ABS Range	0	- 0		
Linear Range	1	- 280		mg/L
Reagent Blank	0	- 0		
Linear Limit	0	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	mg/L	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Kayto 800

Test No	#			Full Name	Phosphorus

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	2	0	1
Increment	4	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	200
Reagent 2	0
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point			
Wavelength 1	340	700		
Blank type	Pre-Blank	32	-	50
Test time	120	150	Sec	
Calculation	Multi-pt.Linear Calibration			

Limited Range

ABS Range	0	-	0		
Linear Range	0.1	-	20	mg/dl	
Reagent Blank	0	-	0		
Linear Limit	0	Substrat Exhaust	0		
Inc. Judgment	0	Dec. Judgment	0		

Result Parameter

Unit	mg/dl	B	0
Precision	X.XX	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 800

Test No	#	Full Name	Triglycerides

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	2	0	1
Increment	4	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	200
Reagent 2	0
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point			
Wavelength 1	505	700		
Blank type	Pre-Blank	32	-	50
Test time	510	530	Sec	
Calculation	Multi-pt.Linear Calibration			

Limited Range

ABS Range	0	-	0		
Linear Range	1	-	600	mg/dl	
Reagent Blank	0	-	0		
Linear Limit	0	Substrat Exhaust	0		
Inc. Judgment	0	Dec. Judgment	0		

Result Parameter

Unit	mg/dl	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 800

Test No	#			Full Name	Total Protein

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	5	0	1
Increment	10	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	200
Reagent 2	0
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point			
Wavelength 1	540	700		
Blank type	Pre-Blank	32	-	50
Test time	280	300	Sec	
Calculation	Multi-pt.Linear Calibration			

Limited Range

ABS Range	0	-	0	
Linear Range	0	-	12	g/dl
Reagent Blank	0	-	0	
Linear Limit	0	Substrat Exhaust	0	
Inc. Judgment	0	Dec. Judgment	0	

Result Parameter

Unit	g/dl	B	0
Precision	X.XX	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Kayto 800

Test No	#			Full Name	Uric Acid

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	8	0	1
Increment	16	0	1
Decrement	4	0	1

Reagent Volume (µl)

Reagent 1	150
Reagent 2	150
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point		
Wavelength 1	540	700	
Blank type	NO	-	
Test time	180	200	Sec
Calculation	Multi-pt.Linear Calibration		

Limited Range

ABS Range	0	- 0		
Linear Range	1	- 35		mg/dl
Reagent Blank	0	- 0		
Linear Limit	0		Substrat Exhaust	0
Inc. Judgment	0		Dec. Judgment	0

Result Parameter

Unit	mg/dl	B	0
Precision	X.X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی ۰۹۱۲۹۵۷۹۴۰۸

BIOMEDIC

Rayto 800

Test No	#			Full Name	Urea

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	2	0	1
Increment	4	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	160
Reagent 2	40
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	Kinetiec			
Wavelength 1	340	405		
Blank type	NO	-		
Test time	30	120	Sec	
Calculation	Multi-pt.Linear Calibration			

Limited Range

ABS Range	0	-	0		
Linear Range	1	-	300	mg/dl	
Reagent Blank	0	-	0		
Linear Limit	20	Substrat Exhaust	0		
Inc. Judgment	0	Dec. Judgment	0		

Result Parameter

Unit	mg/dl	B	0
Precision	X.X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی 09129579408

BIOMEDIC

Rayto 800

Test No	#			Full Name	Zinc

Sample Volume (µl)

	Test Volume	Dilution Volume	Dilution Ratio
Standard	10	0	1
Increment	20	0	1
Decrement	2	0	1

Reagent Volume (µl)

Reagent 1	200
Reagent 2	0
Reagent 3	0
Reagent 4	0

Test Parameter

Method type	End-point			
Wavelength 1	570	700		
Blank type	Pre-Blank	32	-	50
Test time	150	170	Sec	
Calculation	Multi-pt.Linear Calibration			

Limited Range

ABS Range	0	-	0		
Linear Range	1	-	400	µg/dl	
Reagent Blank	0	-	0		
Linear Limit	0	Substrat Exhaust	0		
Inc. Judgment	0	Dec. Judgment	0		

Result Parameter

Unit	µg/dl	B	0
Precision	X	K	1

Prozone

Q1	0	Q2	0	Q3	0	Q4	0
PC	0			ABS	0		

شرکت توسعه کیمیای سعادت www.tksmed.ir
پشتیبانی / شکرالهی 09129579408