

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	On
Load images to graphic segments	On
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Dist. factor	20 %
Position	L1.1 A15.9 F6.2 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	128
FoV read	260 mm
FoV phase	100.0 %
Slice thickness	1.6 mm
TR	3.25 ms
TE	1.53 ms
Averages	1
Concatenations	1
Filter	B1 filter
Coil elements	AC

Contrast - Common

TR	3.25 ms
TE	1.53 ms
Flip angle	16.0 deg

Contrast - Dynamic

Averages	1
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1

Resolution - Common

FoV read	260 mm
FoV phase	100.0 %
Slice thickness	1.6 mm
Base resolution	160
Phase resolution	100 %
Slice resolution	69 %
Phase partial Fourier	6/8
Slice partial Fourier	6/8
Trajectory	Cartesian

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	3
Ref. lines PE	24
Accel. factor 3D	1

Resolution - iPAT

Reference scan mode	Integrated
---------------------	------------

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	On
Unfiltered images	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slab group	1
Slabs	1
Dist. factor	20 %
Position	L1.1 A15.9 F6.2 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice oversampling	0.0 %
Slices per slab	128
FoV read	260 mm
FoV phase	100.0 %
Slice thickness	1.6 mm
TR	3.25 ms
Multi-slice mode	Sequential
Series	Ascending
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	L1.1 A15.9 F6.2 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Transversal

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off

SIEMENS MAGNETOM Investigational_Device_7T

System - Miscellaneous

Coil Select Mode	Default
------------------	---------

System - Adjustments

B0 Shim mode	Tune up
B1 Shim mode	TrueForm
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	297.183273 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - PACE

Resp. control	Off
Concatenations	1

Inline - Common

Flip angle	16.0 deg
Measurements	1
Time to center	6.3 s

Inline - Inline

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Inline Composing	Off
Distortion Corr.	Off

Inline - MapIt

Save original images	On
MapIt	None
Flip angle	16.0 deg
Measurements	1
Contrasts	1

Inline - MapIt

TR	3.25 ms
TE	1.53 ms

Sequence - Part 1

Introduction	On
Dimension	3D
Asymmetric echo	Weak
Contrasts	1
Multi-slice mode	Sequential
Bandwidth	540 Hz/Px

Sequence - Part 2

RF pulse type	Fast
Gradient mode	Normal
Excitation	Non-sel.
RF spoiling	On

Sequence - Assistant

Mode	Off
------	-----

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	On
Load images to graphic segments	On
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Dist. factor	20 %
Position	L1.1 A15.9 F6.2 mm
Orientation	Transversal
Phase enc. dir.	A >> P
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	128
FoV read	260 mm
FoV phase	100.0 %
Slice thickness	1.6 mm
TR	3.25 ms
TE	1.53 ms
Averages	1
Concatenations	1
Filter	B1 filter
Coil elements	AC

Contrast - Common

TR	3.25 ms
TE	1.53 ms
Flip angle	16.0 deg

Contrast - Dynamic

Averages	1
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1

Resolution - Common

FoV read	260 mm
FoV phase	100.0 %
Slice thickness	1.6 mm
Base resolution	160
Phase resolution	100 %
Slice resolution	69 %
Phase partial Fourier	6/8
Slice partial Fourier	6/8
Trajectory	Cartesian

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	3
Ref. lines PE	24
Accel. factor 3D	1

Resolution - iPAT

Reference scan mode	Integrated
---------------------	------------

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	On
Unfiltered images	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slab group	1
Slabs	1
Dist. factor	20 %
Position	L1.1 A15.9 F6.2 mm
Orientation	Transversal
Phase enc. dir.	A >> P
Slice oversampling	0.0 %
Slices per slab	128
FoV read	260 mm
FoV phase	100.0 %
Slice thickness	1.6 mm
TR	3.25 ms
Multi-slice mode	Sequential
Series	Ascending
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	L1.1 A15.9 F6.2 mm
Orientation	Transversal
Phase enc. dir.	A >> P
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Transversal

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off

SIEMENS MAGNETOM Investigational_Device_7T

System - Miscellaneous

Coil Select Mode	Default
------------------	---------

System - Adjustments

B0 Shim mode	Tune up
B1 Shim mode	TrueForm
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	297.183273 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - PACE

Resp. control	Off
Concatenations	1

Inline - Common

Flip angle	16.0 deg
Measurements	1
Time to center	6.3 s

Inline - Inline

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Inline Composing	Off
Distortion Corr.	Off

Inline - MapIt

Save original images	On
MapIt	None
Flip angle	16.0 deg
Measurements	1
Contrasts	1

Inline - MapIt

TR	3.25 ms
TE	1.53 ms

Sequence - Part 1

Introduction	On
Dimension	3D
Asymmetric echo	Weak
Contrasts	1
Multi-slice mode	Sequential
Bandwidth	540 Hz/Px

Sequence - Part 2

RF pulse type	Fast
Gradient mode	Normal
Excitation	Non-sel.
RF spoiling	On

Sequence - Assistant

Mode	Off
------	-----

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Position	R1.9 A6.0 F6.3 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
Slab Scale	-10 %
Slices per slab	80
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	2.62 mm
TR 1	58.3 ms
TR 2	3453 ms
TE 1	9.59 ms
TE 2	27.79 ms
TE 3	45.99 ms
Averages	1
Filter	None
Coil elements	AC

Contrast - Common

TR 1	58.3 ms
TR 2	3453 ms
TE 1	9.59 ms
TE 2	27.79 ms
TE 3	45.99 ms
Multi-echo spacing	18.20 ms
Magn. preparation	Non-sel. HSN IR
TI 1	993 ms
TI 2	1867.5 ms
TI 3	2742 ms
Flip angle	15 deg
Fat suppr.	Fat sat.
Magn. Prep. Shots	1

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	240
Pause after meas.	0.0 s

Resolution - Common

FoV read	220 mm
FoV phase	100.0 %
Slice thickness	2.62 mm

Resolution - Common

Base resolution	84
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	7/8
Slice partial Fourier	6/8
Interpolation	Off

Resolution - IPAT

PAT mode	CAIPIRINHA
Acc. factor PE	3
Ref. lines PE	63
Acc. factor 3D	4
Ref. lines 3D	36
CAIPI 3D Shift	2
Reference Scan Mode	EPI/separate
CAIPI Mode (tooltip)	Skipped-CAIPI
Total PAT factor	12

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slab group	1
Slabs	1
Position	R1.9 A6.0 F6.3 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slab Scale	-10 %
Slices per slab	80
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	2.62 mm
TR 1	58.3 ms
TR 2	3453 ms

Geometry - AutoAlign

Slab group	1
Position	R1.9 A6.0 F6.3 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	R1.9 A6.0 F6.3
R	1.9 mm
A	6.0 mm
F	6.3 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Saturation

Saturation mode	Standard
Fat suppr.	Fat sat.

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

! Position	R0.5 P6.9 H15.4 mm
! Orientation	Sagittal
! Rotation	19.33 deg
! A >> P	185 mm
! F >> H	110 mm
! R >> L	151 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Slab-sel.

System - Tx/Rx

Frequency 1H	297.183273 MHz
Correction factor	1
Gain	Low
Img. Scale Cor.	1.000
Reset	Off
! Ref. amplitude 1H	200.000 V

Sequence - Part 1

Introduction	On
Dimension	3D
Reordering	Linear
Contrasts	3
Echo spacing	0.74 ms
Bandwidth	1860 Hz/Px

Sequence - Part 2

EPI factor	24
Segmentation	1
RF pulse type	Normal
Gradient mode	Fast
Excitation	Slab-sel.
RF spoiling	On

Sequence - Part 2

Turbo factor	15
--------------	----

Sequence - Special

PATRef FA	3 deg
RF duration	600 us
RF BWT product	8
Ernst T1	1200 ms
PATRef prep. shots	10
Volume dummy shots	0
Dummy Measurements	0
ETL per RTEB	1
Invert PE	Off
Min. TE if PF	On
Echo Time Shift	On
Ramp Sampling	Off
NORDIC	Off
SVDPC	On
Sym VASO	Off
Dual-pol. EPI	Off
Invert RO	Off
Invert 3D	Off
Disable PF reco	Off
Disable PF reco	Off
Save sampling	Off
PE VComp	Off
Water Exc.	Long bino-11
External PC	per Series
Saturation RF	per Shot
FIDNavs	-none-
EPI rise time factor	1.15
Mosaic DICOMs	On
Modify Ice Config	On
G-factor map	Off
GRAPPA Regularization	50000 10^-6
HSN RF power scale	3.00
Inversion Delay	400 ms
Relaxation Delay	0 ms
Var. FA /MAGEC	0

Sequence - Assistant

Mode	Off
------	-----

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Position	R1.9 A6.0 F6.3 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
Slab Scale	-10 %
Slices per slab	80
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	2.62 mm
TR 1	58.3 ms
TR 2	3453 ms
TE 1	9.59 ms
TE 2	27.79 ms
TE 3	45.99 ms
Averages	1
Filter	None
Coil elements	AC

Contrast - Common

TR 1	58.3 ms
TR 2	3453 ms
TE 1	9.59 ms
TE 2	27.79 ms
TE 3	45.99 ms
Multi-echo spacing	18.20 ms
Magn. preparation	Non-sel. HSN IR
TI 1	993 ms
TI 2	1867.5 ms
TI 3	2742 ms
Flip angle	15 deg
Fat suppr.	Fat sat.
Magn. Prep. Shots	1

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	3
Pause after meas. 1	0.0 s
Pause after meas. 2	0.0 s

Resolution - Common

FoV read	220 mm
FoV phase	100.0 %

Resolution - Common

Slice thickness	2.62 mm
Base resolution	84
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	7/8
Slice partial Fourier	6/8
Interpolation	Off

Resolution - iPAT

PAT mode	CAIPIRINHA
Acc. factor PE	3
Ref. lines PE	63
Acc. factor 3D	4
Ref. lines 3D	36
CAIPI 3D Shift	2
Reference Scan Mode	EPI/separate
CAIPI Mode (tooltip)	Skipped-CAIPI
Total PAT factor	12

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slab group	1
Slabs	1
Position	R1.9 A6.0 F6.3 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slab Scale	-10 %
Slices per slab	80
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	2.62 mm
TR 1	58.3 ms
TR 2	3453 ms

Geometry - AutoAlign

Slab group	1
Position	R1.9 A6.0 F6.3 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	R1.9 A6.0 F6.3
R	1.9 mm
A	6.0 mm
F	6.3 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Saturation

Saturation mode	Standard
Fat suppr.	Fat sat.

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

! Position	R0.5 P6.9 H15.4 mm
! Orientation	Sagittal
! Rotation	19.33 deg
! A >> P	185 mm
! F >> H	110 mm
! R >> L	151 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Slab-sel.

System - Tx/Rx

Frequency 1H	297.183273 MHz
Correction factor	1
Gain	Low
Img. Scale Cor.	1.000
Reset	Off
! Ref. amplitude 1H	200.000 V

Sequence - Part 1

Introduction	On
Dimension	3D
Reordering	Linear
Contrasts	3
Echo spacing	0.74 ms
Bandwidth	1860 Hz/Px

Sequence - Part 2

EPI factor	24
Segmentation	1
RF pulse type	Normal
Gradient mode	Fast
Excitation	Slab-sel.
RF spoiling	On

Sequence - Part 2

Turbo factor	15
--------------	----

Sequence - Special

PATRef FA	3 deg
RF duration	600 us
RF BWT product	8
Ernst T1	1200 ms
PATRef prep. shots	10
Volume dummy shots	0
Dummy Measurements	0
ETL per RTEB	1
Invert PE	Off
Min. TE if PF	On
Echo Time Shift	On
Ramp Sampling	Off
NORDIC	Off
SVDPC	On
Sym VASO	Off
Dual-pol. EPI	Off
Invert RO	Off
Invert 3D	Off
Disable PF reco	Off
Disable PF reco	Off
Save sampling	Off
PE VComp	Off
Water Exc.	Long bino-11
External PC	per Series
Saturation RF	per Shot
FIDNavs	-none-
EPI rise time factor	1.15
Mosaic DICOMs	On
Modify Ice Config	On
G-factor map	Off
GRAPPA Regularization	50000 10^-6
HSN RF power scale	3.00
Inversion Delay	400 ms
Relaxation Delay	0 ms
Var. FA /MAGEC	0

Sequence - Assistant

Mode	Off
------	-----

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Position	R1.9 P2.5 F6.3 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
Slab Scale	-10 %
Slices per slab	80
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	2.62 mm
TR 1	58.3 ms
TR 2	3043 ms
TE 1	9.59 ms
TE 2	27.79 ms
TE 3	45.99 ms
Averages	1
Filter	None
Coil elements	AC

Contrast - Common

TR 1	58.3 ms
TR 2	3043 ms
TE 1	9.59 ms
TE 2	27.79 ms
TE 3	45.99 ms
Multi-echo spacing	18.20 ms
Magn. preparation	Non-sel. HSN IR
TI 1	993 ms
TI 2	1867.5 ms
TI 3	2742 ms
Flip angle	15 deg
Fat suppr.	None
Magn. Prep. Shots	1

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	50
Pause after meas. 1	0.0 s
Pause after meas. 2	0.0 s
Pause after meas. 3	0.0 s
Pause after meas. 4	0.0 s
Pause after meas. 5	0.0 s
Pause after meas. 6	0.0 s

Contrast - Dynamic

Pause after meas. 7	0.0 s
Pause after meas. 8	0.0 s
Pause after meas. 9	0.0 s
Pause after meas. 10	0.0 s
Pause after meas. 11	0.0 s
Pause after meas. 12	0.0 s
Pause after meas. 13	0.0 s
Pause after meas. 14	0.0 s
Pause after meas. 15	0.0 s
Pause after meas. 16	0.0 s
Pause after meas. 17	0.0 s
Pause after meas. 18	0.0 s
Pause after meas. 19	0.0 s
Pause after meas. 20	0.0 s
Pause after meas. 21	0.0 s
Pause after meas. 22	0.0 s
Pause after meas. 23	0.0 s
Pause after meas. 24	0.0 s
Pause after meas. 25	0.0 s
Pause after meas. 26	0.0 s
Pause after meas. 27	0.0 s
Pause after meas. 28	0.0 s
Pause after meas. 29	0.0 s
Pause after meas. 30	0.0 s
Pause after meas. 31	0.0 s
Pause after meas. 32	0.0 s
Pause after meas. 33	0.0 s
Pause after meas. 34	0.0 s
Pause after meas. 35	0.0 s
Pause after meas. 36	0.0 s
Pause after meas. 37	0.0 s
Pause after meas. 38	0.0 s
Pause after meas. 39	0.0 s
Pause after meas. 40	0.0 s
Pause after meas. 41	0.0 s
Pause after meas. 42	0.0 s
Pause after meas. 43	0.0 s
Pause after meas. 44	0.0 s
Pause after meas. 45	0.0 s
Pause after meas. 46	0.0 s
Pause after meas. 47	0.0 s
Pause after meas. 48	0.0 s
Pause after meas. 49	0.0 s

Resolution - Common

FoV read	220 mm
FoV phase	100.0 %
Slice thickness	2.62 mm
Base resolution	84
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	7/8
Slice partial Fourier	6/8
Interpolation	Off

Resolution - iPAT

PAT mode	CAIPIRINHA
Acc. factor PE	3
Ref. lines PE	63
Acc. factor 3D	4

SIEMENS MAGNETOM Investigational_Device_7T

Resolution - IPAT

Ref. lines 3D	36
CAIPI 3D Shift	2
Reference Scan Mode	EPI/separate
CAIPI Mode (tooltip)	Skipped-CAIPI
Total PAT factor	12

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slab group	1
Slabs	1
Position	R1.9 P2.5 F6.3 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slab Scale	-10 %
Slices per slab	80
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	2.62 mm
TR 1	58.3 ms
TR 2	3043 ms

Geometry - AutoAlign

Slab group	1
Position	R1.9 P2.5 F6.3 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	R1.9 P2.5 F6.3
R	1.9 mm
P	2.5 mm
F	6.3 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Saturation

Saturation mode	Standard
Fat suppr.	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares

System - Miscellaneous

Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

! Position	R0.5 P5.6 H3.0 mm
! Orientation	Sagittal
! Rotation	19.33 deg
! A >> P	185 mm
! F >> H	110 mm
! R >> L	151 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Slab-sel.

System - Tx/Rx

Frequency 1H	297.183273 MHz
Correction factor	1
Gain	Low
Img. Scale Cor.	1.000
Reset	Off
! Ref. amplitude 1H	200.000 V

Sequence - Part 1

Introduction	On
Dimension	3D
Reordering	Linear
Contrasts	3
Echo spacing	0.74 ms
Bandwidth	1860 Hz/Px

Sequence - Part 2

EPI factor	24
Segmentation	1
RF pulse type	Normal
Gradient mode	Fast
Excitation	Slab-sel.
RF spoiling	On
Turbo factor	15

Sequence - Special

PATRef FA	3 deg
RF duration	600 us
RF BWT product	8
Ernst T1	1200 ms
PATRef prep. shots	10
Volume dummy shots	0
Dummy Measurements	0
ETL per RTEB	1
Invert PE	Off
Min. TE if PF	On
Echo Time Shift	On

SIEMENS MAGNETOM Investigational_Device_7T

Sequence - Special

Ramp Sampling	Off
NORDIC	Off
SVDPC	On
Sym VASO	Off
Dual-pol. EPI	Off
Invert RO	Off
Invert 3D	Off
Disable PF reco	Off
Disable PF reco	Off
Save sampling	Off
PE VComp	Off
Water Exc.	Long bino-11
External PC	per Series
FIDNavs	-none-
EPI rise time factor	1.15
Mosaic DICOMs	On
Modify Ice Config	On
G-factor map	Off
GRAPPA Regularization	50000 10^-6
HSN RF power scale	3.00
Inversion Delay	400 ms
Relaxation Delay	0 ms
Var. FA /MAGEC	0

Sequence - Assistant

Mode	Off
------	-----